Cambridge Instruments & Engg. Co.

EST. 1963 · HOWRAH, INDIA

INSULATION TESTER · HAND-DRIVEN, UP TO 10KV

CIE/666

Hand-driven · single-range units, 100 V to 10,000 V · IS: 2992-1987

01 KEY SPECIFICATIONS

ACCURACY	IS: 2992-1987
GENERATOR	Hand-driven · centrifugal governor
ROTATION SPEED	160 r.p.m. or more
RESPONSE TIME	3 seconds
SCALE LENGTH	100 mm (approx.)
BODY	Seasoned wood, sunmica finish

Supplied as base instrument only - carrying case, test leads, and range options available at extra cost.



CIE/666 · INSULATION TESTER

02 TECHNICAL SPECIFICATIONS

PARAMETER	SPECIFICATION
MEASUREMENT	
Scale Length	100 mm (approx.)
Accuracy	Conforms to IS: 2992-1987
Response Time	3 seconds
Range Configuration	Multiple range options available - specify at order
Guard Terminal & Infinity Adjuster	Provided on instruments with insulation range 1000 MΩ and above
GENERATOR	
Generator Type	Hand-driven generator, centrifugal governor
Rotation Speed	160 r.p.m. or more
TEST & SAFETY	
High Voltage Test	Min. 2000V or (rated V + 1000)V A.C. r.m.s., 1 min
Insulation Resistance Test	Better than 50 MΩ at 500V D.C.
GENERAL	
Body	Seasoned wooden, sunmica finish, detachable lid & lock
Dimensions (L×W×H)	265 × 165 × 185 mm (approx.)
Weight	4.1 kg (approx.)
Warranty	12 months against manufacturing defects only

Custom ranges and configurations can also be built against specific requirements.

03 APPLICATIONS

- Motors, generators, and other rotating machinery
- Insulation control and switching equipment checks
- HV switchgear and transformer testing to 10,000V
- Locating shorts, defects, and insulation failures

04 OPTIONAL ACCESSORIES

+ Carrying case
+ Test leads, 1m/2m
+ H.V. test leads, 2m
+ Dual voltage/range option

All specifications are subject to change without prior notice. © Cambridge Instruments & Engg. Co. - CIE/666 Product Catalogue.