Cambridge Instruments
& Engg. Co.

EST. 1963 · HOWRAH, INDIA

INSULATION TESTER · HAND-DRIVEN, UP TO 10KV

CIE/666

500 V / 0–500 MΩ

Hand-driven insulation tester · centrifugal governor · seasoned wooden body

01 KEY FEATURES

- Hand-cranked generator, no battery
- Centrifugal governor
- Seasoned wooden body, sunmica finish
- Detachable lid, handle and lock

Supplied as base instrument only - carrying case, test leads, and range options available at extra cost.



CIE/666 · INSULATION TESTER · ILLUSTRATIVE IMAGE

02 TECHNICAL SPECIFICATIONS

PARAMETER	SPECIFICATION
CONFIGURATION	
Rated Test Voltage	500 V D.C.
Insulation Resistance Range	0–500 MΩ
MEASUREMENT	
Scale Length	100 mm (approx.)
Accuracy	Conforms to IS: 2992-1987
Response Time	3 seconds
GENERATOR	
Generator Type	Hand-driven generator, centrifugal governor
Rotation Speed	160 r.p.m. or more
TEST & SAFETY	
High Voltage Test	Min. 2000V or (rated V + 1000)V A.C. r.m.s., 1 min
Insulation Resistance Test	Better than 50 MΩ at 500V D.C.
GENERAL	
Body	Seasoned wooden, sunmica finish, detachable lid & lock
Dimensions (L×W×H)	265 × 165 × 185 mm (approx.)
Weight	4.1 kg (approx.)
Warranty	12 months against manufacturing defects only

Dual-voltage, dual-range options and custom ranges are available against specific requirements - enquire.

03 APPLICATIONS

- Motors, generators, and other rotating machinery
- Insulation control and switching equipment checks
- HV switchgear and transformer testing to 10,000V
- Locating shorts, defects, and insulation failures

04 OPTIONAL ACCESSORIES

+ Carrying case
+ Test leads, 1m/2m
+ H.V. test leads, 2m
+ Dual voltage/range option

All specifications are subject to change without prior notice. © Cambridge Instruments & Engg. Co. - CIE/666 500 V, 0-500 MΩ Product Catalogue.