Cambridge Instruments
& Engg. Co.

EST. 1963 · HOWRAH, INDIA

INSULATION TESTER · HAND-DRIVEN, BAKELITE BODY

# CIE/555

## 1000 V / 0–200 MΩ

Hand-driven insulation tester · clutch-type generator · bakelite body

## 01 KEY FEATURES

- Silent clutch-type mechanism
- High-sensitivity cross-coil indicator
- Moulded bakelite body, rubber shoes
- Collapsible driving handle




CIE/555 · INSULATION TESTER · ILLUSTRATIVE IMAGE

Supplied as base instrument only - carrying case and test leads available at extra cost.

## 02 OPTIONAL ACCESSORIES (extra cost)

+ Carrying case (bakelite body)
+ Testing leads, 1m

## 03 TECHNICAL SPECIFICATIONS

| PARAMETER | SPECIFICATION |
|---|---|
| **CONFIGURATION** | |
| Rated Test Voltage | 1000 V D.C. |
| Insulation Resistance Range | 0–200 MΩ |
| **MEASUREMENT** | |
| Indicator | High-sensitivity cross-coil |
| Scale Length | 75 mm (approx.) |
| Accuracy | Conforms to IS: 2992-1987 |
| Response Time | 3 seconds |
| **GENERATOR** | |
| Generator Type | Hand-driven, clutch-type generator |
| Rotation Speed | 160 r.p.m. |
| **GENERAL** | |
| Body | Moulded high-grade bakelite |
| Dimensions (L×W×H) | 155 × 104 × 70 mm (approx.) |
| Weight | 1.4 kg (approx.) |
| Warranty | 12 months against manufacturing defects only |

*Dual-voltage, dual-range options and custom ranges are available against specific requirements - enquire.*



All specifications are subject to change without prior notice. © Cambridge Instruments & Engg. Co. - CIE/555 1000 V, 0-200 MΩ Product Catalogue.