Cambridge Instruments
& Engg. Co.

EST. 1963 · HOWRAH, INDIA

INSULATION TESTER · HAND-DRIVEN

CIE/444

5000 V / 0–10000 MΩ

Hand-driven insulation tester · centrifugal governor · cross-coil indicator

01 KEY FEATURES

- Centrifugal governor
- Cross-coil indicator
- Anti-parallax mirror scale
- Anti-magnetic aluminium, die-cast body
- Collapsible generator handle

Supplied as base instrument only - carrying case, test leads, and range options available at extra cost.



CIE/444 · INSULATION TESTER · ILLUSTRATIVE IMAGE

02 TECHNICAL SPECIFICATIONS

PARAMETER	SPECIFICATION
CONFIGURATION	
Rated Test Voltage	5000 V D.C.
Insulation Resistance Range	0–10000 MΩ
Guard Terminal & Infinity Adjuster	Provided
MEASUREMENT	
Scale Length	90 mm (approx.)
Accuracy	Conforms to IS: 2992-1987
Response Time	3 seconds
GENERATOR	
Generator Type	Hand-driven generator, centrifugal governor
Rotation Speed	160 r.p.m. or more
TEST & SAFETY	
High Voltage Test	Min. 2000V or (rated V + 1000)V A.C. r.m.s., 1 min
Insulation Resistance Test	Better than 50 MΩ at 500V D.C.
GENERAL	
Body	Anti-magnetic aluminium
Dimensions (L×W×H)	110 × 128 × 195 mm (approx.)
Weight	2.5 kg (approx.)
Warranty	12 months against manufacturing defects only

Dual-voltage, dual-range and continuity options, and custom ranges, are available against specific requirements - enquire.

03 APPLICATIONS

- Motors, generators, and other rotating machinery
- Insulation control and switching equipment checks
- Distribution network and cable insulation testing
- Transformer, insulator, and installation checks
- Locating shorts, defects, and insulation failures

04 OPTIONS & ACCESSORIES (extra cost)

+ Carrying case
+ Test leads, 1m/2m
+ H.V. test leads, 2m
+ Dual voltage & dual insulation range option
+ Single voltage, dual insulation range option
+ Continuity range option (500 V / 1000 V only)

All specifications are subject to change without prior notice. © Cambridge Instruments & Engg. Co. - CIE/444 5000 V, 0-10000 MΩ Product Catalogue.