



0152GBO, couveuse, fevrier (active)
14 journées, plaque 200 W, 8 critères tenus sur 10

NRMSE 10,2 % $\leq$ 20,0	LDC 4,1 % $\leq$ 20,0	FFT 28,4 % $>$ 18,8	Énergie −0,8 % $\leq$ 20,0	Pointe −7,6 % $\leq$ 20,0
Fact. de charge 5,2 % $\leq$ 15,9	CORR_h 5,1 % $\leq$ 10,0	TVD_h 2,5 % $\leq$ 10,0	EXT_prof 20,1 % $\leq$ 50,0	Heure des pics 3,0 h $>$ 2,0
Forme 1 h 0,8 $\sigma \leq$ 3,4	Forme 2 h 0,6 $\sigma \leq$ 3,1	Pire écart 1 h 19,9 % $\leq$ 91,3	Verdict non vérifiable	Pointe jour. P95 200 / 224 W



Courbe classée, LDC 4,1 %



Harmoniques, FFT 28,4 %

