



National Metrology Laboratory of the Philippines

NATIONAL METROLOGY LABORATORY OF THE PHILIPPINES

Providing International Traceability to Measurements in the Country

CALIBRATION CERTIFICATE

No. 05-2022-FORC-0132

Date of Calibration : December 19, 2022 Calibration Item : Tensile Testing Machine Capacity > 100 lbf Measurement Range : 0 lbf to 90 lbf Resolution 0.01 Make / Model > Yuelian / STS-100kg Serial No. : 14143 (Indicator) Customer : MANILA CORDAGE COMPANY No. 116 Progress Ave., Carmelray Industrial Park I Canlubang, Calamba City, Laguna

Calibration Item : **Tensile Testing Machine**

Capacity : **100 lbf**

Measurement Range : **0 lbf to 90 lbf**

Resolution : **0.01 lbf**

Make / Model : **Yuelian / STS-100kg**

Serial No. : **J974953; 14143 (Indicator)**

Customer : **MANILA CORDAGE COMPANY**

**No. 116 Progress Ave., Carmelray Industrial Park I
Canlubang, Calamba City, Laguna**

MEASUREMENT RESULTS:

Applied Force	Indicated Force	Relative Expanded Uncertainty	Relative Measurement Error	Relative Repeatability Error
lbf	lbf	%	%	%
0.0	0.0	0.00	0.00	0.00
20.0	20.2	0.89	-1.12	1.37
40.0	40.4	0.16	-1.11	0.22
60.0	60.7	0.12	-1.18	0.16
80.0	81.0	0.39	-1.21	0.63
90.0	91.1	0.31	-1.19	0.49

UNCERTAINTY OF MEASUREMENT:

The uncertainty stated is the expanded uncertainty obtained by multiplying the standard

uncertainty by the coverage factor $k = 2$. It has been determined in accordance with the "JCGM 100:2008 Evaluation of measurement data- Guide to the Expression of Uncertainty in measurement". The value of the measurand lies within the assigned range of values with a probability of approximately 95%.

NATIONAL METROLOGY LABORATORY OF THE PHILIPPINES A Division of DOST Under 2

NATIONAL METROLOGY LABORATORY OF THE PHILIPPINES

Metrology Building, DOST Compound, Gen. Santos Avenue, TRO DOST Avenue
Bicutan, Taguig City, Metro Manila 1631 Philippines

Telefax: +63 2 8683 7750 to 69 local 2272 Website: www.nml.gov.ph

Website: www.nml.gov.ph

Page 1 of 2

A Division of



Under



01355

This report shall not be reproduced in any form, except in full, without written approval of the National Metrology Laboratory, Industrial Technology Development Institute.