

## CALIBRATION CERTIFICATE

No. 02-2025-FORC-0040

Date of Calibration : February 24, 2025  
Calibration Item : Axle Weighing Scale  
Capacity : 15 000 kgf  
Measurement Range : 0 kgf to 15 000 kgf  
Resolution : 50 kgf  
Make / Model : Intercomp;  
Serial No. : 1122YL23003  
Customer : SAVVICE CORPORATION  
3rd Floor BMWC Bldg., Cagayan Valley Road  
Sta. Rita, Guiginto, Bulacan

### MEASUREMENT RESULTS:

Applied Force	Indicated Force	Deviation (Indicated Force - Applied Force)	Relative Expanded Uncertainty	Relative Accuracy Error
kgf	kgf	kgf	%	%
0.00	0.00	0	0.00	0.00
3 000	2 850	-150	1.04	5.26
6 000	5 683	-317	0.81	5.57
9 000	8 500	-500	0.41	5.88
12 000	11 333	-667	0.46	5.88
15 000	14 133	-867	0.39	6.13

### UNCERTAINTY OF MEASUREMENT:

The uncertainty stated is the expanded uncertainty obtained by multiplying the standard uncertainty by the coverage factor  $k = 2$ . It has been determined in accordance with the "JCGM 100:2008 Evaluation of measurement data- Guide to the expression of uncertainty in measurement". The value of the measurand lies within the assigned range of values with a probability of approximately 95%.

# CALIBRATION CERTIFICATE

No. 02-2025-FORC-0040

Date of Calibration : **February 24, 2025**  
Calibration Item : **Axle Weighing Scale**  
Capacity : **15 000 kgf**  
Measurement Range : **0 kgf to 15 000 kgf**  
Resolution : **50 kgf**  
Make / Model : **Intercomp;**  
Serial No. : **1122YL23003**  
Customer : **SAVVICE CORPORATION**  
**3rd Floor BMWC Bldg., Cagayan Valley Road**  
**Sta. Rita, Guiginto, Bulacan**

## MEASUREMENT RESULTS:

Applied Force	Indicated Force	Deviation (Indicated Force - Applied Force)	Relative Expanded Uncertainty	Relative Accuracy Error
kgf	kgf	kgf	%	%
0.00	0.00	0	0.00	0.00
3 000	2 850	-150	1.04	5.26
6 000	5 683	-317	0.81	5.57
9 000	8 500	-500	0.41	5.88
12 000	11 333	-667	0.46	5.88
15 000	14 133	-867	0.39	6.13

## UNCERTAINTY OF MEASUREMENT:

The uncertainty stated is the expanded uncertainty obtained by multiplying the standard uncertainty by the coverage factor  $k = 2$ . It has been determined in accordance with the "JCGM 100:2008 Evaluation of measurement data- Guide to the expression of uncertainty in measurement". The value of the measurand lies within the assigned range of values with a probability of approximately 95%.