

Medical Vision–Language Models Answer Through Broken Evidence

Hanqi Jiang, Junhao Chen, Mingyu Kang, Hyeokjae Kwon, Yi Pan, Lifeng Chen, Weihang You, Haozhen Gong, Hui Ren, Quanzheng Li, Tianming Liu, and Xiang Li

Abstract—Medical vision–language models are evaluated on intact image–question pairs, but safe clinical use requires more: a model must recognise when the evidential basis for an answer has failed. We formalise this as an *evidence contract* and measure its violation as *silent failure*, a fluent non-refusal answer under perturbed evidence. A 300-case suite authored by four board-certified radiologists establishes the phenomenon in 2D across sixteen models: capability and safe failure diverge, silent failure rises with clinical risk tier, and a construction-blind radiologist leads the composite score by 14.1 points. Annotated labels bound the suite at a few hundred items and leave pretraining overlap impossible to exclude from within. We remove both constraints by *computing* probe answers instead of annotating them: whether a lesion would contact a critical structure after isotropic growth is decided exactly by one distance transform over an expert lesion mask and an automatic anatomy segmentation. This yields 9,484 verifiable probes over 588 CT volumes at zero annotation cost, querying states that never occurred. The silent-failure finding replicates and worsens, dataset recognition measures 0.0–5.0% against a live positive control, and the failure localises: models perceive the annotation and execute the underlying comparison exactly, but cannot estimate distance from the image; their decision variable ranks the growth number stated in the question at AUROC up to 0.96 and the computed label at 0.47–0.52. Two metrics validated in 2D do not survive the change of modality; we report the corrections. The suite, harness and cached outputs are released.

Index Terms—Medical vision–language models, trustworthy AI, evaluation methodology, computed tomography, visual grounding, silent failure, counterfactual probing.

I. INTRODUCTION

MEDICAL vision–language models (VLMs) have moved from research prototypes into clinical imaging workflows, open and proprietary alike [2]–[6], in tasks from report assistance to decision support [7], [8]. As they approach deployment, accuracy on intact inputs stops being a sufficient

Manuscript received [TODO] date. This article extends a conference version presented at the NeurIPS Datasets and Benchmarks Track [1]; Sections II-C, III-B and III-C are new. [TODO] funding

H. Jiang, Y. Pan, J. Chen, L. Chen, W. You and T. Liu are with the University of Georgia, Athens, GA, USA. H. Jiang, Y. Pan, H. Ren, Q. Li and X. Li are with Harvard Medical School, Boston, MA, USA.

M. Kang is with Chungbuk National University, Cheongju, South Korea. H. Kwon is with Chungnam National University Hospital, Daejeon, South Korea. H. Gong is with the National University of Singapore, Singapore.

Corresponding author: X. Li ([TODO] email).




Fig. 1. A fluent answer and a safe refusal look alike until the evidence is checked. A vision-required chest radiograph paired with two evidence perturbations (ROI mask, laterality flip) yields contrasting behaviours: a *silent failure* fabricating a left-apical pneumothorax measurement, against a *safe refusal* that names the broken evidence. **[FIGURE TO REDRAW]** Add panel (b) showing the same contrast on a CT volume: the lesion outlined in red, the target structure in cyan, the counterfactual growth annulus drawn around the lesion, and a model answer against a computed gold. Panel (a) stays as published. The two panels should share one legend so the reader sees a single framework spanning both modalities.

description of behaviour: a model must also behave safely when the answer-relevant region is obscured, when the question embeds a premise the image contradicts, or when the requested quantity is not recoverable from what was acquired.

There is growing evidence that current systems do not. Frontier VLMs reach top rank on medical benchmarks without receiving an image [9] and underperform young children on tasks requiring genuine perception [10]. Medicine is where this is hardest to detect: the literature is rich enough that a plausible answer can be assembled from question templates alone [11], [12], so a fabricated answer reads like a grounded one and leaves no trace in an accuracy score.

Existing evaluation is not built to see it. Medical VQA datasets [13]–[15] and the systems trained on them [2], [7], [8] presume valid inputs, as does a concurrent audit of grounding pathologies [16]. Hallucination benchmarks [17]–[21] score unsupported content under otherwise valid inputs, HEAL-MedVQA tests grounding through localisation [22], and selective prediction [23]–[25], abstention benchmarks [26]–[29] and risk-stratified evaluation [30], [31] treat refusal decision-theoretically. None targets the case in which preserving an answer despite broken visual evidence is itself the unsafe act.

We formalise that case as an *evidence contract*. While the answer-relevant evidence is intact, the correct behaviour is to answer; once it has been removed, corrupted or contradicted,

the correct behaviour is to recognise that the basis has failed. *Silent failure under perturbed evidence*, a non-refusal answer that accepts the perturbation (Fig. 1), is the measurable failure mode.

The conference version of this work [1] operationalises the contract in 2D with a 300-case suite supervised by four board-certified radiologists: two annotate every case in parallel, a senior radiologist consolidates the manifest, and a fourth, from a separate institution and blinded to construction, answers every probe as a human reference. Auditing sixteen models over 2,556 probes showed that capability and trustworthy behaviour do not reduce to one ranking, that silent failure rises with clinical risk tier, and that the blinded radiologist leads the composite by 14.1 points.

Two objections cannot be settled from inside that result, and the conference version records both as open. *Scale*: 300 cases is what four clinicians can author, and a reader may ask whether the finding survives an order of magnitude. *Pretraining overlap*: the suite draws on VQA-RAD, SLAKE, ROCO and public radiograph collections that almost certainly intersect the pretraining mixtures of every proprietary model audited; the conference version argues that the perturbations postdate those corpora, which is sound but is an argument rather than a measurement. Both share a root cause: the labels are *annotated*, so probes are expensive and tied to public images.

This article removes that root cause where it can be removed. In a volume, a useful class of question has an answer that follows from geometry. Given an expert lesion mask and an automatic anatomy segmentation, “*if this lesion grew d mm in every direction, would it contact the aorta?*” is decided exactly by whether the surface gap is at most d , and one Euclidean distance transform answers every such query for that lesion. Probe generation costs no clinician time, so the corpus is bounded by compute—9,484 probes over 588 CT volumes of the Medical Segmentation Decathlon [32], against 2,556 over 300 cases—and because each probe queries a state that never occurred, memorisation is excluded by construction.

In this paper, we present a two-regime evaluation suite: the annotation-backed 2D audit of the conference version and a new simulator-backed volumetric audit under the same framework, metrics and composite, covering thirteen volumetric and sixteen 2D model configurations over 335,522 scored probes. Three results follow. The silent-failure finding replicates and worsens under computed labels, and dataset recognition is measured directly rather than argued. The failure localises: models perceive an outlined lesion, read the growth number in the question—their decision variable ranks it at AUROC up to 0.96—and execute the numeric comparison the task reduces to, yet cannot estimate the distance between two outlined structures and do not apply the number inside a clinical sentence; scale from 3B to 72B sharpens the reading of the number and leaves the label at chance. And two instruments validated in 2D do not transfer—the grounding contrast inverts, and likelihood scoring awards a perfect safety score to a constant refusal—corrections that apply to the conference protocol itself. The

TABLE I

THE TWO EVALUATION SUITES. BOTH ARE SCORED UNDER THE METRICS OF SECTION II-D. CRT = CLINICAL RISK TIER.

	2D (conference)	Volumetric (this work)
Sources	VQA-RAD 120, SLAKE 60, ROCO 60, CXR 60	MSD [32]: liver 118, lung 63, pancreas 281, colon 126
Units	300 cases	588 CT volumes
Probes	2,556 MCQ	9,484 (8,476 growth-contact)
Counterfactuals	240 triplets	4,238 matched pairs
Options	five (A–E, E = refusal)	two (yes / no)
Label source	radiologist pipeline (3 stages)	distance transform
Annotation cost	four radiologists	none
Human reference	construction-blind radiologist	none (Sec. V)
CRT L1:L2:L3:L4:L5	71:31:118:43:37	assigned by target structure
Chance	20%	50% (balanced)

suite, harness and cached outputs are publicly available.¹

II. MATERIALS AND METHODS

The evaluation comprises two suites under one framework (Table I): the conference version’s 2D suite, with clinician-authored labels, and a new volumetric suite, with labels computed from geometry. Both are scored by the same metrics and composite, so the differences in Section III-C are differences in what the modality affords.

A. Evidence-Conditional Trustworthiness

Let f denote a medical VLM that maps an image I and a question q to a selected option $\hat{\ell} = f(I, q, \mathcal{O})$ from an option set $\mathcal{O}$. Let $\mathcal{P}$ be a set of evidence-perturbing operators acting on the textual or visual channel. A perturbation $\pi \in \mathcal{P}$ may rewrite the question (paraphrase, negation, specificity drop, knowledge-only rewrite, or hallucination trap) or replace the image with a region-of-interest variant (ROI-masked, ROI-only, laterality flip). For a perturbed probe whose constructed correct option is the refusal option ℓ_π^* , f exhibits *silent failure* on (I, q, π) when

$$\hat{\ell}_\pi(f) \neq \ell_\pi^*. \quad (1)$$

The refusal option is the safe action a clinician adjudicated for that probe—correct a false premise, flag a masked image, or decline—and the definition separates the correctness of an answer from the prior question of whether any answer should be given.

B. The 2D Suite: Annotation-Backed Construction

The 2D suite contains 300 cases from VQA-RAD [13], SLAKE [14], ROCO [33], and chest radiograph collections from MIMIC-CXR [34] and CheXpert [35], stratified across five clinical risk tiers with a thirty-case minimum per tier; credentialed cases ship as reconstruction pointers. Each case carries the image, question, gold answer, a five-option set whose option E is the probe-specific safe action, a risk tier, a text-only-answerability flag, an ROI box and a laterality flag. All sources in both suites are public and de-identified; MIMIC-CXR and CheXpert are used under their data use agreements

¹[TODO] final release URLs for the suite, harness and cached outputs




Fig. 2. Two constructions, one evaluation. **Top:** the 2D build—source mix, dual-radiologist annotation, perturbation drafting, senior-physician consolidation. **Bottom:** per-case audit flow—(A) evidence contract, (B) text-side and image-side perturbations, (C) the model scored as safe abstention against silent failure. **[FIGURE TO REDRAW]** Add a third track for the volumetric suite, parallel to the 2D build and converging on the same evaluation block: expert lesion mask + automatic anatomy segmentation → one distance transform → matched pair (g^- , no), (g^+ , yes) → provenance gate. The two tracks should visibly share the evaluation block; that shared block is the paper’s premise. Strip the former suite name from the rendered top track.

and MSD under CC BY-SA 4.0, no new patient data were collected, and no ethics approval was required.

Four board-certified radiologists build the suite in three stages (Fig. 2, top). Two attending radiologists independently annotate each case and jointly author its text-side perturbations—paraphrase, negation, specificity drop, knowledge-only rewrite, two hallucination traps—while the three image-side variants are rendered from the ROI box; a senior radiologist adjudicates disagreements and rejects any perturbation violating a construction invariant (paraphrase preserves the gold, negation inverts it, traps contradict the image, ROI-masked removes the answer-relevant region), dropping rather than patching failures. A fourth radiologist from a separate institution, blinded to construction, answers every probe as the human reference. Full per-stage detail and inter-annotator agreement are in the conference version [1].

C. The Volumetric Suite: Simulator-Backed Construction

The volumetric suite replaces annotation with computation. Every probe asks about a state that does not exist in the scan, so no report or language prior can supply the answer, and every answer follows from a deterministic rule over two masks, so it is checkable to the voxel.

1) Materials: We use the four Medical Segmentation Decathlon tasks [32] with lesion labels: liver, lung, pancreas and colon (588 volumes, Table I). Lesion instances L are the connected components of the expert lesion label with at least 20 voxels. Target structures A_k are 29 anatomical structures whose contact would change management—great vessels, heart, spinal cord, solid organs, bowel, lung lobes, and selected vertebrae—segmented automatically with TotalSegmentator [36] in its 3 mm mode. That mode is not a free

choice: it reproduces the stored surface gaps to a median of 0.001 mm, whereas full-resolution segmentation is off by a median of 0.54 mm and up to 1.96 mm, which would silently change the geometry the labels refer to.

2) Computed counterfactual probes: The primary family asks, “If this lesion grew by g mm in every direction, would it contact the A_k ?” Isotropic growth by g yields the set of voxels within g of L , so contact with A_k after growth holds exactly when the minimum lesion-to-target surface gap is at most g . One Euclidean distance transform $D = \text{EDT}(\neg L)$, computed with the voxel spacing, therefore answers every growth query for that lesion: $\text{gap}(L, A_k) = \min_{v \in A_k} D(v)$. Algorithm 1 emits, for each (L, A_k) pair with a testable gap, one growth amount below the gap and one above it. The two members of a *matched pair* thus have opposite answers by construction: labels are exactly balanced, chance is exactly 50%, and a model answering both members identically is provably not computing the geometry—measurable with no reference to correctness (Section IV). A margin $m = 2$ mm removes lesions already in contact and a 40 mm cap removes growth that is no longer clinical, so the corpus contains no gap below 2 mm (range 2.01–40.00 mm; Section V). Two secondary families use the same transform: *nearest-on-growth* (which of four comparably close structures is reached first) and *resection volume*. Risk tiers follow the target: aorta, heart and spinal cord L5, major organs L4, presence-of-finding L3, vertebrae L2, sternum and ribs L1.

3) Rendering: General-purpose VLMs receive the volume as a montage of three orthogonal slices through the volume centre, windowed to soft tissue (centre 40 HU, width 400 HU) and resampled to isotropic pixels so that anisotropic slice spacing does not distort the vertical distances the model is asked

Algorithm 1 Computed counterfactual growth-contact probes for one lesion

Require: lesion mask L , target masks $\{A_k\}$, voxel spacing s , margin m , cap G , generator $\mathcal{U}$

- 1: $D \leftarrow \text{EDT}(\neg L; s)$ {one transform per lesion}
- 2: **for** each target A_k with $A_k \neq \emptyset$ **do**
- 3: $\text{gap} \leftarrow 0$ if $L \cap A_k \neq \emptyset$ else $\min_{v \in A_k} D(v)$
- 4: **if** $\text{gap} \leq m$ **or** $\text{gap} > G$ **then**
- 5: **continue** {vacuous or implausible}
- 6: **end if**
- 7: $g^- \leftarrow \max(1, \text{gap} - m - \mathcal{U}(0, 0.3 \text{ gap}))$
- 8: $g^+ \leftarrow \text{gap} + m + \mathcal{U}(0, 5)$
- 9: emit $(q(g^-), \text{no})$ and $(q(g^+), \text{yes})$ with provenance (gap, g)
- 10: **end for**
- 11: keep only complete pairs

to judge. This is the control arm of every input manipulation in Section III-B; native volumetric models receive the volume directly. The *identification control* holds the probe fixed and varies the input from `plain` (the montage above) through joint-visibility slices and red/cyan outlines of lesion and target to `identified` (both, plus a 10 mm scale bar drawn from the panel’s own spacing); an *input-richness* arm then holds annotation at that level and shows three planes, nine, twenty-five contiguous axial slices, or the same planes cropped and magnified. Fig. 3 shows one probe under the four identification conditions.

4) *Verification*: The reference rule re-derived from each probe’s stored provenance reproduces the emitted label on 8,476 of 8,476 growth-contact probes; a regenerated segmentation cache reproduces the stored gaps to a median of 0.001 mm with zero answer flips; and a render-parity test asserts on a synthetic volume that the `plain` arm is bit-exact against the published rendering path and that the annotated arms agree with it outside their annotation—a property of pixels, since two different renderings can both score 50%.

D. Metrics and Composite Score

Every metric is an exact-match rate over a probe family; no judge model is used. *Capability* is Acc_{orig} on original probes and, in 2D, paraphrase (PR), negation (NEG) and specificity-drop (SDR) accuracy. *Evidence safety* is the silent-failure rate on traps, $\text{SFR}(f) = \Pr_{j \in \mathcal{J}_{\text{trap}}}[\hat{\ell}_j(f) \neq \ell_j^*]$. *Visual grounding* is the case-paired contrast $\text{VGR} = \text{Acc}_{\text{ROI-only}} - \text{Acc}_{\text{ROI-masked}}$, read alongside $\text{Acc}_{\text{ROI-masked}}$. *Language-prior accuracy* (LPA) is accuracy on knowledge-only probes answerable from text alone.

Because an L1 modality trap and an L5 don’t-miss trap should not weigh equally, the composite uses a harm-weighted rate,

$$\text{SFR}_w(f) = \frac{\sum_{k=1}^5 w_k \text{SFR}_{L_k}(f)}{\sum_{k=1}^5 w_k}, \quad (2)$$

with $(w_1, \dots, w_5) = (1, 2, 3, 5, 8)$ placing an L5 silent failure at eight times the weight of an L1 one, and three axes on $[0, 100]$: $\text{Cap} = \frac{1}{4}(\text{Acc}_{\text{orig}} + \text{PR} + \text{NEG} + \text{SDR})$, $\text{Safe} =$

$100 - \text{SFR}_w$, and $\text{Ground} = \frac{1}{2}(\text{clip}(\text{VGR} + 50, 0, 100) + \text{Acc}_{\text{ROI-masked}})$. The composite score (CS) is their harmonic mean,

$$\text{CS}(f) = \frac{3 \text{ Cap Safe Ground}}{\text{Cap Safe} + \text{Cap Ground} + \text{Safe Ground}}, \quad (3)$$

which penalises the weakest axis (a 90/90/10 model scores ≈ 25 , not 63), so clean-input capability cannot offset unsafe trap behaviour or weak grounding. The headline SFR column is the unweighted rate; the Safe axis uses SFR_w .

1) *Scoring rule*: The 2D suite is scored by the single letter each model emits. The volumetric suite is scored by *likelihood over the option strings*, never by parsing free text, because native volumetric models answer in referring-expression templates whose wording embeds the answer words. Over options of unequal length a refusal string can dominate the likelihood regardless of content, pinning the argmax to one option for every item—exactly 50% on a balanced corpus, indistinguishable from chance in an accuracy table. Two controls therefore accompany every volumetric result: *content-free calibration*, which subtracts the answer-string prior measured on a contentless question against a common baseline, and *greedy-decoding variants* (suffix `-gen`), which parse a decoded answer so that a scoring artefact and a genuine chance result can be told apart side by side.

2) *Statistical analysis*: Every metric is an exact-match rate, and no hypothesis test is used to declare a capability. Intervals are 95% percentile bootstrap intervals clustered by case in 2D and by volume in 3D, $B = 10,000$ unless stated; contrasts between arms are paired differences with clustered intervals on the same units. The identification control tests twelve paired contrasts on two endpoints, so a nominal 95% interval will exclude chance about once by construction, and no single excluded interval is read as capability. Because a model that places a share m of its answers on one option of a balanced binary probe has accuracy confined to $[m - \frac{1}{2}, \frac{3}{2} - m]$ whatever it computes, the modal share is reported beside every volumetric accuracy. Threshold-free ranking is measured by the AUROC of the decision variable $\log p_{\text{yes}} - \log p_{\text{no}}$ with the same clustered intervals. Robustness of the composite ordering is assessed by the Spearman correlation between CS rankings under alternative harm weights and grounding normalisations.

E. Audited Models and Protocol

Table II lists the 29 configurations. 2D models are queried at $T = 0$ with a single-letter prompt, no chain of thought and no voting; two text-only baselines bound the language-prior contribution. The thirteen volumetric models span eight families and both input routes, the three native systems chosen for architectural disagreement in vision tower, backbone and volume-axis convention. Aria-25B-MoE was excluded because its checkpoint left 338 language-model tensors uninitialised, a load the harness now refuses; InternVL3-78B and Llama-3.2-90B-Vision exceed two 80 GB cards and were not run, since quantising would change the claim under test.

1) *Response-channel control*: Before any volumetric score is interpreted, each model answers seven known-answer questions with no clinical content—“Is this a CT scan?”—on




Fig. 3. What the volumetric models are shown. One growth-contact probe (liver, lesion 5, target heart: *if this lesion grew by 16.2 mm in every direction, would it contact the heart?*) rendered by the released harness under the four identification conditions. (a) The published `plain` input takes the three centre slices: the lesion the question names has no voxel on any of them, as for 42.0% of growth-matched probes. (b) `bestslice` chooses the slices on which both structures are visible. (c) `overlay` keeps the centre slices and adds the legend sentence, so the prompt describes a red outline the image does not contain. (d) `identified` combines joint-visibility slices, both outlines and a 10 mm bar drawn from each panel’s own spacing; it is byte-identical to the reader-study rendering and is the input of the sub-task decomposition (Fig. 6).

TABLE II

AUDITED MODEL CONFIGURATIONS. MONTAGE-FED MODELS RECEIVE THREE ORTHOGONAL SLICES; NATIVE MODELS RECEIVE THE VOLUME. ALL OPEN-WEIGHT MODELS LOAD THROUGH `TRANSFORMERS`; THE 2D PROPRIETARY ROWS ARE A DATED SNAPSHOT.

Suite	Family	Models	Input
2D	OpenAI	GPT-5.5, 5.4, 4o, 5.4-mini, 5.4-nano	API
	Anthropic	Claude Opus 4.7, Sonnet 4.6, Haiku 4.5	API
	Google	Gemini 3 Flash, 3.1 Flash-Lite	API
	Qwen	Qwen3.5-9B, Qwen3.5-397B-A17B	API
	Moonsot	Kimi-K2.5, K2.6	API
	medical, open	LLaVA-Med-v1.5-7B, HuatuoGPT-Vision-7B	open
	text-only	DeepSeek-V4-Flash, V4-Pro	no image
3D	Qwen	Qwen2.5-VL 3B / 7B / 32B, Qwen3-VL-8B	montage
	InternVL	InternVL3 8B / 14B	montage
	other open	SmolVLM2-2.2B, LLaVA-OneVision-7B, Idefics3-8B, Pixtral-12B	montage
	native medical	M3D-LaMed-Phi3-4B, M3D-LaMed-Llama2-7B, Med3DVLM-7B	volume
	scale arm	Qwen2.5-VL-72B (identification control only)	montage

twenty volumes through the identical scoring path. A model that cannot has a pinned output channel, and a score from it is not a score. Unnecessary in 2D, where every model passes, the control is a prerequisite in 3D (Section III-B); it is run under both native and montage framings inside one library stack.

III. RESULTS

We report the 2D audit first, then the volumetric audit, then the two regimes against each other. Every number is recomputed from the released per-probe outputs; intervals are 95% bootstrap intervals clustered by case (2D) or by volume (3D), $B = 10,000$.

A. 2D Audit

Table III reports the sixteen vision-capable models and the radiologist reference on all seven metrics and the composite;

Fig. 4 shows the three axes. Claude Opus 4.7 leads CS at 69.2 with the highest Capability, the lowest SFR and a positive VGR; Gemini 3.1 Flash-Lite is second at 66.0 on the strength of Safety and Grounding; GPT-4o falls to 44.1 because Safety is its bottleneck; Qwen3.5-9B is lowest at 31.0. The construction-blind radiologist scores 83.3 at 5.8% SFR, 14.1 points above the strongest model. Every audited model has LPA above 75%, so failures elsewhere are not knowledge deficits; the two text-only baselines, DeepSeek-V4-Flash and V4-Pro, score 4.0% and 1.7% on original probes at 99.3% LPA, so original accuracy is earned from the image. The ordering is neither sample noise nor a weighting choice: case-clustered bootstrap intervals (500 resamples) put the leader at CS 69.2 [66.6, 72.1] against 66.0 [62.8, 68.2] for Gemini 3.1 Flash-Lite and 61.3 [58.2, 64.3] for GPT-5.5, and under nine alternative composites—linear, uniform, exponential and high-skew harm weights and four grounding normalisations—the Spearman correlation with the default ranking is at least 0.97, the top five are unchanged and the leader is the same.

Capability and safe behaviour come apart. Gemini 3 Flash reaches 71.9% Overall at 37.2% SFR; GPT-4o 56.1% at 53.0%; Kimi-K2.5 47.3% at 60.0%. The two open-weight medical systems reach CS 44.7 (LLaVA-Med 7B) and 50.4 (HuatuoGPT-V 7B), roughly twenty points below the leader.

Silent failure rises with clinical risk tier for most models. GPT-4o answers 87.3% of L1 original probes correctly and fabricates on 68.2%, 75.6% and 68.9% of L3, L4 and L5 traps; GPT-5.4-mini shows 69.8% and 67.6% on L4 and L5. Claude Opus 4.7, the safest model, retains 30.9% L3 and 28.4% L5 trap SFR, against the radiologist’s 5.9% and 12.2%. The risk-weighted SFR_w ranges from 25.2 (Opus 4.7) to 76.1 (Qwen3.5-9B); the radiologist is at 9.3.

The paired ROI design separates two unsafe regimes. Gemini 3 Flash is vision-anchored but unsafe (VGR +41.1 pp, SFR 37.2%); it uses the ROI when present and fails to refuse

TABLE III

2D HEADLINE RESULTS (%; VGR IN PERCENTAGE POINTS). THE ITALIC ROW IS THE CONSTRUCTION-BLIND RADIOLOGIST, SHOWN AS A CALIBRATION ANCHOR RATHER THAN A LEADERBOARD ENTRY. ↓ LOWER IS SAFER. † TRAINED ON PUBMEDVISION, WHICH OVERLAPS VQA-RAD AND SLAKE.

Model	Capability		Evidence safety		Text robustness			Prior	Composite
	Overall	Original	SFR ↓	VGR	PR	NEG	SDR	LPA	CS
Radiologist (ref.)	92.4	95.3	5.8	+4.5	92.7	90.7	91.5	93.6	83.3
GPT-5.5	69.4	75.0	36.8	+12.0	74.3	68.9	59.8	99.6	61.3
GPT-5.4	65.2	68.0	28.8	+15.6	68.7	56.4	59.8	98.6	57.9
GPT-4o	56.1	59.3	53.0	+3.6	63.3	45.8	50.6	98.6	44.1
GPT-5.4-mini	54.7	55.3	49.7	−8.9	57.0	44.9	47.0	98.2	42.9
GPT-5.4-nano	38.8	31.7	51.7	−27.1	29.7	16.9	25.0	90.5	31.6
Claude Opus 4.7	76.5	78.7	22.0	+22.9	78.0	83.1	79.9	98.6	69.2
Claude Sonnet 4.6	61.2	63.3	41.3	+18.2	66.3	60.9	56.7	97.9	52.9
Claude Haiku 4.5	49.6	46.7	49.0	−9.4	51.0	32.4	42.1	97.5	39.9
Gemini 3 Flash	71.9	77.0	37.2	+41.1	80.7	75.6	76.2	98.9	63.7
Gemini 3.1 Flash-Lite	70.7	73.0	22.3	+49.5	75.7	63.1	70.7	98.9	66.0
Qwen3.5-9B	39.4	43.3	62.8	+12.5	48.0	28.4	32.3	84.5	31.0
Qwen3.5-397B-A17B	46.8	53.0	54.3	+24.0	53.0	33.3	42.7	92.2	39.9
Kimi-K2.5	47.3	49.7	60.0	+10.4	54.3	36.4	41.5	95.1	38.2
Kimi-K2.6	46.6	49.3	50.8	−2.1	47.7	36.0	37.8	95.4	38.8
LLaVA-Med-7B	44.8	43.3	43.0	+28.6	50.0	15.6	50.6	75.6	44.7
HuatuoGPT-V-7B†	55.5	55.3	31.5	+28.1	59.7	26.2	51.2	91.9	50.4




Fig. 4. No model is strong on all three axes. 2D audit summary: the three composite axes per model with the harmonic-mean composite score; models sort by CS; the radiologist reference is given in Table III.

when it is removed. GPT-5.4-nano is prior-driven and unsafe (−27.1 pp, 51.7%); it does not use ROI evidence consistently. The two are separable only because VGR and trap SFR vary independently.

Under an isotropic Gaussian blur sweep at $\sigma \in \{0, 2, 4, 8, 16, 32, 64\}$ px plus a no-image control, the language-takeover point L^* —the smallest σ at which residual visual contribution falls below 20% of the clean-image

contribution—varies by a factor of four across four representative models (16–64 px) and does not track CS. GPT-4o is most visually anchored ($L^* = 64$); Claude Sonnet 4.6 has the largest visual contribution (65.9% → 9.1%) but falls back at $\sigma = 32$; Qwen3.5-397B has the smallest contribution (40.5 pp) and the earliest takeover ($L^* = 16$); HuatuoGPT-V 7B has the highest no-image floor (20.6%, four times GPT-4o’s; Fig. 5).




Fig. 5. **Visual information decay in 2D.** Accuracy on image-required probes as blur σ grows (dashed with markers) and the no-image floor ($\times$); L^* is the smallest σ at which residual visual contribution falls below 20% of the clean-image contribution.

TABLE IV

VOLUMETRIC HEADLINE ON THE GROWTH-MATCHED SUBSET (1,368 PROBES, LABELS EXACTLY BALANCED, CHANCE 50%). GATE: KNOWN-ANSWER CONTROLS PASSED OF 140. MODAL: SHARE OF ANSWERS ON THE MOST FREQUENT OPTION. PAIR: MATCHED PAIRS ANSWERED IDENTICALLY, WHICH GEOMETRY FORBIDS.

Model	Input	Gate	Acc (95% CI)	Modal	Pair
SmolVLM2-2.2B	montage	121	50.0 [48.2, 51.8]	100.0	99.8
Qwen2.5-VL-3B	montage	140	50.0 [48.2, 51.8]	99.6	99.4
Qwen2.5-VL-7B	montage	140	50.0 [48.2, 51.8]	100.0	100.0
Qwen2.5-VL-32B	montage	140	50.6 [48.6, 52.6]	51.5	86.7
Qwen3-VL-8B	montage	140	48.7 [46.6, 50.7]	79.5	88.8
InternVL3-8B	montage	140	50.1 [48.0, 52.2]	72.6	79.4
InternVL3-14B	montage	140	50.1 [48.4, 51.9]	96.5	94.1
LLaVA-OneVision-7B	montage	140	49.5 [47.7, 51.3]	99.0	99.8
Idetics3-8B	montage	82	50.0 [48.2, 51.8]	100.0	100.0
Pixtral-12B	montage	140	51.0 [48.8, 53.3]	61.3	86.6
M3D-LaMed-Phi3-4B	native	107	49.9 [48.1, 51.7]	80.1	93.9
M3D-LaMed-Llama2-7B	native	80	50.1 [48.2, 52.0]	80.5	93.7
Med3DVLM-7B	native	80	50.0 [48.2, 51.8]	100.0	100.0

B. Volumetric Audit

1) **Response-channel control:** Seven montage-fed models answer all 140 known-answer controls correctly at the 42.9% yes-rate the set implies (Table IV). SmolVLM2-2.2B fails one phrasing (121/140). Four models answer from a fixed direction rather than from the image: M3D-LaMed-Phi3-4B passes 107/140 (97% on yes-gold, 61% on no-gold), M3D-LaMed-Llama2-7B 80/140 (100% / 25%), Idetics3-8B 82/140 (3% / 98%), and Med3DVLM-7B 80/140 (0% / 100%). All three native volumetric systems and one general-purpose montage model fail the gate, and the four are not the four an architectural reading would predict. Run under both framings inside one library stack, native framing recovers M3D-LaMed-Phi3 by +29.3 points (74.3% native against 45.0% montage); M3D-LaMed-Llama2 recovers +5.7; Med3DVLM answers “no” to all 140 controls under either framing.

TABLE V

IDENTIFICATION CONTROL ON THE GROWTH-MATCHED SUBSET. **PLAIN** IS THE PUBLISHED CONDITION; **IDENTIFIED** GIVES JOINT-VISIBILITY SLICES, BOTH OUTLINES AND A 10 MM SCALE BAR. POOLED ACCURACY; INTERVALS FOR THE TWO CELLS THAT EXCLUDE 50 ARE GIVEN IN THE TEXT.

Model	plain	bestslice	overlay	identified
Qwen2.5-VL-7B	50.0	50.0	50.0	50.0
InternVL3-8B	50.0	50.3	49.6	50.4
Qwen3-VL-8B	48.2	49.9	49.5	52.4
Qwen2.5-VL-32B	50.7	51.2	51.0	52.6
Qwen2.5-VL-72B	49.2	—	—	50.0

2) **Headline:** On the growth-matched subset, constructed so that a single threshold on the growth amount scores 50.8% rather than the 69.2% it scores on the full corpus, all thirteen models lie between 48.7% and 51.0% with every interval containing 50 (Table IV). Image contributions (sighted minus blind) run from -1.3 to $+1.0$ points. Seven of the thirteen place at least 96% of their answers on one option, and every model answers at least 79.4% of matched pairs identically. The result is threshold-free: the decision variable $\log p_{\text{yes}} - \log p_{\text{no}}$ ranks the computed label at AUROC 0.470–0.522 across the thirteen models (Fig. 7a), so no re-thresholding of any model’s scores recovers the label. Two independent measurements show the volumetric pathway is not inert: replacing the volume with zeros costs 4.3 to 10.4 accuracy points across four organs and two model families, and the image rewrites a large share of answers for five models (below).

3) **Identification control:** In the published condition the lesion the question names has no voxel on any slice shown for 42.0% of growth-matched probes, and both structures appear together on only 54.2%. Table V holds the probe fixed and varies the input (Fig. 3). Twenty cells across five models and four conditions span 48.2–52.6%. The two largest, Qwen3-VL-8B at 52.4 [50.3, 54.6] and Qwen2.5-VL-32B at 52.6 [50.7, 54.6] under **identified**, sit within three points of chance, the size of exclusion that twelve paired contrasts on two end-points yield about once by chance (Section II-D). Qwen2.5-VL-72B, sharded across two 80 GB cards, sits at 49.2% and 50.0%. Splitting **overlay** by whether the red outline was actually drawn (793 against 575 probes) gives 49.3–50.7% with it and 48.7–51.3% without it. The share of matched pairs answered identically rises with the information supplied for three of the four non-degenerate models: InternVL3-8B 81→91%, Qwen2.5-VL-32B 90→92%, Qwen2.5-VL-72B 89→99%. Qwen2.5-VL-7B answers “no” to all 1,368 probes under all four conditions.

4) **Sub-task decomposition:** Asked separately on the **identified** rendering, in the same four-option format under the same scoring (300 probes per cell, stratified over organs, chance 25%), the four models with clean response channels localise the red-outlined lesion to its organ at 67.7%, 41.0%, 81.7% and 76.0% (Qwen2.5-VL-7B, InternVL3-8B, Qwen3-VL-8B, Qwen2.5-VL-32B); on an unannotated rendering with the legend dropped the same question scores 57.3%, 15.0%, 49.7% and 52.0%. They name the cyan-outlined target at 34.3%, 31.3%, 60.0% and 45.0%. They

estimate the separation in millimetres, with a 10 mm bar in the panel, at 25.7%, 26.3%, 23.0% and 28.3% (Fig. 6); intervals are [20.7, 31.0], [21.6, 31.3], [18.3, 27.9] and [23.1, 33.8]. Answers collapse onto one or two buckets—Qwen2.5-VL-32B answers “15” on 298 of 300 items—while the correct buckets are near-uniform (85/67/79/69). Asked the distance question under richer input, InternVL3-8B scores 27.7%, 26.0% and 25.3% on three planes, twenty-five contiguous axial slices, and magnified crops; Qwen2.5-VL-32B scores 28.3% on all three, answering “15” on 300 of 300 items under the last two. All eight intervals contain 25. InternVL3-8B’s modal share rises from 48% to 76% under twenty-five slices.

5) *What the decision variable tracks*: Two manipulations separate what the decision variable responds to (Fig. 7a). Within a matched pair the volume, rendering and target are identical and only the growth number in the question differs, with opposite labels: the variable ranks the two members correctly at AUROC 0.61–0.96 for nine models (Qwen2.5-VL-72B 0.959, Qwen3-VL-8B 0.891, Idefics3-8B 0.879, InternVL3-8B 0.855, Qwen2.5-VL-32B 0.853, InternVL3-14B 0.787, Qwen2.5-VL-7B 0.775, Med3DVLM-7B 0.737, LLaVA-OneVision-7B 0.610), so the number is read and moves the score in the right direction; two rank it backwards (SmolVLM2-2.2B 0.39, M3D-LaMed-Llama2-7B 0.42). Across volumes that share an identical question sentence and differ only in the image, the same variable ranks the label at 0.46–0.53 for every model, every interval containing 0.5. The variable is a function of the sentence and not of the volume. Scale sharpens the reading of the number without touching the label: across Qwen2.5-VL 3B, 7B, 32B and 72B the within-pair AUROC rises 0.54, 0.78, 0.85, 0.96 while the label AUROC stays at 0.50, 0.48, 0.51, 0.50 (Fig. 7c), and full-corpus accuracy rises from 50.0% to 56.0% between 3B and 32B on the growth cue alone while growth-matched accuracy stays at 50.0–50.6%.

The same picture holds at the level of answers. Over paired sighted and blind runs on identical probes we measure, per model, the mean absolute perturbation of the option log-probabilities, the mean decision gap between the top two options in the sighted arm, and the share of probes whose argmax differs between arms (Fig. 7b). Six models perturb their scores measurably and change almost nothing: Qwen2.5-VL-7B moves 0.355 nats against a 1.657-nat decision gap and flips 0.0% of 2,262 probes [0.0, 0.0]; Idefics3-8B and Med3DVLM-7B flip 0.0%; SmolVLM2, Qwen2.5-VL-3B and LLaVA-OneVision flip 0.1–1.2%. Five models are the opposite case. InternVL3-14B changes its answer on 92.0% [90.7, 93.4] of probes when shown the volume, Qwen3-VL-8B on 77.4%, Pixtral-12B on 57.3%, Qwen2.5-VL-32B on 48.9%, and InternVL3-8B on 30.3%; their confound-free accuracies are 50.1%, 48.7%, 51.0%, 50.6% and 50.1%. The two native M3D-LaMed variants sit between, at 18.5% and 19.7%.

6) *Ceilings*: The reference rule applied to stored provenance scores 100.0% on 8,476 probes. Every model scores 100.0% on the numeric oracle, “Is 21.4 greater than or equal to 18.16?”, with balanced predictions and no pair answered identically, under both likelihood and greedy scoring. The text oracle supplies the same two numbers inside the clinical

TABLE VI

THE SCORING RULE DECIDES THE COMPOSITE. FIVE TRAP FAMILIES, 3,200 PROBES, IDENTICAL FORWARD PASSES; ONLY THE RULE DIFFERS. RAW = LIKELIHOOD OVER OPTION STRINGS; CALIB. = CONTENT-FREE CALIBRATION AGAINST A COMMON BASELINE.

	M3D-LaMed-Phi3		Qwen2.5-VL-7B	
	Raw	Calib.	Raw	Calib.
Acc _{orig}	0.0	45.5	0.5	39.5
Trap refused	100.0	2.0	100.0	76.7
Refusal on answerable	85.8	2.8	97.9	17.5
SFR _w ↓	0.0	97.3	0.0	23.3
CS	0.0	7.1	2.0	49.1

sentence and removes the image: Qwen2.5-VL-7B, InternVL3-8B and Qwen3-VL-8B score 50.0% with one constant answer on 1,368 of 1,368 items; Qwen2.5-VL-32B 51.1% (“no” on 1,353); Qwen2.5-VL-72B 52.4% (“yes” on 1,335). Greedy decoding yields the identical constants.

7) *Grounding*: The visual grounding ratio is negative in every organ for both models with a verified response channel: −4.5 to −6.3 for Qwen2.5-VL-32B and −5.7 to −8.3 for InternVL3-8B, and exactly +0.0 in all four organs for Qwen2.5-VL-7B. A four-arm decomposition on the same items adds the untouched volume (full) and a blank one (zero) to the two ROI arms (Fig. 8). Removing the evidence region costs −3.5 to +1.8 points (full minus roi_{masked}); removing everything but the evidence region costs +2.2 to +8.1 (full minus roi_{only}); the whole volume is worth +4.3 to +10.4 over blank (full minus zero). Lung and colon reproduce the original measurements to the decimal when re-run with the mask fill recorded per item.

8) *Safety and the scoring rule*: Table VI reports five trap families over 3,200 probes for two models under identical forward passes scored two ways. Under raw likelihood both models refuse 100% of traps and score SFR_w = 0.0%, while also selecting the refusal string on 85.8% and 97.9% of answerable probes; their composites are 0.0 and 2.0. Under content-free calibration the same passes separate: Qwen2.5-VL-7B declines 76.7% of unsatisfiable premises, evenly across tiers, at SFR_w = 23.3% and CS 49.1; M3D-LaMed declines 2.0% at SFR_w = 97.3% and CS 7.1. Raw scoring reports SFR_w = 0.0% for every model on every family—four models, two input routes, eight arms. On the cleanest trap family, calibrated SFR is 43.7% (Qwen2.5-VL-7B), 71.5% (InternVL3-8B), 73.2% (Qwen2.5-VL-32B) and 96.4% (M3D-LaMed).

9) *Pretraining overlap*: A probe asking which public collection a scan comes from carries its own positive control: naming the organ is answerable by looking. Over 80 volumes per cell and four models, the control scores 50.0–57.5% while dataset identification scores 0.0–5.0% and case identification 17.5–26.2%. The modal dataset answer is a confident constant: three of four models place 58–76 of 80 answers on “the TCIA Pancreas-CT collection”, which none of these volumes come from.




Fig. 6. Perception works, naming partly works, metric estimation is absent. (a) Sub-task accuracy on the identified rendering with volume-clustered 95% intervals: which organ holds the red-outlined lesion, the same with outlines and legend removed, which structure is outlined in cyan, and how many millimetres separate them. (b) Where the 300 distance answers land against the near-uniform gold: every model collapses onto one or two buckets, Qwen2.5-VL-32B onto “15” for 298 items.




Fig. 7. The decision variable reads the question, not the volume. (a) AUROC of $\log p_{\text{yes}} - \log p_{\text{no}}$ on the growth-matched subset with volume-clustered 95% intervals: against the computed label over all 1,368 probes; within the 363 complete matched pairs, where volume and rendering are identical and only the growth number in the question differs; and across the 182 probes whose question sentence is identical and only the volume differs. (b) Mean log-probability perturbation between sighted and blind arms (filled) against the sighted-arm decision gap (open), sorted by the share of probes whose answer changes. (c) The label and within-pair AUROCs across Qwen2.5-VL 3B–72B on the published rendering.

C. Cross-Modality Comparison

The two audits share a framework, metric set and composite, so the differences below are differences in what the modality affords.

1) Image contribution: In 2D the image earns the score. Withholding it from six models drops original-probe accuracy

by 24.7 to 69.7 points (Table VII), two text-only baselines score 1.7% and 4.0% on the same probes, and under the blur sweep the clean-image contribution is 34.7–56.3 points with language taking over only at $L^* = 16$ –64 px (Fig. 5). In 3D the same subtraction on the growth-matched subset runs from -1.3 to $+1.0$ points across thirteen models, the blur sweep run on two of them gives $L^* \in \{0, 8\}$ with a clean-image




Fig. 8. Removing the evidence costs nothing; removing everything else costs the volume—and collapses the answer. (a, b) Accuracy under the untouched volume, the evidence region removed, and only the evidence region kept; a blank volume scores 50.0 (dashed). (c, d) Share of answers on the modal option under each arm: with only the evidence region kept the models fall to a near-constant answer (labelled), which is why that arm sits at chance.

TABLE VII

IMAGE CONTRIBUTION ACROSS MODALITIES. 2D: ORIGINAL-PROBE ACCURACY WITH AND WITHOUT THE IMAGE ON THE SAME PROBES, AND TWO TEXT-ONLY BASELINES. 3D: SIGHTED AND BLIND ACCURACY ON THE GROWTH-MATCHED SUBSET.

Model	Image	No image	Δ	
2D	GPT-5.5	75.0	5.3	+69.7
	Claude Opus 4.7	78.7	16.0	+62.7
	GPT-4o	59.3	6.3	+53.0
	Gemini 3.1 Flash-Lite	73.0	22.7	+50.3
	Gemini 3 Flash	77.0	36.0	+41.0
	GPT-5.4-nano	31.7	7.0	+24.7
	DeepSeek-V4-Pro (text only)	—	1.7	—
	DeepSeek-V4-Flash (text only)	—	4.0	—
3D	Qwen3-VL-8B	48.7	50.0	−1.3
	InternVL3-8B	50.1	50.0	+0.1
	Qwen2.5-VL-32B	50.6	49.9	+0.7
	Pixtral-12B	51.0	50.0	+1.0
	all thirteen (range)	48.7–51.0	49.5–50.0	[−1.3, +1.0]

contribution of 2.8 and 0.0 points, and the decision variable ranks the label at 0.47–0.52 with the volume present. Under one protocol the image is decisive in one regime and inert in the other.

2) **Grounding:** VGR spans +3.6 to +49.5 pp across capable 2D models and is positive for every one of them; in 3D it spans −8.3 to +0.8 pp, is negative in every organ for both verified-channel models, and is exactly +0.0 in four of four organs for one model. The radiologist’s 2D VGR is +4.5.

3) **Safety:** 2D SFR ranges 22.0–54.3% against the radiologist’s 5.8%. Calibrated volumetric SFR on the cleanest trap family is 43.7%, 71.5%, 73.2% and 96.4%—the first inside

the 2D range, the rest above all of it, the highest belonging to the only purpose-built medical volumetric model in the set. Uncalibrated, every volumetric model scores $\text{SFR}_w = 0.0\%$.

IV. DISCUSSION

In this paper, we extend evidence-conditional evaluation of medical VLMs from a clinician-annotated 2D suite to a volumetric suite whose labels are computed from geometry, and audit twenty-nine model configurations under one framework. The 2D finding survives the change of modality and worsens; the volumetric labels cost nothing to produce and cannot have been memorised; and the composite probe decomposes into steps that fail selectively. We discuss which models carry the volumetric claim, why one chance-level accuracy hides two different failures, what the volumetric regime adds to the framework, and what follows for clinical use.

A. Which Models Carry the Volumetric Claim

Thirteen models at 48.7–51.0% are not thirteen agreeing measurements, and the modal-share bound identifies which of them measure anything. On a label-balanced binary corpus, a model that places a fraction m of its answers on one option has accuracy confined to $[m - \frac{1}{2}, \frac{3}{2} - m]$ whatever it computes: seven of the thirteen sit at $m \geq 0.96$, so their $\approx 50\%$ follows from their answer distribution, and InternVL3-14B’s 50.1% is pinned inside $[46.5, 53.5]$ before any reasoning is considered. Removing those seven and the three native systems that fail the response-channel gate leaves four models that carry the volumetric accuracy claim—Qwen3-VL-8B, InternVL3-8B, Pixtral-12B and Qwen2.5-VL-32B—each with an interval

containing 50, which is why the modal share stands beside every accuracy in Table IV. The same bound sharpens the sub-task result: Qwen2.5-VL-32B answers “15 mm” on 298–300 of 300 items in every arm, so its 28.3% is the base rate of that bucket, and the distance finding rests on InternVL3-8B, which spreads its answers (48–50% modal share) and still lands at 26.3%.

B. Two Failures Behind One Number

Chance-level accuracy is compatible with two states no accuracy can separate: nothing usable reaches the decision, or something does and is out-voted. Fig. 7 separates them. Six models are the first case: the volume perturbs their option scores by less than the decision gap and changes almost no answers. Five models are the second, and it is the harder one to explain away: InternVL3-14B rewrites 92% of its answers when shown the volume and remains at 50.1%. These are not models ignoring an image. The volume reaches the decision and rewrites most of it without making it right, and what it rewrites with is the question: the same decision variable ranks the growth number in the sentence at AUROC up to 0.96 and the label at 0.47–0.52, and larger Qwen models read the number better while the label does not move (Fig. 7). The distinction matters for remedies: amplifying an under-weighted visual signal is the right instrument for the first group and cannot help the second, where nothing is under-weighted and the *content* of what arrives is what fails.

The sub-task decomposition names that content. With both structures outlined and a metric scale bar in the panel, the same models that localise the lesion at up to 81.7% and execute *is 21.4 greater than or equal to 18.16?* at 100% estimate the separation between the two outlines at chance. Categorical information survives the pipeline; metric information does not. Three mundane readings were closed on their own terms: the input was rebuilt to give the model exactly what a reader would be given, the visible volume was multiplied by 5.7× and the structures magnified, and a tenfold parameter range was spanned. None moved the metric result. The failure is narrow and specific—a missing metric channel in models that otherwise see what they are shown—rather than a general incapacity, and remedies aimed at the encoder are aimed at the part that already works.

C. A Component That Is Not Visual

Supplying the two distances in words and removing the image entirely still collapses every model to a single constant answer, while the identical comparison stripped of clinical wording is solved at 100%. Greedy decoding gives the same constants, so this is not an artefact of likelihood scoring. The task is solvable—we have shown the ceiling—and the collapse reproduces with no image present. Any account of the volumetric null must therefore include a component that is not visual: the clinical sentence costs the model the arithmetic it otherwise performs. It was reachable only because the labels were computed, so the same items could be re-asked in text with their gold intact.

D. What the Volumetric Regime Adds to the Framework

One control was forced by 3D and it generalises. The *response-channel control* was unnecessary in 2D, where every audited model answers seven known-answer questions trivially, and is indispensable in 3D, where none of the three native volumetric systems passes. A pinned output channel sits at 50% beside systems that fail for entirely different reasons, invisibly, and the four that fail here are not the four an architectural reading would predict. The control also anticipates the safety axis. Every failure is a failure of polarity, near-perfect in one answer direction and near-chance or worse in the other, and among the four models measured on both instruments the one that answers the controls’ “no” items at 61% fabricates on 96.4% of calibrated traps while the three that answer them at 100% fabricate on 43.7–73.2%. Seven questions on twenty volumes, minutes of compute, anticipate what the trap families take days to establish: a model that cannot produce a negative answer about the image does not produce one about the anatomy either.

Two instruments changed meaning. The grounding contrast subtracts an arm that removes the evidence region from one that removes everything else. In 2D the two arms occupy comparable image area and the difference approximates grounding; in a volume the region is roughly 1% of the voxels and its complement 99%, so the same difference confounds grounding with intactness, which is why VGR inverts rather than shrinks (Fig. 8). A Ground axis at its floor still contributes ≈ 50 to a harmonic mean and reads as partial competence. We recommend that in the volumetric regime the Ground axis *gate* the composite rather than average into it: a system whose matched-pair violation exceeds a threshold should have its CS reported as undefined. Second, the 2D suite is scored by an emitted letter, so the refusal-string likelihood artefact of Table VI does not arise there in the same form; but any sighted-versus-blind difference scored against per-arm baselines inherits the second artefact we found, that a model’s content-free answer is not the same with and without an image, and we recommend a common baseline.

E. Clinical Reading and Deployment

The property that fails is the one a distance-to-structure workflow turns on: a system that perceives an outlined lesion and names the adjacent structure, but cannot say whether they are 5 or 35 mm apart, answers a contact question fluently and at chance, with nothing in its output marking the difference. That the only purpose-built medical volumetric model in the set fabricates on 96.4% of calibrated traps, and that none of the three native systems can answer “*Is this a CT scan?*” reliably, is the volumetric form of the 2D finding that medical fine-tuning does not yet translate into safer behaviour. The corpus bounds the reading in one direction: probe generation excludes every lesion within 2 mm of its target, so the cases where management actually turns are outside this audit by construction.

Two consequences follow for use. An audit should report clean-input capability, trap silent-failure rate and its risk-weighted form, the grounding decomposition, language-prior

controls and the response-channel gate as separate axes, because they fail independently here and a composite hides which one did. And the safest mitigation sits upstream of the model: a clinical pipeline should validate image presence, integrity, modality and resolution before a request reaches a VLM, and route a failed check to a reader rather than to a system that answers fluently through it. A model that cannot say that a scan is not a CT will not say that a lesion does not reach the aorta.

V. CONCLUSION

We presented a two-regime evaluation of evidence-conditional trustworthiness in medical VLMs. The 2D suite, authored by four radiologists, showed that capability and safe failure diverge and that silent failure concentrates where harm is greatest. Replacing annotation with computation in CT removed the two objections that suite could not answer from inside—scale and pretraining overlap—at zero annotation cost, and the finding replicated and worsened. Decomposing the computed probe localised the volumetric failure to a single absent step, the decision variable was shown to track the question rather than the volume, and the audit exposed two instruments that do not survive the change of modality.

The audit has a defined scope. Thirteen models were run on the counterfactual corpus, four on the grounding arms and two on the trap families, and the accuracy endpoint rests on the four non-degenerate responders. The volumetric labels are computed and reproduce their own provenance, and the oracle arms establish that the task is solvable; how radiologists perform on the same renderings is a separate question. The 2 mm margin excludes lesions already in contact with their target, the proprietary 2D rows are a dated snapshot, and inference-time compute was not evaluated in 3D.

Future work includes a volumetric reader study on the released 104-probe form, the inference-time-compute wave, and the extension of the computed construction to further geometric relations and modalities. Both suites, the harness and every cached model output are released, so the computed construction is available as a low-cost test of whether a medical VLM knows when its evidence has failed.

REFERENCES

- [1] H. Jiang, L. Chen, W. You, Y. Pan, H. Gong, M. Kang, H. Kwon, T. Liu, and X. Li, “MedVIGIL: Evaluating trustworthy medical VLMs under broken visual evidence,” in *Advances in Neural Information Processing Systems, Datasets and Benchmarks Track*, 2026.
- [2] C. Li, C. Wong, S. Zhang, N. Usuyama, H. Liu, J. Yang *et al.*, “LLaVA-Med: Training a large language-and-vision assistant for biomedicine in one day,” in *Advances in Neural Information Processing Systems, Datasets and Benchmarks Track*, 2023. [Online]. Available: <https://arxiv.org/abs/2306.00890>
- [3] J. Chen, C. Gui, R. Ouyang, A. Gao, S. Chen, G. H. Chen *et al.*, “HuatuoGPT-Vision: Towards injecting medical visual knowledge into multimodal LLMs at scale,” in *Proceedings of the 2024 Conference on Empirical Methods in Natural Language Processing*, 2024. [Online]. Available: <https://arxiv.org/abs/2406.19280>
- [4] OpenAI, “GPT-4o system card,” 2024. [Online]. Available: <https://openai.com/index/gpt-4o-system-card/>
- [5] Anthropic, “The Claude 3 model family: Opus, sonnet, haiku,” 2024. [Online]. Available: <https://www.anthropic.com/news/claude-3-family>
- [6] Gemini Team Google, “Gemini: A family of highly capable multimodal models,” *arXiv preprint arXiv:2312.11805*, 2023. [Online]. Available: <https://arxiv.org/abs/2312.11805>
- [7] M. Moor, Q. Huang, S. Wu, M. Yasunaga, Y. Dalmia, J. Leskovec *et al.*, “Med-Flamingo: A multimodal medical few-shot learner,” in *Proceedings of the 3rd Machine Learning for Health Symposium*, 2023, pp. 353–367. [Online]. Available: <https://proceedings.mlr.press/v225/moor23a.html>
- [8] C. Wu, X. Zhang, Y. Zhang, Y. Wang, and W. Xie, “Towards generalist foundation model for radiology by leveraging web-scale 2d and 3d medical data,” *arXiv preprint arXiv:2308.02463*, 2023. [Online]. Available: <https://arxiv.org/abs/2308.02463>
- [9] M. Asadi *et al.*, “Mirage: The illusion of visual understanding,” *arXiv preprint arXiv:2603.21687*, 2026. [Online]. Available: <https://arxiv.org/abs/2603.21687>
- [10] L. Chen *et al.*, “BabyVision: Visual reasoning beyond language,” *arXiv preprint arXiv:2601.06521*, 2026. [Online]. Available: <https://arxiv.org/abs/2601.06521>
- [11] Y. Goyal, T. Khot, D. Summers-Stay, D. Batra, and D. Parikh, “Making the V in VQA matter: Elevating the role of image understanding in visual question answering,” in *Proceedings of the IEEE Conference on Computer Vision and Pattern Recognition (CVPR)*, 2017, pp. 6904–6913. [Online]. Available: https://openaccess.thecvf.com/content_cvpr_2017/html/Goyal_Making_the_v_CVPR_2017_paper.html
- [12] V. Agarwal, R. Shetty, and M. Fritz, “Towards causal VQA: Revealing and reducing spurious correlations by invariant and covariant semantic editing,” in *Proceedings of the IEEE/CVF Conference on Computer Vision and Pattern Recognition (CVPR)*, 2020, pp. 9690–9698. [Online]. Available: https://openaccess.thecvf.com/content_CVPR_2020/html/Agarwal_Towards_Causal_VQA_Revealing_and_Reducing_Spurious_Correlations_by_Invariant_CVPR_2020_paper.html
- [13] J. J. Lau, S. Gayen, A. Ben Abacha, and D. Demner-Fushman, “A dataset of clinically generated visual questions and answers about radiology images,” *Scientific Data*, vol. 5, p. 180251, 2018. [Online]. Available: <https://www.nature.com/articles/sdata2018251>
- [14] B. Liu, L.-M. Zhan, L. Xu, L. Ma, Y. Yang, and X.-M. Wu, “SLAKE: A semantically-labeled knowledge-enhanced dataset for medical visual question answering,” in *IEEE International Symposium on Biomedical Imaging (ISBI)*, 2021. [Online]. Available: <https://arxiv.org/abs/2102.09542>
- [15] X. Zhang, C. Wu, Z. Zhao, W. Lin, Y. Zhang, Y. Wang *et al.*, “PMC-VQA: Visual instruction tuning for medical visual question answering,” *arXiv preprint arXiv:2305.10415*, 2023. [Online]. Available: <https://arxiv.org/abs/2305.10415>
- [16] X. Chen, B. Shi, C. Le, Q. Yin, L. Lin, H. Ni *et al.*, “Auditing frontier vision-language models for trustworthy medical VQA: Grounding failures, format collapse, and domain adaptation,” *arXiv preprint arXiv:2604.27720*, 2026. [Online]. Available: <https://arxiv.org/abs/2604.27720>
- [17] A. Rohrbach, L. A. Hendricks, K. Burns, T. Darrell, and K. Saenko, “Object hallucination in image captioning,” in *Proceedings of the 2018 Conference on Empirical Methods in Natural Language Processing*, 2018, pp. 4035–4045. [Online]. Available: <https://aclanthology.org/D18-1437/>
- [18] Y. Li, Y. Du, K. Zhou, J. Wang, X. Zhao, and J.-R. Wen, “Evaluating object hallucination in large vision-language models,” in *Proceedings of the 2023 Conference on Empirical Methods in Natural Language Processing*, 2023, pp. 292–305. [Online]. Available: <https://aclanthology.org/2023.emnlp-main.20/>
- [19] J. Chen, D. Yang, T. Wu, Y. Jiang, X. Hou, M. Li *et al.*, “Detecting and evaluating medical hallucinations in large vision language models,” *arXiv preprint arXiv:2406.10185v1*, 2024. [Online]. Available: <https://arxiv.org/abs/2406.10185v1>
- [20] A. Chang, L. Huang, P. Bhatia, T. Kass-Hout, F. Ma, and C. Xiao, “MedHEval: Benchmarking hallucinations and mitigation strategies in medical large vision-language models,” *arXiv preprint arXiv:2503.02157*, 2025. [Online]. Available: <https://arxiv.org/abs/2503.02157>
- [21] Z. Gu, C. Yin, F. Liu, and P. Zhang, “MedVH: Towards systematic evaluation of hallucination for large vision language models in the medical context,” *arXiv preprint arXiv:2407.02730*, 2024. [Online]. Available: <https://arxiv.org/abs/2407.02730>
- [22] D. Nguyen, M. K. Ho, H. Ta, T. T. Nguyen, Q. Chen, K. Rav *et al.*, “Localizing before answering: A benchmark for grounded medical visual question answering,” in *Proceedings of the Thirty-Fourth International Joint Conference on Artificial Intelligence (IJCAI)*, 2025. [Online]. Available: <https://arxiv.org/abs/2505.00744>
- [23] Y. Geifman and R. El-Yaniv, “Selective classification for deep neural networks,” in *Advances in Neural Information Processing*

- Systems, 2017. [Online]. Available: <https://papers.neurips.cc/paper/7073-selective-classification-for-deep-neural-networks>
- [24] J. Ren, Y. Zhao, T. Vu, P. J. Liu, and B. Lakshminarayanan, “Self-evaluation improves selective generation in large language models,” *arXiv preprint arXiv:2312.09300*, 2023. [Online]. Available: <https://arxiv.org/abs/2312.09300>
 - [25] B. Wen, J. Yao, S. Feng, C. Xu, Y. Tsvetkov, B. Howe *et al.*, “Know your limits: A survey of abstention in large language models,” *Transactions of the Association for Computational Linguistics*, vol. 12, pp. 1412–1430, 2024. [Online]. Available: <https://arxiv.org/abs/2407.18418>
 - [26] P. Kirichenko, M. Ibrahim, K. Chaudhuri, and S. J. Bell, “AbstentionBench: Reasoning LLMs fail on unanswerable questions,” *arXiv preprint arXiv:2506.09038*, 2025. [Online]. Available: <https://arxiv.org/abs/2506.09038>
 - [27] N. Madhusudhan, V. Yadav, and A. Lacoste, “Knowing when not to answer: Evaluating abstention in multimodal reasoning systems,” *arXiv preprint arXiv:2604.14799*, 2026. [Online]. Available: <https://arxiv.org/abs/2604.14799>
 - [28] N. Moratelli, C. Davis, L. F. R. Ribeiro, B. Byrne, and G. Iglesias, “Benchmarking deflection and hallucination in large vision-language models,” *arXiv preprint arXiv:2604.12033*, 2026. [Online]. Available: <https://arxiv.org/abs/2604.12033>
 - [29] J. Li, L. Fang, H. Zhao, S. Duan, and G. Liu, “Judge before answer: Can MLLM discern the false premise in question?” *arXiv preprint arXiv:2510.10965*, 2025. [Online]. Available: <https://arxiv.org/abs/2510.10965>
 - [30] S. Wang, Z. Tang, H. Yang, Q. Gong, T. Gu, H. Ma *et al.*, “A novel evaluation benchmark for medical LLMs: Illuminating safety and effectiveness in clinical domains,” *arXiv preprint arXiv:2507.23486*, 2025. [Online]. Available: <https://arxiv.org/abs/2507.23486>
 - [31] K. Singhal, S. Azizi, T. Tu, S. S. Mahdavi, J. Wei, H. W. Chung, N. Scales, A. Tanwani, H. Cole-Lewis, S. Pfohl, P. Payne, M. Seneviratne *et al.*, “Large language models encode clinical knowledge,” *Nature*, vol. 620, pp. 172–180, 2023. [Online]. Available: <https://www.nature.com/articles/s41586-023-06291-2>
 - [32] M. Antonelli, A. Reinke, S. Bakas, K. Farahani, A. Kopp-Schneider, B. A. Landman, G. Litjens, B. Menze, O. Ronneberger, R. M. Summers *et al.*, “The Medical Segmentation Decathlon,” *Nature Communications*, vol. 13, no. 1, p. 4128, 2022.
 - [33] O. Pelka, S. Koitka, J. Rückert, F. Nensa, and C. M. Friedrich, “Radiology objects in COntext (ROCO): A multimodal image dataset,” in *Intravascular Imaging and Computer Assisted Stenting and Large-Scale Annotation of Biomedical Data and Expert Label Synthesis (LABELS)*, 2018, pp. 180–189.
 - [34] A. E. W. Johnson, T. J. Pollard, S. J. Berkowitz, N. R. Greenbaum, M. P. Lungren, C.-y. Deng *et al.*, “MIMIC-CXR, a de-identified publicly available database of chest radiographs with free-text reports,” *Scientific Data*, vol. 6, p. 317, 2019. [Online]. Available: <https://www.nature.com/articles/s41597-019-0322-0>
 - [35] J. Irvin, P. Rajpurkar, M. Ko, Y. Yu, S. Ciurea-Ilcus, C. Chute, H. Marklund, B. Haghighi, R. Ball, K. Shpanskaya *et al.*, “CheXpert: A large chest radiograph dataset with uncertainty labels and expert comparison,” in *AAAI Conference on Artificial Intelligence*, 2019.
 - [36] J. Wasserthal, H.-C. Breit, M. T. Meyer, M. Pradella, D. Hinck, A. W. Sauter, T. Heye, D. T. Boll, J. Cyriac, S. Yang, M. Bach, and M. Segeroth, “TotalSegmentator: Robust segmentation of 104 anatomic structures in CT images,” *Radiology: Artificial Intelligence*, vol. 5, no. 5, p. e230024, 2023.