

## Available Calibration Certificates For Newly Purchased Products

New Changes Highlighted In Yellow

Model #	Item #	STANDARD CAL CERTS WHEN NEW					OPTIONAL CAL CERTS WHEN NEW			Comments
		Statement Of Cal Practices	Default National Lab of Traceability	Generic Traceable Cal Cert with data	Z540 Compliant Cal Cert	17025 Accredited Calibration	Z540 Compliant Cal Cert	17025 Accredited Calibration	17025 Accredited By	
80	1583376	X					optional	N/A		
81	1583383	X					optional	N/A		
396	2140629	X					optional	N/A		
397	2423260	X					optional	N/A		
908	1583650	X					N/A	optional	A2LA	CF - Contact Factory for specific pricing
909	1583692	X					N/A	optional	A2LA	
910	1583737	X					optional	optional	A2LA	
910R	1583801	X					optional	optional	A2LA	
2600A-101	664323			X			N/A	optional	NVLAP	Vendor provides NIST traceable cert w/mainframe.
2620A	827550	X					optional	N/A		
2620A/05	877837	X					optional	N/A		
2620A-100	889709	X					N/A	N/A		
2620T	664331	Non RTD Functions	NIST - USA			RTD Probe & Unit - NVLAP	optional	N/A		Standard with this product the mainframe & RTD probe are calibrated together for an accredited NVLAP Cert. for ONLY the RTD Temperature Function. For other functions the formal calibration is optional
2620T-2	1608394	Non RTD Functions	NIST - USA			RTD Probe & Unit - NVLAP	optional	N/A		Standard with this product the mainframe & RTD probe are calibrated together for an accredited NVLAP Cert. for ONLY the RTD Temperature Function. For other functions the formal calibration is optional
2625A	889696	X					optional	N/A		
2635A	931985	X					optional	N/A		
2635A-1MB	949503	X					optional	N/A		
2635A-2MB	949508	X					optional	N/A		
2635A-4MB	606308	X					optional	N/A		
2635T	664349	Non RTD Functions	NIST - USA			RTD Probe & Unit - NVLAP	optional	N/A		
2635T-2	1608382	Non RTD Functions	NIST - USA			RTD Probe & Unit - NVLAP	optional	N/A		
2640A	939686	X					optional	N/A		
2640A-912	620713	X					optional	N/A		
2645A	939694	X					optional	N/A		
2680A	1608309	X					N/A	N/A		
2680A-FAI	1620366	X					optional	N/A		
2680A-PAI	1620375	X					optional	optional	A2LA	
2686A	1620184	X					N/A	N/A		
271-U 115V	2423224	X					optional	N/A		
281-U 115V	2423135	X					optional	N/A		
282-U 115V	2423164	X					optional	N/A		
284-U 115V	2423199	X					optional	N/A		
5000A-RH/T	1606205		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
5000A-RH/T1	1626385		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
5000A-RH/T2	1607995		N/A				N/A	N/A		No Calibration Certificate Included in this version
525A-P02	1631369		NIST - USA	X			N/A	N/A		Vendor provides NIST traceable cert w/mainframe.
525A-P04	2074781		NIST - USA	X			N/A	N/A		Vendor provides NIST traceable cert w/mainframe.
525A-P05	1631378		NIST - USA	X			N/A	N/A		Vendor provides NIST traceable cert w/mainframe.
525A-P06	1631384		NIST - USA	X			N/A	N/A		Vendor provides NIST traceable cert w/mainframe.
525A-P07	1631391		NIST - USA	X			N/A	N/A		Vendor provides NIST traceable cert w/mainframe.
525A-P08	1631406		NIST - USA	X			N/A	N/A		Vendor provides NIST traceable cert w/mainframe.
525A-P29	1631414		NIST - USA	X			N/A	N/A		Vendor provides NIST traceable cert w/mainframe.
525A-PA4	1631461		NIST - USA	X			N/A	N/A		Vendor provides NIST traceable cert w/mainframe.
525A-PA6	1631423		NIST - USA	X			N/A	N/A		Vendor provides NIST traceable cert w/mainframe.
525A-PA7	1631438		NIST - USA	X			N/A	N/A		Vendor provides NIST traceable cert w/mainframe.
525A-PA8	1631445		NIST - USA	X			N/A	N/A		Vendor provides NIST traceable cert w/mainframe.
525A-PV4	1631450		NIST - USA	X			N/A	N/A		Vendor provides NIST traceable cert w/mainframe.
525B 120V	3064199		NIST - USA	X			optional	optional	A2LA	Vendor provides NIST traceable cert w/mainframe.
5320A AP	2719019		PTB - Germany	X			N/A	optional	DKD	
5320A EU	2718978		PTB - Germany	X			N/A	optional	DKD	
5320A UK	2719141		PTB - Germany	X			N/A	optional	DKD	

## Available Calibration Certificates For Newly Purchased Products

New Changes Highlighted In Yellow

Model #	Item #	STANDARD CAL CERTS WHEN NEW					OPTIONAL CAL CERTS WHEN NEW			Comments
		Statement Of Cal Practices	Default National Lab of Traceability	Generic Traceable Cal Cert with data	Z540 Compliant Cal Cert	17025 Accredited Calibration	Z540 Compliant Cal Cert	17025 Accredited Calibration	17025 Accredited By	
5320A US	2719107		PTB - Germany	X			N/A	optional	DKD	
5320A ZA	2719055		PTB - Germany	X			N/A	optional	DKD	
5320A/40 AP	2719037		PTB - Germany	X			N/A	optional	DKD	
5320A/40 EU	2718991		PTB - Germany	X			N/A	optional	DKD	
5320A/40 UK	2718957		PTB - Germany	X			N/A	optional	DKD	
5320A/40 US	2719129		PTB - Germany	X			N/A	optional	DKD	
5320A/40 ZA	2719070		PTB - Germany	X			N/A	optional	DKD	
5320A/VLC AP	2719028		PTB - Germany	X			N/A	optional	DKD	
5320A/VLC EU	2718984		PTB - Germany	X			N/A	optional	DKD	
5320A/VLC UK	2718940		PTB - Germany	X			N/A	optional	DKD	
5320A/VLC US	2719118		PTB - Germany	X			N/A	optional	DKD	
5320A/VLC ZA	2719062		PTB - Germany	X			N/A	optional	DKD	
5320A/VLC/40 AP	2719043		PTB - Germany	X			N/A	optional	DKD	
5320A/VLC/40 EU	2719004		PTB - Germany	X			N/A	optional	DKD	
5320A/VLC/40 UK	2718969		PTB - Germany	X			N/A	optional	DKD	
5320A/VLC/40 US	2719134		PTB - Germany	X			N/A	optional	DKD	
5320A/VLC/40 ZA	2719081		PTB - Germany	X			N/A	optional	DKD	
5320A-LOAD	3050121		PTB - Germany	X			N/A	optional	DKD	CF - Contact Factory for specific pricing
5500A 100	945167		NIST - USA		X		N/A	optional	NVLAP	
5500A 120	945162		NIST - USA		X		N/A	optional	NVLAP	
5500A 240	945170		NIST - USA		X		N/A	optional	NVLAP	
5500A/3 100	673688		NIST - USA	Scope Option	Mainframe		Scope optional	optional	NVLAP + A2LA	For optional 17025: Mainframe is NVLAP; Scope Option is A2LA
5500A/3 120	673673		NIST - USA	Scope Option	Mainframe		Scope optional	optional	NVLAP + A2LA	For optional 17025: Mainframe is NVLAP; Scope Option is A2LA
5500A/3 240	673699		NIST - USA	Scope Option	Mainframe		Scope optional	optional	NVLAP + A2LA	For optional 17025: Mainframe is NVLAP; Scope Option is A2LA
5500A/6 100	673696		NIST - USA		Mainframe & Scope		N/A	optional	NVLAP + A2LA	For optional 17025: Mainframe is NVLAP; Scope Option is A2LA
5500A/6 120	673681		NIST - USA		Mainframe & Scope		N/A	optional	NVLAP + A2LA	For optional 17025: Mainframe is NVLAP; Scope Option is A2LA
5500A/6 240	673704		NIST - USA		Mainframe & Scope		N/A	optional	NVLAP + A2LA	For optional 17025: Mainframe is NVLAP; Scope Option is A2LA
5500A/COIL	200664	X					optional	N/A		
5520A 100	627007		NIST - USA		X		N/A	optional	NVLAP	
5520A 120	627015		NIST - USA		X		N/A	optional	NVLAP	
5520A 240	627023		NIST - USA		X		N/A	optional	NVLAP	
5520A/1GHZ 100	1547290		NIST - USA		Mainframe & Scope		N/A	optional	NVLAP + A2LA	For optional 17025: Mainframe is NVLAP; Scope Option is A2LA
5520A/1GHZ 120	1547309		NIST - USA		Mainframe & Scope		N/A	optional	NVLAP + A2LA	For optional 17025: Mainframe is NVLAP; Scope Option is A2LA
5520A/1GHZ 240	1547311		NIST - USA		Mainframe & Scope		N/A	optional	NVLAP + A2LA	For optional 17025: Mainframe is NVLAP; Scope Option is A2LA
5520A/3 100	673715		NIST - USA	Scope Option	Mainframe		Scope optional	optional	NVLAP + A2LA	For optional 17025: Mainframe is NVLAP; Scope Option is A2LA
5520A/3 120	673707		NIST - USA	Scope Option	Mainframe		Scope optional	optional	NVLAP + A2LA	For optional 17025: Mainframe is NVLAP; Scope Option is A2LA
5520A/3 240	673723		NIST - USA	Scope Option	Mainframe		Scope optional	optional	NVLAP + A2LA	For optional 17025: Mainframe is NVLAP; Scope Option is A2LA
5520A/6 100	673720		NIST - USA		Mainframe & Scope		N/A	optional	NVLAP + A2LA	For optional 17025: Mainframe is NVLAP; Scope Option is A2LA
5520A/6 120	673712		NIST - USA		Mainframe & Scope		N/A	optional	NVLAP + A2LA	For optional 17025: Mainframe is NVLAP; Scope Option is A2LA
5520A/6 240	673731		NIST - USA		Mainframe & Scope		N/A	optional	NVLAP + A2LA	For optional 17025: Mainframe is NVLAP; Scope Option is A2LA
5520A-PQ 100	1605600		NIST - USA	PQ Option	Mainframe		N/A	optional	NVLAP	For optional 17025 accredited calibration the mainframe is NVLAP; 17025 accredited cal does not include the PQ Option (PQ is NIST Traceable with data)
5520A-PQ 120	1605617		NIST - USA	PQ Option	Mainframe		N/A	optional	NVLAP	For optional 17025 accredited calibration the mainframe is NVLAP; 17025 accredited cal does not include the PQ Option (PQ is NIST Traceable with data)
5520A-PQ 240	1605621		NIST - USA	PQ Option	Mainframe		N/A	optional	NVLAP	For optional 17025 accredited calibration the mainframe is NVLAP; 17025 accredited cal does not include the PQ Option (PQ is NIST Traceable with data)
5520A-PQ/1GHZ 100	1605639		NIST - USA	PQ Option	Mainframe & Scope		N/A	optional	NVLAP + A2LA	For optional 17025 accredited calibration the mainframe is NVLAP; the scope option A2LA; 17025 accredited cal does not include the PQ Option (PQ is NIST Traceable with data)

## Available Calibration Certificates For Newly Purchased Products

New Changes Highlighted In Yellow

Model #	Item #	STANDARD CAL CERTS WHEN NEW					OPTIONAL CAL CERTS WHEN NEW			Comments
		Statement Of Cal Practices	Default National Lab of Traceability	Generic Traceable Cal Cert with data	Z540 Compliant Cal Cert	17025 Accredited Calibration	Z540 Compliant Cal Cert	17025 Accredited Calibration	17025 Accredited By	
5520A-PQ/1GHZ 120	1605656		NIST - USA	PQ Option	Mainframe & Scope		N/A	optional	NVLAP + A2LA	For optional 17025 accredited calibration the mainframe is NVLAP; the scope option A2LA; 17025 accredited cal does not include the PQ Option (PQ is NIST Traceable with data)
5520A-PQ/1GHZ 240	1605674		NIST - USA	PQ Option	Mainframe & Scope		N/A	optional	NVLAP + A2LA	For optional 17025 accredited calibration the mainframe is NVLAP; the scope option A2LA; 17025 accredited cal does not include the PQ Option (PQ is NIST Traceable with data)
5520A-PQ/3 100	1605688		NIST - USA	PQ & Scope Option	Mainframe		Scope optional	optional	NVLAP + A2LA	For optional 17025 accredited calibration the mainframe is NVLAP; the scope option A2LA; 17025 accredited cal does not include the PQ Option (PQ is NIST Traceable with data)
5520A-PQ/3 120	1605695		NIST - USA	PQ & Scope Option	Mainframe		Scope optional	optional	NVLAP + A2LA	For optional 17025 accredited calibration the mainframe is NVLAP; the scope option A2LA; 17025 accredited cal does not include the PQ Option (PQ is NIST Traceable with data)
5520A-PQ/3 240	1605707		NIST - USA	PQ & Scope Option	Mainframe		Scope optional	optional	NVLAP + A2LA	For optional 17025 accredited calibration the mainframe is NVLAP; the scope option A2LA; 17025 accredited cal does not include the PQ Option (PQ is NIST Traceable with data)
5520A-PQ/6 100	1605718		NIST - USA	PQ Option	Mainframe & Scope		N/A	optional	NVLAP + A2LA	For optional 17025 accredited calibration the mainframe is NVLAP; the scope option A2LA; 17025 accredited cal does not include the PQ Option (PQ is NIST Traceable with data)
5520A-PQ/6 120	1605729		NIST - USA	PQ Option	Mainframe & Scope		N/A	optional	NVLAP + A2LA	For optional 17025 accredited calibration the mainframe is NVLAP; the scope option A2LA; 17025 accredited cal does not include the PQ Option (PQ is NIST Traceable with data)
5520A-PQ/6 240	1605734		NIST - USA	PQ Option	Mainframe & Scope		N/A	optional	NVLAP + A2LA	For optional 17025 accredited calibration the mainframe is NVLAP; the scope option A2LA; 17025 accredited cal does not include the PQ Option (PQ is NIST Traceable with data)
5700A 100	775544		NIST - USA		X		N/A	optional	NVLAP	
5700A 115	775585		NIST - USA		X		N/A	optional	NVLAP	
5700A 230	791822		NIST - USA		X		N/A	optional	NVLAP	
5700A/01 100	674090		NIST - USA		X		N/A	optional	NVLAP	
5700A/01 115	674077		NIST - USA		X		N/A	optional	NVLAP	
5700A/01 230	674108		NIST - USA		X		N/A	optional	NVLAP	
5700A/03 100	674093		NIST - USA		Mainframe & Widenban Option		N/A	optional	NVLAP	
5700A/03 115	674082		NIST - USA		Mainframe & Widenban Option		N/A	optional	NVLAP	
5700A/03 230	674116		NIST - USA		Mainframe & Widenban Option		N/A	optional	NVLAP	
5700A/04 100	674101		NIST - USA		Mainframe & Widenban Option		N/A	optional	NVLAP	
5700A/04 115	674085		NIST - USA		Mainframe & Widenban Option		N/A	optional	NVLAP	
5700A/04 230	674119		NIST - USA		Mainframe & Widenban Option		N/A	optional	NVLAP	
5720A 100	600699		NIST - USA		X		N/A	optional	NVLAP	
5720A 115	600731		NIST - USA		X		N/A	optional	NVLAP	
5720A 230	600742		NIST - USA		X		N/A	optional	NVLAP	
5720A/01 100	674135		NIST - USA		X		N/A	optional	NVLAP	
5720A/01 115	674124		NIST - USA		X		N/A	optional	NVLAP	
5720A/01 230	674150		NIST - USA		X		N/A	optional	NVLAP	
5720A/03 100	674140		NIST - USA		Mainframe & Widenban Option		N/A	optional	NVLAP	
5720A/03 115	674127		NIST - USA		Mainframe & Widenban Option		N/A	optional	NVLAP	
5720A/03 230	674157		NIST - USA		Mainframe & Widenban Option		N/A	optional	NVLAP	
5720A/04 100	674143		NIST - USA		Mainframe & Widenban Option		N/A	optional	NVLAP	
5720A/04 115	674132		NIST - USA		Mainframe & Widenban Option		N/A	optional	NVLAP	
5720A/04 230	674165		NIST - USA		Mainframe & Widenban Option		N/A	optional	NVLAP	
5725A 100	849091		NIST - USA		X		N/A	optional	NVLAP	
5725A 115	827139		NIST - USA		X		N/A	optional	NVLAP	
5725A 230	849096		NIST - USA		X		N/A	optional	NVLAP	

## Available Calibration Certificates For Newly Purchased Products

New Changes Highlighted In Yellow

Model #	Item #	STANDARD CAL CERTS WHEN NEW					OPTIONAL CAL CERTS WHEN NEW			Comments
		Statement Of Cal Practices	Default National Lab of Traceability	Generic Traceable Cal Cert with data	Z540 Compliant Cal Cert	17025 Accredited Calibration	Z540 Compliant Cal Cert	17025 Accredited Calibration	17025 Accredited By	
5725A/1 100	674173		NIST - USA		X		N/A	optional	NVLAP	
5725A/1 115	674168		NIST - USA		X		N/A	optional	NVLAP	
5725A/1 230	675173		NIST - USA		X		N/A	optional	NVLAP	
5790A 100	900402		NIST - USA	X			optional	optional	NVLAP	
5790A 115	880760		NIST - USA	X			optional	optional	NVLAP	
5790A 230	900407		NIST - USA	X			optional	optional	NVLAP	
5790A/3 100	661089		NIST - USA	X			optional	optional	NVLAP	
5790A/3 115	661030		NIST - USA	X			optional	optional	NVLAP	
5790A/3 230	661105		NIST - USA	X			optional	optional	NVLAP	
57LFC/AN	2094997		NIST - USA	X			optional	N/A		
5820A-1C 100	673183		NIST - USA		X		N/A	optional	A2LA	
5820A-1C 120	673186		NIST - USA		X		N/A	optional	A2LA	
5820A-1C 240	673178		NIST - USA		X		N/A	optional	A2LA	
5820A-1C-GHZ100	673225		NIST - USA		X		N/A	optional	A2LA	
5820A-1C-GHZ120	673228		NIST - USA		X		N/A	optional	A2LA	
5820A-1C-GHZ240	673217		NIST - USA		X		N/A	optional	A2LA	
5820A-5C 100	673209		NIST - USA		X		N/A	optional	A2LA	
5820A-5C 120	673159		NIST - USA		X		N/A	optional	A2LA	
5820A-5C 240	673202		NIST - USA		X		N/A	optional	A2LA	
5820A-5C-GHZ100	673236		NIST - USA		X		N/A	optional	A2LA	
5820A-5C-GHZ120	673241		NIST - USA		X		N/A	optional	A2LA	
5820A-5C-GHZ240	673233		NIST - USA		X		N/A	optional	A2LA	
6100A	1887434		NPL - UK		X		N/A	optional	UKAS	
6100A/80A	2101438		NPL - UK		X		N/A	optional	UKAS	
6100A/E	2111095		NPL - UK		X		N/A	optional	UKAS	
6100A/E/80A	2111175		NPL - UK		X		N/A	optional	UKAS	
6101A	1887465		NPL - UK		X		N/A	optional	UKAS	
6101A/80A	2111130		NPL - UK		X		N/A	optional	UKAS	
6120A	1887490		NPL - UK		X		N/A	optional	UKAS	
6120A/80A	2111148		NPL - UK		X		N/A	optional	UKAS	
6120A/E	2111109		NPL - UK		X		N/A	optional	UKAS	
6120A/E/80A	2111182		NPL - UK		X		N/A	optional	UKAS	
6130A	1887526		NPL - UK		X		N/A	optional	UKAS	
6130A/80A	2111153		NPL - UK		X		N/A	optional	UKAS	
6130A/E	2111111		NPL - UK		X		N/A	optional	UKAS	
6130A/E/80A	2111194		NPL - UK		X		N/A	optional	UKAS	
6140A	1887559		NPL - UK		X		N/A	optional	UKAS	
6140A/80A	2111166		NPL - UK		X		N/A	optional	UKAS	
6140A/E	2111127		NPL - UK		X		N/A	optional	UKAS	
6140A/E/80A	2111203		NPL - UK		X		N/A	optional	UKAS	
7000	1571329		NPL - UK	X			optional	optional	UKAS	CF - Contact Factory for specific pricing
7001	1571307		NPL - UK	X			optional	optional	UKAS	CF - Contact Factory for specific pricing
7000S	1571318	N/A	N/A	X		N/A	N/A	N/A		Certificates of Calibration are not available for this item; item cannot be calibrated it does not measure or indicate values.
7004N	1571294		NPL - UK	X			optional	optional	UKAS	CF - Contact Factory for specific pricing
7004T	1571282		NPL - UK				optional	optional	UKAS	CF - Contact Factory for specific pricing
7010N	1571334		NPL - UK	X			optional	optional	UKAS	CF - Contact Factory for specific pricing
7010T	1571341		NPL - UK				optional	optional	UKAS	CF - Contact Factory for specific pricing
720A	201228	X					optional	N/A	NVLAP	
732B 100	869222	X					N/A	N/A		Calibration not available for this configuration
732B 120	869144	X					N/A	N/A		Calibration not available for this configuration
732B 220	869227	X					N/A	N/A		Calibration not available for this configuration
732B 240	869230	X					N/A	N/A		Calibration not available for this configuration
732B/C 100	674069		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
732B/C 120	674058		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
732B/C 220	675074		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
732B/C 240	675089		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
732B/CI 120	1551416		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
732B/H 100	674066		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add two weeks for calibration
732B/H 120	674051		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add two weeks for calibration
732B/H 220	674074		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add two weeks for calibration
732B/H 240	675082		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add two weeks for calibration
734A 100	869235	X					N/A	N/A		Calibration not available for this configuration
734A 120	869243	X					N/A	N/A		Calibration not available for this configuration

## Available Calibration Certificates For Newly Purchased Products

New Changes Highlighted In Yellow

Model #	Item #	STANDARD CAL CERTS WHEN NEW					OPTIONAL CAL CERTS WHEN NEW			Comments
		Statement Of Cal Practices	Default National Lab of Traceability	Generic Traceable Cal Cert with data	Z540 Compliant Cal Cert	17025 Accredited Calibration	Z540 Compliant Cal Cert	17025 Accredited Calibration	17025 Accredited By	
734A 220	869248	X					N/A	N/A		Calibration not available for this configuration
734A 240	869250	X					N/A	N/A		Calibration not available for this configuration
734A/01 100	665866		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/01 120	665791		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/01 220	673936		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/01 240	674002		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/02 100	665869		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/02 120	665794		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/02 220	673939		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/02 240	674009		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/03 100	665874		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/03 120	665802		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/03 220	673944		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/03 240	674010		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/04 100	673878		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/04 120	665809		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/04 220	673947		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/04 240	674017		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/10 100	665841		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add two weeks for calibration
734A/10 120	665775		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add two weeks for calibration
734A/10 220	673910		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add two weeks for calibration
734A/10 240	673985		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add two weeks for calibration
734A/11 100	673886		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/11 120	665810		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/11 220	673951		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/11 240	674025		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/12 100	673889		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/12 120	665817		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/12 220	673954		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/12 240	674028		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/13 100	673894		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/13 120	665825		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/13 220	673962		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/13 240	674033		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/20 100	665844		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add two weeks for calibration
734A/20 120	665778		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add two weeks for calibration
734A/20 220	673913		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add two weeks for calibration
734A/20 240	673988		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add two weeks for calibration

## Available Calibration Certificates For Newly Purchased Products

New Changes Highlighted In Yellow

Model #	Item #	STANDARD CAL CERTS WHEN NEW					OPTIONAL CAL CERTS WHEN NEW			Comments
		Statement Of Cal Practices	Default National Lab of Traceability	Generic Traceable Cal Cert with data	Z540 Compliant Cal Cert	17025 Accredited Calibration	Z540 Compliant Cal Cert	17025 Accredited Calibration	17025 Accredited By	
734A/21 100	673897		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/21 120	665828		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/21 220	673969		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/21 240	674036		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/22 100	673902		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/22 120	665833		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/22 220	673970		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/22 240	674041		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/30 100	665851		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add two weeks for calibration
734A/30 120	665783		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add two weeks for calibration
734A/30 220	673921		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add two weeks for calibration
734A/30 240	673993		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add two weeks for calibration
734A/31 100	673905		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/31 120	665836		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/31 220	673977		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/31 240	674044		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add six weeks for characterization
734A/40 100	665858		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add two weeks for calibration
734A/40 120	665786		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add two weeks for calibration
734A/40 220	673928		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add two weeks for calibration
734A/40 240	673996		NIST - USA	X			optional	optional	NVLAP	Optional Cert is both Z540 and 17025, add two weeks for calibration
742A-1	833293		NIST - USA	X			optional	optional	NVLAP	
742A-1.9	833376		NIST - USA	X			optional	optional	NVLAP	
742A-10	833301		NIST - USA	X			optional	optional	NVLAP	
742A-100	833319		NIST - USA	X			optional	optional	NVLAP	
742A-100K	833343		NIST - USA	X			optional	optional	NVLAP	
742A-10K	833335		NIST - USA	X			optional	optional	NVLAP	
742A-10M	833368		NIST - USA	X			optional	optional	NVLAP	
742A-19K	833384		NIST - USA	X			optional	optional	NVLAP	
742A-19M	833392		NIST - USA	X			optional	optional	NVLAP	
742A-1K	833327		NIST - USA	X			optional	optional	NVLAP	
742A-1M	833350		NIST - USA	X			optional	optional	NVLAP	
742A-25	2072673		NIST - USA	X			optional	optional	NVLAP	
752A	590539	X					N/A	N/A		Only an optional performance check is available for the 752A
792A 100	883046		NIST - USA	X			optional	optional	NVLAP	
792A 120	871710		NIST - USA	X			optional	optional	NVLAP	
792A 220	871947		NIST - USA	X			optional	optional	NVLAP	
792A 240	883053		NIST - USA	X			optional	optional	NVLAP	
8508A 120	1886175		NPL - UK		X		N/A		UKAS	
8508A 240	1886166		NPL - UK		X		N/A		UKAS	
8508A/01 120	1886148		NPL - UK		X		N/A		UKAS	
8508A/01 240	1886153		NPL - UK		X		N/A		UKAS	
8508A-7000K	1886226		NPL - UK	X			optional	optional	UKAS	
8508A-PRT	1886194		NIST - USA			NVLAP	N/A	N/A		Vendor provides NVLAP cert (using comparison method) w/PRT.
8508A-SPRT	1886182		NIST - USA			NVLAP	N/A	N/A		Vendor provides fixed-point NVLAP cert w/SPRT.
8808A 100V	2802360	X					optional			
8808A 120V	2570235	X					optional			
8808A 220V	2802372	X					optional			
8808A 240V	2802385	X					optional			
8808A/SU 100V	2802444	X					optional			
8808A/SU 120V	2802432	X					optional			
8808A/SU 220V	2802459	X					optional			



## Available Calibration Certificates For Newly Purchased Products

New Changes Highlighted In Yellow

Model #	Item #	STANDARD CAL CERTS WHEN NEW					OPTIONAL CAL CERTS WHEN NEW			Comments
		Statement Of Cal Practices	Default National Lab of Traceability	Generic Traceable Cal Cert with data	Z540 Compliant Cal Cert	17025 Accredited Calibration	Z540 Compliant Cal Cert	17025 Accredited Calibration	17025 Accredited By	
8808A/SU 240V	2802467	X					optional			
8845A 120V	2441853	X					optional	optional	NVLAP	
8845A 220V	2577352	X					optional	optional	NVLAP	
8845A 100V	2577341	X					optional	optional	NVLAP	
8845A 240V	2577365	X					optional	optional	NVLAP	
8845A/C 120V	2675277		NIST - USA	X			optional	optional	NVLAP	
8845A/C 220V	2720674		NIST - USA	X			optional	optional	NVLAP	
8845A/C 100V	2720688		NIST - USA	X			optional	optional	NVLAP	
8845A/C 240V	2720695		NIST - USA	X			optional	optional	NVLAP	
8845A/SU 120V	2675292	X					optional	optional	NVLAP	
8845A/SU 220V	2675315	X					optional	optional	NVLAP	
8845A/SU 100V	2675326	X					optional	optional	NVLAP	
8845A/SU 240V	2675332	X					optional	optional	NVLAP	
8845A/CSU 120V	2675303		NIST - USA	X			optional	optional	NVLAP	
8845A/CSU 220V	2722675		NIST - USA	X			optional	optional	NVLAP	
8845A/CSU 100V	2722682		NIST - USA	X			optional	optional	NVLAP	
8845A/CSU 240V	2722694		NIST - USA	X			optional	optional	NVLAP	
8846A 120V	2454021	X					optional	optional	NVLAP	
8846A 220V	2577383	X					optional	optional	NVLAP	
8846A 100V	2577376	X					optional	optional	NVLAP	
8846A 240V	2577390	X					optional	optional	NVLAP	
8846A/C 120V	2675289		NIST - USA	X			optional	optional	NVLAP	
8846A/C 220V	2720707		NIST - USA	X			optional	optional	NVLAP	
8846A/C 100V	2720718		NIST - USA	X			optional	optional	NVLAP	
8846A/C 240V	2720729		NIST - USA	X			optional	optional	NVLAP	
8846A/SU 120V	2675344	X					optional	optional	NVLAP	
8846A/SU 220V	2675367	X					optional	optional	NVLAP	
8846A/SU 100V	2675371	X					optional	optional	NVLAP	
8846A/SU 240V	2675380	X					optional	optional	NVLAP	
8846A/CSU 120V	2675359		NIST - USA	X			optional	optional	NVLAP	
8846A/CSU 220V	2722701		NIST - USA	X			optional	optional	NVLAP	
8846A/CSU 100V	2722712		NIST - USA	X			optional	optional	NVLAP	
8846A/CSU 240V	2722720		NIST - USA	X			optional	optional	NVLAP	
9100	1583873		NPL - UK		X		N/A		UKAS	CF - Contact Factory for specific pricing
9100-200	1583963		NPL - UK		X		N/A		UKAS	CF - Contact Factory for specific pricing
9100-250	1583974		NPL - UK		X		N/A		UKAS	CF - Contact Factory for specific pricing
9100-600	1583988		NPL - UK		X		N/A		UKAS	
9500B/1100	1623989		NPL - UK		X		N/A		UKAS	CF - Contact Factory for specific pricing
9500B/3200	1623992		NPL - UK		X		N/A		UKAS	
9500B/600	1623977		NPL - UK		X		N/A		UKAS	
9510 FLK	1625165		NPL - UK		X		N/A		UKAS	CF - Contact Factory for specific pricing
9530 FLK	1625176		NPL - UK		X		N/A		UKAS	
9550 FLK	1625183		NPL - UK		X		N/A		UKAS	CF - Contact Factory for specific pricing
9560 FLK	1625004		NPL - UK		X		N/A		UKAS	CF - Contact Factory for specific pricing
9640A-STD 115V	2464734		NPL - UK		X		N/A			
9640A-STD 230V	2543221		NPL - UK		X		N/A			
9640A-STD/75 115V	2464741		NPL - UK		X		N/A			
9640A-STD/75 230V	2543239		NPL - UK		X		N/A			
9640A-LPN 115V	2464752		NPL - UK		X		N/A			
9640A-LPN/75 115V	2464765		NPL - UK		X		N/A			
9640A-LPN 230V	2543242		NPL - UK		X		N/A			
9640A-LPN/75 230V	2543256		NPL - UK		X		N/A			
A40-010MA	199141	X					N/A	optional	NVLAP	Confirm amplitude & frequency points needed
A40-020MA	199133	X					N/A	optional	NVLAP	Confirm amplitude & frequency points needed
A40-030MA	199125	X					N/A	optional	NVLAP	Confirm amplitude & frequency points needed
A40-050MA	199117	X					N/A	optional	NVLAP	Confirm amplitude & frequency points needed
A40-100MA	199109	X					N/A	optional	NVLAP	Confirm amplitude & frequency points needed
A40-1A	199067	X					N/A	optional	NVLAP	Confirm amplitude & frequency points needed
A40-200MA	199091	X					N/A	optional	NVLAP	Confirm amplitude & frequency points needed
A40-2A	199059	X					N/A	optional	NVLAP	Confirm amplitude & frequency points needed
A40-300MA	199083	X					N/A	optional	NVLAP	Confirm amplitude & frequency points needed
A40-3A	199042	X					N/A	optional	NVLAP	Confirm amplitude & frequency points needed
A40-500MA	199075	X					N/A	optional	NVLAP	Confirm amplitude & frequency points needed
A40-5A	199034	X					N/A	optional	NVLAP	Confirm amplitude & frequency points needed
A40A-10A	216077	X					N/A	optional	NVLAP	Confirm amplitude & frequency points needed
A40A-20A	216069	X					N/A	optional	NVLAP	Confirm amplitude & frequency points needed
FLUKE-700HTP	662830	N/A					N/A	N/A		Certificates of Calibration are not available for this item; item cannot be calibrated it does not measure or indicates values.
FLUKE-700LTP	1566748	N/A					N/A	N/A		Certificates of Calibration are not available for this item; item cannot be calibrated it does not measure or indicates values.
FLUKE-700P00	667391		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700P01	936005		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.

## Available Calibration Certificates For Newly Purchased Products

New Changes Highlighted In Yellow

Model #	Item #	STANDARD CAL CERTS WHEN NEW					OPTIONAL CAL CERTS WHEN NEW			Comments
		Statement Of Cal Practices	Default National Lab of Traceability	Generic Traceable Cal Cert with data	Z540 Compliant Cal Cert	17025 Accredited Calibration	Z540 Compliant Cal Cert	17025 Accredited Calibration	17025 Accredited By	
FLUKE-700P02	936013		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700P03	936018		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700P04	936021		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700P05	936026		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700P06	936034		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700P07	936039		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700P08	936042		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700P09	617123		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700P22	100138		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700P23	100146		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700P24	100153		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700P27	1549399		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700P29	612157		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700P30	612165		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700P31	612173		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700PA3	612207		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700PA4	612215		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700PA5	612223		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700PA6	612231		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700PD2	612090		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700PD3	612108		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700PD4	606837		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700PD5	612116		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700PD6	612124		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700PD7	612132		NIST - USA	X			N/A	N/A		Vendor provides traceable cert w/mainframe.
FLUKE-700PMP	617131	N/A	N/A	X		N/A	N/A	N/A		Certificates of Calibration are not available for this item. Item cannot be calibrated and does not measure or indicate values.
FLUKE-700PTP	662822		NIST - USA	X			N/A	N/A		
FLUKE-700PV3	612249		NIST - USA	X			N/A	N/A		
FLUKE-700PV4	612256		NIST - USA	X			N/A	N/A		
PM6681/016	101953	x					optional	N/A		
PM6681/056	101979	x					optional	N/A		
PM6681/066	651874	x					optional	N/A		
PM6681/416	101980	x					optional	N/A		
PM6681/456	102001	x					optional	N/A		
PM6681/466	651882	x					optional	N/A		
PM6681/616	102004	x					optional	N/A		
PM6681/656	102027	x					optional	N/A		
PM6681/666	651890	x					optional	N/A		
PM6681/716	2464079	x					optional	N/A		
PM6681/756	2464087	x					optional	N/A		
PM6681/766	2464093	x					optional	N/A		
PM6681R/076	610920	x					optional	N/A		
PM6681R/476	610938	x					optional	N/A		
PM6681R/676	610946	x					optional	N/A		
PM6681R/776	2464101	x					optional	N/A		
PM6685/011	934377	x					optional	N/A		
PM6685/013	934380	x					optional	N/A		
PM6685/016	934385	x					optional	N/A		
PM6685/051	934435	x					optional	N/A		
PM6685/053	934443	x					optional	N/A		
PM6685/056	934448	x					optional	N/A		
PM6685/061	651924	x					optional	N/A		
PM6685/063	655628	x					optional	N/A		
PM6685/066	655636	x					optional	N/A		
PM6685/411	934476	x					optional	N/A		
PM6685/413	934484	x					optional	N/A		
PM6685/416	934489	x					optional	N/A		
PM6685/451	934539	x					optional	N/A		
PM6685/453	934542	x					optional	N/A		
PM6685/456	934547	x					optional	N/A		
PM6685/461	655651	x					optional	N/A		
PM6685/463	655669	x					optional	N/A		
PM6685/466	655677	x					optional	N/A		
PM6685/611	934575	x					optional	N/A		
PM6685/613	934583	x					optional	N/A		
PM6685/616	934588	x					optional	N/A		
PM6685/651	934633	x					optional	N/A		
PM6685/653	934638	x					optional	N/A		
PM6685/656	934641	x					optional	N/A		
PM6685/661	655693	x					optional	N/A		
PM6685/663	655701	x					optional	N/A		
PM6685/666	655719	x					optional	N/A		
PM6685/711	2464112	x					optional	N/A		
PM6685/716	2464120	x					optional	N/A		



## Available Calibration Certificates For Newly Purchased Products

New Changes Highlighted in Yellow

Model #	Item #	STANDARD CAL CERTS WHEN NEW					OPTIONAL CAL CERTS WHEN NEW			Comments
		Statement Of Cal Practices	Default National Lab of Traceability	Generic Traceable Cal Cert with data	Z540 Compliant Cal Cert	17025 Accredited Calibration	Z540 Compliant Cal Cert	17025 Accredited Calibration	17025 Accredited By	
PM6685/751	2464135	x					optional	N/A		
PM6685/756	2464147	x					optional	N/A		
PM6685/761	2464158	x					optional	N/A		
PM6685/766	2464164	x					optional	N/A		
PM6685R/071	102459	x					optional	N/A		
PM6685R/076	102464	x					optional	N/A		
PM6685R/471	102467	x					optional	N/A		
PM6685R/476	102472	x					optional	N/A		
PM6685R/671	102475	x					optional	N/A		
PM6685R/676	102480	x					optional	N/A		
PM6685R/771	2464173	x					optional	N/A		
PM6685R/776	2464186	x					optional	N/A		
PM6690/011	2463886	x					optional	N/A		
PM6690/012	2463899	x					optional	N/A		
PM6690/051	2463900	x					optional	N/A		
PM6690/052	2463917	x					optional	N/A		
PM6690/061	2463921	x					optional	N/A		
PM6690/062	2463939	x					optional	N/A		
PM6690/611	2463942	x					optional	N/A		
PM6690/612	2463956	x					optional	N/A		
PM6690/651	2463963	x					optional	N/A		
PM6690/652	2463974	x					optional	N/A		
PM6690/661	2463988	x					optional	N/A		
PM6690/662	2463995	x					optional	N/A		
PM6690/711	2464005	x					optional	N/A		
PM6690/712	2464010	x					optional	N/A		
PM6690/751	2464022	x					optional	N/A		
PM6690/752	2464031	x					optional	N/A		
PM6690/761	2464046	x					optional	N/A		
PM6690/762	2464054	x					optional	N/A		

Products not listed within this document are provided with a Statement of Calibration Practices, which is not a certificate of calibration. Customers wanting calibration certs for products not listed in this document must pay for that service.

"Vendor provides" means the finished product is one Fluke procures a Fluke-branded product from another source; that source provides the traceable (or other) calibration certificate with the product prior to shipping it to Fluke.

Acronym Guide:	NIST-USA	National Laboratory for traceability in the USA
	NPL-UK	National Laboratory for traceability in the United Kingdom
	PTB-Germany	National Laboratory for traceability in Germany
	NVLAP	17025 Accrediting Body in the USA
	A2LA	17025 Accrediting Body in the USA
	UKAS	17025 Accrediting Body in the United Kingdom
	DKD	17025 Accrediting Body in Germany
	17025	Reference to the International Standard for Calibration - ISO 17025
	Z540	Reference to the USA Standard for Calibration - ANSI/NCSL Z540.1

### ITEM # Description

**1256480** **Z540 Traceable Calibration with Data** - The instrument is calibrated to manufacturer's specifications using test equipment traceable to NIST or other recognized standards laboratories. It is processed in accordance with ANSI/NCSL Z540-1 1994\*, and is returned with a "Report of Calibration"; report includes "as found" and "as left" readings, including test uncertainties. Calibration stickers and seals are included.

\*Reference: "Calibration Laboratories and Measuring and Test Equipment - General Requirements" - ANSI/NCSL Z540-1 1994

**1256990**  
**1883673** **Accredited Calibration** - This calibration is done under the requirements of 17025. Labs are accredited to meet 17025 requirements by any of several accrediting organization (NVLAP, A2LA, DKD, etc.).  
**2724154**