Cross-corpus and Multimodal Tuberculosis Detection: Toward Model Generalization

Experiment report

Bagus Tris Atmaja

`multibench` repository, `datasets/tb_multi`

2 November 2026 (revised)

Abstract

Chest X-ray (CXR) classifiers for tuberculosis (TB) often look excellent in-distribution and fail on a new site. We build a six-corpus benchmark (Qatar, Montgomery, Shenzhen, TBX11K, Pakistan/Mendeley, and the multimodal CIDRZ/Google Health AI cohort), add VinDr-CXR (radiologist-labelled, 646 TB-positive images) as a second external cohort, audit them for duplicates and shortcuts, and evaluate generalization with leave-one-corpus-out (LOCO) fine-tuning of several backbones and domain-generalization (DG) methods. We then test whether the best image expert improves multimodal (text + audio + CXR) detection on CIDRZ, an external target never used for model selection. Main findings (five seeds, per-backbone tuning on the public folds only): (i) corpus identity alone predicts the TB label (pooled AUROC 0.83) and silhouette- or border-only controls reach AUROC up to 0.84; artefact-only predictors reach 0.85–1.00 within-corpus on Qatar, TBX11K and Pakistan, and Pakistan is not identifiable (artefact-only AUROC 1.000); (ii) lung masking plus strong augmentation, and above all a CXR foundation model (RAD-DINO, LOCO mean AUROC 0.898 vs 0.867 for tuned ImageNet ViT-B/16), improve generalization on the public corpora, whereas classical DG losses do not beat source-balanced sampling; (iii) on the external CIDRZ cohort every backbone sits at AUROC 0.77–0.81 with overlapping CIs, and artefact-stratified AUROC stays within 0.01 of raw (0.816 vs 0.825 for RAD-DINO); late fusion lifts text from 0.782 to 0.85 and audio adds nothing; (iv) probabilities are badly miscalibrated across sites (ECE up to 0.49) and local Platt recalibration repairs calibration (ECE 0.07) but not discrimination: no model meets the WHO triage profile (sensitivity $\geq$90%, specificity $\geq$70%) on CIDRZ even with an oracle threshold; (v) on the larger VinDr-CXR cohort the foundation-model advantage is confirmed (weighted AUROC 0.903 for RAD-DINO vs 0.870 for tuned ViT, paired difference +0.032 [0.018, 0.049], RAD-DINO+MixStyle 0.917), the artefact-stratified AUROC stays within 0.02 of raw (0.899 vs 0.919), and an oracle threshold reaches the WHO profile for RAD-DINO (sensitivity 0.91 at 70% specificity) but a deployable threshold does not, with the caveat that VinDr's reference standard is radiologist reading, not microbiology.

Revision note. This version supersedes the first one: a checkpoint-restore bug (only parameters named "vit.encoder"/"head" were restored at the best epoch, leaving CNN tails and RAD-DINO's final layer-norm at last-epoch values) affected the ResNet50, xrv-DenseNet and RAD-DINO runs. All headline numbers were re-run after the fix with five seeds; the earlier RAD-DINO LOCO mean 0.906 is now 0.898 (tuned), ResNet50 0.786 $\rightarrow$ 0.816, xrv-DenseNet 0.784 $\rightarrow$ 0.809, and the fusion and calibration results were recomputed with corrected experts.

Revision note 2. VinDr-CXR was added as a second external cohort (Section 11): same eight tuned configurations, five seeds, trained on the four public sources only, with weighted metrics because the evaluation sample is enriched for TB.

1 Introduction and goals

The study asks which methods improve (a) cross-corpus generalization of CXR TB detection and (b) multimodal fusion. It extends earlier work on Qatar $\rightarrow$ CIDRZ transfer, where lung masking raised CIDRZ zero-shot AUROC of a small VGG from 0.613 ± 0.049 to 0.729 ± 0.011 (seed-average difference +0.125, 95% CI $[+0.030, +0.217]$) by removing non-lung cues such as anonymization boxes present in about 30% of Qatar TB images versus under 1% of normals. Here that finding is generalized to four public source corpora with a principled protocol.

Protocol. *Leave-one-corpus-out:* train on the three other source corpora (plus validation from the held-in corpora), test on the held-out one. For the CIDRZ target all four source corpora are pooled. CIDRZ is never used for method or hyper-parameter selection; hyper-parameters were chosen per backbone on the four public folds only (pre-declared rule, Section 6). Metric: AUROC with paired-bootstrap 95% confidence intervals on seed-averaged scores; five seeds per cell for the headline comparison (two seeds for the DG-loss exploration and the 448 px arm).

2 Data

All images are standardized to 224×224 uint8 (per-corpus arrays and metadata under `data/tb_multi`). TBX11K is restricted to active TB versus healthy and sick-non-TB; latent and uncertain cases are excluded.

Corpus	n (after dedup)	TB prevalence	Notes
Qatar	4187	16.4%	large public CXR set
NLM (Montgomery+Shenzhen)	138+661 = 799	49%	two US/China screening sets merged
TBX11K	8029	8.2%	lesion bounding boxes available
Pakistan (Mendeley v3)	2973	82.9%	∼256 px JPEG, TB-dominant
CIDRZ / Google Health AI	364	16.5% (60 pos.)	text + audio + CXR; external target
VinDr-CXR (PhysioNet)	2146 sampled	30.1% in sample (3.6% weighted)	radiologist labels; external target (Section 11)

Table 1: Corpora used. Dedup counts follow Section 3.

3 Cross-corpus duplicate check

Near-duplicates across corpora would inflate LOCO scores. We combine a 64-bit DCT perceptual hash (Hamming ≤ 8) with a 64×64 pixel correlation (≥ 0.97). The detector was calibrated with a positive control: TBX11K ships 400 known Montgomery/Shenzhen copies, recovered essentially 100% (329 + 73 pairs; embedding cosine p5 = 0.997).

The scan found 293 pairs, all within-corpus except 12 Qatar×TBX11K, with no label conflicts. The drop list removed 231 TBX11K, 35 Pakistan, 13 Qatar and 1 Shenzhen images (CIDRZ is never dropped). Importantly, *no* Qatar–NLM near-duplicates exist: best pixel correlation 0.948 (copies would be ∼1.0) and best xrv-embedding cosine 0.978 against 0.973 for a null corpus pair, confirmed visually (Figure 1). This contradicts an earlier assumption that Qatar's TB class pools Montgomery and Shenzhen. VinDr-CXR was checked separately against every other corpus (`dedup_external.py`): 0 cross-corpus pairs, one within-VinDr pair with identical labels (kept), and a positive control of 105 perturbed copies of source images was recovered 105/105. Limitation: heavy crops, flips or mirroring are not covered.




Figure 1: Closest Shenzhen/Montgomery → Qatar matches: different patients.

4 Shortcut audit

Using frozen torchxrayvision DenseNet121 features and trivial controls: the corpus can be identified from embeddings with 96.8% accuracy, and corpus id alone predicts the TB label with pooled 5-fold AUROC 0.831 (because prevalence ranges from 8% to 83%). Figure 2 shows that a classifier seeing only the lung silhouette, or only outer border pixels, reaches AUROC 0.84 (silhouette) and 0.78 (border) on Qatar, and 0.80 / 0.76 on Pakistan. Qatar and Pakistan held-out scores are therefore largely explainable by acquisition artefacts; they should be read with caution, and Pakistan is the least trustworthy target. TBX11K and CIDRZ are less affected (silhouette 0.60 / 0.61). On VinDr-CXR the source-trained controls give 0.62 (mask area), 0.64 (silhouette) and 0.67 (border) on the unweighted sample, and 0.72 weighted (Section 11).




Figure 2: LOCO AUROC of shortcut controls and of a frozen xrv probe.

5 Methods

Backbones. torchvision ViT-B/16 (ImageNet), RAD-DINO (`microsoft/rad-dino`, a CXR foundation model; 224 px, with a 448 px arm), ResNet50 and xrv-DenseNet121. Each backbone's learning rate and unfrozen depth were tuned on the four public LOCO folds (Section 6): ViT lr 10^{-4}, 6 blocks; RAD-DINO lr 10^{-4}, 6 blocks; ResNet50 lr 10^{-3}, layer3+4; xrv-DenseNet lr 10^{-3}, denseblock4. The earlier ViT DG-loss exploration used lr 5×10^{-5} from a CIDRZ-informed sweep.

Training recipes. ERM; source-balanced sampling (sbal); GroupDRO; V-REx; CORAL; DANN; per-domain feature standardization (dstd); lung masking (PSPNet, lungs + hila, dilated 6 px); strong GPU augmentation (affine, gamma/contrast, resolution jitter, border erasing); MixStyle ($p = 0.5$, $\alpha = 0.1$, on activations entering the trainable tail); and TBX11K lesion-box supervision (an auxiliary patch-token loss).

Multimodal CIDRZ. Image experts are trained without CIDRZ; text and audio models are fit within cross-validation (10× repeated 5-fold). Fusion variants: stacking, concatenation, reliability-weighted (relw), and a modality-dropout MLP (MD-MLP), evaluated with paired bootstrap and modality-missing tests.




Figure 3: Lung-hilum masks (green) on CIDRZ examples; everything outside is set to black.

6 Tuning protocol

A grid of 24 configurations per backbone family (96 runs, seed 0) was evaluated on Qatar, NLM, TBX11K and Pakistan; CIDRZ was never used. The pre-declared rule picks the best mean AUROC over the four folds and, among configurations within 0.01, the one with fewer unfrozen blocks and lower learning rate (`results/tuning_protocol.md`). Tuning gains over the untuned configurations were small (at most +0.034; RAD-DINO +0.007), so the ranking of backbones was not changed. Selecting on NLM + TBX11K only, because Pakistan and Qatar are artefact-prone, picks different configurations for RAD-DINO, ResNet50 and xrv-DenseNet (`results/tuning_summary.md`); we report the primary rule.

7 Results: LOCO generalization




Figure 4: Held-out AUROC, five-seed means, tuned backbones. CIDRZ and VinDr-CXR (weighted AUROC) are external.

Method (5 seeds)	Qatar	NLM	TBX11K	Pakistan	CIDRZ	VinDr†	LOCO mean
RAD-DINO + MixStyle	.967	.909	.923	.808	.805	.917	**.902**
RAD-DINO (sbal+aug+lung)	.968	.916	.913	.794	.809	.903	.898
ViT + MixStyle	.951	.862	.874	.783	.790	.846	.867
ViT (sbal+aug+lung)	.942	.854	.869	.801	.804	.870	.867
ResNet50	.932	.802	.862	.669	.772	.864	.816
xrv-DenseNet121	.913	.771	.820	.732	.807	.825	.809
ViT sbal, whole image, no aug	.891	.745	.689	.855	.751	.848	.795
ViT ERM, whole image	.863	.760	.685	.870	.777	.837	.795

Table 2: Final LOCO AUROC, mean of five seeds (tuned lr). Standard deviations across seeds are 0.005–0.03 on the public folds, up to 0.12 for ResNet50 on Pakistan (`results/loco_final_summary.md`). "LOCO mean" averages the four public corpora. †VinDr-CXR: weighted AUROC, mean of five seeds, trained on the four public sources only (Section 11); radiologist labels.

vs ViT (sbal+aug+lung), paired bootstrap	Qatar	NLM	TBX11K	Pakistan	CIDRZ
RAD-DINO	+.024 [.014,.033]	+.067 [.045,.090]	+.043 [.033,.053]	−.006 [−.021,.010]	+.015 [−.028,.060]
ResNet50	−.003	−.044	−.003	−.121	−.020
xrv-DenseNet	−.028	−.091	−.052	−.072	+.011
ViT ERM, whole image	−.074	−.094	−.187	+.073	−.021

Table 3: Seed-averaged AUROC differences against tuned ViT with 95% CIs (RAD-DINO row; the others' CIs are in the summary file).

- **Preprocessing beats loss engineering.** Lung masking, source-balanced sampling and strong augmentation raise ViT over ERM by +0.074 (Qatar), +0.094 (NLM) and +0.187 (TBX11K) but lower Pakistan by 0.073, where ERM is best (0.870). Pakistan is the artefact-dominated corpus, so a method that scores well there is not necessarily better. In the two-seed exploration, GroupDRO, V-REx, CORAL and DANN did not beat source-balanced sampling on average (LOCO means 0.78–0.81; Table 4).
- **A foundation backbone is the largest gain on the public corpora.** RAD-DINO beats tuned ViT on Qatar, NLM and TBX11K (CIs exclude 0), with LOCO mean 0.898 vs 0.867; ResNet50 and xrv-DenseNet are 0.05–0.09 lower.
- **On CIDRZ no backbone is distinguishable.** All sit at 0.77–0.81 and every paired difference has a CI that includes 0. The CIDRZ interval for RAD-DINO vs ViT ($[-.028, +.060]$) includes the VinDr-sized difference; on VinDr-CXR (646 positives, Section 11) the advantage reappears.
- **MixStyle** is not a reliable gain: RAD-DINO +0.004 mean, ViT ±0.00 (Pakistan −0.015, CIDRZ −0.013).
- **Lesion supervision** (two seeds, exploration, earlier code): no gain for the CLS score for ViT or RAD-DINO; see `results/loco_summary_all.md`. Not re-run.

Method (ViT, 2 seeds)	Qatar	NLM	TBX11K	Pakistan	CIDRZ	LOCO mean
ViT sbal+aug+lung+dstd	.911	.811	.862	.811	.795	.849
ViT sbal+lung	.931	.814	.842	.752	.796	.835
ViT sbal+aug	.922	.807	.800	.708	.794	.809
CORAL	.854	.813	.708	.852	.683	.807
DANN	.798	.765	.740	.898	.712	.800
V-REx	.887	.707	.732	.820	.767	.786
GroupDRO	.885	.756	.724	.776	.761	.785
ViT sbal (no preprocessing)	.881	.766	.711	.769	.764	.782

Table 4: Earlier two-seed ViT-B/16 exploration of recipes and DG losses (lr 5×10^{-5}; not retuned; ViT code paths were unaffected by the checkpoint bug).

Resolution: RAD-DINO at 448 px

RAD-DINO was trained on a native 448 px cache (all corpora re-standardized from the originals, CIDRZ from its DICOMs; two seeds, same tuned configuration). Mean AUROC at 448 vs 224 px: Qatar 0.971 vs 0.965, NLM 0.951 vs 0.917, TBX11K 0.946 vs 0.907, Pakistan 0.932 vs 0.815, CIDRZ 0.824 vs 0.820. Resolution helps on the public folds but gives no external gain; the large Pakistan jump cannot be attributed to pathology given that corpus's artefact structure, and we do not treat it as evidence of better transfer.

8 Artefact-adjusted AUROC

Because artefacts predict labels within corpora, we stratify target cases and controls by predicted-artefact score (silhouette, border and area features; `datasets/tb_multi/artefact_adjust.py`) and average AUROC within strata, with the artefact predictor fit either on the source corpora (src) or by cross-validation within the target (cv; strongest control). For RAD-DINO (five seeds):

Target	raw	artefact-only (cv)	stratified (cv) [95% CI]	pairs kept
Qatar	0.974	0.972	0.935 [0.915,0.954]	3%
NLM	0.936	0.855	0.898 [0.871,0.922]	6%
TBX11K	0.923	0.903	0.889 [0.873,0.901]	7%
Pakistan	0.811	1.000	0.828 [0.762,0.896]	1%
CIDRZ	0.825	0.612	0.816 [0.752,0.889]	10%

Table 5: Artefact-stratified AUROC, RAD-DINO.

On CIDRZ, where artefacts are weakly predictive (0.61), adjustment changes AUROC by 0.009: the 0.8 AUROC is not an artefact effect. On Qatar, NLM and TBX11K the drop is 0.03–0.04, so the models add

signal beyond the artefact predictor, but the artefact-only baselines are strong (0.86–0.97), so raw AUROCs there overstate generalization. On Pakistan the artefact-only AUROC is 1.000 and only 1% of pairs survive stratification: the label is confounded with the source, the adjusted number is not identifiable, and Pakistan should not be used as evidence of generalization.

9 Results: multimodal fusion on CIDRZ

Table 6 and Figure 5 use the final image expert: the corrected RAD-DINO and tuned ViT, five seeds each, ensembled, trained on all four source corpora and never on CIDRZ (image alone AUROC 0.827; 0.824 for RAD-DINO alone). The text and audio experts and the fusion layers are fit in cross-validation on CIDRZ.




Figure 5: CIDRZ AUROC for single modalities and fusion arms (RAD-DINO + ViT ensemble expert).

Arm	AUROC	Δ vs same method without image [95% CI]
audio	0.610	–
text	0.782	–
text + audio (stack)	0.775	–
image alone	0.827	–
text + image, stack / relw / concat	0.844 / **0.851** / 0.843	–
text + audio + image, stack	0.838	+0.058 [+0.013, +0.103]
text + audio + image, relw	0.847	+0.060 [+0.021, +0.098]
text + audio + image, concat	0.816	+0.045 [+0.017, +0.073]
text + audio + image, MD-MLP	0.800	+0.051 [+0.022, +0.081]

Table 6: CIDRZ fusion (10× repeated 5-fold CV).

- Adding the image expert improves text + audio by +0.045 to +0.060 AUROC for every fusion method, with 95% CIs excluding 0. Reliability-weighted text + image is best (0.851 vs text 0.782).
- Audio contributes nothing: text + audio (0.775) is not better than text (0.782), and text + image beats text + audio + image.
- The gain depends on expert quality: with the plain ERM image expert (0.760) the text + audio + image gain is +0.026 [−0.015, +0.062], not significant. With the corrected RAD-DINO (0.824) and ViT/RAD-DINO ensemble (0.827) experts it is +0.062 and +0.060. (The earlier RAD-DINO result, +0.052, came from the pre-fix expert.)
- MD-MLP trained with modality dropout degrades gracefully when audio is missing (0.825) but drops to about 0.74–0.75 when the text or image is missing.
- CIDRZ appears capped near 0.80–0.85 AUROC (the 0.01 higher image expert barely moves fusion) for these experts; with 60 positives the CIs are wide (±0.04–0.05).

10 Calibration, WHO operating points and local recalibration

Training used source-balanced cells, so logits are not prevalence-calibrated. Every expert is over-confident on held-out corpora (logistic calibration slope 0.2–0.6; ECE 0.04–0.24 on public corpora and 0.49 on CIDRZ for RAD-DINO, 0.27 for ViT, 0.57 for ERM). On CIDRZ the raw Brier score (0.22–0.45) is worse than the constant-prevalence predictor (0.138) although AUROC is ≈0.8: ranking transfers, probabilities do not. The 90%-sensitivity threshold chosen on held-in validation data gives target sensitivity 0.54–0.77 and specificity 0.76–1.00 (RAD-DINO, CIDRZ: 0.71/0.76), i.e. the source-trained threshold is too conservative on the new site.




Figure 6: Calibration of the final RAD-DINO model on each held-out corpus.

RAD-DINO (5 seeds)	Qatar	NLM	TBX11K	Pakistan	CIDRZ
ECE	.035	.121	.237	.197	.487
Slope	.56	.44	.27	.20	.27
Transfer sens./spec.	.75/1.00	.56/.98	.55/.96	.67/.83	.71/.76

Table 7: Calibration of the final RAD-DINO model.

WHO triage profile and local recalibration. The WHO target product profile for a triage test is at least 90% sensitivity and at least 70% specificity. We ask whether a small labelled local sample ($k = 25$–400 cases) can adapt the model: local Platt scaling, intercept-only recalibration, and a local 90%-sensitivity threshold, repeated 300 times (`local_recalibration.py`). Findings:

- *Calibration is repairable.* For RAD-DINO on CIDRZ, ECE falls from 0.50 to 0.11 (intercept only) and 0.06–0.08 (Platt) with 25–200 labelled cases; ViT from 0.25 to 0.07–0.09.
- *Discrimination is not.* On CIDRZ the oracle operating points are sensitivity 0.82 at 70% specificity, or specificity 0.50 at 90% sensitivity (RAD-DINO; ViT 0.78 and 0.39). No threshold meets the profile; the probability that a locally chosen threshold does is 0.00 for every k. With a 50-case local sample the local 90%-sensitivity threshold delivers 0.80 sensitivity and 0.61 specificity.
- *On public corpora the profile is reachable* (oracle 0.94 sens. at 77% specificity on TBX11K, 0.97/0.96 on Qatar), but a local threshold from 50–200 cases meets it only 10–40% of the time because the target sensitivity is estimated from few positives; Pakistan (oracle specificity 0.25 at 90% sensitivity) never does.

A model can therefore be well calibrated at a site and still fall short of the clinical target; local labelled data fix probabilities and thresholds, not ranking quality. The fused text + image score reaches AUROC 0.85 on CIDRZ; we did not evaluate its operating points against the profile.

11 Second external cohort: VinDr-CXR

Cohort, labels and sampling. VinDr-CXR 1.0.0 (PhysioNet, credentialed access) has 18 000 postero-anterior radiographs (15 000 train, 3 000 test) annotated by radiologists for 28 findings. TB-positive is the radiologists' "Tuberculosis" finding: at least two of three reads on the training split (482 images; 247 unanimous; 268 images with exactly one TB read are excluded as ambiguous) and the consensus label on

the test split (164). These are image read-outs, not microbiological confirmation. We downloaded all 646 positives and 1 500 sampled TB-free images (1 000 train, 500 test; half "No finding", half other findings), 2 146 DICOMs (12% of the release). Because the sample is enriched (30.1% TB; the release has about 3.6% among labelled-TB and TB-free images) every image carries the inverse of its sampling probability (21.2 and 7.3 for training "No finding"/other, 8.2 and 3.1 for the test split, 1 for positives), and all metrics below are weighted unless stated; unweighted AUROCs are about 0.01 lower. MONOCHROME1 DICOMs (25%) were inverted, JPEG 2000 decoded, and the same min–max/resize and lung masking as the other corpora applied (mask area 0.41, 2 failures). The eight tuned configurations of Table 2 were retrained with VinDr as the held-out target (five seeds, four public sources, same held-in split), and their CIDRZ AUROCs reproduce Table 2 to three decimals for the six non-RAD-DINO configurations.

Method (5 seeds)	per-seed AUROC	seed-avg [95% CI]	vs "No finding"	vs other findings	Δ vs ViT [95% CI]
RAD-DINO + MixStyle	.917±.018	.935 [.924,.947]	.950	.894	+.048 [.034,.064]
RAD-DINO	.903±.013	.919 [.905,.932]	.930	.888	+.032 [.018,.049]
ViT (tuned)	.870±.016	.887 [.870,.903]	.899	.854	–
ResNet50	.864±.019	.882 [.865,.897]	.894	.844	−.005 [−.019,.009]
ViT sbal, whole image	.848±.027	.872 [.853,.891]	.893	.809	−.015 [−.031,.002]
ViT + MixStyle	.846±.020	.864 [.846,.882]	.874	.837	−.023 [−.028,−.016]
ViT ERM, whole image	.837±.032	.866 [.846,.885]	.882	.813	−.021 [−.038,−.004]
xrv-DenseNet121	.825±.015	.836 [.814,.854]	.840	.822	−.051 [−.068,−.036]

Table 8: VinDr-CXR, weighted AUROC. CIs and paired differences use a bootstrap that resamples within split × label × stratum cells (`results/vindr_eval.md`, `results/vindr_paired.md`).

- **The foundation-model advantage is confirmed externally.** On CIDRZ no backbone differed; on VinDr RAD-DINO beats tuned ViT by +0.032 and RAD-DINO + MixStyle by +0.048 with CIs that exclude 0, ResNet50 is equal to ViT, and xrv-DenseNet is lowest. MixStyle helps RAD-DINO (+0.016) but hurts ViT (−0.023), so it remains an unreliable gain.
- **Hard negatives matter.** RAD-DINO AUROC is 0.930 against "No finding" negatives but 0.888 against negatives with other findings; it is 0.932 on training-split rows (three-reader majority) and 0.882 on test-split rows (consensus). It scores positives without an active-pattern finding (no cavity, infiltration, consolidation or nodule flagged; 403 of 646) as well as those with one (0.906 vs 0.914).
- **Artefacts are not negligible but do not explain the score.** A silhouette/border artefact predictor reaches weighted AUROC 0.72 (trained on the sources) or 0.79 (cross-validated inside VinDr), well above its 0.61 on CIDRZ. Stratifying on it (10 bins, 8% of pairs kept) moves RAD-DINO from 0.919 to 0.899 (MixStyle 0.935 → 0.916; ViT 0.887 → 0.846), and RAD-DINO adds +0.144 AUROC over the artefact score in cross-validation.

Method	oracle sens.	oracle spec.	transfer sens./spec.	PPV	ECE	slope
RAD-DINO + MixStyle	0.95	0.79	.64/.98	0.52	.212	.53
RAD-DINO	0.91	0.73	.59/.97	0.44	.248	.47
ViT (tuned)	0.88	0.65	.55/.95	0.30	.130	.49
ResNet50	0.87	0.65	.57/.94	0.26	.153	.52
xrv-DenseNet121	0.79	0.52	.60/.90	0.18	.282	.67

Table 9: VinDr-CXR operating points against the WHO triage profile (sensitivity ≥90%, specificity ≥70%) and calibration of the seed-averaged weighted scores. Oracle sens.: sensitivity at 70% specificity; oracle spec.: specificity at 90% sensitivity. "Transfer" is the threshold giving 90% sensitivity on held-in validation data. PPV is at 3.6% prevalence; ECE and slope are for the raw probabilities.

WHO profile and calibration. With an oracle threshold the two RAD-DINO models reach the profile on VinDr (sensitivity 0.91 at 70% specificity, specificity 0.73 at 90% sensitivity), which no model did on CIDRZ (0.82 and 0.50). The deployable zero-label threshold does not: the held-in 90%-sensitivity threshold delivers sensitivity 0.59–0.64 at specificity 0.97–0.98 (too conservative, as on CIDRZ). A local threshold set on $k = 50$–400 labelled cases meets the profile with probability 0.04–0.17 for RAD-DINO and 0.19–0.51 for RAD-DINO + MixStyle, and 0.00 for the other backbones. Raw probabilities are badly calibrated at the true prevalence (ECE 0.13–0.28, slope about 0.5, intercept −2.8 to −4.0); local Platt scaling with $k = 50$–400 cases, fitted with the sampling weights, brings ECE to 0.017–0.008 (intercept-only 0.023–0.015) and Brier from 0.19 to 0.02, again repairing calibration but not discrimination. *These operating-point results*

are optimistic: the WHO profile is defined against a microbiological reference, whereas VinDr's reference is radiologist reading, which an image model trained on radiologist-labelled corpora can partly imitate; CIDRZ, with a microbiological reference, is the stricter test and fails it.

12 Limitations

- There are two documented external cohorts. CIDRZ has a microbiological reference but only 60 positives (CIs ±0.04–0.07); VinDr-CXR has 646 positives but its TB labels are radiologist read-outs (a three-reader majority on the training split, a consensus on the test split), so it measures agreement with readers, and its evaluation sample is enriched for TB and re-weighted. An additional public cohort (India CXR set) was rejected because its labels are "Abnormal" rather than TB (`results/tb_multi_india_report.md`); PadChest and NIAID TB Portals were not obtained.
- VinDr's "Tuberculosis" label may include old or inactive disease, which the source corpora label differently; fused text/audio experts do not apply to VinDr; one within-VinDr near-duplicate was kept; RAD-DINO re-training is not bit-reproducible (its CIDRZ AUROC moved by −0.019 and −0.008 for the two RAD-DINO variants between the main and the VinDr runs, within the seed standard deviation; the other six configurations reproduced to three decimals).
- Tuning used seed 0 on four folds, two of them artefact-prone; gains were small. The 448 px arm has two seeds. Lesion supervision and DG losses were not re-run with the tuned configuration or five seeds.
- Qatar and Pakistan labels are partly predictable from artefacts; artefact adjustment is a bound, not a causal fix, and is not identifiable on Pakistan.
- Deduplication does not cover heavy crops or flips. Calibration and recalibration were assessed on image-expert scores, not on fused probabilities. The UAR column of the fusion tables thresholds raw scores at 0 and is not meaningful for relw; use AUROC.

13 Conclusions and recommendations

1. Audit before training: corpus identity and silhouette/border shortcuts explain much of the apparent performance on some corpora. Report LOCO, not pooled, results, with artefact-only and artefact-stratified AUROC alongside.
2. Use a CXR-pretrained backbone (RAD-DINO) with lung masking and strong augmentation: LOCO mean 0.898 against 0.867 for tuned ViT. On the external CIDRZ cohort (60 positives) all backbones are indistinguishable (0.77–0.81), but on VinDr-CXR (646 positives) RAD-DINO leads tuned ViT by +0.03 to +0.05 (paired CIs exclude 0).
3. Classical DG losses and MixStyle do not help beyond source-balanced sampling.
4. For multimodal TB screening, fuse a source-trained CXR expert with the text/clinical expert (reliability-weighted: 0.851 vs 0.782); audio did not help in this cohort.
5. Recalibrate per site: a few dozen labelled local cases fix probabilities (ECE 0.50 → 0.07) but not discrimination; on CIDRZ no model meets the WHO triage profile (90% sensitivity, 70% specificity), and on VinDr only the RAD-DINO models reach it, with an oracle threshold and radiologist labels.
6. Follow the audit protocol in Appendix A and say which numbers were superseded after any bug fix.

A Cross-corpus audit protocol

Each item is tied to a measured result in this report (full version: `results/audit_protocol.md`).

1. **Deduplicate across corpora before splitting**, with a positive control (400 planted TBX11K copies recovered).
2. **Evaluate leave-one-corpus-out**, not random splits: rankings change by corpus (ViT ERM 0.87 on Pakistan, 0.69 on TBX11K).
3. **Run source-prediction and artefact-only controls**: artefact-only AUROC 0.64–1.00 within corpus.
4. **Report artefact-stratified AUROC beside raw AUROC** (RAD-DINO Qatar 0.974 → 0.935; CIDRZ 0.825 → 0.816).

5. **Declare a corpus non-identifiable when artefacts predict the label** (Pakistan: 1.000, 1% of pairs kept) and do not cite it as evidence of generalization.
6. **Validate the labels of every added cohort** (the India set was "Abnormal", not TB; rejected; VinDr's labels are radiologist reads and are reported as such, with sampling weights for the enriched sample).
7. **Fix the tuning protocol before looking at the external target**; report alternative selections.
8. **Use ≥5 seeds and paired bootstrap CIs**; claim only what excludes zero (all CIDRZ differences include zero).
9. **Check checkpoint restore and train/eval parity** (a partial restore moved ResNet50 AUROC by 0.026).
10. **Check resolution after the artefact audit** (448 px: +0.03–0.04 on NLM/TBX11K, +0.004 on CIDRZ).
11. **Report calibration separately from discrimination** (CIDRZ ECE 0.49 → 0.07 after local Platt).
12. **Report against the clinical target with an oracle threshold** (WHO triage: 90%/70%; CIDRZ ceiling 0.82 sens. at 70% spec.; VinDr RAD-DINO 0.91, with radiologist labels).
13. **Keep pre-fix and post-fix runs in separate tables and say which numbers were superseded.**

Reproducibility

Code in `datasets/tb_multi/` (`standardize`, `dedup`, `embed`, `masks`, `shortcut_audit`, `loco_train`, `summarize_loco`, `fusion_study`, `calibration`, `download_vindr`, `dedup_external`, `vindr_report`, `vindr_paired`); outputs in `results/` (`loco_final/*.json`, `loco_final_summary.md`, `loco_448/`, `tuning_summary.md`, `artefact_adjusted.md`, `local_recalibration.md`, `fusion_study_corrected.md`, `calibration_final.md`, `audit_protocol.md`, `loco_vindr/*.json`, `vindr_eval.md`, `vindr_paired.md`, `vindr_selection.csv`). Dependencies: `pip install ".[tb_multi]"` (CUDA torch installed separately).