Certificate No. A7145

Opti-Dry
Certificate of Calibration

The equipment described below has been tested in accordance with the requirements of EN26330, BSEN ISO 6330, AATCC 135, MARKS & SPENCER PLC, NEXT PLC using calibration equipment and a calibration system traceable to international standards.

Customer Details Sell Quick

Sales Order No. 24587

Equipment Details

Manufacurer: Roaches International Ltd **Type:** MK4
Machine: Opti-Dry **Serial No:** 21840

Table of Measurements - For: Opti-Dry

Item	Required Measurement	Recorded Measurement
Internal Temp	40 ºC ± 2 ºC	40.0 ºC
External Temp	40 ºC ± 2 ºC	39.5 ºC
Internal Temp	70 ºC ± 2 ºC	70.0 ºC
External Temp	70 ºC ± 2 ºC	69.6 ºC
Timer Accuracy	Cooling 5.00 Minutes ± 5 Seconds	5.00
	Heating 30.00 Minutes ± 5 Seconds	30.00
Drum Rotation Speed (No Load)	52 RPM ± 2 RPM	51.0 RPM

Calibration Equipment Details
The equipment was checked using the following apparatus:

Digital Thermometer **RC 053** Due Date: Mar-26
Digital Tachometer **RC 066** Due Date: Mar-26

Name: Wayne Gibbons **Calibration Due Date:**
Position: Service Technician 05/11/2026
Date of Calibration: 05/11/2025

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.



