

actiCHamp TurboLink Bossdevice接続について

Recorder PC Setting

④ Use the following IP address:

IP address:

192.168.5.1


Subnet mask:

255.255.255.0

Default gateway:




On the Recorder PC,
open a browser and go to 192.168.5.21:8000



RecorderPC



actiCHamp/Plus




MatLab PC Setting

IP Address: 192.168.7.10
Subnet mask: 255.255.255.0
Default gateway:
DNS server:



Bossdevice



IP 192.168.200.255
(srltExplorerにて設定)
IP 192.168.7.5
(デフォルト)

TurboLink License
TLxxxx

must be in the 192.168.200 subnet

BOSSDEVICE RESEARCH Pre-requisites

- Windows 10 operating system
- MatLab R2021a
- Simulink for MatLab R2021a update release 5
- Simulink Coder add-on
- Simulink Real-time add on
- Simulink Real-Time target support package (add on)
- The bossdevice RESEARCH API (download from website).
 - API には、デバイスを制御するために必要な Matlab クラスと、デバイスで動作するように実装できるコードの例としていくつかの MatLab デモ スクリプトが含まれています。Bossdevice RESEARCH ファームウェア。sync2brain.com/downloads からアクセスできます。
 - Hardware drivers named as Speedgoat I/O Blockset R2021a (v9.2.2.1) drivers, download from website.
- Optional: a screen (recommended in case of troubleshooting)