

TABLE I. TOF SENSOR PERFORMANCE UNDER DIFFERENT MEDICATION PLACEMENT CONDITIONS

Test Condition	Sensor 1	Sensor 2	Sensor 3	Sensor 4	Average Sensitivity
2 Medicine (Neat)	100%	100%	100%	100%	100%
2 Medicines (Scattered)	0%	0%	0%	0%	0%
4 Medicines (Neat)	100%	100%	100%	100%	100%
4 Medicines (Scattered)	100%	100%	100%	90%	98%

TABLE II. OVERALL PERFORMANCE

Overall Performance	Value
Total Samples	1000
True Positive(TP)	745
False Negative(FN)	255
Overall Sensitivity	74,5%




Fig. 3. System flowchart

TABLE III. STEPPER MOTOR PERFORMANCE

Compartment	Successful Trials	Failed Trials	Success Rate
1	20	0	100%
2	20	0	100%
3	11	9	55%
4	18	2	90%
5	15	5	75%
Average	84	16	85%



Fig. 4. Appearance of Smart Pillbox System