

BOOKING SHEET FOR BOYA RANGE BARCODE STAFF CALIBRATION

- 1) LEVEL BUBBLE CHECKED? (Circle one) YES/NO ADJUSTED? YES/NO
- 2) STAFF BUBBLE CHECKED? (Circle one) YES/NO ADJUSTED? YES/NO
- 3) COLLIMATION TEST OF DIGITAL LEVEL

See Boya Range plan for location of collimation test spikes 1- 4

WEATHER: CLOUDY
COOL.

Use instrument internal software to calculate and store collimation constants

DATE/TIME	MONDAY 15/6/2020 9:23am.	17.3°C	17.1°C	15.9°C	16.5°C.
INSTRUMENT MAKE/MODEL SERIAL #	LEICA LS15 # 702272				
STAFF MAKE/MODEL SERIAL#	INVAR STAFF # 26296.				
COLLIMATION DIFFERENCE	4.3"	Old: -25.3"	New: -21.0"	RETICLE: 1.65304	
ACCEPT?/STORE?	YES.				

4) RANGE OBSERVATION DETAILS

See Boya Range plan for location of Pillar MV 83 and Pins 1- 21

Set digital level to record most precise readings with multiple observations, means and standard deviations if possible. Shade instrument unless overcast.

Book field information in these tables to complete documentation for calibration.

Observe a BS on Pin 1, IS on Pins 2-20 and a FS on Pin 21 for 1 complete set for a 4 m staff.
Observe a BS on Pin 1, IS on Pins 2-14 and a FS on Pin 15 for 1 complete set for a 3 m staff.

If instrument can't mean multiple observations, do at least 3 complete sets for redundancy.

DATE/TIME	MONDAY 15/6/2020 9:39am.	MONDAY 15/6/2020 10:08am.		
TEMPERATURE START/FINISH	ST: 16.3°C 16.7°C. FN: 16.7°C 16.3°C.	ST: 16.5°C 16.8°C. FN: 16.4°C 16.7°C.		
INSTRUMENT MAKE/MODEL SERIAL #	LEICA LS15 # 702272.	LEICA LS15 # 702272		
STAFF MAKE/MODEL SERIAL#	INVAR STAFF # 26296.	INVAR STAFF # 26296		
#OBS FOR MEAN (minimum 5)	10	10		
NUMBER OF SETS	1	1		

pin # 1-15

pin # 7-21.

I certify that the above observations were made by me at the Boya barcode staff calibration range.

SURVEYOR..... V.UNG

DATE..... 15/6/2020.

#1
pin 100.000
pin # 21. 96.19699.
3.80301.