

## Test Report

Test bench serial number: 0183  
 Transducer type: WIKA S-20  
 Ordering number: -  
 Range: 0-1000 bar  
 Output signal: 4-20 mA  
 Power supply: -  
 Transducer serial number: 1A01DGO34I1  
 Digital indicator: PLC+PC  
 Line: B  
 Station: 1

Test report n. **14\_07\_2022\_TR\_00000664**  
 Customer: JC FABRICA DE VALVULAS S.A.U

### Certificate

**LAT 093 170719P**

### Climbing measure

Pressure	Measure	Error limits measured (%)
0.00	0.00	0.00
100.00	100.00	0.00
200.00	200.00	0.00
300.00	300.00	0.00
500.00	500.00	0.00
700.00	700.00	0.00

### Downhill measure

Pressure	Measure	Error limits measured (%)
700.00	700.00	0.00
500.00	500.00	0.00
300.00	299.00	-0.10
200.00	200.00	0.00
100.00	100.00	0.00
0.00	0.00	0.00

Date: 14/07/2022

Firma HPSC:

