

Test Report

Test bench serial number: 0183
Transducer type: WIKA S-20
Ordering number: -
Range: 0-1000 bar
Output signal: 4-20 mA
Power supply: -
Transducer serial number: 1A01DGO8IIK
Digital indicator: PLC+PC
Line: A
Station: 1

Test report n. **14 07 2022 TR 00000659**
Customer: JC FABRICA DE VALVULAS S.A.U

Certificate

LAT 093 170719P

Climbing measure

Pressure	Measure	Error limits measured (%)
0.00	0.00	0.00
100.00	100.00	0.00
200.00	200.00	0.00
300.00	300.00	0.00
500.00	501.00	0.10
700.00	701.00	0.10

Downhill measure

Pressure	Measure	Error limits measured (%)
700.00	701.00	0.10
500.00	500.00	0.00
300.00	301.00	0.10
200.00	200.00	0.00
100.00	100.00	0.00
0.00	0.00	0.00

Date: 14/07/2022

Firma HPSC:


