

Discussion

Comparisons of ^{67}Ga and ^{201}Tl

Details

- Gallium-67
 - 3.2613 (5) d
 - 100% EC
 - 184.576 keV γ
 - Hodgkin's disease
 - Lymphoma
 - Bronchogenic Cacinoma
 - Infection Imaging
- Thallium-201
 - 3.0421 (17) d
 - 100% EC
 - 167.45 keV γ
 - Myocardial Perfusion Imaging

Key comparison BIPM.RI(II)-K1.Ga-67

MEASURAND : Equivalent activity of ^{67}Ga

Key comparison reference value: the SIR reference value for this radionuclide x_R is 116.2 MBq, with a standard uncertainty u_R of 0.6 MBq (see Final Report updated in June 2006).

The degree of equivalence of each laboratory with respect to the reference value is given by a pair of terms:

$D_i = (x_i - x_R)$ and U_i , its expanded uncertainty ($k = 2$), both expressed in MBq. With n the number of laboratories,


$U_i = 2[(1 - 2/n)u_i^2 + (1/n^2)\sum u_j^2]^{1/2}$ when each laboratory has contributed to the computation of x_R , see [Metrologia, 42, 140-144](#).

The degree of equivalence between two laboratories is given by a pair of terms:

$D_{ij} = D_i - D_j = (x_i - x_j)$ and U_{ij} , its expanded uncertainty ($k = 2$), both expressed in MBq.

The approximation $U_{ij} \sim 2(u_i^2 + u_j^2)^{1/2}$ is used in the following table.

Lab i ↓			Lab j →															
	D_i	U_i			D_{ij}	U_{ij}			D_{ij}	U_{ij}			D_{ij}	U_{ij}			D_{ij}	U_{ij}
	/ MBq				/ MBq				/ MBq				/ MBq				/ MBq	
CMI-IIR	2.6	2.0			2.8	3.6	2.4	2.3	3.6	2.5	2.6	2.3	3.6	2.4	0.8	3.0	5.0	2.3
NPL	-0.2	2.5	-2.8	3.6			-0.4	2.9	0.8	3.0	-0.2	2.9	0.8	2.9	-2.0	3.5	2.2	2.9
CSIR-NML	0.2	0.7	-2.4	2.3	0.4	2.9			1.2	1.3	0.2	0.9	1.2	1.0	-1.5	2.2	2.6	0.9
OMH	-1.0	1.2	-3.6	2.5	-0.8	3.0	-1.2	1.3			-1.0	1.4	0.0	1.5	-2.8	2.4	1.4	1.4
NIST	0.0	0.9	-2.6	2.3	0.2	2.9	-0.2	0.9	1.0	1.4			1.0	1.1	-1.7	2.2	2.4	1.0
NMIJ	-1.0	0.9	-3.6	2.4	-0.8	2.9	-1.2	1.0	0.0	1.5	-1.0	1.1			-2.8	2.3	1.4	1.1
CIEMAT	1.8	1.9	-0.8	3.0	2.0	3.5	1.5	2.2	2.8	2.4	1.7	2.2	2.8	2.3			4.1	2.2
LNE-LNHB	-2.4	0.8	-5.0	2.3	-2.2	2.9	-2.6	0.9	-1.4	1.4	-2.4	1.0	-1.4	1.1	-4.1	2.2		



Key comparison BIPM.RI(II)-K1.TI-201

MEASURAND : Equivalent activity of ^{201}TI

Key comparison reference value: the SIR reference value for this radionuclide is $x_R = 312.8 \text{ MBq}$

with a standard uncertainty $u_R = 1.3 \text{ MBq}$.

x_R is computed as the mean of the results obtained by primary methods.

The degree of equivalence of each laboratory with respect to the reference value is given by a pair of terms:

$D_i = (x_i - x_R)$ and U_i , its expanded uncertainty ($k = 2$), both expressed in MBq, and with n the number of laboratories

$U_i = 2((1 - 2/n)u_i^2 + (1/n^2)\sum u_j^2)^{1/2}$ when each laboratory has contributed to the reference value (see Final Report dated 21 April 2008)


The degree of equivalence between two laboratories is given by a pair of terms:

$D_{ij} = D_i - D_j = (x_i - x_j)$ and U_{ij} , its expanded uncertainty ($k = 2$), both expressed in MBq.

The approximation $U_{ij} \sim 2(u_i^2 + u_j^2)^{1/2}$ is used in the following table.

			Lab $j \Rightarrow$													
Lab $i \Downarrow$																

BIPM.RI(II)-K1.TI-201
Degrees of equivalence for equivalent activity of ^{201}Tl



Note: The right-hand axis is approximate only