

Standardisations at NPL

Wednesday, 02 September
2009


Lena Johansson
Radioactivity Group, NPL

Introduction



- Tc-99m
- Tl-201
- Cu-64
(on-going)

The nuclear decay



Elution of the generator

- Activity concentration vs. salt concentration
- No Mo-99 impurities (or Nb-92m)
- 9 $\mu\text{g/g}$ NaCl in water
- Mo-99 (0.0010) %; $u = (0.0005)$ %


Standardisation

- $4\pi\beta(\text{LS})-\gamma$ using DCC
- $4\pi\beta(\text{PC})-\gamma$ with HV variation

Efficiencies:

20% - 4%

both methods



Standardisation

$4\pi\beta(\text{LS})-\gamma$ using DCC


7630 kBq/g \pm 0.54%

$4\pi\beta(\text{PC})-\gamma$ with HV variation

7600 kBq/g \pm 0.42%


SIR Tc-99m

Figure Graph of degrees of equivalence with the KCRV for $^{99}\text{Tc}^{\text{m}}$
(as it appears in Appendix B of the MRA)



N.B. The right-hand axis gives approximate relative values only

The nuclear decay



Impurities: Tl-200 (1.088d)
 Tl-202 (12.23d)

Ion chamber calibration factor

^{200}Tl 52.1 pA/MBq $\pm 10\%$ **$52.4 \pm 1.7\%$**

^{201}Tl 6.9537 pA/MBq $\pm 0.5\%$

^{202}Tl 17.2 pA/MBq $\pm 10\%$ **$21.7 \pm 2.3\%$**

Ion chamber calibration factor



Impurities

Determined by gamma-ray spectrometry:

^{200}Tl 0.133 %; $u = 0.003$ %


^{202}Tl 0.206 %; $u = 0.003$ % @ ref date

Source preparation

Solvent : water carrier : 700 $\mu\text{g/g}$ NaCl 1.001 g/cm^3

Strong solution


10.8 MBq/g



BIPM ampoule
VYNS sources
IC ampoules

Weaker solution

1.12 MBq/g




VYNS sources
IC ampoules

Coincidence counting

				WEAK foil		Half lives	3.0421	1.088	12.23	d		
Reference date		07/11/2006 12:00						Impurities ratio				
Activity concentration		1000.00 kBq/g						0.0013146	0.0020595	@ ref. date		
Measurement time		300 s								bkg subtracted		
Source	mass	Start	Finish	t1 (days)	t2 (days)	εβ	N(TI-201)	N(TI-200)	N(TI-202)	N(TI-201)	N(TI-201) _{corr}	
s06210	26.411	13/11/2006 17:18	13/11/2006 17:23	6.22	6.22	0.7161	1374358	142	8212	1512872	1503732	
	26.411	13/11/2006 17:23	13/11/2006 17:28	6.22	6.23	0.7161	1373253	141	8211	1507178	1498067	
	26.411	13/11/2006 17:28	13/11/2006 17:33	6.23	6.23	0.7161	1372149	141	8209	1508712	1499586	
	26.411	13/11/2006 17:33	13/11/2006 17:38	6.23	6.24	0.7161	1371046	141	8208	1503316	1494218	
	26.411	13/11/2006 17:38	13/11/2006 17:43	6.24	6.24	0.7161	1369944	140	8206	1503901	1494794	
S06206	23.193	14/11/2006 10:01	14/11/2006 10:06	6.92	6.92	0.6761	972227	75	6545	1063303	1056111	
	23.193	14/11/2006 10:06	14/11/2006 10:11	6.92	6.92	0.6761	971445	75	6544	1060689	1053511	
	23.193	14/11/2006 10:11	14/11/2006 10:16	6.92	6.93	0.6761	970664	75	6543	1060898	1053714	
	23.193	14/11/2006 10:16	14/11/2006 10:21	6.93	6.93	0.6761	969883	75	6541	1060898	1053710	
	23.193	14/11/2006 10:21	14/11/2006 10:26	6.93	6.94	0.6761	969104	75	6540	1058229	1051055	
s06210	26.411	14/11/2006 10:51	14/11/2006 10:58	6.95	6.96	0.5954	1289576	98	8734	1346167	1337009	
	26.411	14/11/2006 10:58	14/11/2006 11:04	6.96	6.96	0.5954	1288199	98	8732	1345990	1336827	
	26.411	14/11/2006 11:05	14/11/2006 11:11	6.96	6.97	0.5954	1286824	98	8730	1344649	1335488	

Coincidence counting




Coincidence counting

Results


10.799 MBq/g $\pm 0.5\%$ strong solution

10.776 MBq/g $\pm 0.5\%$ weak solution with dil. factor

IC Calibration



SIR TI-201



Ongoing medical nuclides

