

63X\_1\_4\_610

## Microscope info:

| Image                                |                      | Image4_bead3            |             |            |                       |                      |                     |  |
|--------------------------------------|----------------------|-------------------------|-------------|------------|-----------------------|----------------------|---------------------|--|
| image's                              | date                 | 2024-10-17 10:22:37     |             |            |                       |                      |                     |  |
| creation                             | method used          | from file creation date |             |            |                       |                      |                     |  |
| Actual image depth                   |                      | 16                      |             |            |                       |                      |                     |  |
| Microscope type                      |                      | WideField               |             |            |                       |                      |                     |  |
|                                      | NA                   | 1.4                     | 1.4         |            |                       |                      |                     |  |
| Objective                            | im. refractive index | 1.518                   |             |            |                       |                      |                     |  |
|                                      |                      | Wavel                   | engths      |            | sampling (X,Y,Z)      |                      |                     |  |
| Chan                                 | nel(s)               | Ex.<br>(nm)             | Em.<br>(nm) | Saturation | Nyquist (µm)          | Found (µm)           | Nyquist/found ratio |  |
| Channel 0                            |                      |                         | 610.0       | none       | 0.109x0.109x0.<br>328 | 0.063x0.063x0.<br>06 | 0.6, 0.6, 0.2       |  |
| Bead original coordinates(in pixels) |                      | 1193.0,                 | 110.0       |            | ·                     |                      |                     |  |

### Warnings:

(No saturated pixels detected). (All channels sampled following Shannon-Nyquist criterion). (A subresolution bead is used for all channels).

### Resolution table:

| Channel                    | Sig/Backgn<br>d ratio | Dimension | Measured<br>FWHM<br>(µm) | theory (µm) | Fit<br>Goodness | Mes./theory ratio |
|----------------------------|-----------------------|-----------|--------------------------|-------------|-----------------|-------------------|
|                            |                       | X         | 0.28                     | 0.222       | 1.0             | 1.26              |
| Channel 0 (em.<br>610.0nm) |                       | Υ         | 0.278                    | 0.222       | 1.0             | 1.25              |
|                            |                       | Z         | 0.714                    | 0.836       | 0.98            | 0.85              |

Green: within specifications, red: outside specifications (ie. XY ratios above 1.5 or Z ratio above 2.0)

#### Lateral asymmetry ratios:

| Channel                 | Ratio |
|-------------------------|-------|
| Channel 0 (em. 610.0nm) | 0.99  |


## Detailed channel detection info:

## Channel #0





| Channel 0 (em. 610.0nm) |                 |   |       |              |  |
|-------------------------|-----------------|---|-------|--------------|--|
| Sig./Backgnd ratio      | Backgnd LAR Dim |   | FWHM  | Fit goodness |  |
|                         |                 | X | 0.28  | 1.0          |  |
| 12.4                    | 0.99            | Υ | 0.278 | 1.0          |  |
|                         |                 | Z | 0.714 | 0.98         |  |



## Analysis parameters

|                                 | Tool   | PSF Profiler (batch)  |  |  |
|---------------------------------|--|---|--|--|
| Tool &<br>Operator              | Versions   | MetroloJ_QC v1.3.1.1, ImageJ v2.14.0/1.54f, Java v1.8.0_322, OS Mac OS X  |  |  |
|                                 | Operator & date  | SO, October 25, 2024 2:37 PM  |  |  |
| data                            | result folder  | /Users/oggsc/Documents/OM/ImageAnalysis/QC/Elyra/PSFs/20241014/63X_1_4/610//Processed/63X_1_4_610/Image4/bead3/ |  |  |
| data                            | Type of saved data                                     | .pdf, .jpg, .xls  |  |  |
|                                 | Input data bit depth                                   | 16  |  |  |
| Dim                             | ension order   | XY-(C)Z   |  |  |
| Discard s                       | aturated samples                                       | false   |  |  |
|                                 | Bead detection threshold                               | Legacy  |  |  |
|                                 | Center detection method                                | Legacy Maximum Intensity  |  |  |
|                                 | Discard bead if more than one particle are thresholded | true  |  |  |
|                                 | Background annulus<br>thickness in µm                  | 0.5   |  |  |
| Beads                           | Background annulus<br>distance to bead edges<br>in µm  | 0.5   |  |  |
| Deads                           | Multiple beads in image                                | true  |  |  |
|                                 | Bead identification method                             | Using Find Maxima (prominence of 1000.0)  |  |  |
|                                 | Bead size (µm)   | 0.1   |  |  |
|                                 | Bead crop Factor                                       | 5.0   |  |  |
|                                 | Cropped ROI size in µm                                 | 2.31x2.31 (using bead size & background annulus parameters)   |  |  |
|                                 | Bead rejection distance to top/bottom                  | 2.0 μm  |  |  |
| Square Root PSF Image displayed |  | true  |  |  |
|                                 | Applied in this report                                 | true  |  |  |
| Tolerance                       | X & Y FWHM ratios valid if below                       | 1.5   |  |  |
|                                 | Z FWHM ratio valid if below                            | 2.0   |  |  |

# Analysis log

| image name    | creation<br>date       | saturation | sampling<br>density | status   |
|---------------|------------------------|------------|---------------------|----------|
| Image 4_bead3 | 2024-10-17<br>10:22:37 | none       | correct             | analysed |

#### Formulas used:

Lateral  $(res_{x,y}^o)$  and axial  $(res_z^o)$  theoretical resolution values used for widefield microscopes are calculated as defined in Wilhelm, S. Confocal Laser Scanning Microscopy, 2011:

$$res_{x,y}^o = \frac{0.51*\lambda_{em}}{NA}$$
  $res_z^o = \frac{1,77n*\lambda_{em}}{NA^2}$ 

NA: numerical aperture,  $\lambda_{em}$ : emission wavelength, n: refractive index of the lens immersion & mounting media.

Axis profiles are fitted using ImageJ Gaussian Curve Fitter and the following formula  $y = a + (b - a) * e^{\frac{-(x-c)^2}{2d^2}}$  (Gaussian fitting).

Measured lateral and axial resolution (Full Width at Half Maximum, FWHM) values are derived using FWHM =  $2d\sqrt{2ln(2)}$ 

Compliance with the Shannon-Nyquist criterion uses the following formulas for Shannon-Nyquist distances calculation:

$$\alpha = \arcsin(\frac{NA}{n})$$

$$\Delta_{x,y} = \frac{\lambda_{em}}{4.NA} \qquad \Delta_z = \frac{\lambda_{em}}{2.n. (1-\cos{(\alpha)})}$$