

2. Landscape Section

This section must render as A4 landscape with its own header and footer rules.

Equipment	Id	Range	Cal due	Std	Traceability	Location	Status
Balance	EQ-01	0–220 g	2026-11-01	Class E2	NPL	Lab 1	OK
Pipette	EQ-07	1–10 mL	2026-12-15	Gravimetric	NPL	Lab 1	OK
Oven	EQ-11	ambient–300 °C	2027-01-20	PT-100 probe	NPL	Lab 2	OK
pH meter	EQ-03	0–14 pH	2026-10-30	Buffer 4/7/9	NPL	Lab 1	OK
Conductivity	EQ-04	0–200 mS	2026-11-11	KCl std	NPL	Lab 1	OK
Incubator	EQ-09	20–60 °C	2027-02-01	PT-100 probe	NPL	Lab 3	OK