PROCEDURE STEP NO: 1

TIMEBASE ACCURACY

Conditions:

Reference: Internal.

NOMINAL VALUE (MHz)	*LOWER LIMIT dev(Hz)*	*BEFORE ADJ. dev(Hz)*	*AFTER ADJ. dev(Hz)*	*UPPER LIMIT dev(Hz)*	*UNCERTAINTY ±(Hz)*
10	-100	23		100	29

PROCEDURE STEP NO: 2

INPUT RESISTANCE DC

Conditions:

Input Resistance: 1 MOhm

Channel 1

VERTICAL SCALE (V/div)	*LOWER LIMIT (MOhm)*	*BEFORE ADJ. (MOhm)*	*AFTER ADJ. (MOhm)*	*UPPER LIMIT (MOhm)*	*UNCERTAINTY ±(MOhm)*
0.01	0.9900	0.9965		1.0100	0.002
1	0.9900	0.9981		1.0100	0.002

Channel 2

VERTICAL SCALE (V/div)	*LOWER LIMIT (MOhm)*	*BEFORE ADJ. (MOhm)*	*AFTER ADJ. (MOhm)*	*UPPER LIMIT (MOhm)*	*UNCERTAINTY ±(MOhm)*
0.01	0.9900	0.9977		1.0100	0.002
1	0.9900	0.9974		1.0100	0.002

Channel 3

VERTICAL SCALE (V/div)	*LOWER LIMIT (MOhm)*	*BEFORE ADJ. (MOhm)*	*AFTER ADJ. (MOhm)*	*UPPER LIMIT (MOhm)*	*UNCERTAINTY ±(MOhm)*
0.01	0.9900	0.9973		1.0100	0.002
1	0.9900	0.9974		1.0100	0.002

Channel 4

VERTICAL SCALE (V/div)	*LOWER LIMIT (MOhm)*	*BEFORE ADJ. (MOhm)*	*AFTER ADJ. (MOhm)*	*UPPER LIMIT (MOhm)*	*UNCERTAINTY ±(MOhm)*
0.01	0.9900	0.9972		1.0100	0.002
1	0.9900	0.9976		1.0100	0.002

PROCEDURE STEP NO: 3.1

DC GAIN ACCURACY

Channel 1

VERT SCALE (V/div)	*CAL VOLTAGE (V)*	*LOWER LIMIT dev(%)*	*BEFORE ADJ. dev(%)*	*AFTER ADJ. dev(%)*	*UPPER LIMIT dev(%)*	*UNCERTAINTY ±(%)*
100	±300	-1.5	-0.82		1.5	0.25
50	±150	-1.5	-0.86		1.5	0.25
40	±120	-1.5	-0.88		1.5	0.25
20	±60	-1.5	-0.88		1.5	0.25
10	±30	-1.5	-0.67		1.5	0.25
5	±15	-1.5	-0.65		1.5	0.25
4	±12	-1.5	-0.91		1.5	0.25
2	±6	-1.5	-0.89		1.5	0.25
1	±3	-1.5	-0.91		1.5	0.25
0.5	±1.5	-1.5	-0.82		1.5	0.25
0.4	±1.2	-1.5	-0.84		1.5	0.25
0.2	±0.6	-1.5	-0.74		1.5	0.25
0.1	±0.3	-1.5	-0.64		1.5	0.25
0.05	±0.15	-1.5	-0.88		1.5	0.25
0.04	±0.12	-1.5	-0.70		1.5	0.25
0.02	±0.06	-1.5	-0.81		1.5	0.25
0.01	±0.03	-1.5	-0.93		1.5	0.25
0.005	±0.015	-2.0	-1.13		2.0	0.25
0.004	±0.012	-2.0	-1.10		2.0	0.25
0.002	±0.004	-2.5	-1.30		2.5	0.25

PROCEDURE STEP NO: 3.2

DC GAIN ACCURACY

Channel 2

VERT SCALE (V/div)	*CAL VOLTAGE (V)*	*LOWER LIMIT dev(%)*	*BEFORE ADJ. dev(%)*	*AFTER ADJ. dev(%)*	*UPPER LIMIT dev(%)*	*UNCERTAINTY ±(%)*
100	±300	-1.5	-0.24		1.5	0.25
50	±150	-1.5	-0.35		1.5	0.25
40	±120	-1.5	-0.37		1.5	0.25
20	±60	-1.5	-0.47		1.5	0.25
10	±30	-1.5	-0.28		1.5	0.25
5	±15	-1.5	-0.33		1.5	0.25
4	±12	-1.5	-0.48		1.5	0.25
2	±6	-1.5	-0.48		1.5	0.25
1	±3	-1.5	-0.62		1.5	0.25
0.5	±1.5	-1.5	-0.58		1.5	0.25
0.4	±1.2	-1.5	-0.24		1.5	0.25
0.2	±0.6	-1.5	-0.53		1.5	0.25
0.1	±0.3	-1.5	-0.38		1.5	0.25
0.05	±0.15	-1.5	-0.52		1.5	0.25
0.04	±0.12	-1.5	-0.53		1.5	0.25
0.02	±0.06	-1.5	-0.57		1.5	0.25
0.01	±0.03	-1.5	-0.76		1.5	0.25
0.005	±0.015	-2.0	-0.91		2.0	0.25
0.004	±0.012	-2.0	-1.00		2.0	0.25
0.002	±0.004	-2.5	-1.72		2.5	0.25

PROCEDURE STEP NO: 3.3

DC GAIN ACCURACY

Channel 3

VERT SCALE (V/div)	*CAL VOLTAGE (V)*	*LOWER LIMIT dev(%)*	*BEFORE ADJ. dev(%)*	*AFTER ADJ. dev(%)*	*UPPER LIMIT dev(%)*	*UNCERTAINTY ±(%)*
100	±300	-1.5	-0.80		1.5	0.25
50	±150	-1.5	-0.70		1.5	0.25
40	±120	-1.5	-0.67		1.5	0.25
20	±60	-1.5	-0.72		1.5	0.25
10	±30	-1.5	-0.51		1.5	0.25
5	±15	-1.5	-0.64		1.5	0.25
4	±12	-1.5	-0.74		1.5	0.25
2	±6	-1.5	-0.77		1.5	0.25
1	±3	-1.5	-0.64		1.5	0.25
0.5	±1.5	-1.5	-0.89		1.5	0.25
0.4	±1.2	-1.5	-0.62		1.5	0.25
0.2	±0.6	-1.5	-0.56		1.5	0.25
0.1	±0.3	-1.5	-0.79		1.5	0.25
0.05	±0.15	-1.5	-0.74		1.5	0.25
0.04	±0.12	-1.5	-0.81		1.5	0.25
0.02	±0.06	-1.5	-0.88		1.5	0.25
0.01	±0.03	-1.5	-0.95		1.5	0.25
0.005	±0.015	-2.0	-0.91		2.0	0.25
0.004	±0.012	-2.0	-1.25		2.0	0.25
0.002	±0.004	-2.5	-1.53		2.5	0.25

PROCEDURE STEP NO: 3.4

DC GAIN ACCURACY

Channel 4

VERT SCALE (V/div)	*CAL VOLTAGE (V)*	*LOWER LIMIT dev(%)*	*BEFORE ADJ. dev(%)*	*AFTER ADJ. dev(%)*	*UPPER LIMIT dev(%)*	*UNCERTAINTY ±(%)*
100	±300	-1.5	-0.59		1.5	0.25
50	±150	-1.5	-0.70		1.5	0.25
40	±120	-1.5	-0.66		1.5	0.25
20	±60	-1.5	-0.76		1.5	0.25
10	±30	-1.5	-0.40		1.5	0.25
5	±15	-1.5	-0.40		1.5	0.25
4	±12	-1.5	-0.55		1.5	0.25
2	±6	-1.5	-0.64		1.5	0.25
1	±3	-1.5	-0.61		1.5	0.25
0.5	±1.5	-1.5	-0.73		1.5	0.25
0.4	±1.2	-1.5	-0.58		1.5	0.25
0.2	±0.6	-1.5	-0.50		1.5	0.25
0.1	±0.3	-1.5	-0.53		1.5	0.25
0.05	±0.15	-1.5	-0.57		1.5	0.25
0.04	±0.12	-1.5	-0.75		1.5	0.25
0.02	±0.06	-1.5	-0.73		1.5	0.25
0.01	±0.03	-1.5	-0.77		1.5	0.25
0.005	±0.015	-2.0	-0.99		2.0	0.25
0.004	±0.012	-2.0	-1.27		2.0	0.25
0.002	±0.004	-2.5	-1.39		2.5	0.25

PROCEDURE STEP NO: 4.1

OFFSET ACCURACY

Conditions:

RTH Position set to 0 div

Channel 1

VERT SCALE *(V/div)*	*OFFSET & CAL VOLTAGE (V)*	*LOWER LIMIT dev(mV)*	*BEFORE ADJ. dev(mV)*	*AFTER ADJ. dev(mV)*	*UPPER LIMIT dev(mV)*	*UNCERTAINTY ±(mV)*
0.002	0	-1.7	-0.447		1.7	0.02
0.002	4	-21.7	-1.285		21.7	0.25
0.002	-4	-21.7	0.925		21.7	0.25
0.05	0	-6.5	-0.23		6.5	0.45
0.05	4	-26.5	-0.80		26.5	0.45
0.05	-4	-26.5	1.18		26.5	0.45
0.1	0	-11.5	-2.57		11.5	0.90
0.1	4	-31.5	-3.87		31.5	0.90
0.1	-4	-31.5	-1.21		31.5	0.90
0.5	0	-51.5	-1.3		51.5	4.5
0.5	4	-71.5	-3.1		71.5	4.5
0.5	-4	-71.5	-0.8		71.5	4.5
1	0	-101.5	-35.0		101.5	9.0
1	200	-1101.5	-8.0		1101.5	16.1
1	-200	-1101.5	-43.6		1101.5	16.1
10	0	-1002	-63		1002	90
10	200	-2002	-35		2002	91
10	-200	-2002	-81		2002	91
20	0	-2000	-60		2000	180
20	200	-3000	-100		3000	180
20	-200	-3000	-10		3000	180

PROCEDURE STEP NO: 4.2

OFFSET ACCURACY

Conditions:

RTH Position set to 0 div

Channel 2

VERT SCALE *(V/div)*	*OFFSET & CAL VOLTAGE (V)*	*LOWER LIMIT dev(mV)*	*BEFORE ADJ. dev(mV)*	*AFTER ADJ. dev(mV)*	*UPPER LIMIT dev(mV)*	*UNCERTAINTY ±(mV)*
0.002	0	-1.7	-0.216		1.7	0.02
0.002	4	-21.7	-0.060		21.7	0.25
0.002	-4	-21.7	0.020		21.7	0.25
0.05	0	-6.5	0.22		6.5	0.45
0.05	4	-26.5	0.31		26.5	0.45
0.05	-4	-26.5	0.41		26.5	0.45
0.1	0	-11.5	-1.30		11.5	0.90
0.1	4	-31.5	-1.32		31.5	0.90
0.1	-4	-31.5	-1.50		31.5	0.90
0.5	0	-51.5	-1.9		51.5	4.5
0.5	4	-71.5	-0.6		71.5	4.5
0.5	-4	-71.5	-2.5		71.5	4.5
1	0	-101.5	-9.1		101.5	9.0
1	200	-1101.5	40.7		1101.5	16.1
1	-200	-1101.5	-55.2		1101.5	16.1
10	0	-1002	4		1002	90
10	200	-2002	79		2002	91
10	-200	-2002	-33		2002	91
20	0	-2000	10		2000	180
20	200	-3000	240		3000	180
20	-200	-3000	-240		3000	180

PROCEDURE STEP NO: 4.3

OFFSET ACCURACY

Conditions:

RTH Position set to 0 div

Channel 3

VERT SCALE *(V/div)*	*OFFSET & CAL VOLTAGE (V)*	*LOWER LIMIT dev(mV)*	*BEFORE ADJ. dev(mV)*	*AFTER ADJ. dev(mV)*	*UPPER LIMIT dev(mV)*	*UNCERTAINTY ±(mV)*
0.002	0	-1.7	0.134		1.7	0.02
0.002	4	-21.7	-1.159		21.7	0.25
0.002	-4	-21.7	1.766		21.7	0.25
0.05	0	-6.5	0.17		6.5	0.45
0.05	4	-26.5	-1.37		26.5	0.45
0.05	-4	-26.5	1.76		26.5	0.45
0.1	0	-11.5	0.96		11.5	0.90
0.1	4	-31.5	-0.11		31.5	0.90
0.1	-4	-31.5	1.44		31.5	0.90
0.5	0	-51.5	-2.5		51.5	4.5
0.5	4	-71.5	-2.6		71.5	4.5
0.5	-4	-71.5	-2.5		71.5	4.5
1	0	-101.5	-9.3		101.5	9.0
1	200	-1101.5	24.3		1101.5	16.1
1	-200	-1101.5	-30.9		1101.5	16.1
10	0	-1002	-36		1002	90
10	200	-2002	-26		2002	91
10	-200	-2002	-100		2002	91
20	0	-2000	-80		2000	180
20	200	-3000	-40		3000	180
20	-200	-3000	-250		3000	180

PROCEDURE STEP NO: 4.4

OFFSET ACCURACY

Conditions:

RTH Position set to 0 div

Channel 4

VERT SCALE *(V/div)*	*OFFSET & CAL VOLTAGE (V)*	*LOWER LIMIT dev(mV)*	*BEFORE ADJ. dev(mV)*	*AFTER ADJ. dev(mV)*	*UPPER LIMIT dev(mV)*	*UNCERTAINTY ±(mV)*
0.002	0	-1.7	-0.109		1.7	0.02
0.002	4	-21.7	0.222		21.7	0.25
0.002	-4	-21.7	-0.315		21.7	0.25
0.05	0	-6.5	0.14		6.5	0.45
0.05	4	-26.5	0.26		26.5	0.45
0.05	-4	-26.5	-0.29		26.5	0.45
0.1	0	-11.5	-0.38		11.5	0.90
0.1	4	-31.5	-0.75		31.5	0.90
0.1	-4	-31.5	-0.19		31.5	0.90
0.5	0	-51.5	0.1		51.5	4.5
0.5	4	-71.5	-0.1		71.5	4.5
0.5	-4	-71.5	1.2		71.5	4.5
1	0	-101.5	-3.1		101.5	9.0
1	200	-1101.5	17.3		1101.5	16.1
1	-200	-1101.5	-17.2		1101.5	16.1
10	0	-1002	33		1002	90
10	200	-2002	68		2002	91
10	-200	-2002	7		2002	91
20	0	-2000	260		2000	180
20	200	-3000	40		3000	180
20	-200	-3000	280		3000	180

PROCEDURE STEP NO: 5.1

ANALOG BANDWIDTH – CHANNEL 1

Conditions:

Vertical Scale: 10 mV/div
Horizontal Scale: 5 ns/div up to 200 MHz and 1 ns/div at 350 MHz

TEST FREQUENCY (MHz)	*LOWER LIMIT dev(dB)*	*BEFORE ADJ. dev(dB)*	*AFTER ADJ. dev(dB)*	*UPPER LIMIT dev(dB)*	*UNCERTAINTY ±(dB)*
60	-3.0	0.13		~	0.20
100	-3.0	0.23		~	0.20
200	-3.0	0.72		~	0.30
350	-3.0	-0.72		~	0.50

Conditions:

Vertical Scale: 100 mV/div
Horizontal Scale: 5 ns/div up to 200 MHz and 1 ns/div at 350 MHz

TEST FREQUENCY (MHz)	*LOWER LIMIT dev(dB)*	*BEFORE ADJ. dev(dB)*	*AFTER ADJ. dev(dB)*	*UPPER LIMIT dev(dB)*	*UNCERTAINTY ±(dB)*
60	-3.0	0.24		~	0.20
100	-3.0	0.33		~	0.20
200	-3.0	0.67		~	0.30
350	-3.0	-0.85		~	0.50

Conditions:

Vertical Scale: 1 V/div
Horizontal Scale: 5 ns/div up to 200 MHz and 1 ns/div at 350 MHz

TEST FREQUENCY (MHz)	*LOWER LIMIT dev(dB)*	*BEFORE ADJ. dev(dB)*	*AFTER ADJ. dev(dB)*	*UPPER LIMIT dev(dB)*	*UNCERTAINTY ±(dB)*
60	-3.0	0.13		~	0.20
100	-3.0	0.47		~	0.20
200	-3.0	0.41		~	0.30
350	-3.0	-1.60		~	0.50

PROCEDURE STEP NO: 5.2

ANALOG BANDWIDTH – CHANNEL 2

Conditions:

Vertical Scale: 10 mV/div
Horizontal Scale: 5 ns/div up to 200 MHz and 1 ns/div at 350 MHz

TEST FREQUENCY (MHz)	*LOWER LIMIT dev(dB)*	*BEFORE ADJ. dev(dB)*	*AFTER ADJ. dev(dB)*	*UPPER LIMIT dev(dB)*	*UNCERTAINTY ±(dB)*
60	-3.0	0.31		~	0.20
100	-3.0	0.34		~	0.20
200	-3.0	0.73		~	0.30
350	-3.0	-0.59		~	0.50

Conditions:

Vertical Scale: 100 mV/div
Horizontal Scale: 5 ns/div up to 200 MHz and 1 ns/div at 350 MHz

TEST FREQUENCY (MHz)	*LOWER LIMIT dev(dB)*	*BEFORE ADJ. dev(dB)*	*AFTER ADJ. dev(dB)*	*UPPER LIMIT dev(dB)*	*UNCERTAINTY ±(dB)*
60	-3.0	0.12		~	0.20
100	-3.0	0.48		~	0.20
200	-3.0	0.83		~	0.30
350	-3.0	-0.40		~	0.50

Conditions:

Vertical Scale: 1 V/div
Horizontal Scale: 5 ns/div up to 200 MHz and 1 ns/div at 350 MHz

TEST FREQUENCY (MHz)	*LOWER LIMIT dev(dB)*	*BEFORE ADJ. dev(dB)*	*AFTER ADJ. dev(dB)*	*UPPER LIMIT dev(dB)*	*UNCERTAINTY ±(dB)*
60	-3.0	0.13		~	0.20
100	-3.0	0.56		~	0.20
200	-3.0	0.71		~	0.30
350	-3.0	-1.32		~	0.50

PROCEDURE STEP NO: 5.3

ANALOG BANDWIDTH – CHANNEL 3

Conditions:

Vertical Scale: 10 mV/div
Horizontal Scale: 5 ns/div up to 200 MHz and 1 ns/div at 350 MHz

TEST FREQUENCY (MHz)	*LOWER LIMIT dev(dB)*	*BEFORE ADJ. dev(dB)*	*AFTER ADJ. dev(dB)*	*UPPER LIMIT dev(dB)*	*UNCERTAINTY ±(dB)*
60	-3.0	0.17		~	0.20
100	-3.0	0.44		~	0.20
200	-3.0	0.71		~	0.30
350	-3.0	-0.63		~	0.50

Conditions:

Vertical Scale: 100 mV/div
Horizontal Scale: 5 ns/div up to 200 MHz and 1 ns/div at 350 MHz

TEST FREQUENCY (MHz)	*LOWER LIMIT dev(dB)*	*BEFORE ADJ. dev(dB)*	*AFTER ADJ. dev(dB)*	*UPPER LIMIT dev(dB)*	*UNCERTAINTY ±(dB)*
60	-3.0	0.16		~	0.20
100	-3.0	0.49		~	0.20
200	-3.0	0.88		~	0.30
350	-3.0	-0.33		~	0.50

Conditions:

Vertical Scale: 1 V/div
Horizontal Scale: 5 ns/div up to 200 MHz and 1 ns/div at 350 MHz

TEST FREQUENCY (MHz)	*LOWER LIMIT dev(dB)*	*BEFORE ADJ. dev(dB)*	*AFTER ADJ. dev(dB)*	*UPPER LIMIT dev(dB)*	*UNCERTAINTY ±(dB)*
60	-3.0	0.23		~	0.20
100	-3.0	0.49		~	0.20
200	-3.0	0.88		~	0.30
350	-3.0	-1.42		~	0.50

PROCEDURE STEP NO: 5.4

ANALOG BANDWIDTH – CHANNEL 4

Conditions:

Vertical Scale: 10 mV/div
Horizontal Scale: 5 ns/div

TEST FREQUENCY (MHz)	*LOWER LIMIT dev(dB)*	*BEFORE ADJ. dev(dB)*	*AFTER ADJ. dev(dB)*	*UPPER LIMIT dev(dB)*	*UNCERTAINTY ±(dB)*
60	-3.0	0.32		~	0.20
100	-3.0	0.60		~	0.20
200	-3.0	1.00		~	0.30
350	-3.0	-0.36		~	0.50

Conditions:

Vertical Scale: 100 mV/div
Horizontal Scale: 5 ns/div

TEST FREQUENCY (MHz)	*LOWER LIMIT dev(dB)*	*BEFORE ADJ. dev(dB)*	*AFTER ADJ. dev(dB)*	*UPPER LIMIT dev(dB)*	*UNCERTAINTY ±(dB)*
60	-3.0	0.18		~	0.20
100	-3.0	0.49		~	0.20
200	-3.0	0.99		~	0.30
350	-3.0	-0.28		~	0.50

Conditions:

Vertical Scale: 1 V/div
Horizontal Scale: 5 ns/div

TEST FREQUENCY (MHz)	*LOWER LIMIT dev(dB)*	*BEFORE ADJ. dev(dB)*	*AFTER ADJ. dev(dB)*	*UPPER LIMIT dev(dB)*	*UNCERTAINTY ±(dB)*
60	-3.0	0.25		~	0.20
100	-3.0	0.61		~	0.20
200	-3.0	0.86		~	0.30
350	-3.0	-1.21		~	0.50