

Gamma Analysis QA Report

Patient ID:	Demo_Patient_001	Date:	2026-03-05 09:30
Reference:	Reference_Plan_Target.dcm	Machine:	LINAC
Evaluation:	Evaluation_Dose_Log.dcm	Physicist:	Default Physicist
Criteria:	3.0%, 2.0mm, TH: 10.0%	Overall GPR:	98.76 %

STATUS: PASS (GPR = 98.76 %)

ROI Gamma Analysis

ROI Name	GPR (%)	Voxels	Mean	Median	Max
PTV	99.10	1000	0.300	0.250	1.200
Rectum	100.00	500	0.100	0.050	0.800

Optimization & Warnings

Mode:	3d
Best Shift (x, y, z):	(1.00, 0.50, -0.50) mm
Absolute Geometry:	N/A
Warnings:	None

Reproducibility Information

Python	3.12.10	Platform	Windows-11-10.0.26100-SP0		
pydicom	2.4.4	numpy	1.26.4	scipy	1.16.1
numba	0.61.2	matplotlib	3.10.3	reportlab	4.4.9

Command: generate_sample.py

Reviewed By: _____ Date: _____