

# MINISTRY OF PETROLEUM RESOURCES

## DEPARTMENT OF PETROLEUM RESOURCES

7 KOFO-ABAYOMI STREET, VICTORIA ISLAND, LAGOS



P.M.B. No **12650**

Telephone **01-2612320, 2618228, 3200440-9**

Website: [www.dprnigeria.com](http://www.dprnigeria.com)



Ref. No.: **DPR/DS/CTO/2160/V-2/09**

Date: **February 13, 2014**

The Managing Director  
Seplat Petroleum Development Company Limited  
25a, Lugard Avenue, Ikoyi,  
Lagos.

Dear Sir,

### RE: METER INSTALLATION, COMMISSIONING & CALIBRATION AT SEPLAT'S RAPELE LACT UNIT

We refer to the recent recalibration of Rapele Lease Automatic Custody Transfer (LACT) unit proving system and fiscal instruments.

After a review of the report, we wish to convey approval for use of **11.91930 Barrels (1895.01360 Litres)** and **11.91981 Barrels (1895.09408 Litres)** at **60°F/opsig** between detector switches 1&4 and 2&3 respectively as the New Base Prover Volumes at RAPELE for custody transfer purposes.

However, further linearity tests should be carried out, which must be witnessed and signed by DPR representative, within two (2) months from the date of this approval on Meters FT-002 and FT-003 to achieve a repeatability within the statutory  $\pm 0.02\%$ .

Accordingly, two (2) copies of the approved calibration report are forwarded herewith for your use as statutorily required.

Yours faithfully,

**O.C. Okparaojiako**

For: Director of Petroleum Resources

**Cc: ZONAL OPSCON, Warri**

The above is for your information and record. In addition, two (2) copies of the approved tables are forwarded herewith for your use.

**O.C. Okparaojiako**

For: Director of Petroleum Resources