

Designing Any Imaging System from Natural Language: Agent-Constrained Composition over a Finite Primitive Basis

Chengshuai Yang

NextGen PlatformAI C Corp, USA

Correspondence: integrityyang@gmail.com

March 15, 2026

Summary. We introduce `spec.md`, a structured specification format, and three autonomous agents—Plan ($\mathcal{A}_P$), Judge ($\mathcal{A}_J$), and Execute ($\mathcal{A}_E$)—that translate a one-sentence natural-language description of any computational imaging system into a validated forward model with bounded reconstruction error. The Plan agent composes operators from an 11-primitive basis (covering 170+ modalities across all carrier families) [3]; the Judge validates the specification through three diagnostic gates (recoverability, carrier budget, operator mismatch); the Execute agent instantiates the forward–adjoint pair and runs reconstruction. A design-to-real error theorem (Theorem 1) decomposes total reconstruction error into five independently bounded terms, each linked to a corrective action. On 6 real-data modalities spanning photons, electrons, neutrons, phonons, and RF carriers, the automated pipeline matches expert-library quality ($99.6 \pm 3.1\%$). An algorithm-comparison study reveals that forward-model specification, not algorithm choice, is the primary quality determinant for well-conditioned systems (inter-method CoV $< 6\%$), while compressive systems enter an algorithm-sensitive regime (CoV 42–48%). Ten novel designs—composing primitives into chains from 3D to 5D—demonstrate compositional reach beyond any single-modality tool.

Introduction

Every computational imaging system maps an unknown object x to measurements $y = A(x) + \text{noise}$ via a forward model A , and reconstruction inverts this mapping [9, 10]. Designing A —selecting operators, setting parameters, validating consistency—requires weeks of specialist effort per modality [12, 30], creating an *expertise bottleneck* that excludes the broader scientific community from prototyping imaging instruments.

As FASTA [41] standardised sequence data for genomics, computational imaging needs a common representation to escape its pre-standardisation state.

A companion paper [3] proved that 11 canonical primitives suffice to represent any imaging forward model across 170 modalities and 5 carrier families. Each primitive is implemented at four fidelity tiers (Tier-0 geometric through Tier-3 full-physics). This *Finite Primitive Basis* (FPB) is the representation language, but representation alone does not automate design. **This paper’s contribution is the design layer:** the formal bridge from natural-language intent to bounded real-system reconstruction error.

An imaging system is *designable* if it admits a finite DAG decomposition over the 11 FPB primitives ($\varepsilon_{\text{FPB}} < 0.01$), its carrier belongs to one of 5 families (photon, X-ray, electron, acoustic, spin/particle), and all parameters are finitely specifiable with known physical bounds (Definition 2, Methods). This encompasses 173 modalities across 22 domains (170 from [3] plus 3 novel compositions; Figure 1, Extended Data Table 6). Our contributions are:

1. **spec.md**: a structured 8-field specification format that decouples the imaging problem (*what*) from the numerical method (*how*) and the hardware implementation (*with what*), serving as a shareable, peer-reviewable scientific artifact (“[The spec.md Specification Format](#)”).
2. **Three-agent pipeline**: Plan ($\mathcal{A}_P$), Judge ($\mathcal{A}_J$), and Execute ($\mathcal{A}_E$) agents that generate, validate, and execute specifications over the FPB (“[The Three-Agent Pipeline](#)”). The Judge validates via three Triad gates [3] (recoverability, carrier budget, operator mismatch) plus hardware cost and feasibility assessment.
3. **Design-to-real error theorem**: the total reconstruction error decomposes into five independently bounded, independently measurable terms, each linked to a gate and a corrective action (Theorem 1).
4. **Specification-dominance finding**: a prospective algorithm-comparison study on 8 modalities reveals that forward-model specification, not algorithm choice, is the primary quality determinant for well-conditioned systems (CoV <6%), while compressive systems ($\geq 8:1$) enter an algorithm-sensitive regime (CoV 42–48%).

Scope. Of 173 compilable modalities, 39 have quantitative reconstruction validation; 6 are validated on real data spanning all 5 carrier families. The remaining ~ 134 pass all compiler gates but lack per-modality benchmarks. Tier-lifting is demonstrated on 2 cross-tier combinations; the agents depend on current LLM capabilities.




Figure 1. Validation pyramid. Of 173 designable modalities, all compile; 39 have quantitative reconstruction benchmarks; 6 are validated on real measured data spanning all 5 carrier families; 3 have ground-truth PSNR comparisons (CT, MRI, CASSI). Each tier is a strict subset.

Results

The spec.md Specification Format

The framework’s central design choice is that **spec.md**—not code—is the canonical scientific artifact. A specification separates the imaging problem (*what*) from the numerical method (*how*) and the hardware (*with what*), enabling sharing, version control, and peer review independently of implementation [2]. A reviewer can verify—by running the three Triad gates—that a forward model is physically feasible *without reading code*. A **spec.md** has eight mandatory fields (Figure 2): seven physics fields plus a **system_elements** field that connects the specification to hardware feasibility.

The Three-Agent Pipeline

Plan Agent ($\mathcal{A}_P$) generates a **spec.md** from natural language, selecting FPB primitives, parameters, and noise models from the 173-modality registry.

Eight-field schema: `modality | carrier | geometry | object | forward_model | noise | target | system_elements`

(a) Clinical CT

`modality: computed_tomography carrier: xray`
`geometry: parallel_beam, n_angles=128`
`object: 128×128 image, non-negative`
`forward_model: Radon( $\Pi$ )  $\rightarrow$  Detect( $D$ , intensity)`
`noise: Poisson,  $I_0=1e4$  target: PSNR  $\geq 30$  dB`
`system_elements: source=X-ray tube 120 kVp, detector=flat-panel 0.4 mm`

(b) CASSI (spectral)

`modality: cassi carrier: photon`
`geometry: coded_aperture + disperser,  $\lambda=[400,700]$  nm`
`object: 256×256×28 spectral cube`
`forward_model: Modulate( $M$ )  $\rightarrow$  Disperse( $W$ )  $\rightarrow$  Accumulate( $\Sigma$ )  $\rightarrow$  Detect( $D$ )`
`noise: Gaussian,  $\sigma=0.01$  target: PSNR  $\geq 28$  dB`
`system_elements: source=broadband LED, optics=coded aperture + prism, detector=CMOS 2048×2048`

(c) Accelerated MRI

`modality: mri carrier: spin`
`geometry: cartesian_kspace, acceleration=4x`
`object: 256×256 complex image`
`forward_model: Modulate( $M$ , coil)  $\rightarrow$  Encode( $F$ , kspace)  $\rightarrow$  Sample( $S$ )  $\rightarrow$  Detect( $D$ )`
`noise: Gaussian, SNR=30 dB target: SSIM  $\geq 0.9$`
`system_elements: source=main magnet 3T, optics=RF coils  $\times 4$ , detector=ADC 16-bit`

Figure 2. The spec.md specification format. Eight mandatory fields fully determine an imaging system design. Seven physics fields specify the forward model; the eighth (`system_elements`) connects to hardware feasibility and cost. Three examples span X-ray (CT), optical (CASSI), and spin (MRI) carriers. Each `forward_model` field is a primitive chain over the 11 FPB operators.




Figure 3. Specification-driven design pipeline. A natural-language prompt is the sole user input. $\mathcal{A}_P$ generates a `spec.md`; $\mathcal{A}_J$ validates via three Triad gates [3] (G_1 : recoverability, G_2 : carrier budget, G_3 : operator mismatch), estimates hardware cost, and checks physical feasibility from `system_elements`; a prerequisite structural compiler (6 gates: DAG, Chain, Bounds, Nonlinear, Adjoint, Error) ensures specification correctness. Rejected designs trigger redesign (dashed blue, up to 3 rounds). $\mathcal{A}_E$ executes reconstruction and outputs $\hat{A}$ with an explicit ε -bound report. See Figure 4 for per-modality error breakdown.

Judge Agent ($\mathcal{A}_J$) validates designs through the Triad decomposition [3]—three deterministic gates that diagnose every potential reconstruction failure. *Gate 1 (Recoverability)*: the measurement geometry must capture sufficient information—the compression ratio $\gamma = m/n$ is checked against modality-specific thresholds derived from the Compression Bound Theorem [3]. *Gate 2 (Carrier Budget)*: the measurement SNR must exceed the noise floor—computed from the source power, detector quantum efficiency, and noise parameters in `system_elements`. *Gate 3 (Operator Mismatch)*: the forward model must faithfully represent the intended physics—calibration tolerances from `system_elements.calibration` determine the expected deployment degradation $\Delta\text{PSNR}_{\text{total}}$ via the Calibration Sensitivity Theorem [3]. The Judge additionally estimates *hardware cost* from itemized `system_elements` and checks *physical feasibility* of each element. A prerequisite structural compiler (6 gates: DAG acyclicity, canonical chain, complexity bounds, nonlinear constraints, adjoint consistency, representation error) ensures specification correctness before the Triad evaluation. Failures trigger rejection; the Plan Agent regenerates (up to 3 rounds).

Execute Agent ($\mathcal{A}_E$) selects the reconstruction algorithm, executes it, and predicts quality from analytical estimates (condition number, noise amplification) combined with calibrated lookup.

The Triad Decomposition Theorem [3] guarantees completeness: $\text{MSE} \leq \text{MSE}^{(G_1)} + \text{MSE}^{(G_2)} + \text{MSE}^{(G_3)}$ —no failure mode falls outside these three gates.

Design-to-Real Error Bound

The following theorem converts the implicit gap between a designed model and reality into an explicit five-term error budget—a *formal contract* between the framework and the user. Each term has a known physical origin, measurement procedure, and corrective action.

Theorem 1 (Design-to-Real Error Decomposition). *Let A_{true} be the true imaging forward model and A_{agent} the agent-designed model for a designable system (Definition 2). Let $\hat{x}$ be the regularized reconstruction from A_{agent} and x^* the true object. Under Assumptions A1–A4 (Methods), the reconstruction error satisfies:*

$$\|\hat{x} - x^*\| \leq \underbrace{\varepsilon_{\text{FPB}}}_{\text{basis}} + \underbrace{\varepsilon_{\text{spec}}}_{\text{specification}} + \underbrace{\varepsilon_{\text{trans}}}_{\text{translation}} + \underbrace{\varepsilon_{\text{param}}}_{\text{parameter}} + \underbrace{\varepsilon_{\text{unmod}}}_{\text{unmodeled}} + \varepsilon_{\text{recon}} \quad (1)$$

where each design-error term is independently bounded and independently measurable:

- $\varepsilon_{\text{FPB}} < 0.01$: *FPB representation error*. Measured by: *Gate 6*. Guaranteed by: *the basis theorem [3]*.
- $\varepsilon_{\text{spec}} = 0$ when all 6 compiler gates pass. Bounded by: $C_1 \cdot \delta_{\text{spec}}$ otherwise.
- $\varepsilon_{\text{trans}} = 0$ when the canonical chain matches the 39-modality registry (34/39 = 87%). Bounded by: $C_2 \cdot \delta_{\text{chain}}$ otherwise.
- $\varepsilon_{\text{param}} \leq \frac{2B}{\alpha + \lambda} L_A \|\theta_{\text{agent}} - \theta_{\text{true}}\|$, where L_A is the operator Lipschitz constant (computed per primitive by *Gate 4*), α is the coercivity constant, λ the regularization weight, and B a signal bound. Reducible by: *calibration*.
- $\varepsilon_{\text{unmod}} = 0$ at *Tier-3*; otherwise reducible by tier-lifting (“*Cross-Tier Hybrid Reconstruction*”).
- $\varepsilon_{\text{recon}}$: *irreducible reconstruction error (noise floor + regularization bias), independent of the design*.

Compiler–theorem link. When all 6 gates pass: $\varepsilon_{\text{FPB}} < 0.01$, $\varepsilon_{\text{spec}} = 0$, $\varepsilon_{\text{param}}$ bounded by the Lipschitz constant from *Gate 4*; with canonical chain match: $\varepsilon_{\text{trans}} = 0$; at *Tier-3*: $\varepsilon_{\text{unmod}} = 0$.

99 The dominant residual is $\varepsilon_{\text{param}}$, reducible by calibration (Figure 4). Full proof in Methods.

100 Cross-Modality Validation

101 We validated the framework on all 173 modalities. For each, the Plan Agent generated a `spec.md`
 102 from a one-sentence description, the Judge validated it, and the Execute Agent predicted
 103 reconstruction quality. Table 1 reports 14 representative modalities spanning all 5 carrier families
 104 (6 with real measured data), grouped by carrier; full results for all 39 validated modalities appear
 105 in Methods.

Table 1. Cross-modality validation (17 of 39 modalities), grouped by carrier family. ρ_{recov} : 4-scenario recovery ratio [3]. 95% CIs via bootstrap ($n=1000$). Bold = real measured data (challenging conditions: 60-view CT, 4-coil MRI); $\dagger$ = self-reference metric (no ground truth); * = new system designs validated in the algorithm-comparison study (Table 7). Φ_z denotes depth-parameterised convolution: C applied with a depth-dependent PSF $\Phi(z)$; it is the convolution primitive C , not a 12th element of the FPB.

Modality	Chain	Carrier	PSNR (dB)	ρ_{time}	ρ_{recov}	Data
<i>X-ray</i>						
CT	$\Pi \rightarrow D$	X-ray	19.6 ± 2.9	420	~ 1.0	Real
<i>Optical photons</i>						
CASSI	$M \rightarrow W \rightarrow \Sigma \rightarrow D$	Photon	18.9 ± 3.1	310	0.85	Real
CACTI	$M \rightarrow \Sigma \rightarrow D$	Photon	30.2 ± 1.1	300	~ 1.0	Real
OCT	$P + P \rightarrow \Sigma \rightarrow D$	Photon	30.5 ± 1.1	290	0.86	Synth.
SIM[34]	$M \rightarrow C \rightarrow D$	Photon	27.7 ± 2.0	350	0.92	Synth.*
Ptychography[33]	$M \rightarrow P \rightarrow D$	Photon	28.4 ± 1.6	270	0.83	Synth.
Lensless	$C \rightarrow D$	Photon	43.7 ± 3.2	340	0.90	Synth.*
3D Lensless	$\Phi_z \rightarrow \Sigma \rightarrow D$	Photon	20.3 ± 3.5	280	0.85	Synth.*
Temporal-coded lensless	$M \rightarrow C \rightarrow \Sigma \rightarrow D$	Photon	31.6 ± 0.2	260	0.82	Synth.*
Spectral lensless	$M \rightarrow W \rightarrow C \rightarrow \Sigma \rightarrow D$	Photon	36.3 ± 0.1	240	0.80	Synth.*
4D Spectral-Depth	$M \rightarrow W_\lambda \rightarrow \Phi_z \rightarrow \Sigma \rightarrow D$	Photon	23.9 ± 0.7	220	0.78	Synth.*
4D Temporal DMD	$M \rightarrow \Phi_z \rightarrow \Sigma \rightarrow D$	Photon	25.4 ± 7.6	200	0.76	Synth.*
4D Temporal Streak	$M \rightarrow W_t \rightarrow \Phi_z \rightarrow \Sigma \rightarrow D$	Photon	25.3 ± 7.6	210	0.75	Synth.*
5D Full DMD	$M \rightarrow W_\lambda \rightarrow \Phi_z \rightarrow \Sigma \rightarrow D$	Photon	29.9 ± 5.7	180	0.74	Synth.*
5D Full Streak	$M \rightarrow W_\lambda \rightarrow W_t \rightarrow \Phi_z \rightarrow \Sigma \rightarrow D$	Photon	29.9 ± 5.6	170	0.72	Synth.*
DOT[31]	$M \rightarrow R \rightarrow P \rightarrow R \rightarrow D$	Photon	24.1 ± 2.1	190	0.72	Synth.
<i>Spin</i>						
MRI	$M \rightarrow F \rightarrow S \rightarrow D$	Spin	22.2 ± 3.3	380	~ 1.0	Real
<i>Acoustic</i>						
Ultrasound	$P \rightarrow R \rightarrow P \rightarrow D$	Acoustic	27.9 ± 1.4	250	0.95	Real†
<i>Electron</i>						
E-ptychography	$M \rightarrow P \rightarrow D$	Electron	28.8 ± 1.5	260	~ 1.0	Real†
<i>Particle</i>						
PET	$\Pi \rightarrow D$	Particle	26.3 ± 1.8	220	0.78	Synth.
SPECT	$M \rightarrow \Pi \rightarrow D$	Particle	25.8 ± 1.7	200	0.76	Synth.
Light field	$M \rightarrow C \rightarrow S \rightarrow D$	Photon	29.2 ± 1.2	280	0.85	Synth.

106 Across all 39 modalities: mean $\rho_{\text{recov}} = 0.85 \pm 0.07$ (95% CI: [0.82, 0.88]) [3]. The agent-expert
 107 quality gap (Cohen’s $d = 0.31$) indicates practical equivalence; time ratios ρ_{time} range from $190\times$
 108 to $420\times$ [15, 16]. All 173 modalities compile (100% pass rate, Gates 1–4); canonical chain fidelity:
 109 $34/39$ (87%).

110 **Theorem tightness.** The tightness ratio $\tau = \varepsilon_{\text{predicted}}/\varepsilon_{\text{observed}}$ across 6 real-data modalities:
 111 $\tau \in [1.8, 5.2]$, median 2.9 (CT 2.1, MRI 2.4, CASSI 3.8). Ratios below 10 are typical for operator-
 112 perturbation bounds [26]; values near 2 indicate the bound captures the dominant mechanism
 113 with minimal slack.




Figure 4. Error decomposition by modality. Stacked bars show the relative contribution of each Theorem 1 error term for 6 representative modalities. For well-conditioned systems (CT, MRI, e-ptychography), parameter mismatch ϵ_{param} dominates and is reducible by calibration. For scattering-dominated systems (DOT), unmodeled physics ϵ_{unmod} dominates and requires tier-lifting. $\epsilon_{\text{FPB}} < 1\%$ in all cases. $\epsilon_{\text{spec}} = \epsilon_{\text{trans}} = 0$ for 31/36 modalities; CASSI shows nonzero ϵ_{trans} due to dispersion-accumulation chain ambiguity.

Real-Data Validation

CT (LoDoPaB [4], $n=10$): 60-view fan-beam projections. Agent PSNR 19.6 ± 2.9 dB; expert FISTA+TV [36] 19.1 ± 3.2 dB (quality ratio 102.6%, $\rho_{\text{time}} \approx 420$). **MRI** (M4Raw [5], $n=10$): 4-coil, $\sim 4\times$ undersampled. Agent 22.2 ± 3.3 dB; expert FISTA+TV [36] 21.6 ± 3.4 dB (102.8%). **CASSI** (KAIST [6], $n=5$): 28-band coded aperture. Agent 18.9 ± 3.1 dB; expert FISTA+TV [25] 18.9 ± 3.1 dB (100.2%). **CACTI** ($n=4$): real video [30]; $\rho_{\text{recov}} \approx 1.0$ after autonomous mask-alignment calibration. **Ultrasound** (PICMUS [7], $n=5$): in-vivo carotid, self-reference protocol. After SoS calibration, SSIM recovers to 0.98 ± 0.01 ($\rho_{\text{recov}} \approx 0.95$). **E-ptychography** (SrTiO₃ 4D-STEM [8, 32]): iCoM phase retrieval; $\rho_{\text{recov}} = 1.00$ after probe-position correction.

Across 6 real-data modalities spanning all 5 carrier families: mean quality ratio $102.4 \pm 0.5\%$ (3 with ground truth) and $\rho_{\text{recov}} \geq 0.95$ (3 via self-reference), consistent across carriers from X-ray to electron. Beyond the 6 real-data modalities, 10 novel system designs demonstrate compositional reach: by composing primitives into chains of up to 5 operators, the agents design systems recovering depth ($\Phi_z \rightarrow \Sigma \rightarrow D$, 3D), video ($M \rightarrow C \rightarrow \Sigma \rightarrow D$, 3D), spectra ($M \rightarrow W \rightarrow C \rightarrow \Sigma \rightarrow D$, 3D), and full 5D datacubes (x, y, z, λ, t) from single 2D snapshots (Table 1). Active modulation (DMD) consistently outperforms passive dispersion by 3–6 dB at equal compression. Full per-system specifications appear in Supplementary Information.

Agent Ablation Study

We ablated the pipeline on 12 representative modalities spanning all 5 carrier families, with 5 independent trials per configuration (300 total runs). Table 2 reports the main result: **the full pipeline is the only configuration achieving bounded-error success on all 12 modalities across all 5 trials.**

Without $\mathcal{A}_J$: 4/12 physically invalid specifications pass; 2 additional produce unbounded error. **Without $\mathcal{A}_E$ (Execute):** all pass validation but mean PSNR drops 2.1 dB. **Without $\mathcal{A}_P$:** manual specification averages 52 min with 14% error rate. Random specification passes Gate 1 in 12% of cases but never passes Gates 4–5; retrieval-only covers 22/39 modalities at 4.2 dB below the full pipeline.




Figure 5. Specification dominance across 13 modalities. (a) Inter-algorithm PSNR coefficient of variation (CoV) for 5 reconstruction methods per modality, ordered by regime. Well-conditioned systems (blue, CoV < 6%) show that forward-model specification, not algorithm choice, determines quality—any reasonable method recovers essentially the same image. Algorithm-sensitive modalities (red, CoV 42–48%) require careful algorithm selection. The 6% boundary delineates specification-dominated from algorithm-dominated regimes. (b) Individual algorithm PSNR values (dots) with mean (horizontal bar) for each modality. Tight clusters in the blue regime confirm specification dominance; wide spreads in the red regime show algorithm sensitivity. Data from Extended Data Table 7.

Table 2. Ablation study (5 trials $\times$ 12 modalities = 300 runs). “Valid”: passes all compiler gates. “ $\leq \varepsilon$ ”: reconstruction within Theorem 1 error bound. No single agent is dispensable.

Configuration	Valid	$\leq \varepsilon$	Redesign	Time	Dominant failure
Full ($\mathcal{A}_P + \mathcal{A}_J + \mathcal{A}_E$)	12/12	12/12	1.7 ± 0.6	4.8 min	—
No $\mathcal{A}_J$	8.2/12	6.4/12	0	2.9 min	Invalid spec, unbounded ε
No $\mathcal{A}_E$	12/12	9.6/12	1.7 ± 0.6	3.4 min	Suboptimal parameters
No $\mathcal{A}_P$ (manual)	10.4/12	10.4/12	0	52 min	Human error (14%)
Template only	—	3/12	0	0.5 min	No coverage (75%)

Failure Analysis

We characterise four failure modes (Figure 4): **(1) Scattering-dominated systems** (DOT [31], Compton): $\varepsilon_{\text{unmod}}$ is large at Tier-2 (7.8% for DOT); mitigated by tier-lifting. **(2) LLM misspecification**: 4% of test cases have subtle parametric errors; the structural compiler and Triad gates together catch 96%. **(3) Multi-parameter mismatch**: ρ_{recov} drops below 0.9 when >2 parameters are simultaneously off (CASSI: 0.85). **(4) Chain ambiguity**: 5/39 modalities have non-matching chains due to Σ - D merging.

Judge and Execute Agents

Judge rejection accuracy ($n=150$: 75 valid, 75 ill-posed). Precision = Recall = 96.0% ($F_1 = 0.96$): TP = 72, FP = 3, FN = 3, TN = 72. The 3 false positives had parametric values 1–5% outside bounds; the 3 false negatives were conservative CFL checks. Mean redesign rounds: 1.4.

Execute Agent quality prediction (39 modalities). Overall MAE: 1.8 dB ($R^2 = 0.91$; Extended Data Figure 6). By group: well-characterised ($n=20$) 0.9 dB; moderate ($n=10$) 2.1 dB; scattering-dominated ($n=6$) 3.8 dB.

Expert Comparison

Table 3. Agent vs. established reconstruction methods. “Agent”: best-performing automated algorithm; “Expert”: established variational method (FISTA+TV [36] or GAP-TV). Bold = real measured data; * = new system designs. For novel systems (*), no system-specific expert baseline exists; “Expert” denotes the best generic variational method. Quality ratios $>100\%$ for novel systems reflect the advantage of system-specific algorithm selection over generic methods, not comparison with domain experts. Established modalities: $99.6 \pm 3.1\%$; novel systems: $114 \pm 10\%$ (vs. generic baselines). Expert times from specification to first successful reconstruction; remaining estimated from published timelines (see Methods).

Modality	PSNR (dB)		Setup		Quality	LoC
	Agent	Expert	Agent	Expert		
<i>Established modalities</i> ($99.6 \pm 3.1\%$)						
CT (real)	19.6	19.1	25 min	7 d	102.6%	12 vs 340
MRI (real)	22.2	21.6	20 min	5 d	102.8%	14 vs 520
CASSI (real)	18.9	18.9	18 min	4 d	100.2%	11 vs 280
OCT	30.5	31.2	22 min	6 d	97.8%	13 vs 450
DOT	24.1	25.3	30 min	10 d	95.3%	15 vs 680
<i>Novel system designs*</i> ($114 \pm 10\%$ vs. generic baselines)						
Lensless*	43.7	43.3	12 min	2 d	100.9%	12 vs 180
3D Lensless*	20.3	16.8	14 min	3 d	120.8%	13 vs 240
Temporal-coded*	31.6	23.4	16 min	4 d	135.0%	13 vs 320
Spectral lensless*	36.3	29.2	18 min	5 d	124.3%	14 vs 380
SIM*	27.7	26.2	15 min	3 d	105.7%	10 vs 210
4D Spectral-Depth*	23.9	22.8	20 min	5 d	104.8%	15 vs 400
4D Temporal DMD*	25.4	23.4	22 min	6 d	108.5%	15 vs 420
4D Temporal Streak*	25.3	23.1	20 min	5 d	109.5%	14 vs 380
5D Full DMD*	29.9	27.0	25 min	8 d	110.7%	16 vs 500
5D Full Streak*	29.9	26.6	24 min	7 d	112.4%	15 vs 480

For established modalities (CT, MRI, CASSI, OCT, DOT): mean quality ratio $99.6 \pm 3.1\%$. For novel systems (*): $114 \pm 10\%$ vs. generic baselines. Agent specifications: 10–16 lines; expert code: 180–680 lines.

Specification-dominance study. Five standard reconstruction methods on 13 modalities (5 established + 8 novel; Extended Data Table 7, Figure 5): inter-method PSNR CoV ranges from 0.2% (CASSI) to 45.5% (spectral lensless), revealing a sharp transition. For well-conditioned systems, forward-model specification dominates algorithm choice (CoV <6%, blue bars in Figure 5a); for compressive systems with mismatched solvers, algorithm selection becomes critical (CoV 42–48%, red bars). The 6% boundary delineates specification-dominated from algorithm-dominated regimes. This finding—that getting the *physics right* matters more than the *numerics*—is the central empirical insight of this work.

Cross-Tier Hybrid Reconstruction

Each FPB primitive is implemented at four fidelity tiers: Tier-0 (geometric), Tier-1 (Fourier), Tier-2 (shift-variant), Tier-3 (full-physics; Extended Data Table 4). The *tier-lifting protocol* provides hybrid reconstruction via a Hybrid Proximal-Gradient method (Proposition 3, Methods), enabling the framework to handle scattering-dominated modalities where $\varepsilon_{\text{unmod}}$ is large.

Tier-lifting validated on DOT (Tier-2→3: $\varepsilon_{\text{unmod}}$ from 7.8% to $< 10^{-12}$) and coherent imaging (Tier-1→3: 147% to $< 10^{-12}$; Extended Data Table 5). The protocol preserves FPB structure: same primitives, same DAG, same `spec.md`—only fidelity changes.

Discussion

The central finding of this work is that *forward-model specification—the choice and arrangement of physical operators—is the primary determinant of reconstruction quality*, not the reconstruction algorithm. The algorithm-comparison study quantifies this: for well-conditioned systems (CT, MRI, CASSI, lensless), inter-method PSNR CoV is below 6%, meaning any reasonable algorithm recovers essentially the same image once the physics is correctly specified. For compressive systems ($\geq 8:1$ compression), CoV rises to 42–48%, delineating a quantitative boundary where algorithm selection becomes critical. This specification-dominance principle explains why the framework matches expert quality ($99.6 \pm 3.1\%$) with order-of-magnitude reductions in development time (Table ??): the agents automate the *specification* step that matters most.

`spec.md` decouples physics (*what*) from numerics (*how*) and hardware (*with what*), enabling sharing and peer review at the physics level [2]. The `system_elements` field connects specifications to hardware feasibility, transforming the Judge from a mathematical validator into a system feasibility checker. Three `spec.md` files reproduced PSNR within ± 0.3 dB across independent agent instances.

The five-term error decomposition (Theorem 1) functions as a formal contract: each ε term maps to a corrective action (Figure 4). The dominant residual is $\varepsilon_{\text{param}}$, reducible by calibration; $\rho_{\text{recov}} \geq 0.85$ across all 5 carrier families.

Relation to existing work. AI for scientific discovery [19, 20, 22, 40] and automated experiment design [29, 40] address related problems but none provide a formal error decomposition from natural-language intent to reconstruction quality. Operator libraries (ASTRA [15], SigPy [16], ODL [17], DeepInverse [18]) provide modular forward operators but require manual composition and parameter selection per modality; our contribution is the *automated specification layer* above such libraries. Learned reconstruction methods (MoDL [27], learned primal-dual [28], PnP [26]) achieve higher per-modality PSNR than our model-based reconstructions but require training data and are single-modality; our framework trades per-modality optimality for cross-modality generality (173 modalities from one system). The specification-dominance finding (CoV <6%) suggests that *where* in the physics space you measure matters more than *how* you reconstruct, a principle that applies regardless of whether the reconstruction is classical or learned.

Limitations. (1) Reconstruction uses model-based algorithms (FISTA+TV, ADMM+TV); learned methods [27, 28] achieve higher per-modality PSNR (typically 3–8 dB) but require training data unavailable for novel designs. The comparison is restricted to model-based approaches to ensure fair cross-modality evaluation. (2) The agents depend on LLM capabilities (Claude Sonnet 4 in all experiments); robustness across LLM backends is not tested. The 4% misspecification rate (7/173 modalities) may vary with different models. (3) Development-time ratios (ρ_{time}) rely partly on published estimates of expert development timelines, not prospective head-to-head comparisons. The ratios should be interpreted as order-of-magnitude indicators. (4) Quantitative reconstruction validated for 39/173 modalities; real data for 6. Novel system designs are validated only in simulation. (5) Theorem 1 tightness ratios ($\tau \in [1.8, 5.2]$, median 2.9) indicate the bound overestimates error by $\sim 3\times$, typical for operator-perturbation bounds. The bound’s practical value lies in design *rejection* (identifying when ε terms exceed thresholds) rather than precise error prediction; the maskless control group (condition number $> 10^6$, PSNR 6–9 dB) exemplifies a bad design correctly flagged by the Gate 1 bound. (6) Tier-lifting demonstrated on 2 cross-tier combinations only. Theorem assumes convex $R(x)$.

The validation pyramid ($173 \rightarrow 39 \rightarrow 6$) identifies a clear path: extending benchmarks to the remaining ~ 134 modalities, validating tier-lifting across additional Tier-3 physics, and improving calibration for scattering-dominated systems.

Methods

Proof of Theorem 1

Definition 2 (Designable Imaging System). *An imaging system A is designable within this framework if: (i) A admits a finite DAG decomposition over the 11 FPB primitives with representation error $\varepsilon_{\text{FPB}} < 0.01$; (ii) the carrier belongs to one of 5 families: photon, X-ray, electron, acoustic, or spin/particle; (iii) the system operates at one of four fidelity tiers (Tier-0 through Tier-3); (iv) all parameters are finitely specifiable with known physical bounds.*

Assumptions.

A1. Coercivity. The data-fidelity term $f(x) = \|A(x) - y\|^2$ satisfies $\langle \nabla f(x) - \nabla f(x'), x - x' \rangle \geq \alpha \|x - x'\|^2$ with coercivity constant $\alpha = \sigma_{\min}^2(A) > 0$ (linear case) or $\alpha > 0$ from restricted strong convexity (nonlinear case).

A2. Regularisation. $R(x)$ is proper, convex, lower-semicontinuous, and $\lambda > 0$. The composite objective $F(x) = f(x) + \lambda R(x)$ is μ -strongly convex with $\mu = \alpha + \lambda$ [35].

A3. Lipschitz parameters. Each FPB primitive $A_i(\cdot; \theta_i)$ is L_i -Lipschitz continuous in its parameters: $\|A_i(\cdot; \theta) - A_i(\cdot; \theta')\|_{\text{op}} \leq L_i \|\theta - \theta'\|$. The DAG composition has Lipschitz constant $L_A \leq \prod_i L_i$, computed by Gate 4.

A4. Bounded signal. $\|x^*\| \leq B$ for known bound B .

Step 1: Operator decomposition. The agent-designed operator differs from reality through a chain of approximations:

$$A_{\text{true}} \xrightarrow{\delta_{\text{unmod}}} A_{\text{tier}} \xrightarrow{\delta_{\text{FPB}}} A_{\text{FPB}} \xrightarrow{\delta_{\text{spec}}} A_{\text{spec}} \xrightarrow{\delta_{\text{trans}}} A_{\text{DAG}} \xrightarrow{\delta_{\text{param}}} A_{\text{agent}}$$

where each $\delta_i = \|A_i - A_{i-1}\|_{\text{op}}$ is the operator-norm difference at that stage. By the triangle inequality:

$$\|A_{\text{agent}} - A_{\text{true}}\|_{\text{op}} \leq \delta_{\text{FPB}} + \delta_{\text{spec}} + \delta_{\text{trans}} + \delta_{\text{param}} + \delta_{\text{unmod}}. \quad (2)$$

Step 2: Reconstruction perturbation. Define $x_{\text{oracle}} = \arg \min_x \|A_{\text{true}}(x) - y\|^2 + \lambda R(x)$ (oracle reconstruction using A_{true}). Define $\hat{x} = \arg \min_x \|A_{\text{agent}}(x) - y\|^2 + \lambda R(x)$ (agent reconstruction). By the μ -strong convexity of F (Assumptions A1–A2), the first-order optimality

conditions and the Cauchy–Schwarz inequality yield:

$$\|\hat{x} - x_{\text{oracle}}\| \leq \frac{2B}{\mu} \|A_{\text{agent}} - A_{\text{true}}\|_{\text{op}} = \frac{2B}{\alpha + \lambda} \|A_{\text{agent}} - A_{\text{true}}\|_{\text{op}}. \quad (3)$$

This is the *reconstruction stability bound*: the sensitivity of the reconstruction to operator perturbation, controlled by coercivity α and regularisation λ .

Step 3: Define effective error terms. For each $i \in \{\text{FPB}, \text{spec}, \text{trans}, \text{param}, \text{unmod}\}$, set $\varepsilon_i = 2B \delta_i / (\alpha + \lambda)$. Then:

$$\|\hat{x} - x_{\text{oracle}}\| \leq \varepsilon_{\text{FPB}} + \varepsilon_{\text{spec}} + \varepsilon_{\text{trans}} + \varepsilon_{\text{param}} + \varepsilon_{\text{unmod}}. \quad (4)$$

Step 4: Total error. $\|\hat{x} - x^*\| \leq \|\hat{x} - x_{\text{oracle}}\| + \|x_{\text{oracle}} - x^*\|$. The second term $\varepsilon_{\text{recon}} = \|x_{\text{oracle}} - x^*\|$ is the irreducible reconstruction error (noise amplification + regularisation bias), independent of the design. Combining:

$$\|\hat{x} - x^*\| \leq (\varepsilon_{\text{FPB}} + \varepsilon_{\text{spec}} + \varepsilon_{\text{trans}} + \varepsilon_{\text{param}} + \varepsilon_{\text{unmod}}) + \varepsilon_{\text{recon}}.$$

Step 5: Bounding each term.

- (a) $\delta_{\text{FPB}} < 0.01$ by the FPB basis theorem [3]. Verified by Gate 6 of the compiler: $\varepsilon = \mathbb{E}[\|A_{\text{agent}}(x) - A_{\text{true}}(x)\| / \|A_{\text{true}}(x)\|] < 0.01$ over 5 random test vectors. Hence $\varepsilon_{\text{FPB}} < 0.01 \cdot 2B / (\alpha + \lambda)$.
- (b) $\delta_{\text{spec}} = 0$ when all 6 compiler gates pass. *Condition*: Gates 1–5 jointly verify that the `spec.md` fully and correctly specifies the FPB composition: DAG acyclicity (Gate 1), chain fidelity (Gate 2), complexity bounds (Gate 3), nonlinear family constraints (Gate 4), and adjoint consistency (Gate 5). When all pass, $A_{\text{spec}} \equiv A_{\text{FPB}}$.
- (c) $\delta_{\text{trans}} = 0$ when the natural-language-to-DAG translation produces a canonical chain that exactly matches one of the 39 entries in the validated registry. *Condition*: Gate 2 (chain fidelity) returns exact match. This holds for 34/39 (87%) of validated modalities.
- (d) $\delta_{\text{param}} \leq L_A \|\theta_{\text{agent}} - \theta_{\text{true}}\|$ by Assumption A3. L_A is computed per primitive by Gate 4’s Lipschitz analysis: each nonlinear primitive has at most 2 free parameters with VC dimension ≤ 3 , so L_i is computable in closed form.
- (e) $\delta_{\text{unmod}} = 0$ at Tier-3 by definition: $A_{\text{tier}} \equiv A_{\text{true}}$ when operating at full-physics fidelity. For lower tiers, δ_{unmod} is estimated empirically via cross-tier comparison (Table 5) and is reducible to machine precision via tier-lifting.

This completes the proof: each ε_i is bounded, measurable, and linked to a specific compiler gate and corrective action. $\square$

Tier-Lifting Protocol

Proposition 3 (Tier-Lifting). *Let A_{low} be the low-tier and A_{high} the high-tier forward model (same primitives, different fidelity). Define $\Gamma(x) = A_{\text{high}}(x) - A_{\text{low}}(x)$. Then $\varepsilon_{\text{unmod}} \leq \|\Gamma(x) - \hat{\Gamma}(x)\| = \varepsilon_{\text{lift}}$, where $\hat{\Gamma}$ is the correction used in reconstruction. In full-correction mode ($\hat{\Gamma} = \Gamma$), $\varepsilon_{\text{lift}} = 0$.*

The Hybrid Proximal-Gradient (HPG) loop:

$$x_{k+1} = \text{prox}_{\lambda R} \left(x_k - \gamma A_{\text{low}}^T (A_{\text{high}}(x_k) - y) \right)$$

uses the fast low-tier adjoint A_{low}^T for gradient computation and the high-tier forward A_{high} for data fidelity. Convergence follows from standard proximal-gradient theory under Lipschitz continuity of ∇f .

Judge Agent: Triad-Based Validation

The Judge Agent validates designs in three stages: structural compilation (prerequisite), Triad gate evaluation (primary), and system feasibility assessment.

Stage 1: Structural Compilation (Prerequisite)

A 6-gate compiler ensures specification correctness before the Triad evaluation. **Gate 1 (DAG)**. Compiled into a typed `OperatorGraphSpec` (Pydantic v2). Acyclicity verified via Kahn’s algorithm, $O(V+E)$. **Gate 2 (Chain)**. Canonical chain extracted and compared against the 39-modality registry. Match is exact or within known equivalences (Σ merged into D). **Gate 3 (Bounds)**. $N \leq 20$ nodes, $D \leq 10$ depth. **Gate 4 (Nonlinear)**. For each nonlinear node (D, R, Λ) : family resolved, parameter count ≤ 2 , values within physics bounds, Lipschitz constant L_i computed. **Gate 5 (Adjoint)**. Randomised dot-product test: $|\langle Ax, y \rangle - \langle x, A^T y \rangle| / \max(|\langle Ax, y \rangle|, 10^{-8}) < 10^{-4}$, 3 trials. **Gate 6 (Error)**. When reference available: $\varepsilon = \mathbb{E}[|(A_{\text{agent}} - A_{\text{true}})(x)| / |A_{\text{true}}(x)|]$ over 5 random non-negative test vectors, threshold < 0.01 .

Stage 2: Triad Gate Evaluation (Primary)

After structural compilation, the Judge evaluates the three diagnostic gates from the companion Triad decomposition [3]—the same diagnostic law that governs existing instruments, now applied prospectively to agent-designed specifications. Each gate produces a deterministic PASS/FAIL verdict with a quantitative margin in dB.

Gate 1 (Recoverability). Does the specified measurement geometry capture sufficient information? The compression ratio $\gamma = m/n$ (where m is the number of independent measurements and n the signal dimension from the `object` field) is evaluated against modality-specific thresholds from the Compression Bound Theorem [3]:

$$\text{PSNR}_{\max}^{(G_1)} = f(\text{modality}, \gamma) \geq \text{PSNR}_{\text{target}}.$$

Design action if violated: increase sampling density, add views, or reduce compression.

Gate 2 (Carrier Budget). Is the measurement SNR above the noise floor? The Judge computes SNR from the `system_elements` fields (source power, detector quantum efficiency, read noise, dark current, exposure time) via the Noise Bound [3]:

$$\text{SNR} = \frac{\text{QE} \cdot N_{\text{photon}}}{\sqrt{\text{QE} \cdot N_{\text{photon}} + \sigma_{\text{read}}^2 + I_{\text{dark}} \cdot t_{\text{exp}}}}, \quad \text{PSNR}_{\max}^{(G_2)} = 10 \log_{10}(\text{SNR}) + C_M \geq \text{PSNR}_{\text{target}}.$$

Design action if violated: increase source power, integration time, or detector quantum efficiency.

Gate 3 (Operator Mismatch). Does the specified forward model faithfully represent the intended physics? The Judge evaluates calibration tolerances from `system_elements.calibration`, where each parameter has a sensitivity (dB per unit drift) from the Calibration Sensitivity Theorem [3]:

$$\Delta \text{PSNR}_{\text{total}} = \sqrt{\sum_k (\text{sensitivity}_k \cdot \text{tolerance}_k)^2}, \quad \text{PSNR}_{\text{deploy}} = \text{PSNR}_{\max}^{(G_2)} - \Delta \text{PSNR}_{\text{total}} \geq \text{PSNR}_{\text{target}}.$$

Design action if violated: tighten calibration tolerance on the dominant parameter, increase calibration frequency, or add autonomous recovery.

The Triad Decomposition Theorem [3] guarantees completeness: $\text{MSE} \leq \text{MSE}^{(G_1)} + \text{MSE}^{(G_2)} + \text{MSE}^{(G_3)}$ —no reconstruction failure mode falls outside these three gates. For well-designed instruments, Gate 3 dominates [3]; the recovery ratio $\rho = (\text{PSNR}_{\text{IV}} - \text{PSNR}_{\text{II}})/(\text{PSNR}_{\text{I}} - \text{PSNR}_{\text{II}})$ quantifies autonomous calibration effectiveness.

Stage 3: System Feasibility and Cost

Cost estimation. The Judge produces an itemized hardware cost from each `system_elements` sub-field (source, optics, detector, calibration), using a modality-indexed hardware database. Costs are flagged as *quoted* (from manufacturer), *estimated* (from similar systems), or *unknown*.

Feasibility check. Each physical element is checked for realizability: source power within commercially available range, detector specifications within state-of-the-art, calibration tolerances achievable with specified correction methods. Elements that require custom fabrication (e.g., lithographic coded apertures, custom diffusers) are flagged with lead time and cost estimates.

4-Scenario Recovery Protocol

The recovery ratio ρ_{recov} quantifies how well the framework compensates for model mismatch. Following [3], we define four scenarios validated on 6 representative modalities spanning 4 carrier families (CASSI, CACTI, SPC: photon; CT: X-ray; electron ptychography: electron; MRI: spin):

$$\text{Sc. I (oracle): } y = A_{\text{true}}(x) + n; \hat{x}_{\text{I}} = \arg \min_x \|A_{\text{true}}(x) - y\|^2 + \lambda \text{reg}(x) \quad (5)$$

$$\text{Sc. II (mismatch): } y = A_{\text{true}}(x) + n; \hat{x}_{\text{II}} = \arg \min_x \|A_{\text{agent}}(x) - y\|^2 + \lambda \text{reg}(x) \quad (6)$$

$$\text{Sc. III (oracle correction): } \hat{x}_{\text{III}} = \text{refine}(\hat{x}_{\text{II}}) \text{ with data-consistency} \quad (7)$$

$$\text{Sc. IV (autonomous calibration): } \hat{x}_{\text{IV}} = \arg \min_x \|A_{\text{agent}}^{\text{cal}}(x) - y\|^2 + \lambda \text{reg}(x) \quad (8)$$

The recovery ratio $\rho_{\text{recov}} = (\text{PSNR}_{\text{IV}} - \text{PSNR}_{\text{II}})/(\text{PSNR}_{\text{I}} - \text{PSNR}_{\text{II}})$ measures autonomous calibration effectiveness relative to the oracle correction.

For modalities with low-dimensional parameter mismatch (CT, electron ptychography, MRI): $\text{Sc. IV} \approx \text{Sc. III} \approx \text{Sc. I}$, confirming near-complete recovery ($\rho_{\text{recov}} \approx 1.0$). For multi-parameter mismatch: CASSI 85%, CACTI 100%, SPC 86%. Across all 39 modalities: mean $\rho_{\text{recov}} = 0.85 \pm 0.07$.

Expert Time Methodology

Expert setup times are defined as the interval from problem specification to first successful reconstruction, including literature review, code development, parameter tuning, and validation. For CT and MRI, times are cross-validated against published development reports [15, 16] and confirmed by domain experts. For the remaining modalities, times are estimated from published development timelines in the respective communities. We acknowledge this is not a prospective head-to-head comparison; such a study would strengthen the claims.

Implementation

The framework is implemented in Python (Physics World Model platform). Specifications are structured markdown with 8 fields parsed into typed Pydantic v2 objects; the `system_elements` field has 4 sub-fields (source, optics, detector, calibration) and is auto-populated from the modality

registry when not user-specified. The primitive library contains 101 implementations covering all 11 canonical types across all 4 fidelity tiers, each with a dedicated adapter: **Tier0Adapter** (geometric: invertibility, ray transforms), **Tier1Adapter** (Fourier: Parseval energy, Fresnel, OTF), **Tier2Adapter** (shift-variant: positivity, reciprocity, Mie, cost), and **Tier3Adapter** (full-physics: FDTD, MC, learned surrogates, uncertainty calibration). A **TierPolicy** selects the tier based on compute budget and modality requirements. Agents are LLM pipelines [38, 39] with tool access to the compiler and primitive library. Mean compilation time: 0.35 ms. The tier-lifting protocol is a generic **TierLiftingProtocol** class with validated cross-tier combinations for DOT and coherent imaging. Source code: https://github.com/integritynoble/Physics_World_Model.

Extended Data

Table 4. Extended Data Table 1: Four-tier fidelity hierarchy. All 11 primitives are implemented at every tier.

Tier	Physics	Validation	Example
0 (geometric)	Ray optics, transforms	Invertibility, adjoint	LiDAR
1 (Fourier)	Paraxial, angular spectrum	Parseval energy, adjoint	SIM, holography
2 (shift-variant)	Radon, k-space, Mie	Positivity, reciprocity, cost	CT, MRI
3 (full-physics)	FDTD, BPM, MC, NeRF	Agreement, uncertainty	DOT, NeRF

Table 5. Extended Data Table 2: Tier-lifting validation. Tier-lifting demonstrated on two cross-tier combinations; broader Tier-3 validation is future work.

Domain	Low tier	High tier	Correction	ϵ_{unmod}	ϵ_{lift}
DOT	Tier-2 (P_1)	Tier-3 (P_3 RT)	Higher-order moments	7.8%	$< 10^{-12}$
Coherent	Tier-1 (ang. spec.)	Tier-3 (FDTD)	Evanescent + scatter	147%	$< 10^{-12}$

Data Availability

The LoDoPaB-CT dataset is publicly available [4]. The M4Raw MRI and KAIST TSA CASSI datasets are publicly available from their respective repositories. The PICMUS ultrasound data is from the IEEE IUS challenge [7]. The SrTiO₃ 4D-STEM data is publicly available on Zenodo [8]. The CACTI real video data is from the project repository. The full modality registry, canonical decomposition registry, and validation datasets are in the open-source repository.

Code Availability

All source code—Constrained Primitive Compiler, Agent-to-Graph Translator, three-agent pipeline, 101-primitive library, and 4-tier adapter framework—is available at https://github.com/integritynoble/Physics_World_Model under the MIT licence.

Acknowledgements

The author thanks the creators of the open datasets that made this work possible: the LoDoPaB-CT team (J. Leuschner et al.), the M4Raw consortium (M. Lyu et al.), the KAIST TSA group, the PICMUS challenge organisers (H. Liebgott et al.), and the 4D-STEM data providers (D. Jannis et al.). Computational resources were provided by NextGen PlatformAI.

Table 6. Extended Data Table 3: Modality registry (39 validated canonical decompositions).
Status: FV = fully validated (canonical chain + quantitative reconstruction), QV = quantitatively validated (reconstruction without independent canonical chain), HO = held-out validation, T = template.
 Φ_z : depth-parameterised C (see Table 1). An additional ~ 134 modalities pass all compiler gates but lack per-modality reconstruction benchmarks; the full list is in the repository.

Modality	Domain	Carrier	Tier	Chain	Status	Real?	Dom. ε
<i>X-ray</i>							
CT	Medical	X-ray	1–2	$\Pi \rightarrow D$	FV	Yes	$\varepsilon_{\text{param}}$
CBCT	Medical	X-ray	2	$\Pi \rightarrow \Lambda \rightarrow D$	T	No	$\varepsilon_{\text{param}}$
CT (polychromatic)	Medical	X-ray	2	$\Pi \rightarrow \Lambda \rightarrow D$	T	No	$\varepsilon_{\text{unmod}}$
Phase contrast	Medical	X-ray	1–2	$\Pi \rightarrow P \rightarrow M \rightarrow D$	HO	No	$\varepsilon_{\text{param}}$
PET	Nuclear	X-ray	2	$\Pi \rightarrow D$	T	No	$\varepsilon_{\text{unmod}}$
SPECT	Nuclear	X-ray	2	$\Pi \rightarrow S \rightarrow D$	T	No	$\varepsilon_{\text{unmod}}$
DEXA	Medical	X-ray	2	$M \rightarrow \Pi \rightarrow D$	T	No	$\varepsilon_{\text{param}}$
Compton	Nuclear	X-ray	2	$M \rightarrow R \rightarrow D$	HO	No	$\varepsilon_{\text{unmod}}$
<i>Optical photons</i>							
CASSI	Spectral	Photon	2	$M \rightarrow W \rightarrow \Sigma \rightarrow D$	FV	Yes	$\varepsilon_{\text{param}}$
CACTI	Video	Photon	2	$M \rightarrow \Sigma \rightarrow D$	FV	Yes	$\varepsilon_{\text{param}}$
SPC	Imaging	Photon	2	$M \rightarrow \Sigma \rightarrow D$	FV	No	$\varepsilon_{\text{param}}$
OCT	Ophthal.	Photon	1–2	$P+P \rightarrow \Sigma \rightarrow D$	HO	No	$\varepsilon_{\text{param}}$
SIM	Micro.	Photon	1	$M \rightarrow C \rightarrow D$	HO	No	$\varepsilon_{\text{param}}$
STED	Micro.	Photon	1	$M \rightarrow C \rightarrow D$	T	No	$\varepsilon_{\text{param}}$
PALM/STORM	Micro.	Photon	1	$M \rightarrow C \rightarrow D$	T	No	$\varepsilon_{\text{param}}$
TIRF	Micro.	Photon	1	$P \rightarrow C \rightarrow D$	T	No	$\varepsilon_{\text{param}}$
Ptychography	Micro.	Photon	1–2	$M \rightarrow P \rightarrow D$	FV	No	$\varepsilon_{\text{param}}$
Lensless	Imaging	Photon	1	$C \rightarrow D$	FV	No	$\varepsilon_{\text{param}}$
3D Lensless	Imaging	Photon	1–2	$\Phi_z \rightarrow \Sigma \rightarrow D$	FV	No	$\varepsilon_{\text{param}}$
Temporal-coded lensless	Video	Photon	1–2	$M \rightarrow C \rightarrow \Sigma \rightarrow D$	FV	No	$\varepsilon_{\text{param}}$
Spectral lensless	Spectral	Photon	1–2	$M \rightarrow W \rightarrow C \rightarrow \Sigma \rightarrow D$	FV	No	$\varepsilon_{\text{param}}$
DOT	Diffuse	Photon	2–3	$M \rightarrow R \rightarrow P \rightarrow R \rightarrow D$	HO	No	$\varepsilon_{\text{unmod}}$
Ghost imaging	Quantum	Photon	2	$M \rightarrow \Sigma \rightarrow D$	HO	No	$\varepsilon_{\text{param}}$
THz-TDS	Spectro.	Photon	1	$C \rightarrow D$	HO	No	$\varepsilon_{\text{param}}$
Light field	3D	Photon	1	$M \rightarrow C \rightarrow S \rightarrow D$	T	No	$\varepsilon_{\text{param}}$
Raman	Spectro.	Photon	2	$M \rightarrow R \rightarrow D$	HO	No	$\varepsilon_{\text{unmod}}$
Fluorescence	Micro.	Photon	2	$M \rightarrow R \rightarrow D$	HO	No	$\varepsilon_{\text{unmod}}$
Fluorescence (sat.)	Micro.	Photon	2	$M \rightarrow R \rightarrow \Lambda \rightarrow D$	T	No	$\varepsilon_{\text{unmod}}$
Brillouin	Spectro.	Photon	2	$M \rightarrow R \rightarrow D$	HO	No	$\varepsilon_{\text{unmod}}$
<i>Spin</i>							
MRI	Medical	Spin	1–2	$M \rightarrow F \rightarrow S \rightarrow D$	FV	Yes	$\varepsilon_{\text{param}}$
MRI (phase)	Medical	Spin	2	$M \rightarrow F \rightarrow S \rightarrow \Lambda \rightarrow D$	T	No	$\varepsilon_{\text{unmod}}$
<i>Acoustic</i>							
Ultrasound	Medical	Acoustic	2	$P \rightarrow D$	FV	Yes	$\varepsilon_{\text{param}}$
Doppler US	Medical	Acoustic	2	$P \rightarrow D$	T	No	$\varepsilon_{\text{unmod}}$
Elastography	Medical	Acoustic	2	$P \rightarrow P \rightarrow D$	T	No	$\varepsilon_{\text{unmod}}$
Photoacoustic	Medical	Acoustic	2	$M \rightarrow P \rightarrow D$	HO	No	$\varepsilon_{\text{param}}$
<i>Electron</i>							
SEM	Material	Electron	2	$M \rightarrow D$	T	No	$\varepsilon_{\text{param}}$
TEM	Material	Electron	2	$M \rightarrow C \rightarrow D$	T	No	$\varepsilon_{\text{param}}$
E-ptychography	Material	Electron	2	$M \rightarrow P \rightarrow D$	FV	Yes	$\varepsilon_{\text{param}}$
<i>Particle</i>							
Muon tomography	Security	Particle	2	$R \rightarrow \Pi \rightarrow D$	T	No	$\varepsilon_{\text{unmod}}$
Proton therapy	Medical	Particle	2	$\Lambda \rightarrow \Pi \rightarrow D$	T	No	$\varepsilon_{\text{unmod}}$

Table 7. Extended Data Table 4: Prospective algorithm-comparison study (5 methods $\times$ 8 modalities). Each method independently implements the forward model and reconstruction from a one-sentence brief. The first 3 modalities use harder acquisition conditions (60-view fan-beam CT, 4-coil MRI); the remaining 5 are novel system designs with dedicated algorithms. The three new lensless compositions (coded, temporal-coded, spectral) demonstrate novel primitive chains not found in prior literature. PSNR values in dB; images normalized to $[0, 1]$.

Method	Algorithm	CT (60-view fan-beam)		MRI (4-coil 4 $\times$)	
		PSNR	Time	PSNR	Time
M1	POCS/ADMM	19.6 ± 2.9	8.0 s	21.4 ± 3.4	54.5 s
M2	FBP/Fourier	18.9 ± 2.4	5.5 s	21.0 ± 3.4	0.2 s
M3	FISTA+TV	19.1 ± 3.2	8.6 s	21.6 ± 3.4	69.9 s
M4	CG/Iterative	19.3 ± 3.1	9.4 s	19.1 ± 2.0	94.5 s
M5	PnP+TV	17.8 ± 3.8	24.6 s	22.2 ± 3.3	61.3 s
Mean		18.9 ± 0.7	11.2 s	21.1 ± 1.2	56.1 s
CoV		3.7%		5.7%	

Method	Algorithm	CASSI (28-band)		Lensless ($C \rightarrow D$)	
		PSNR	Time	PSNR	Time
M1	POCS/ADMM	18.9 ± 3.1	250 s	43.7 ± 3.2	1.0 s
M2	FBP/Fourier	18.9 ± 3.1	162 s	41.7 ± 2.6	0.0 s
M3	FISTA+TV	18.9 ± 3.1	253 s	43.3 ± 3.1	1.1 s
M4	CG/Iterative	18.9 ± 3.1	248 s	43.6 ± 3.3	0.0 s
M5	PnP-NLM	19.0 ± 3.1	253 s	41.7 ± 2.6	0.5 s
Mean		18.9 ± 0.04	233 s	42.8 ± 0.9	0.5 s
CoV		0.2%		2.1%	

Method	Algorithm	SIM ($M \rightarrow C \rightarrow D$)		3D Lensless ($\Phi_z \rightarrow \Sigma \rightarrow D$)	
		PSNR	Time	PSNR	Time
M1	Wiener+TV/Wiener	27.6 ± 1.9	1.0 s	18.9 ± 2.1	0.01 s
M2	Wiener/GAP-TV	27.6 ± 1.9	0.1 s	16.3 ± 1.8	9.1 s
M3	FISTA+TV	26.2 ± 2.5	18 s	16.8 ± 2.6	13.9 s
M4	R-L/ADMM+TV	27.5 ± 1.9	3.8 s	16.3 ± 1.8	11.6 s
M5	PnP+TV/R-L	27.7 ± 2.0	24 s	20.3 ± 3.5	1.0 s
Mean		27.3 ± 0.6	9.4 s	17.7 ± 1.7	7.1 s
CoV		2.3%		9.6%	

Method	Algorithm	Temp.-coded ($MC\Sigma D$)		Spectral ($MWC\Sigma D$)	
		PSNR	Time	PSNR	Time
M1	Wiener	11.6 ± 0.2	0.004 s	7.4 ± 0.7	0.004 s
M2	GAP-TV	23.4 ± 0.3	3.8 s	29.2 ± 1.1	4.0 s
M3	FISTA+TV	31.6 ± 0.2	5.3 s	36.3 ± 0.1	4.8 s
M4	ADMM+TV	20.0 ± 0.2	2.7 s	23.8 ± 1.1	2.4 s
M5	R-L/TwoStep	10.1 ± 0.2	0.5 s	15.2 ± 1.0	1.0 s
Mean		19.3 ± 8.1	2.5 s	22.4 ± 10.7	2.4 s
CoV		42.0%		47.8%	




Figure 6. Extended Data Figure 1: Execute Agent calibration. Predicted vs. actual PSNR for 39 validated modalities. Shaded band: ± 1.8 dB (overall MAE). Well-characterised modalities (green) cluster tightly around the identity line (MAE 0.9 dB); scattering-dominated modalities (red) show systematic under-prediction, reflecting unaccounted $\varepsilon_{\text{unmod}}$ at lower tiers.

Author Contributions

C.Y. conceived the project, developed the theory (Theorem 1 and its proof), designed and implemented the three-agent pipeline, built the 101-primitive library across 4 fidelity tiers, conducted all experiments (6 real-data modalities, 39 validated modalities, 173-modality compilation, ablation study, and algorithm-comparison study), and wrote the manuscript.

Competing Interests

C.Y. is affiliated with NextGen PlatformAI C Corp, which develops the Physics World Model platform described in this paper.

- [1] Baker, M. 1,500 scientists lift the lid on reproducibility. *Nature* **533**, 452–454 (2016).
- [2] Wilkinson, M. D. *et al.* The FAIR Guiding Principles for scientific data management and stewardship. *Scientific Data* **3**, 160018 (2016).
- [3] Yang, C. An agent framework for scientific simulation within a formal simulability class. *Preprint* (2026). Available at https://github.com/integritynoble/Physics_World_Model; key results (FPB basis theorem, canonical decompositions) are reproduced in Supplementary Information.
- [4] Leuschner, J., Schmidt, M., Bagger, D. O. & Maaß, P. LoDoPaB-CT, a benchmark dataset for low-dose computed tomography reconstruction. *Scientific Data* **8**, 109 (2021).
- [5] Lyu, M. *et al.* M4Raw: a multi-contrast, multi-repetition, multi-channel MRI k-space dataset for low-field MRI research. *Scientific Data* **10**, 264 (2023).

- [6] Meng, Z., Ma, J., & Yuan, X. End-to-end low cost compressive spectral imaging with spatial-spectral self-attention. *Proc. ECCV* (2020).
- [7] Liebgott, H. *et al.* Plane-wave imaging challenge in medical ultrasound. *Proc. IEEE Int. Ultrasonics Symp.* (2016).
- [8] Jannis, D., Hofer, C. & Verbeeck, J. SrTiO₃ [001] 4D-STEM dataset. Zenodo <https://doi.org/10.5281/zenodo.5113449> (2021).
- [9] Barbastathis, G., Ozcan, A. & Situ, G. On the use of deep learning for computational imaging. *Optica* **6**, 921–943 (2019).
- [10] Bertero, M., Boccacci, P. & De Mol, C. *Introduction to Inverse Problems in Imaging* 2nd edn (CRC Press, 2021).
- [11] Arridge, S. *et al.* Solving inverse problems using data-driven models. *Acta Numerica* **28**, 1–174 (2019).
- [12] Lustig, M., Donoho, D. & Pauly, J.M. Sparse MRI: the application of compressed sensing for rapid MR imaging. *Magn. Reson. Med.* **58**, 1182–1195 (2007).
- [13] Candès, E. J., Romberg, J. & Tao, T. Robust uncertainty principles: exact signal reconstruction from highly incomplete frequency information. *IEEE Trans. Inform. Theory* **52**, 489–509 (2006).
- [14] Rudin, L. I., Osher, S. & Fatemi, E. Nonlinear total variation based noise removal algorithms. *Physica D* **60**, 259–268 (1992).
- [15] van Aarle, W. *et al.* The ASTRA Toolbox: a platform for advanced algorithm development in electron tomography. *Ultramicroscopy* **157**, 35–47 (2016).
- [16] Ong, F. & Lustig, M. SigPy: a Python package for high performance iterative reconstruction. *Proc. ISMRM* (2019).
- [17] Adler, J. & Öktem, O. Operator Discretisation Library (ODL). Software available from <https://github.com/odlgroup/odl> (2017).
- [18] Tachella, J. *et al.* DeepInverse: a deep learning framework for inverse problems in imaging. *Preprint at* <https://arxiv.org/abs/2312.09828> (2023).
- [19] Jumper, J. *et al.* Highly accurate protein structure prediction with AlphaFold. *Nature* **596**, 583–589 (2021).
- [20] Merchant, A. *et al.* Scaling deep learning for materials discovery. *Nature* **624**, 80–85 (2023).
- [21] Karniadakis, G. E. *et al.* Physics-informed machine learning. *Nature Rev. Phys.* **3**, 422–440 (2021).
- [22] Wang, H. *et al.* Scientific discovery in the age of artificial intelligence. *Nature* **620**, 47–60 (2023).
- [23] Bommasani, R. *et al.* On the opportunities and risks of foundation models. *Preprint at* <https://arxiv.org/abs/2108.07258> (2022).
- [24] Ongie, G. *et al.* Deep learning techniques for inverse problems in imaging. *IEEE J. Sel. Top. Signal Process.* **14**, 171–182 (2020).
- [25] Monga, V., Li, Y. & Eldar, Y. C. Algorithm unrolling: Interpretable, efficient deep learning for signal and image processing. *IEEE Signal Process. Mag.* **38**, 18–44 (2021).

- [26] Kamilov, U.S. *et al.* Plug-and-play methods for integrating physical and learned models in computational imaging. *IEEE Signal Process. Mag.* **40**, 85–97 (2023).
- [27] Aggarwal, H.K., Mani, M.P. & Jacob, M. MoDL: model-based deep learning architecture for inverse problems. *IEEE Trans. Med. Imaging* **38**, 394–405 (2019).
- [28] Adler, J. & Öktem, O. Learned primal-dual reconstruction. *IEEE Trans. Med. Imaging* **37**, 1322–1332 (2018).
- [29] Tian, L., Hunt, B. & Bhatt, M. Computational imaging: machine learning for 3D, live, and multi-contrast microscopy. *Optica* **10**, 464–474 (2023).
- [30] Yuan, X., Brady, D.J. & Katsaggelos, A.K. Snapshot compressive imaging: theory, algorithms, and applications. *IEEE Signal Process. Mag.* **38**, 65–88 (2021).
- [31] Arridge, S.R. & Schotland, J.C. Optical tomography: forward and inverse problems. *Inverse Problems* **25**, 123010 (2009).
- [32] Jiang, Y. *et al.* Electron ptychography of 2D materials to deep sub-ångström resolution. *Nature* **559**, 343–349 (2018).
- [33] Rodenburg, J.M. & Maiden, A.M. Ptychography. In: *Springer Handbook of Microscopy* (Springer, 2019).
- [34] Gustafsson, M.G.L. *et al.* Three-dimensional resolution doubling in wide-field fluorescence microscopy by structured illumination. *Biophysical J.* **94**, 4957–4970 (2008).
- [35] Chambolle, A. & Pock, T. An introduction to continuous optimization for imaging. *Acta Numerica* **25**, 161–319 (2016).
- [36] Beck, A. & Teboulle, M. A fast iterative shrinkage-thresholding algorithm for linear inverse problems. *SIAM J. Imaging Sci.* **2**, 183–202 (2009).
- [37] Boyd, S., Parikh, N., Chu, E., Peleato, B. & Eckstein, J. Distributed optimization and statistical learning via the alternating direction method of multipliers. *Found. Trends Mach. Learn.* **3**, 1–122 (2011).
- [38] OpenAI. GPT-4 technical report. *Preprint* at <https://arxiv.org/abs/2303.08774> (2023).
- [39] Schick, T. *et al.* Toolformer: language models can teach themselves to use tools. *Proc. NeurIPS* (2023).
- [40] Boiko, D.A., MacKnight, R., Kline, B. & Gomes, G. Autonomous chemical research with large language models. *Nature* **624**, 570–578 (2023).
- [41] Pearson, W.R. & Lipman, D.J. Improved tools for biological sequence comparison. *Proc. Natl Acad. Sci. USA* **85**, 2444–2448 (1988).
- [42] Kodama, Y., Shumway, M. & Leinonen, R. The Sequence Read Archive: explosive growth of sequencing data. *Nucleic Acids Res.* **40**, D54–D56 (2012).