

Paper Summary for Prof. Steve Jiang

Physics World Models for Computational Imaging:
Clinical Translation and CT Quality Assurance

Chengshuai Yang Xin Yuan

February 2026 — Prepared for Nature submission

Paper Overview

This paper introduces **Physics World Models (PWM)**, a universal diagnostic and correction framework for computational imaging. The theoretical backbone is the **Triad Decomposition**, which decomposes every imaging failure into three gates:

- **Gate 1** (Recoverability): Does the measurement encode sufficient information?
- **Gate 2** (Carrier Budget): Is the signal-to-noise ratio sufficient?
- **Gate 3** (Operator Mismatch): Does the assumed forward model match the true physics?

Across 7 validated modalities, Gate 3 dominates in *every* case, including clinical CT and MRI. The framework has been translated to clinical quality assurance through the **CT QC Copilot**.

Sections Most Relevant to Your Expertise

1. CT QC Copilot: Clinical Translation of the Triad Decomposition

The Triad Decomposition maps directly to clinical CT failure modes:

Gate	Research	Clinical CT
Gate 1	Information deficiency	Protocol design inadequacy (insufficient projections, FOV)
Gate 2	Carrier budget (SNR)	Dose budget (noise vs. diagnostic need)
Gate 3	Operator mismatch	Scanner calibration drift (HU drift, CoR offset, gain)

Consistent with the research findings, **Gate 3 dominates in clinical practice**: most QA failures trace to calibration drift, not protocol or dose issues.

2. ACR Metric Validation (9 Metrics, GE Revolution Apex)

Nine ACR-aligned QC metrics are computed automatically from CT phantom (Gammex 464) scans:

Metric	Copilot	Console	Dev.	% Tol.
CT# Water (HU)	0.3	0.2	0.1	2%
CT# Bone (HU)	952	953	1.0	5%
CT# Air (HU)	−997	−998	1.0	5%
CT# Acrylic (HU)	120	121	1.0	7%
CT# Polyethylene (HU)	−96	−96	0.0	0%
Geometric accuracy (mm)	0.15	0.10	0.05	2.5%
Slice thickness (mm)	5.08	5.02	0.06	4%
Uniformity (HU)	1.2	1.1	0.1	2%
Noise std. dev. (HU)	4.6	4.5	0.1	10%
Spatial resolution (lp/cm)	6.2	6.0	0.2	4%

Agreement: ≤ 1.2 HU for CT number accuracy, ≤ 0.15 mm for geometric accuracy — well within clinical tolerance bands.

3. Automated Drift Detection (30-Scanner Fleet)

Statistical process control (SPC) using Western Electric rules over a 30-scanner simulated fleet:

- **Sensitivity:** 100% (4/4 drifting scanners detected; 95% CI: 39.8–100%)
- **Specificity:** 100% (0/26 false positives; 95% CI: 86.8–100%)
- **Early warning:** 3–6 months *before* ACR threshold exceedance
- Drift types: noise drift (2 scanners), uniformity drift (1), CT number drift (1)

4. Workflow Efficiency (94% Time Reduction)

Stage	Manual (min)	Copilot (min)
DICOM ingestion	5.0	0.01
Metric computation	25.0	0.01
Threshold evaluation	5.0	<0.01
Trend analysis	10.0	<0.01
Report generation	15.0	0.02
Physicist review/sign-off	7.0	4.0
Total	67 $\pm$ 12	4.2 $\pm$ 0.8

Savings: **377 physicist-hours/year** for a 30-scanner fleet (0.18 FTE reallocation).

5. MRI Validation

For MRI, the paper validates on fastMRI data with coil sensitivity mismatch (5–15% multiplicative deviation). A 5% coil sensitivity error produces 6.94 dB degradation; PWM correction recovers +48.25 dB, the largest correction gain among all modalities.

6. CT Correction Results

For CT, a sub-degree center-of-rotation offset produces characteristic ring artifacts. PWM correction recovers +10.68 dB. Six artifact signatures (ring, cupping, streaks, HU drift, noise ratio, geometric distortion) correspond directly to the OperatorGraph mismatch parameter families.

How Your Expertise Would Strengthen This Paper

1. **Prospective clinical validation:** Your access to clinical CT scanners and physics staff enables validating the CT QC Copilot with *real* ACR phantoms and physical scanner fleets, replacing the simulated fleet.
2. **Clinical MRI validation:** Your expertise in medical AI and radiation oncology provides the pathway to validate the MRI correction pipeline on clinical data with real coil sensitivity variations.
3. **Regulatory pathway:** Your experience with clinical AI deployment informs the regulatory strategy for the CT QC Copilot as a clinical decision support tool.
4. **Domain authority:** Co-authorship from a clinical medical physicist and AI researcher validates the clinical translation claims.

Paper Statistics

- Main paper: 34 pages (Nature review format), 7 figures
- Supplementary: 22 pages, including Supplementary Note 7 (Clinical CT QC Validation with 4 tables)
- Clinical metrics: 9 ACR-aligned QC metrics, 94% workflow time reduction
- MRI correction: +48.25 dB (largest gain), CT correction: +10.68 dB