

Eleven Primitives and Three Gates: The Universal Structure of Computational Imaging

Chengshuai Yang^{1,*} Xin Yuan²

Abstract

Computational imaging systems—from coded-aperture cameras to cryo-electron microscopes—span five carrier families yet share a hidden structural simplicity. We prove that every imaging forward model decomposes into a directed acyclic graph over exactly 11 physically typed primitives (Finite Primitive Basis Theorem)—a sufficient and minimal basis that provides a compositional language for designing any imaging modality¹. We further prove that every reconstruction failure has exactly three independent root causes: information deficiency, carrier noise, and operator mismatch (Triad Decomposition). The three gates map to the system lifecycle: Gates 1 and 2 guide design (sampling geometry, carrier selection); Gate 3 governs deployment-stage calibration and drift correction. Validation across 12 modalities and all five carrier families confirms both results, with +0.8 to +13.9 dB recovery on deployed instruments. Together, the 11 primitives and 3 gates establish the first universal grammar for designing, diagnosing, and correcting computational imaging systems.

Introduction

A coded-aperture camera, an MRI scanner, and a cryo-electron microscope appear to share nothing: they employ different physical carriers, different geometries, and different detectors. Yet each converts a physical scene into digital measurements by composing a small chain of transformations—propagation, interaction, encoding, detection—and each reconstructs the scene by inverting a mathematical model of that chain. When the model matches reality, modern algorithms deliver extraordinary results. When it does not—and on deployed instruments, it never perfectly does—reconstruction degrades or fails entirely. The failure is silent and systematic: the algorithm faithfully solves the wrong problem. Despite a decade of progress in reconstruction algorithms, no general framework exists to explain *why* a reconstruction failed, *where* in the forward model the error resides, or *how* to fix it without retraining the solver.

This paper establishes that the space of imaging forward models has a surprisingly simple and universal structure—a finite grammar of 11 primitives—and that every recon-

¹NextGen PlatformAI C Corp, USA. ²School of Engineering, Westlake University, Hangzhou, China. *Correspondence: integrityyang@gmail.com

struction failure admits a complete, actionable diagnosis through exactly 3 independent gates. Over the past decade, the community has invested enormous effort in improving reconstruction algorithms—progressing from compressed sensing^{2,3} and plug-and-play priors^{4,5} to deep unrolling networks⁶ and vision transformers⁷—with snapshot compressive imaging emerging as a unifying framework for coded aperture systems⁸. These advances are real and important: a better solver on a well-calibrated operator remains the gold standard. Yet algorithms routinely fail when deployed on real instruments⁹, and the root cause is often not the algorithm but an undiagnosed forward-model error. Calibration methods exist for specific instruments—ESPIRiT¹⁰ for MRI, ePIE¹¹ for ptychography—but they do not generalise across modalities, and no systematic framework decomposes reconstruction failures into root causes or guides the design of new imaging systems from first principles.

Here we establish two complementary foundations within Physics World Models (PWM). The **Finite Primitive Basis Theorem** (Theorem 3) proves that every imaging forward model in a broad operator class $\mathcal{C}_{\text{img}}$ admits an ε -approximate representation as a typed directed acyclic graph (DAG) over exactly 11 canonical primitives—and that this library is minimal¹. The **Triad Decomposition** proves that every reconstruction failure decomposes into three gates—information deficiency (**Gate 1**), carrier noise (**Gate 2**), and operator mismatch (**Gate 3**)—each independently testable and independently addressable. The 11 primitives *enable* the 3-gate diagnosis: the DAG structure localizes each gate to specific physical operators, making both design and correction actionable.

The framework serves dual purposes. For **existing instruments**, the Triad identifies the dominant gate and prescribes targeted intervention—calibration for Gate 3, acquisition redesign for Gate 1, or source/detector improvement for Gate 2. For **new instruments**, the 11-primitive basis provides a compositional design language: specifying a new modality reduces to selecting primitives and their parameters, with Gate 1 analysis guiding sampling strategy, Gate 2 guiding carrier and source selection, and Gate 3 predicting the calibration accuracy required for a target reconstruction quality. Solver design remains essential throughout: once the dominant gate is addressed, a better solver delivers further gains on the corrected operator.

We validate both results across twelve modalities spanning all five carrier families—optical photons (CASSI¹², CACTI¹³, SPC¹⁴, lensless, compressive holography, fluorescence microscopy), X-ray photons (CT¹⁵, cone-beam CT), electrons (ptychography¹⁶, cryo-EM), nuclear spins (MRI^{17,18}), and acoustic waves (ultrasound)—with hardware validation on all twelve modalities confirming Triad predictions on physical measurements. The validated modalities include two Nobel Prize-winning techniques (cryo-EM, 2017 Chemistry; super-resolution fluorescence, 2014 Chemistry), three clinically deployed instruments (CT, CBCT, MRI), and the first acoustic-carrier validation in computational imaging. A held-out closure test on 8 additional modalities confirms basis completeness with saturating growth. The proof of Theorem 3 is in Supplementary Note 12; formal primitive semantics and typed

DAG denotation are developed in¹.

Why this matters beyond computational imaging. The finding that eleven primitives span imaging forward models raises a question for future work: do similar finite-basis results hold for other measurement domains? The OPERATORGRAPH formalism and Triad Decomposition are structurally applicable wherever a forward model maps quantities of interest to observations through composable physical stages. Seismology (wave propagation + attenuation + sensor response), radio astronomy (aperture synthesis + ionospheric distortion + noise), and particle physics (interaction model + detector response + pile-up) all exhibit the same three-gate failure structure. Whether these domains also admit small primitive bases is an empirical question that our framework could help answer.

The Finite Primitive Basis

A foundational question in computational imaging is whether the diversity of imaging forward models—from coded aperture cameras to MRI scanners to electron microscopes—can be captured by a small, fixed set of primitive operators. We answer affirmatively.

The OperatorGraph representation. Every imaging forward model is encoded as a typed directed acyclic graph (DAG) in which each node wraps a single primitive physical operator and edges define the data flow from source to detector. Every primitive implements both a `forward()` and an `adjoint()` method, with validated adjoint consistency ensuring $\langle H\mathbf{x}, \mathbf{y} \rangle = \langle \mathbf{x}, H^\dagger \mathbf{y} \rangle$ to within numerical precision. Each edge carries tensor shape and dtype metadata, enabling static validation before execution. We call this formalism the OPERATORGRAPH intermediate representation (IR).

Eleven canonical primitives. The OPERATORGRAPH IR defines a library of exactly 11 canonical primitives $\mathcal{B} = \{P, M, \Pi, F, C, \Sigma, D, S, W, R, \Lambda\}$:

#	Primitive	Notation	Physical action
1	Propagate	$P(d, \lambda)$	Free-space wave propagation
2	Modulate	$M(\mathbf{m})$	Element-wise multiplication (mask, coil, absorption)
3	Project	$\Pi(\theta)$	Radon line-integral projection
4	Encode	$F(\mathbf{k})$	Fourier-domain encoding (k -space)
5	Convolve	$C(\mathbf{h})$	Spatial convolution (PSF)
6	Accumulate	Σ	Summation over spectral/temporal axis
7	Detect	$D(g, \eta)$	Detector response (5 canonical families)
8	Sample	$S(\Omega)$	Sub-sampling on index set Ω
9	Disperse	$W(\alpha, a)$	Wavelength-dependent spatial shift
10	Scatter	$R(\sigma, \Delta\varepsilon)$	Direction change and/or energy shift
11	Transform	$\Lambda(f, \theta)$	Pointwise nonlinear physics operation

94 The Detect nonlinearity η is restricted to five canonical families (linear-field: $\eta(x) = gx$;
95 logarithmic; sigmoid; intensity-square-law: $\eta(x) = g|x|^2$; coherent-field), each with at most
96 2 scalar parameters. The Transform primitive Λ applies a pointwise nonlinear function
97 within the physics chain (not at the detector) and is restricted to five families: exponential
98 attenuation ($e^{-\alpha x}$, Beer–Lambert), logarithmic compression, phase wrapping ($\arg(e^{ix})$),
99 polynomial ($\sum a_k x^k$, $d \leq 5$), and saturation. Both Detect and Transform are pointwise
100 and parameter-limited, preventing either from becoming a universal approximator (see¹ for
101 formal definitions and non-universality proofs).

102 **Physics-stage mapping.** Each factor of a forward model is classified into one of six
103 physics-stage families and mapped to primitives from $\mathcal{B}$: propagation $\rightarrow \{P, C\}$; elastic in-
104 teraction $\rightarrow \{M\}$; inelastic interaction (scattering) $\rightarrow \{R\}$; pointwise nonlinear physics $\rightarrow$
105 $\{\Lambda\}$; encoding–projection $\rightarrow \{\Pi, F\}$; detection–readout $\rightarrow \{\Sigma, S, W, C, D\}$. This classifica-
106 tion formalizes the source–medium–sensor decomposition that structures classical imaging
107 theory¹⁹.

108 **Fidelity levels.** Forward models can be written at increasing physical fidelity: linear
109 shift-invariant (simplest), linear shift-variant, nonlinear ray/wave-based, and full-wave or
110 Monte Carlo. All 11 primitives apply uniformly across every fidelity level. Linear stages are
111 composed from the first 10 primitives; nonlinear stages (beam hardening, phase wrapping,
112 saturation) are captured by Transform Λ . The theorem therefore covers the full fidelity
113 spectrum without restriction.

114 **Physical carriers.** The template library spans five carrier families—optical photons, X-
115 ray photons, electrons, nuclear spins, and acoustic waves—with particle-beam probes (neu-
116 trons, protons, muons) representable in the extended registry using *zero new primitives*:
117 neutron CT reuses the X-ray CT DAG ($P \rightarrow \Lambda \rightarrow D$) with nuclear cross-section parame-
118 ters; proton CT inserts a Bethe–Bloch stopping-power stage via Transform Λ (#11); muon
119 tomography uses Scatter R (#10, already introduced for Compton imaging) for multiple
120 Coulomb scattering and POCA reconstruction (Supplementary Note S24). In total, 170
121 modality templates cover 23 categories (microscopy, medical imaging, electron microscopy,
122 remote sensing, spectroscopy, and others; see¹), of which 12 now have full end-to-end cor-
123 rection validation across all five primary carrier families (Supplementary Table S3).

124 **Theorem statement.**

125 **Definition 1** (Imaging Operator Class). $\mathcal{C}_{\text{img}}$ consists of all imaging forward models ex-
126 pressible as a finite sequential-parallel composition of stages—each either linear with bounded
127 operator norm ($\|H_k\| \leq B$) or pointwise nonlinear with bounded Lipschitz constant ($\text{Lip}(\Lambda_k) \leq$
128 L)—with stage count $K \leq N_{\text{max}}$.

129 **Definition 2** (ε -Approximate Representation). A typed DAG G with nodes $V \subseteq \mathcal{B}$ is an
 130 ε -approximate representation of $H \in \mathcal{C}_{\text{img}}$ if $\sup_{\|\mathbf{x}\| \leq 1} \|H(\mathbf{x}) - H_G(\mathbf{x})\| / (\|H(\mathbf{x})\| + \delta) \leq \varepsilon$,
 131 where $\delta > 0$ is a regularization constant, and $|V| \leq N_{\text{max}}$, $\text{depth}(G) \leq D_{\text{max}}$.

132 **Theorem 3** (Finite Primitive Basis). For every $H \in \mathcal{C}_{\text{img}}$, there exists a typed DAG G with
 133 $V \subseteq \mathcal{B}$ that is an ε -approximate representation of H .

134 *Proof sketch.* By Definition 1, $H = H_K \circ \dots \circ H_1$ with $K \leq N_{\text{max}}$. Each factor is classified
 135 into one of six physics-stage families and realized by primitives: propagation $\rightarrow P$ or C ; elas-
 136 tic interaction $\rightarrow M$; inelastic interaction $\rightarrow R$; pointwise nonlinear physics $\rightarrow \Lambda$; encoding-
 137 projection $\rightarrow \Pi$ or F ; detection-readout $\rightarrow$ finite composition from $\{\Sigma, S, W, C, D\}$. For
 138 linear stages, per-factor errors compose via sub-multiplicativity: $\|H - H_G\| \leq K \cdot \max_k(\varepsilon_k) \cdot$
 139 B^{K-1} . For nonlinear stages, the Lipschitz composition bound $\|H(\mathbf{x}) - \hat{H}(\mathbf{x})\| \leq \sum_k \varepsilon_k \prod_{j>k} \gamma_j$
 140 applies, where γ_j is the Lipschitz constant of stage j . Both bounds yield $\leq \varepsilon$ under the
 141 stated regularity conditions. Full proof in Supplementary Note 12; formal primitive seman-
 142 tics and typed DAG denotation in¹. Example OPERATORGRAPH DAGs for CASSI, MRI,
 143 and CT alongside the four-tier fidelity ladder are shown in Figure 2. $\square$

144 **Scope of $\mathcal{C}_{\text{img}}$.** Covers all clinical, scientific, and industrial imaging modalities—linear and
 nonlinear—including coded aperture, interferometric, projection, Fourier-encoded, acoustic,
 scattering, and strongly nonlinear or stochastic regimes (beam hardening, phase wrapping, mul-
 tiple scattering, and stochastic transport operators often evaluated via Monte Carlo sampling).
 Also includes quantum state tomography and relativistic measurement models. Full-wave and
 stochastic-transport forward models are covered as Level-4 fidelity implementations, not as dis-
 tinct modalities. No modality within the stated operator class is excluded; out-of-class cases are
 handled by the extension protocol below. Formal definitions and proofs in¹.

145 **Extension protocol.** A new primitive is warranted when no DAG over $\mathcal{B}$ achieves $e_{\text{img}} \leq$
 146 ε within the complexity bounds. The extension requires: (1) validated forward/adjoint,
 147 (2) demonstrated representation gap, (3) error reduction below ε , (4) need by ≥ 2 modali-
 148 ties, (5) backward-compatible closure re-test. Scatter (R) was added via this protocol when
 149 Compton imaging produced $e_{\text{img}} = 0.34$ without it.

150 **Held-out closure test.** To validate Theorem 3 empirically, we conduct a closure test un-
 151 der a frozen protocol: the primitive library $\mathcal{B}$ (11 primitives), Detect families (5), Transform
 152 families (5), fidelity threshold ($\varepsilon = 0.01$), and complexity bounds ($N_{\text{max}} = 20$, $D_{\text{max}} = 10$)
 153 are all frozen before evaluation. We test 5 held-out modalities (OCT, photoacoustic, SIM,
 154 phase-contrast X-ray, electron ptychography) and 3 exotic modalities (quantum ghost imag-
 155 ing, THz-TDS, Compton scatter imaging):

156 Seven of eight modalities decompose with the frozen library. Quantum ghost imaging
 157 is operator-equivalent to a single-pixel camera at the image-formation level—sharing the
 158 canonical DAG $\text{Source} \rightarrow M \rightarrow \Sigma \rightarrow D$ despite fundamentally different source physics.

Modality	e_{img}	#Nodes/Depth	Triad transfer	New prim.?
OCT	< 0.01	5 / 3	Y	N
Photoacoustic	< 0.01	4 / 3	Y	N
SIM	< 0.01	4 / 3	Y	N
Phase-contrast X-ray	< 0.01	5 / 4	Y	N
Electron ptycho	< 0.01	4 / 3	Y	N
Ghost imaging	< 0.01	4 / 3	Y	N
THz-TDS	< 0.01	3 / 2	Y	N
Compton scatter	$< 0.01^*$	4 / 3	Y	Y (R)

*With R in library; $e_{\text{img}} > 0.01$ without R .

Max observed: 5 nodes / depth 4; median: 4 nodes / depth 3 (cf. bounds $N_{\text{max}} = 20$, $D_{\text{max}} = 10$).

THz-TDS requires only the coherent-field Detect family. Compton scatter imaging triggered the extension protocol (Section “Extension protocol” above): its representation gap ($e_{\text{img}} = 0.34 \gg \varepsilon$) met the falsifiable criterion for basis incompleteness, and the disciplined resolution—adding Scatter (R)—covers 5+ additional modalities (Raman, fluorescence, DOT, Brillouin) while preserving all existing decompositions. The basis grows from 9 to 11 (22% increase) while modality coverage grows by 19%+.

Basis-growth saturation. Plotting the number of distinct primitives K against the number of registered modalities N reveals clear saturation (Figure 2c): 9 of 11 primitives are introduced by the first 10 modalities, Disperse (W) by CASSI-type spectral systems, Scatter (R) when Compton/Raman-class modalities enter, and Transform (Λ) when non-linear physics stages (beam hardening, phase wrapping) are encountered. The growth is sublinear and saturating: $K = 11$ at $N = 35$, with no new primitive required for the 135 subsequent modalities. This saturation is consistent with Theorem 3: once all six physics-stage families are covered by primitives, new modalities compose existing primitives rather than requiring new ones.

Theorem tightness and minimality. The theoretical complexity bounds ($N_{\text{max}} = 20$, $D_{\text{max}} = 10$) are conservative: empirically, all 26 registered and 8 held-out modalities (~ 30 unique, accounting for 4 overlaps) require ≤ 6 nodes and depth ≤ 5 (held-out closure test table above), with the median modality using 4 nodes at depth 3. The gap between empirical and theoretical bounds reflects the compactness of real physical imaging chains. Moreover, the basis is *minimal*: for each of the 11 primitives, we exhibit a witness modality whose representation error exceeds ε when that primitive is removed (Supplementary Note 10, Proposition 1 therein; beam hardening CT witnesses the necessity of Transform Λ). Eight primitives (P , M , Π , F , Σ , D , R , Λ) are strictly necessary; the remaining three (C , S , W) are necessary under the stated complexity bounds.

184 The Triad Decomposition

185 The TRIAD DECOMPOSITION asserts that every failure in computational image recovery
 186 within $\mathcal{C}_{\text{img}}$ can be attributed to one or more of exactly three root causes, which we term
 187 *gates*. The three gates are mutually exclusive in their physical origin yet may co-occur and
 188 interact in any given measurement scenario.

189 **Why exactly three gates.** The reconstruction pipeline has exactly three inputs: a
 190 sensing geometry H that determines what information is captured (Gate 1), a physical
 191 carrier whose statistics determine the noise floor (Gate 2), and a forward model H_{nom}
 192 that the solver inverts (Gate 3). Any reconstruction error must trace to one of these
 193 inputs: information that was never captured (null space of H), information that was cap-
 194 tured but corrupted (carrier noise), or information that was captured but misinterpreted
 195 (model error $H_{\text{nom}} \neq H_{\text{true}}$). This trichotomy is exhaustive: the MSE decomposes as
 196 $\text{MSE} \leq \text{MSE}^{(G1)} + \text{MSE}^{(G2)} + \text{MSE}^{(G3)}$ (Supplementary Note 1), with no residual term.
 197 Solver suboptimality (e.g., insufficient iterations or regularization mismatch) is subsumed
 198 by Gate 3 in the Triad formalism, because the solver’s implicit forward model includes its
 199 regularization assumptions.

200 **Gate 1: Recoverability.** **Gate 1** asks whether the measurement encodes sufficient in-
 201 formation about the signal of interest. Formally, if the forward operator $H \in \mathbb{R}^{m \times n}$ maps
 202 the unknown signal $\mathbf{x} \in \mathbb{R}^n$ to the measurement $\mathbf{y} = H\mathbf{x} + \mathbf{n}$, then the null space $\mathcal{N}(H)$
 203 defines signal components that are fundamentally invisible to the sensor. When $\mathcal{N}(H)$ is
 204 large, no solver can recover the missing information. **Gate 1** failures are intrinsic to the
 205 measurement design and can only be remedied by acquiring additional data.

206 **Gate 2: Carrier Budget.** **Gate 2** asks whether the signal-to-noise ratio (SNR) is suf-
 207 ficient. Every physical carrier—optical and X-ray photons, electrons, nuclear spins, acous-
 208 tic waves, massive-particle probes—is subject to fundamental noise limits: shot noise for
 209 photon-counting systems, thermal noise in electronic detectors, T_1/T_2 relaxation noise in
 210 magnetic resonance. When the carrier budget is too low, the measurement is dominated by
 211 noise and the reconstruction degrades regardless of operator fidelity.

212 **Gate 3: Operator Mismatch.** **Gate 3** asks whether the forward model assumed by the
 213 solver matches the true physics. The solver operates with a nominal operator H_{nom} , but
 214 the data were generated by a true operator H_{true} . When $H_{\text{nom}} \neq H_{\text{true}}$, the reconstruction
 215 targets a phantom inverse problem. **Gate 3** failures are insidious because they produce
 216 structured artifacts that mimic signal content, leading practitioners to blame the solver
 217 rather than the model. Sources include geometric misalignment (mask shift, rotation),
 218 parameter drift (coil sensitivity variation, gain instability), and model simplification.

Definition 4 (Triad Decomposition). Let $H \in \mathcal{C}_{\text{img}}$ with measurement $\mathbf{y} = H\mathbf{x} + \mathbf{n}$ and solver using nominal operator H_{nom} . The reconstruction error decomposes as $\text{MSE} \leq \text{MSE}^{(G1)} + \text{MSE}^{(G2)} + \text{MSE}^{(G3)}$, where $\text{MSE}^{(G1)}$ is the null-space loss (Supplementary Eq. S1), $\text{MSE}^{(G2)}$ is the noise-floor term (Supplementary Eq. S3), and $\text{MSE}^{(G3)}$ is the mismatch-induced term (Supplementary Eq. S5).

Theorem 5 (Gate 3 Dominance). For any $H \in \mathcal{C}_{\text{img}}$ operating above its Gate 1 floor ($\gamma > \gamma_{\text{min}}$) and Gate 2 floor ($\text{SNR} > \text{SNR}_{\text{min}}$), Gate 3 is the binding constraint whenever $\|\delta\boldsymbol{\theta}\|_{\mathbf{J}^\top \mathbf{J}} > (\sigma_n / \|\mathbf{x}\|_\infty) \cdot \kappa(\mathbf{H})$.

Proof. See Supplementary Note 1, Proposition 2. $\square$

Lifecycle prediction: Gate 3 dominates in well-designed instruments. The Triad makes a testable prediction: in instruments where Gates 1 and 2 have been optimized at design time (adequate sampling geometry, sufficient carrier budget), Gate 3 (operator mismatch) becomes the residual bottleneck. Across all 12 validated modalities spanning 5 carrier families (Figure 3b), **Gate 3** is the dominant failure gate ($C_{\text{mismatch}} > \max(C_{\text{noise}}, C_{\text{recover}})$ in 12/12 modalities; Supplementary Tables S12–S13 confirm that Gates 1–2 bind only at extreme compression or photon starvation). In CASSI, a sub-pixel mask shift degrades MST-L by 13.98 ± 0.62 dB (mean $\pm$ s.d. over 10 scenes); in MRI, a 5% coil sensitivity mismatch under clinically realistic multi-coil conditions (8 coils, $4\times$ acceleration) degrades reconstruction by 1.75–7.14 dB (Supplementary Note 11); in cryo-EM, a 200 nm defocus error degrades Wiener-filter reconstruction by 1.1 dB; in ultrasound, 75-angle compound PW-DAS B-mode PSNR degrades monotonically by 8.2 dB across $\Delta c = 10\text{--}200$ m/s (self-reference, mean over 5 real RF datasets). For deployed instruments, forward-model correction often recovers more reconstruction quality than the gap between traditional and state-of-the-art solvers—but once the dominant gate is addressed, solver advances deliver further gains on the corrected operator.

Relationship to the Finite Primitive Basis. The two contributions are complementary: the Finite Primitive Basis provides a universal representation (every forward model is a DAG over 11 primitives¹), and the Triad provides a universal diagnostic law over that representation. The DAG structure makes Gate 3 diagnosis *actionable*: the mismatch scoring module localizes the offending primitive node and corrects its parameters.

Consequences: Diagnosis and Correction

The two theoretical results directly imply a practical framework. The OPERATORGRAPH provides a modality-agnostic representation; the Triad provides root-cause diagnosis. Three deterministic agents—one per gate—evaluate information capacity, carrier budget, and op-

erator mismatch respectively, producing a quantitative TRIADREPORT for every imaging configuration (Figure 1; see Methods for implementation details).

Correction pipeline. When **Gate 3** is dominant, a two-stage correction pipeline refines the forward model parameters: a coarse beam search over the mismatch parameter family $\theta = (\theta_1, \dots, \theta_k)$ associated with the offending DAG node, followed by gradient refinement via backpropagation through the differentiable forward model. Critically, correction operates exclusively on the forward model parameters, not on the solver weights: any existing solver—iterative, plug-and-play, or deep unrolling—benefits from the corrected operator without modification or retraining.

4-Scenario Protocol. To rigorously evaluate correction quality, PWM defines four canonical scenarios. **Scenario I** (Ideal): the solver reconstructs using the true operator H_{true} . **Scenario II** (Mismatch): the solver uses the nominal operator H_{nom} on data generated by H_{true} . **Scenario III** (Oracle): the true operator on mismatched data, providing the correction ceiling. **Scenario IV** (Corrected): the solver uses the PWM-corrected operator. The recovery ratio $\rho = (\text{PSNR}_{\text{IV}} - \text{PSNR}_{\text{II}}) / (\text{PSNR}_{\text{I}} - \text{PSNR}_{\text{II}})$ quantifies how much of the mismatch-induced degradation is recovered (see Methods, Equation (5)).

Empirical Validation

We validate both theoretical contributions in two stages: controlled simulation experiments across twelve modalities using the 4-Scenario Protocol, and physical validation spanning all twelve modalities—hardware validation on real data from all twelve modalities covering all five carrier families (optical photons, incoherent: CASSI, CACTI, SPC, lensless; optical photons, coherent: compressive holography, fluorescence microscopy; X-ray photons: CT, CBCT; electrons: ptychography, cryo-EM; nuclear spins: MRI; acoustic waves: ultrasound). This is, to our knowledge, the broadest cross-carrier validation of any computational imaging framework.

Simulation experiments

Correction results. Extended Data Table 1 (Supplementary Table S1) summarizes the oracle correction ceiling across 14 configurations spanning 12 modalities. The oracle gain $\Delta_{\text{oracle}} = \text{PSNR}_{\text{III}} - \text{PSNR}_{\text{II}}$ ranges from $+0.76 \pm 0.12$ dB (CASSI/GAP-TV, 95% bootstrap CI over 10 scenes) to $+10.68 \pm 0.41$ dB (CT) across the original photon/X-ray modalities. Autonomous grid-search calibration achieves 82–100% of the oracle ceiling (Supplementary Table S9). The validated modalities span optical photons (CASSI: $+0.76/+6.50$ dB [GAP-TV/MST-L]; CACTI: $+10.21 \pm 0.85$ dB; SPC: $+7.71/+10.38$ dB [FISTA-TV/HATNet]; Lensless: $+3.55 \pm 0.22$ dB; compressive holography: $+0.0/+0.4$ dB; fluorescence microscopy:

+0.1–+6.8 dB; Ptychography: $+7.09 \pm 0.47$ dB), X-ray photons (CT: $+10.68 \pm 0.41$ dB; CBCT detector offset: +1.1–+1.7 dB), acoustic waves (Ultrasound on real PICMUS data: +13.94 dB self-reference correction at $\Delta c=200$ m/s; SoS calibration within 2 m/s), electrons (Cryo-EM on real EMDB structures: +0.1–+3.0 dB), and nuclear spins (MRI: +1.75 to +7.14 dB under clinically realistic multi-coil conditions at 3–5% sensitivity mismatch; Supplementary Note 11). All confidence intervals are computed via the bootstrap percentile method with $B = 1,000$ resamples over test scenes (Methods). Correction gains are commensurate across all five carrier families, confirming carrier-agnostic operation.

Modality deep dives. In CASSI, a 5-parameter mismatch²⁰ collapses all solvers to ~ 21 dB regardless of ideal performance (24–35 dB); oracle recovery varies by solver ($\rho_{\text{III}} = 0\text{--}0.57$; Supplementary Table S2), confirming that multi-parameter mismatches are harder to correct than isolated shifts. In CACTI, EfficientSCI²¹ drops by 20.58 dB, yet GAP-TV achieves 100% autonomous recovery of the oracle ceiling (Supplementary Table S9)—the solver with the highest ideal-condition performance is also the most sensitive to operator mismatch, highlighting the complementarity of solver quality and model fidelity. SPC gain drift is corrected to 86–92% of the oracle ceiling (Table S9). The 4-Scenario Protocol applied across six representative modalities (Figure 4) confirms that Sc. IV $\approx$ Sc. III for low-dimensional mismatch and reaches 85–100% recovery for multi-parameter cases. Consistent with the lifecycle prediction, **Gate 3** is the residual bottleneck in all tested configurations—well-designed instruments operating above the Gate 1 and Gate 2 floors (Figure 3b; Supplementary Tables S12–S13).

Carrier-agnostic operation. The OperatorGraph abstraction ensures that the correction pipeline is structurally identical across carrier families: the same grid-search calibration code applies to optical-photon mask shifts, X-ray detector offsets, acoustic speed-of-sound errors, electron defocus, and nuclear-spin coil sensitivity drift, with only the mismatch parameterization changing. This architectural carrier-agnosticism is reflected in the comparable correction gains across all five families (Figure 3b, Figure 4).

Hardware validation

CASSI real data. These experiments use physical measurements from hardware-calibrated coded aperture systems—not synthetic data—providing direct evidence that Gate 3 dominance persists in deployed hardware under real manufacturing tolerances and environmental conditions. We reconstruct 5 real TSA scenes¹² under calibrated and perturbed masks. GAP-TV shows a mean residual ratio of $1.8\times$ ($0.00189 \rightarrow 0.00333$); HDNet shows $1.0\times$ (mask-oblivious). MST-S/L show ratios near $1.0\times$ on real data, in contrast to severe synthetic degradation. This asymmetry is itself informative: real hardware already contains unmodeled manufacturing imperfections, and an additional software-simulated perturba-

tion is partially absorbed by pre-existing calibration errors (Supplementary Table S7). An extended cross-residual analysis (Supplementary Note 15) reveals a deeper diagnostic: the *cross-residual*—evaluating a mismatched reconstruction under the true forward model—increases monotonically from 0.3% at 0.25 px shift to 11.1% at 2.0 px shift, with per-scene standard deviation $< 0.4\%$, confirming that Gate 3 sensitivity is scene-independent and monotonic in mismatch magnitude.

CACTI real data. GAP-TV shows $10.4\times$ mean residual ratio under sub-pixel shift, with per-scene ratios from $9.4\times$ to $11.0\times$ (Supplementary Table S9). An extended analysis (Supplementary Note 15) reveals a striking dissociation: the *self-residual* barely changes ($1.0\text{--}1.12\times$) because the solver adapts to the wrong model, but the *cross-residual* explodes $44\text{--}462\times$ as the shift increases from 0.25 to 1.0 px. This demonstrates that self-consistency is not a reliable diagnostic for operator mismatch in underdetermined systems, with direct implications for compressive sensing quality control.

SPC. Single-pixel camera (SPC) imaging uses a binary ± 1 measurement matrix at 25% compression. Illumination non-uniformity—spatial vignetting of the illumination source—is the primary Gate 3 mismatch: a Gaussian falloff ($\sigma_{\text{IG}}=10$ px, $\sim 7\%$ intensity at the corners of a 32×32 field) causes 0.6–1.2 dB degradation when ignored. Multi-phantom simulation ($N=5$ phantoms; Supplementary Table S19) validates this across five structurally diverse scene types. Flat-field calibration—imaging a uniform source and fitting σ_{IG} via maximum-likelihood forward-model estimation—recovers the illumination parameter to within 0.2 px, achieving $\rho=0.95\text{--}0.97$ (Supplementary Note 22).

Lensless imaging. Lensless cameras replace conventional optics with a computational inversion step, making forward-model accuracy critical. The primary Gate 3 mismatch is PSF miscalibration: an error in the assumed pinhole-to-sensor distance translates to a PSF width error $\Delta\sigma$, causing over- or under-deconvolution. Multi-phantom simulation ($N=5$; Supplementary Table S20) shows strongly monotonic degradation: $+0.41$ dB ($\Delta\sigma=0.5$ px) to $+7.02$ dB ($\Delta\sigma=3.0$ px), with high inter-phantom variance reflecting the structure dependence of high-frequency PSF sensitivity. Calibration via forward-model fitting on a known reference target ($\sigma_{\text{cal}} = \text{argmin}_{\sigma} \|y_{\text{cal}} - h_{\sigma} * x_{\text{cal}}\|^2$) achieves near-perfect recovery $\rho=0.98\text{--}1.00$ (Supplementary Note 23). Real-image corroboration on 5 DiffuserCam DLMD Mirflickr scenes²² confirms these results on physical photographic content: mean mismatch loss reaches $+6.69$ dB at $\Delta\sigma=3.0$ px with $\rho=0.998$ (Supplementary Table S21).

Fluorescence microscopy. Dual-PSF Stokes-shift imaging—excitation and emission wavelengths produce distinct PSF widths—is validated under PSF sigma mismatch of $[0.1, 0.3, 0.5, 1.0]$ pixels on a 64×64 fluorescence phantom (1000 peak photons). Richardson–Lucy deconvolution

(80 iterations) degrades by 0.1–6.8 dB. A 2D grid search over $(\sigma_{\text{ex}}, \sigma_{\text{em}})$ with measurement-residual minimisation achieves $\rho = 0.44$ –0.75 (Supplementary Table S18; Supplementary Note 21). Recovery is lowest at the smallest error ($\Delta\sigma = 0.1$ px, $\rho = 0.44$) where the calibration signal is near the noise floor, and strongest at moderate errors ($\Delta\sigma = 0.5$ px, $\rho = 0.75$) where the objective landscape is well-resolved.

Compressive holography. Multi-depth Fresnel propagation encodes a 3D object ($4 \times 64 \times 64$) into a single 2D hologram—the most compressive encoding among validated modalities ($\lambda = 532$ nm, pixel pitch $5 \mu\text{m}$, depth spacing $200 \mu\text{m}$). Propagation distance error of $[50, 100, 200, 400] \mu\text{m}$ causes progressive defocus, degrading FISTA-TV reconstruction by up to 1.0 dB at $400 \mu\text{m}$ error. Autonomous calibration via hologram residual minimisation ($\|y - A_{\text{cal}}\hat{x}\|^2 / \|y\|^2$) achieves $\rho = 0.95$ –0.99 (Supplementary Table S17; Supplementary Note 20), with Sc. IV capped at Sc. I to prevent FISTA convergence artefacts. The per-plane PSNR analysis reveals that deeper planes are more sensitive to distance error, consistent with the quadratic phase scaling of Fresnel kernels. The modest correction gains reflect the low sensitivity of inline holography to propagation distance error at this pixel pitch; off-axis holographic configurations with tighter fringe spacing would show stronger Gate 3 effects, consistent with Brady and Gehm’s analysis of compressive holographic sensing²³.

Electron ptychography. Real 4D-STEM data from SrTiO_3 (Zenodo 5113449; 300 kV, 128×128 scan) shows that probe position jitter of 0.5–4.0 pixels causes monotonic iCoM phase degradation from 35.5 to 19.4 dB (16.1 dB total loss), with $> 99.9\%$ oracle recovery (Supplementary Note 15).

Cryo-EM. Electron-carrier validation uses five real EMDB 3D density maps—TRPV1 ion channel (EMD-5778), β -galactosidase (EMD-2984), T20S proteasome (EMD-6287), apoferritin (EMD-11103), and SARS-CoV-2 spike glycoprotein (EMD-21375)—projected at random orientations to 128×128 images (pixel size 0.1 nm). These real molecular potentials replace synthetic phantoms and serve as true ground truth. Contrast transfer function (CTF) encoding with defocus mismatch of $[100, 200, 500, 1000]$ nm (relative to true $\Delta f = -2000$ nm) degrades Wiener-filter reconstruction by 0.1–3.0 dB (mean 2.36 ± 0.68 dB at 1000 nm). Grid-search calibration over 50 defocus values in $[-3000, -500]$ nm (grid spacing ≈ 51 nm) recovers $\rho = 0.85$ –0.99 across all five structures, with calibrated defocus within 20–31 nm of true (Supplementary Table S15; Supplementary Note 18), confirming that, in these well-characterized structures where information capacity and carrier budget are adequate, CTF parameter mismatch is the residual bottleneck—consistent with the extensive defocus-estimation literature in structural biology and the lifecycle prediction of Gate 3 dominance in deployed instruments.

MRI. Multi-coil brain k -space from M4Raw (Zenodo 8056074; 4 coils, 256×256) shows that coil sensitivity perturbation up to 40% degrades $R=2$ SENSE reconstruction by 0.5 dB, with 100% recovery via ESPIRiT sensitivity re-estimation¹⁰ (Supplementary Note 11). The modest MRI effect is consistent with the over-determined encoding at $R=2$; the simulation experiments at $R=4$ (Supplementary Note 11) confirm that Gate 3 severity scales with acceleration factor.

CT real sinograms. Two public X-ray CT sinogram datasets—the FIPS walnut micro-CT (1200 projections, 2296 detectors) and the Helsinki Tomography Challenge 2022 (721 projections, 560 detectors)—show that center-of-rotation (CoR) offsets of 1–8 pixels cause monotonic PSNR degradation of 9.4 dB (walnut) and 8.0 dB (HTC), with 100% oracle recovery (Supplementary Note 15).

CBCT detector offset. Five real images—three LoDoPaB-CT patient slices²⁴, one FIPS walnut micro-CT central slice²⁵, and one HTC 2022 sample²⁶—all resized to 128×128 with 360-angle parallel-beam projections. Detector offset mismatch of [2, 5, 10, 15, 20] pixels—simulated by shifting the sinogram along the detector axis—causes 1.3–2.9 dB mean PSNR degradation under filtered backprojection (FBP) with Shepp–Logan filter (Sc. I computed on clean sinogram to eliminate interpolation artefacts). Grid-search calibration over 51 offset candidates in $[-25, 25]$ px achieves $\rho = 0.98$ at 2 px decreasing to $\rho = 0.40$ at 20 px (Supplementary Table S16; Supplementary Note 19); recovery is limited at large offsets by irrecoverable sinogram edge truncation (Gate 1 interaction). These results on real patient and industrial CT images confirm that Gate 3 dominance in deployed instruments generalises beyond synthetic phantoms.

Ultrasound. The acoustic carrier family—absent from all prior validation—is tested on real RF channel data from the PICMUS experimental benchmark²⁷ (4 datasets: resolution phantom, contrast phantom, in-vivo carotid cross-section and longitudinal) plus one DeepUS CIRS-040GSE tissue-mimicking phantom²⁸—all acquired with 128-element linear arrays at 75 plane-wave angles. Speed-of-sound (SoS) mismatch of [10, 25, 50, 100, 200] m/s (relative to nominal $c = 1540$ m/s) is applied to 75-angle compound plane-wave DAS beam-forming with Hilbert-envelope detection. The compound imaging coherently sums all steering angles, so SoS error causes angle-dependent phase shifts that produce measurable destructive interference. Because the true tissue reflectivity is unknown, Scenario I reconstruction (nominal c) serves as pseudo-ground-truth (self-reference). Self-reference PSNR shows clear monotonic degradation: mean Sc. II decreases from 33.1 dB ($\Delta c=10$ m/s) to 24.9 dB ($\Delta c=200$ m/s)—an 8.2 dB gradient across the mismatch range. Grid-search calibration over 51 SoS candidates in [1400, 1700] m/s recovers $c_{\text{cal}} = 1538$ m/s (≤ 2 m/s from nominal) for all 5 datasets, with Sc. IV PSNR of 32.2–42.7 dB (Supplementary Table S14; Supplementary Note 17). This confirms that Gate 3 is the residual bottleneck for acoustic propagation on

430 real experimental data—consistent with the lifecycle prediction for well-designed clinical
431 arrays—completing the five-carrier-family validation.

432 **Autonomous calibration.** Grid-search calibration recovers $\rho = 0.85$ (CASSI, 1,140 s;
433 95% CI [0.79, 0.91]), $\rho = 1.00$ (CACTI, 60 s; CI [0.97, 1.00]), $\rho = 0.86$ –0.92 (SPC gain
434 drift via TV objective; CI width ≤ 0.06), $\rho = 0.95$ –0.97 (SPC illumination, flat-field MLE;
435 Supplementary Table S19), $\rho = 0.98$ –1.00 (Lensless PSF, forward-model fit; Supplementary
436 Table S20), SoS calibration within 2 m/s of nominal (Ultrasound, real PICMUS data),
437 $\rho = 0.85$ –0.99 (Cryo-EM defocus, calibrated within 20–31 nm on all 5 EMDB structures),
438 $\rho = 0.40$ –0.98 (CT detector offset, real images), $\rho = 0.95$ –0.99 (compressive holography),
439 and $\rho = 0.44$ –0.75 (fluorescence PSF) of the oracle correction. Single-parameter modalities
440 (CACTI, cryo-EM, compressive holography, lensless) achieve near-perfect recovery because
441 their mismatch manifolds are low-dimensional with clean objective landscapes. CT CoR
442 calibration on real sinograms also achieves $\rho = 1.00$ (CI [0.99, 1.00]); the multiphantom CT
443 offset recovery ($\rho = 0.40$ –0.98) is lower at large offsets due to irrecoverable sinogram edge
444 truncation (Gate 1 interaction). Multi-parameter modalities (CASSI, fluorescence) show
445 lower but still substantial recovery, reflecting the higher-dimensional search space.

446 **Simulation-to-hardware gap.** Figure 5 summarises residual ratios (panels a–b), the
447 simulation-to-hardware gap (panel c), and autonomous calibration recovery across all three
448 modalities (panel d). CASSI shows smaller real degradation ($1.8\times$) than predicted by
449 simulation, because as-built masks already contain unmodeled imperfections that absorb
450 incremental perturbations. CACTI shows larger real sensitivity ($10.4\times$) than simulation,
451 because the temporal mask pattern replicates errors across all compressed frames. This
452 bidirectional discrepancy—invisible without a unified cross-modality framework—carries a
453 methodological lesson: synthetic mismatch studies can both overestimate marginal impact
454 (CASSI) and underestimate cumulative sensitivity (CACTI).

455 Discussion

456 The central finding is that the space of imaging forward models has finite structural
457 complexity—11 primitives suffice and are necessary—and that every reconstruction fail-
458 ure decomposes into exactly three independent, testable root causes. This closure extends
459 to particle-beam probes (neutron CT, proton CT, muon tomography) with zero additional
460 primitives, as their DAGs decompose entirely into existing library members (Supplementary
461 Note S24). The twelve validated modalities span the breadth of modern imaging science:
462 from coded aperture cameras (CASSI, CACTI) through clinical scanners (CT, CBCT, MRI,
463 ultrasound) to instruments at the frontier of structural biology (cryo-EM, 2017 Nobel Prize
464 in Chemistry) and super-resolution microscopy (fluorescence, 2014 Nobel Prize in Chem-

istry). In every case, the same 11 primitives compose the forward model and the same 3 gates diagnose every failure.

The three gates map naturally to the imaging system lifecycle. At the **design stage**, Gates 1 and 2 are the primary concerns: Gate 1 governs information capacity (sampling geometry, encoding strategy, number of measurements), and Gate 2 governs the carrier budget (photon flux, coherence, penetration depth, detector sensitivity). At the **deployment stage**, Gates 1 and 2 are fixed by the engineering decisions already made; what remains—and what drifts—is Gate 3 (operator mismatch due to calibration error, environmental change, or component aging). In well-designed instruments where Gates 1 and 2 have been optimized, Gate 3 emerges as the residual bottleneck: this is itself a Triad prediction, confirmed by our 12-modality validation with +0.8 to +13.9 dB recovery through forward-model correction. For poorly designed systems (extreme undersampling, photon-starved regimes), the Triad predicts Gate 1 or Gate 2 dominance instead—boundary conditions that define the framework’s scope. Solver design remains essential throughout: once the dominant gate is addressed, a better solver on the corrected operator delivers further gains. The Triad does not replace solver research—it tells the practitioner *where the dB are hiding* so that engineering effort targets the intervention with the largest marginal return.

Implications for autonomous scientific discovery. The bounded complexity of imaging physics has consequences that extend beyond calibration. If every imaging forward model—from a \$10 webcam to a particle accelerator beamline—decomposes into the same 11 primitives, then any autonomous system tasked with “learning to see” faces a *finite search problem* over typed physical operators, not an open-ended function approximation. This reframes the challenge for AI Scientists²⁹: rather than discovering physical laws from scratch, a machine learning system equipped with the OPERATORGRAPH representation needs only to identify which subset of 11 known primitives is active in a new instrument and estimate their parameters from data. The Triad Decomposition further constrains the hypothesis space for failure: an AI Scientist confronting a degraded reconstruction has exactly three root causes to consider, each with a formal test and a known correction strategy. The observation that 9 of 11 primitives are introduced by the first 10 modalities, with the final two (Disperse and Scatter) appearing only for exotic carrier physics, suggests that the primitive manifold is dense at its core and sparse at its boundary—a structure confirmed by the registry’s growth to 170 modalities without requiring a 12th primitive—including particle-beam probes (neutron CT, proton CT, muon tomography) that naive taxonomy places in a distinct carrier family yet decompose fully into existing DAG nodes (Supplementary Note S24). This enables few-shot generalization to new modalities from existing primitive combinations. More broadly, the Finite Primitive Basis Theorem raises a question for computational science beyond imaging: if the forward model of every physical measurement system can be expressed in a small universal grammar, what analogous grammars

503 exist for other physical processes, and what does their existence imply for the long-term
504 trajectory of physics-informed machine learning?

505 **Implications for imaging system design.** The Finite Primitive Basis implies that
506 imaging system design is a combinatorial optimization over 11 typed primitives, not an
507 open-ended search. The OPERATORGRAPH representation makes this explicit: a designer
508 specifies which primitives are active, their parameters, and their interconnection topology.
509 The Triad Decomposition then provides a quantitative sensitivity analysis: for any can-
510 didate design, Gate 3 analysis predicts which parameters are most sensitive to mismatch,
511 enabling tolerance-aware co-design of hardware and reconstruction algorithms. This is di-
512 rectly relevant to multi-scale camera architectures³⁰, where dozens of sub-apertures must be
513 jointly calibrated, and to snapshot compressive imaging systems⁸, where the measurement
514 operator encodes the entire signal recovery problem.

515 **Implications for biology and medicine.** The framework has immediate relevance for
516 the two largest imaging user communities. In structural biology, cryo-EM resolution is
517 rate-limited by contrast transfer function (CTF) estimation—a textbook Gate 3 problem.
518 A 200 nm defocus error at 300 kV limits the achievable resolution from $\sim 2 \text{ \AA}$ to $\sim 3 \text{ \AA}$,
519 and CTFFIND-class estimators require thousands of particles to converge. The PWM
520 Triad diagnoses this as Gate 3 dominance with a 1-parameter correction manifold (Δf),
521 and autonomous defocus calibration recovers $\rho = 1.00\text{--}1.12$ of the oracle (Supplementary
522 Table S15). In clinical imaging, CBCT image quality in radiation therapy is limited by
523 scatter contamination and geometric mismatch—problems that have driven both physics-
524 based correction via Monte Carlo scatter estimation³¹ and data-driven approaches including
525 deep learning³² and reinforcement-learning-based parameter tuning³³. AAPM Task Group
526 142³⁴ mandates monthly mechanical isocentre verification ($\leq 2 \text{ mm}$ tolerance) and AAPM
527 Task Group 66³⁵ specifies CT-simulator geometric accuracy requirements for treatment
528 planning. The Triad Decomposition provides a unifying perspective on these efforts: scat-
529 ter is a Gate 3 forward-model mismatch (the assumed scatter-free model diverges from
530 the scatter-contaminated measurement), geometric drift is a Gate 3 parameter error, and
531 learned CBCT-to-CT mappings implicitly compensate for the aggregate forward-model mis-
532 match without decomposing it. Our validation shows that detector offsets of 5–20 pixels—
533 commensurate with the mechanical tolerances flagged by these guidelines—cause 2.5–3.9 dB
534 PSNR degradation with 100% oracle recovery, suggesting that automated Gate 3 monitoring
535 could complement existing QA protocols. More broadly, ultrasound speed-of-sound inho-
536 mogeneity causes geometric distortion and resolution loss in abdominal imaging, and MRI
537 coil sensitivity drift degrades parallel imaging reconstruction. All three failure modes are in-
538 stances of Gate 3, and all three are correctable through the same forward-model calibration
539 pipeline. The unification of these disparate calibration challenges under a single diagnostic
540 framework suggests a path toward automated quality assurance across the clinical imaging

541 fleet.

542 **Comparison with specialist calibration.** PWM adopts the best available calibration
543 method per modality: for MRI, ESPIRiT¹⁰ is used when sufficient calibration data (ACS
544 ≥ 48 lines) is available, achieving 85–95% of the oracle correction ceiling; for ptychography,
545 ePIE blind position correction¹¹ provides 70–90% recovery; for CT, cross-correlation auto-
546 focus achieves 95–100%; when no specialist method exists (CASSI, CACTI, compressive
547 holography, ultrasound SoS), the modality-agnostic grid-search pipeline provides the first
548 automated calibration (Supplementary Note S14). This principle—use the best domain-
549 specific calibration when available, fall back to physics-model grid-search otherwise—means
550 PWM is not limited to gradient methods but spans the full spectrum from data-driven
551 (ESPIRiT) to model-based (beam-search) calibration. Specialist methods degrade under
552 data-limited conditions (ESPIRiT: -9.29 dB at 24 ACS lines) while PWM maintains pos-
553 itive correction ($+0.75$ dB), demonstrating complementarity: specialist estimates initialize
554 PWM, which refines using forward-model structure across data regimes.

555 **Hardware validation interpretation.** The hardware results reveal that Gate 3 sen-
556 sitivity depends on the *condition number* of the encoding matrix. On real CASSI, per-
557 turbation produces smaller degradation ($1.8\times$) than predicted, because the as-built mask
558 already contains unmodeled manufacturing imperfections. On real CACTI, degradation is
559 severe ($10.4\times$), because the simpler temporal-mask optics lack this pre-existing error buffer.
560 On real CT sinograms, CoR offset causes 8–9 dB loss; electron ptychography shows up to
561 16.1 dB phase degradation; on real PICMUS ultrasound data, compound PW-DAS shows
562 8.2 dB monotonic degradation with SoS calibration within 2 m/s of nominal; on real EMDB
563 cryo-EM structures, defocus error causes up to 3.0 dB degradation with perfect calibration
564 recovery; and on real patient CT images, detector offset produces 1.1–1.7 dB mean loss. This
565 range—from sub-dB (well-conditioned MRI at $R=2$) to > 16 dB (ptychography)—is itself a
566 Triad prediction: the condition number of the encoding matrix determines mismatch sensi-
567 tivity, with well-conditioned systems showing graceful degradation and ill-conditioned sys-
568 tems showing catastrophic failure. The extended cross-residual analysis further reveals that
569 *self-consistency metrics are misleading* in underdetermined systems (CACTI cross-residual
570 is $462\times$ while self-residual is only $1.12\times$), underscoring the need for forward-model-aware
571 diagnostics.

572 **Boundary conditions and failure modes.** The framework has well-defined limitations
573 that inform its scope of applicability:

- 574 • **Multi-parameter mismatch:** CASSI recovery ratios under 5-parameter mismatch
575 are moderate ($\rho = 22\text{--}46\%$), reflecting the inherent difficulty of simultaneously cor-
576 recting multiple coupled parameters on a high-dimensional manifold. Single-parameter

577 correction achieves 85–100%.

578 • **Nonlinear forward models:** The Triad diagnostic has been validated on linear
579 forward models only; beam hardening CT and phase-wrapped MRI require additional
580 validation (the 11-primitive basis formally covers these via Transform Λ^1).

581 • **Undeclared mismatch:** Correction is limited to declared mismatch parameter fami-
582 lies in the mismatch database; undeclared failure modes (e.g., nonlinear detector drift)
583 require extending the database.

584 • **Software-perturbation protocol:** Current real-data experiments inject controlled
585 perturbations into calibrated measurements, isolating the effect of each mismatch
586 parameter. The CT, ptychography, and MRI experiments use genuinely independent
587 public datasets with unknown calibration states, providing complementary evidence.
588 Full hardware-in-the-loop validation with physical parameter displacement is specified
589 in Methods and Supplementary Note 8 and is underway.

590 **Falsifiable predictions.** The Triad framework generates testable predictions across all
591 three gates—including regime transitions that test the decomposition itself, not just indi-
592 vidual gate effects:

593 *Gate 1 dominance* (information deficiency binds):

594 1. **SPC at $< 5\%$ compression:** Gate 1 should dominate—correcting illumination non-
595 uniformity should yield < 1 dB gain because the measurement matrix lacks sufficient
596 rank to recover the signal. *Confirmed:* information-deficiency loss reaches -7.1 dB at
597 5% compression (Supplementary Table S12), matching the $+7.7$ dB Gate 3 correction
598 ceiling and defining the cross-over threshold.

599 2. **Lensless at PSF blur $\sigma > 2$ px:** Gate 1 loss should exceed the Gate 3 correction
600 ceiling ($+3.6$ dB), making forward-model refinement futile regardless of calibration
601 accuracy. *Confirmed:* -11.1 dB at $\sigma=3$ px (Supplementary Table S12).

602 *Gate 2 dominance* (carrier noise binds):

603 3. **MRI at additive noise $\sigma > 1\%$:** Gate 2 should dominate—coil sensitivity correction
604 should be ineffective because Fourier encoding concentrates signal in a few k-space
605 samples that are disproportionately corrupted. *Confirmed:* -11.7 dB loss at $\sigma=0.01$
606 (Supplementary Table S13), matching the $+11.2$ dB Gate 3 ceiling and defining one
607 of the sharpest cross-over thresholds in the framework.

608 4. **Fluorescence at < 10 peak photons:** Shot-noise floor should produce > 4 dB loss,
609 reducing the effective Gate 3 correction margin. *Confirmed:* -6.1 dB at 5 photons
610 (Figure 3b).

611 *Gate 3 dominance* (operator mismatch binds):

- 612 5. **SIM:** Gate 3 should dominate when pattern phase error exceeds 0.05 rad. Predicted
613 correction: +3–8 dB.
- 614 6. **OCT:** Dispersion mismatch should produce +2–5 dB correction gain via the Disperse
615 primitive (W).
- 616 7. **Ultrasound:** SoS mismatch > 100 m/s should produce monotonic B-mode degra-
617 dation. *Confirmed:* 8.2 dB across $\Delta c = 10$ –200 m/s on real PICMUS data, with
618 calibration within 2 m/s of nominal.
- 619 8. **Electron ptychography:** Position errors $> 1/10$ probe diameter should trigger
620 Gate 3 dominance. *Confirmed:* +5 to +16 dB on real SrTiO₃ data (Supplementary
621 Note 15).

622 *Cross- over predictions* (regime transitions):

- 623 9. **SPC G1 $\leftrightarrow$ G3 cross-over at $\sim 5\%$ compression:** Below this threshold, adding
624 measurements yields larger PSNR gains than correcting the forward model; above it,
625 Gate 3 correction dominates. Testable by sweeping compression ratio while simulta-
626 neously applying gain-drift mismatch.
- 627 10. **MRI G2 $\leftrightarrow$ G3 cross-over at $\sigma \approx 1\%$:** The sharpest transition in the framework—
628 MRI switches from G3-dominated ($\sigma < 0.01$, correction gains +11.2 dB) to G2-dominated
629 ($\sigma > 0.01$, noise loss > 11.7 dB) over a single order of magnitude.
- 630 11. **CT: pure Gate 3 regime.** CT should remain Gate 3–dominated even at extreme
631 operating conditions (5 angles, 100 photons), because the many-angle sinogram pro-
632 vides sufficient redundancy that Gates 1–2 never bind. *Confirmed:* G1 loss = -4.3 dB
633 and G2 loss = -3.8 dB at extreme versus G3 gain = $+10.7$ dB (Supplementary Ta-
634 bles S12–S13).
- 635 12. **Ultrasound G2 immunity:** Compound PW-DAS should remain Gate 2–immune
636 at up to 50% additive RF noise due to $\sqrt{N_{\text{angles}}}$ spatial averaging across 75 steering
637 angles. *Confirmed:* 0.0 dB G2 degradation at 50% noise (Figure 3b).

638 These twelve predictions span all three gates and their transitions. They are falsifiable: if a
639 modality shows a different dominant gate under the specified conditions, or if a cross-over
640 threshold differs by more than one order of magnitude, the decomposition would require
641 revision. Eight of twelve are confirmed by simulation or hardware data; the remaining four
642 (SIM, OCT, SPC cross-over, MRI cross-over) are directly testable by any laboratory with
643 the relevant instruments.

644 **Acknowledgements.** We thank the open-source computational imaging community for
645 making reconstruction code and benchmark datasets publicly available. This work was
646 supported by NextGen PlatformAI C Corp.

647 **Author Contributions.** C.Y. conceived the project, designed the TRIAD DECOMPOSI-
648 TION framework, proved the Finite Primitive Basis Theorem, developed the OPERATOR-
649 GRAPH IR, implemented the agent and correction systems, performed all simulation and
650 real-data experiments, and wrote the manuscript. X.Y. developed the GAP-TV reconstruc-
651 tion algorithm used as the primary solver across CASSI, CACTI, and SPC experiments,
652 contributed the EfficientSCI architecture used for CACTI validation, provided the CASSI
653 and CACTI forward model specifications and mismatch parameter characterizations that
654 define the 5-parameter mismatch model, validated the real-data experimental protocols for
655 both CASSI (TSA scenes) and CACTI instruments, and edited the manuscript.

656 **Competing Interests.** C.Y. is an employee of NextGen PlatformAI C Corp, which de-
657 velops the PWM platform. The authors declare no other competing interests.

658 **Ethics Declarations.** This study does not involve human participants, human tissue, or
659 animals. All experiments use publicly available benchmark datasets and simulated data.

660 **Data Availability.** All synthetic measurement data can be regenerated using the OPER-
661 ATORGRAPH templates and mismatch parameters in the Supplementary Information. The
662 KAIST hyperspectral dataset³⁶ and TSA real-data scenes used for CASSI experiments are
663 publicly available. CACTI real-data scenes are available from the EfficientSCI repository²¹.
664 CT sinograms are from the FIPS walnut dataset (Zenodo 1254206) and Helsinki Tomogra-
665 phy Challenge 2022 (Zenodo 6984868). The 4D-STEM ptychography dataset (SrTiO₃ [001])
666 is from Zenodo 5113449. Multi-coil MRI k -space data (M4Raw) is from Zenodo 8056074.
667 Ultrasound, cryo-EM, CBCT, compressive holography, and fluorescence microscopy exper-
668 iments use procedurally generated phantoms with parameters and random seeds fully spec-
669 ified in Methods; all scripts are included in the repository.

670 **Code Availability.** The PWM codebase, including all OPERATORGRAPH templates, agent
671 implementations, real-data validation scripts, and evaluation pipelines, is available at https://github.com/integritynoble/Physics_World_Model
672 under the PWM Noncommercial
673 Share-Alike License v1.0 (see LICENSE in the repository).

674 **Correspondence.** Correspondence and requests for materials should be addressed to
675 C.Y. (integrityyang@gmail.com).

Online Methods

OperatorGraph Specification

Formal definition. The OPERATORGRAPH intermediate representation encodes the forward physics of any computational imaging modality as a directed acyclic graph (DAG) $\mathcal{G} = (\mathcal{V}, \mathcal{E})$. Each node $v_i \in \mathcal{V}$ wraps a *primitive operator* and implements two entry points: `forward( $x$ )  $\rightarrow$   $y$` and `adjoint( $y$ )  $\rightarrow$   $x$` , the latter defined only when the primitive is linear. Edges $e_{ij} \in \mathcal{E}$ encode data flow: the output of node v_i is passed to node v_j . Each node additionally exposes a set of learnable parameters θ_i that may be perturbed during mismatch simulation or optimized during calibration, as well as read-only metadata flags (`is_linear`, `is_stochastic`, `is_differentiable`). The graph is stored as a declarative YAML specification (`OperatorGraphSpec`) and compiled to an executable `GraphOperator` object by the `GraphCompiler`.

Node types. Primitive operators fall into two categories:

- **Linear operators.** Convolution (`conv2d`), mask modulation (`mask_modulate`), subpixel shift (`subpixel_shift_2d`), Radon transform (`radon_fanbeam`), Fourier encoding (`fourier_encode`), spectral dispersion (`spectral_disperse`), Fresnel propagation (`fresnel_propagate`), random projection (`random_project`), and structured illumination (`sim_modulate`). Each implements both `forward()` and `adjoint()`.
- **Nonlinear operators.** Squared magnitude (`magnitude_sq`), Poisson–Gaussian noise (`poisson_gaussian`), saturation clipping (`saturation_clip`), phase retrieval nonlinearity (`phase_abs`), and detector quantization (`quantize`). These set `is_linear` = `False` and raise `NotImplementedError` on `adjoint()`, except where a well-defined pseudo-adjoint exists (*e.g.*, the identity adjoint for magnitude-squared in Gerchberg–Saxton-type algorithms).

Adjoint consistency and Lipschitz bounds as mathematical safety rails. Two structural constraints distinguish the OPERATORGRAPH IR from standard deep-learning forward models and prevent either nonlinear primitive family from becoming a universal approximator.

First, *adjoint consistency*: correctness of every linear primitive is verified by a randomized dot-product test. For a primitive A with forward map $A : \mathbb{R}^n \rightarrow \mathbb{R}^m$, we draw $x \sim \mathcal{N}(0, I_n)$ and $y \sim \mathcal{N}(0, I_m)$ and compute

$$\delta = \frac{|\langle A^* y, x \rangle - \langle y, Ax \rangle|}{\max(|\langle A^* y, x \rangle|, \epsilon)} \quad (1)$$

where $\epsilon = 10^{-12}$ guards against division by zero. The test is repeated $n_{\text{trials}} = 5$ times with independent random draws; the primitive passes if $\delta_{\text{max}} < 10^{-6}$. At the graph level,

a compiled `GraphOperator` composed entirely of linear nodes executes the same test over the composed forward–adjoint chain. A `GraphAdjointCheckReport` records n_{trials} , δ_{max} , and $\bar{\delta}$ for audit. All graph templates that consist solely of linear primitives pass this check. Adjoint consistency is not merely a numerical correctness guarantee: it is a *physical faithfulness certificate*. A model whose adjoint is inconsistent implements a physically impossible energy accounting, making gradient-based calibration unreliable and certifiable reconstruction bounds vacuous. Neural network forward models lack this certificate by construction, since their weight matrices have no physical meaning and their transposes are not computed or validated.

Second, *Lipschitz bounds*: the two nonlinear primitive families, Detect D and Transform Λ , are each restricted to five canonical function families with at most two scalar parameters. This restriction enforces bounded Lipschitz constants ($\text{Lip}(\Lambda_k) \leq L$ for the explicit L derived in¹) and prevents either primitive from becoming a universal approximator in the sense of Cybenko³⁷. Concretely: a generic neural network layer with sigmoid activations can approximate any continuous function on a compact domain; neither Detect nor Transform can, because their function families are closed under composition only within the five enumerated types. This non-universality is the key property that makes the 11-primitive library *falsifiable*: if a modality’s forward physics cannot be represented within the restricted families, the extension protocol (Main Text, “Extension protocol”) fires and a new primitive must be justified physically. The combination of adjoint consistency and bounded Lipschitz nonlinearity provides the mathematical safety rails that make OPERATORGRAPH-based reconstruction both interpretable and certifiable—properties that standard physics-informed neural networks³⁸ achieve only approximately and without modality-agnostic guarantees.

Graph compilation. The compiler executes a four-stage pipeline:

1. **Validate.** Confirm acyclicity via topological sort (Kahn’s algorithm), verify that every `primitive_id` exists in the global `PRIMITIVE_REGISTRY`, reject duplicate `node_id` values, and optionally verify shape compatibility along edges when a `canonical_chain` metadata flag is set.
2. **Bind.** Instantiate each primitive with its parameter dictionary θ_i .
3. **Plan forward.** The topological sort yields a sequential execution plan $(v_{\pi(1)}, \dots, v_{\pi(|V|)})$.
4. **Plan adjoint.** For graphs where `all_linear = True`, the adjoint plan reverses the topological order and applies each node’s individual adjoint in sequence, implementing the chain rule $A^* = A_1^* \circ \dots \circ A_{|V|}^*$ for a composition $A = A_{|V|} \circ \dots \circ A_1$. For graphs containing nonlinear nodes, the adjoint plan is not generated, and any call to `adjoint()` raises `NotImplementedError` at runtime.

The compiled `GraphOperator` is serializable to JSON and hashable via SHA-256 for provenance tracking in RunBundle manifests.

746 **Template library.** The `graph_templates.yaml` registry contains 170 templates spanning
 747 all five carrier families. Of these, 12 modalities carry full end-to-end Scenario I–IV correc-
 748 tion validation across all five carrier families (Supplementary Table S3). Representative
 749 modalities by carrier family include:

- 750 • **Optical and X-ray photons:** CASSI, SPC, CACTI, structured illumination mi-
 751 croscopy (SIM), confocal, light-sheet, holography, ptychography, Fourier ptychographic
 752 microscopy (FPM), optical coherence tomography (OCT), lensless imaging, light field,
 753 integral imaging, neural radiance fields (NeRF), Gaussian splatting, fluorescence life-
 754 time imaging (FLIM), diffuse optical tomography (DOT), phase retrieval, X-ray com-
 755 puted tomography (CT), and cone-beam CT (CBCT).
- 756 • **Electrons:** Electron diffraction, electron backscatter diffraction (EBSD), electron
 757 energy loss spectroscopy (EELS), and electron holography.
- 758 • **Nuclear spins (MRI):** Functional MRI (fMRI), diffusion-weighted MRI (DW-MRI),
 759 and magnetic resonance spectroscopy (MRS).
- 760 • **Acoustic waves:** Ultrasound B-mode, Doppler ultrasound, shear-wave elastogra-
 761 phy, sonar, and photoacoustic tomography (combines optical excitation with acoustic
 762 detection).
- 763 • **Particle-beam probes:** Neutron tomography, proton CT, and muon tomography.
 764 Each decomposes into existing primitives with no extension required: neutron CT
 765 follows the X-ray CT DAG ($P \rightarrow \Lambda \rightarrow D$, Beer–Lambert attenuation with nuclear cross-
 766 section μ_n); proton CT inserts a Bethe–Bloch stopping-power stage as Transform Λ
 767 ($\#11$); muon tomography uses Scatter R ($\#10$) for multiple Coulomb scattering and
 768 POCA vertex reconstruction (Supplementary Note S24).

769 **Fidelity levels.** Each template is parameterized by a fidelity level that controls the degree
 770 of physical realism in the simulated forward model:

771 **Level 1 (Linear, shift-invariant):** The forward model is a linear, spatially uniform operator—
 772 the simplest approximation, suitable for initial diagnostics and rapid prototyping.

773 **Level 2 (Linear, shift-variant):** Spatially varying operator parameters (e.g. non-uniform
 774 illumination, position-dependent PSF, multi-coil sensitivity maps in MRI). Adds a
 775 modality-appropriate noise model (Poisson shot noise plus Gaussian read noise for
 776 photon-counting modalities, Rician noise for MRI, Poisson for CT).

777 **Level 3 (Nonlinear, ray/wave-based):** Includes nonlinear effects such as beam hard-
 778 ening, phase wrapping, and scattering, captured by the Transform Λ primitive. Per-
 779 turbation families and ranges are specified in `mismatch_db.yaml`.

780 **Level 4 (Full-wave / Monte Carlo):** Complete physical simulation including wave-optical
 781 propagation, spatially varying aberrations, detector nonlinearities, and environmental
 782 drift. Currently implemented for holography and ptychography.

783 All 11 primitives in the library $\mathcal{B}$ apply uniformly across every fidelity level; the levels
 784 organize simulation complexity, not theorem scope.

785 Triad Decomposition Formalization

786 The TRIAD DECOMPOSITION asserts that the quality of any computational imaging re-
 787 construction is bounded by three fundamental gates. Rather than a qualitative guideline,
 788 PWM quantifies each gate numerically and uses the resulting scores to diagnose the domi-
 789 nant bottleneck in any imaging configuration.

790 **Gate 1 (Recoverability).** Recoverability measures the information-theoretic capacity
 791 of the sensing geometry. We quantify it via the *effective compression ratio* $r = m/n$, where
 792 m is the number of independent measurements and n the dimension of the signal. The
 793 `compression_db.yaml` registry (1,186 lines) stores, for each modality, a lookup table map-
 794 ping compression ratio to expected reconstruction PSNR under ideal conditions, obtained
 795 from calibration experiments or published benchmarks. Each entry carries a **provenance**
 796 field citing the source (paper DOI, internal experiment ID, or theoretical formula). Addi-
 797 tional recoverability indicators include the effective rank of the measurement matrix (esti-
 798 mated via randomized SVD for large operators), the dimension of the null space, and the
 799 restricted isometry property (RIP) constant where analytically tractable (*e.g.*, for Gaussian
 800 random projections in SPC).

801 **Gate 2 (Carrier Budget).** The carrier budget quantifies the signal-to-noise ratio (SNR)
 802 of the measurement channel. The photon-budget module consumes the `photon_db.yaml`
 803 registry which stores, per modality, a deterministic photon model parameterized by source
 804 power, quantum efficiency, exposure time, and detector characteristics. It classifies the noise
 805 regime into one of three categories: *shot-limited* (Poisson-dominated, $\text{SNR} \propto \sqrt{N_{\text{photon}}}$),
 806 *read-limited* (Gaussian read noise dominates, $\text{SNR} \propto N_{\text{photon}}/\sigma_{\text{read}}$), and *dark-current-*
 807 *limited* (long exposures where dark current accumulation dominates). The output is a
 808 **PhotonReport** containing the estimated SNR in decibels, the noise regime classification, per-
 809 element photon count, and a feasibility verdict (**sufficient**, **marginal**, or **insufficient**).

810 **Gate 3 (Operator Mismatch).** Operator mismatch quantifies the discrepancy between
 811 the assumed forward model H_{nom} and the true physical operator H_{true} . The mismatch
 812 scoring module consults `mismatch_db.yaml` which catalogs, for each modality, the set of
 813 mismatch parameters (spatial shifts, rotational offsets, dispersion errors, PSF deviations,
 814 coil sensitivity errors, center-of-rotation offsets, *etc.*), their typical ranges, and available

correction methods. The mismatch severity score $s \in [0, 1]$ is computed as the normalized ℓ_2 distance $\|\boldsymbol{\theta}_{\text{true}} - \boldsymbol{\theta}_{\text{nom}}\| / \|\boldsymbol{\theta}_{\text{range}}\|$, where $\boldsymbol{\theta}_{\text{range}}$ is the per-parameter dynamic range from the registry. Sensitivity analysis $\partial\text{PSNR}/\partial\theta_k$ is estimated via finite differences on the forward model. The output is a **MismatchReport** containing the severity score, the dominant mismatch parameter, the recommended correction method, and the expected PSNR gain from correction.

Gate binding determination. Given reconstruction results under the four-scenario protocol (the Evaluation Protocol section below), PWM identifies the dominant gate by comparing three cost terms:

$$C_{\text{mismatch}} = \text{PSNR}_{\text{I}} - \text{PSNR}_{\text{II}} \quad (2)$$

$$C_{\text{noise}} = \text{PSNR}_{\text{ideal}} - \text{PSNR}_{\text{noisy}} \quad (3)$$

$$C_{\text{recover}} = \text{PSNR}_{\text{limit}} - \text{PSNR}_{\text{I}} \quad (4)$$

where PSNR_{I} is the reconstruction PSNR under Scenario I (ideal operator), PSNR_{II} under Scenario II (mismatched operator), $\text{PSNR}_{\text{noisy}}$ under the corresponding noisy condition, and $\text{PSNR}_{\text{limit}}$ is the theoretical upper bound from the compression table. The dominant gate is $\arg \max_g C_g$.

TriadReport. For every diagnosis, PWM produces a **TRIADREPORT**: a Pydantic-validated structured artifact comprising **dominant_gate** (enum: **recoverability**, **carrier_budget**, **operator_mismatch**), **evidence_scores** (three floats, one per gate), **confidence_interval** (float, 95% CI width from bootstrap), **recommended_action** (string, *e.g.* “increase compression ratio” or “apply mismatch correction”), and **parameter_sensitivities** (dictionary mapping each mismatch parameter name to its $\partial\text{PSNR}/\partial\theta_k$ value). The **TRIADREPORT** is mandatory—PWM does not permit a reconstruction to be reported without an accompanying diagnosis.

Recovery ratio. We define the *recovery ratio*

$$\rho = \frac{\text{PSNR}_{\text{IV}} - \text{PSNR}_{\text{II}}}{\text{PSNR}_{\text{I}} - \text{PSNR}_{\text{II}}} \quad (5)$$

which lies in $[0, 1]$ under standard convexity conditions (see Supplementary Note 1 for formal analysis; values $\rho > 1$ are possible when the corrected operator provides beneficial regularization). $\rho = 0$ indicates that calibration yields no benefit (mismatch is not the bottleneck), while $\rho = 1$ indicates that calibration fully closes the mismatch gap.

841 Agent System Architecture

842 The three gates are implemented as a deterministic scoring pipeline: an orchestrator maps
843 the user request to a modality from the registry, then dispatches three parallel scoring
844 modules—one per gate—whose reports are fused by a bottleneck classifier that identifies
845 the dominant gate and by a negotiator that enforces cross-gate veto rules (e.g., rejecting
846 reconstruction when photon starvation coincides with aggressive compression). All quanti-
847 tative decisions are deterministic; an LLM is used only as an optional fallback for natural-
848 language modality mapping and is never on the scoring path. The full agent architecture
849 and ablation studies are described in a companion paper.

850 **Contract system.** Inter-agent communication uses 25 Pydantic v2 contract models. All
851 contracts inherit from `StrictBaseModel`, which enforces `extra="forbid"` (no unexpected
852 fields), `validate_assignment=True` (mutations re-validated), and a model validator that
853 rejects NaN and Inf in any float field. Bounded scores use `Field(ge=0.0, le=1.0)`. Enums
854 are string enums for human-readable JSON serialization. This design ensures that pipeline
855 failures surface immediately as validation errors rather than propagating silently.

856 **YAML registries.** The system is driven by 9 YAML registries totalling 7,034 lines:
857 `modalities.yaml` (modality definitions), `graph_templates.yaml` (OperatorGraph skele-
858 tons), `photon_db.yaml` (photon models), `mismatch_db.yaml` (mismatch parameters and
859 correction methods), `compression_db.yaml` (recoverability tables with provenance), `solver_registry.yaml`
860 (solver configurations), `primitives.yaml` (primitive operator metadata), `dataset_registry.yaml`
861 (dataset locations and formats), and `acceptance_thresholds.yaml` (pass/fail thresholds
862 per metric).

863 Correction Algorithms

864 We implement two complementary algorithms for operator mismatch correction. Crucially,
865 both algorithms operate on the forward operator parameters θ rather than the reconstruc-
866 tion solver weights, making them *solver-agnostic*: the corrected operator $H(\hat{\theta})$ benefits any
867 downstream solver (GAP-TV, MST-L, HDNet³⁹, CST, *etc.*) without retraining.

868 **Algorithm 1: Hierarchical Beam Search.** The coarse correction phase employs a
869 hierarchical search strategy to rapidly explore the mismatch parameter space. For CASSI,
870 the five-parameter mismatch model comprises mask affine parameters (spatial shifts dx , dy
871 and rotation θ) and dispersion parameters (slope a_1 and axis angle α); an optional sixth
872 parameter, PSF width σ_{psf} , is available but not used in the primary experiments. The
873 algorithm proceeds as follows:

- 874 1. **1D sweeps.** Each parameter is swept independently over its full range while holding

875 others at nominal values. This produces five 1D cost curves from which coarse optima
876 are extracted.

877 **2. 3D beam search.** The mask affine subspace (dx, dy, θ) is searched over a $5 \times 5 \times 5$
878 grid centered on the 1D optima. The top- k ($k = 5$) candidates by reconstruction
879 PSNR are retained.

880 **3. 2D beam search.** For each retained mask candidate, the dispersion subspace (a_1, α)
881 is searched over a 5×7 grid. The joint top- k candidates are retained.

882 **4. Coordinate descent refinement.** Three rounds of univariate refinement on each
883 parameter, shrinking the search interval by factor 2 at each round, produce the final
884 estimate $\hat{\theta}_{\text{Alg1}}$.

885 Total runtime is approximately 300 seconds per scene on a single GPU. Accuracy is
886 ± 0.1 – 0.2 pixels for spatial parameters and $\pm 0.05^\circ$ for angular parameters.

887 **Algorithm 2: Joint Gradient Refinement.** The fine correction phase uses a differen-
888 tiable forward model to jointly optimize all mismatch parameters via gradient descent. The
889 key components are:

890 **1. Differentiable mask warp.** The binary mask is warped by a continuous affine
891 transformation using bilinear interpolation, implemented as a custom PyTorch module
892 (`DifferentiableMaskWarpFixed`). The mask values are passed through a straight-
893 through estimator (STE) to maintain binary structure while permitting gradient flow.

894 **2. Differentiable forward model.** The CASSI forward model $y = \text{CASSI}(x; \theta)$ is
895 implemented as a differentiable PyTorch module (`DifferentiableCassiForwardSTE`)
896 that accepts mismatch parameters as differentiable inputs.

897 **3. GPU grid initialization.** A full-range 3D grid search over (dx, dy, θ) with $9 \times 9 \times 7 =$
898 567 points provides diverse starting candidates. The top 9 candidates seed multi-start
899 gradient refinement.

900 **4. Staged gradient refinement.** Each of the 9 candidates is refined using Adam
901 optimization (learning rate 10^{-2} , decaying to 10^{-3}) for 200 steps. For each candidate,
902 4 random restarts with jittered initialization guard against local minima. The loss
903 function is the negative PSNR computed via an unrolled K -iteration differentiable
904 GAP-TV solver (`DifferentiableGAPTV`, $K = 10$ unrolled iterations).

905 Total runtime for Algorithm 2 is approximately 3,200 seconds (200 steps $\times$ 4 restarts $\times$
906 9 candidates with early stopping). Accuracy improves to ± 0.05 – 0.1 pixels, a 3–5 $\times$ improve-
907 ment over Algorithm 1. The two algorithms are used sequentially in practice: Algorithm 1
908 provides a warm start, and Algorithm 2 refines to sub-pixel precision.

Evaluation Protocol

Four-Scenario Protocol. We evaluate every modality under four standardized scenarios that isolate different sources of quality degradation:

Scenario I (Ideal): $\mathbf{y}_{\text{obs}} = H_{\text{true}} \mathbf{x}_{\text{gt}}$; reconstruct with H_{true} . In this scenario the system is perfectly calibrated ($H_{\text{true}} = H_{\text{nom}}$), so the operator used for reconstruction matches the one that generated the data. This yields the oracle upper bound on reconstruction quality, limited only by the sensing geometry and solver convergence.

Scenario II (Mismatch): $\mathbf{y}_{\text{obs}} = H_{\text{true}} \mathbf{x}_{\text{gt}}$; reconstruct with H_{nom} ($H_{\text{nom}} \neq H_{\text{true}}$). This is the standard operating condition in practice: the measurement is generated by the true physics, but the reconstruction uses a nominal (potentially mismatched) forward model.

Scenario III (Oracle): Same measurements as Scenario II ($\mathbf{y}_{\text{obs}} = H_{\text{true}} \mathbf{x}_{\text{gt}}$ with $H_{\text{true}} \neq H_{\text{nom}}$); reconstruct with H_{true} instead of H_{nom} . Provides the correction ceiling: the best reconstruction achievable when the true operator is known exactly, applied to data that were sensed by the mismatched system. The gap between Scenario III and Scenario I reveals the irreducible loss from the degraded sensing configuration itself (e.g., a shifted mask pattern is suboptimal even when perfectly known).

Scenario IV (Corrected): $\mathbf{y}_{\text{obs}} = H_{\text{true}} \mathbf{x}_{\text{gt}}$; reconstruct with $\hat{H} = H(\hat{\theta})$ where $\hat{\theta}$ is estimated by Algorithms 1 and 2. This quantifies the benefit of mismatch calibration.

Metrics. Reconstruction quality is assessed using three complementary metrics:

- **PSNR** (peak signal-to-noise ratio, in dB): the primary metric, computed per scene and averaged. For signals normalized to $[0, 1]$, $\text{PSNR} = 10 \log_{10}(1/\text{MSE})$. For SPC data normalized to $[0, 255]$, the peak value is 255.
- **SSIM** (structural similarity index): captures perceptual quality including luminance, contrast, and structural components, computed with a Gaussian window of width 11 and standard deviation 1.5.
- **SAM** (spectral angle mapper): for hyperspectral modalities (CASSI), measures the angle between predicted and true spectral vectors at each spatial location, reported in degrees. Lower is better.

Datasets.

- **CASSI:** 10 scenes from the KAIST dataset³⁶, each a $256 \times 256 \times 28$ spectral cube (28 spectral bands from 450 nm to 650 nm). Data range $[0, 1]$.

941 • **CACTI**: 6 benchmark videos, each $256 \times 256 \times 8$ (8 temporal frames encoded per
942 snapshot). Data range $[0, 1]$.

943 • **SPC**: 11 natural images from the Set11 benchmark, each 256×256 grayscale. Data
944 range $[0, 255]$.

945 All per-scene metrics are reported individually as well as averaged, and all reconstruction
946 arrays are saved as NumPy NPZ files.

947 Experimental Details

948 **Hardware.** All experiments are conducted on a single NVIDIA GPU. Algorithm 1 (beam
949 search) and all solver-based reconstructions use the GPU for matrix–vector products and
950 FFT operations. Algorithm 2 (gradient refinement) additionally uses PyTorch automatic
951 differentiation on the same GPU.

952 **CASSI configuration.** The coded aperture snapshot spectral imaging (CASSI) system
953 uses a TSA-Net binary mask of dimensions 256×256 , with 28 spectral bands dispersed along
954 the spatial dimension. The five-parameter mismatch model $\theta = (dx, dy, \theta, a_1, \alpha)$ describes:
955 mask spatial shift in x (dx , pixels), mask spatial shift in y (dy , pixels), mask rotation angle
956 (θ , degrees), dispersion slope (a_1 , pixels per band), and dispersion axis angle (α , degrees).
957 An optional sixth parameter, PSF blur width (σ_{psf} , pixels), is available but not used in the
958 primary experiments. These mismatch parameter values were determined through system-
959 atic characterization of realistic CASSI assembly errors (Supplementary Note 8). The true
960 mismatch parameters are $\theta_{\text{true}} = (dx = 0.5 \text{ px}, dy = 0.3 \text{ px}, \theta = 0.1^\circ, a_1 = 2.02, \alpha =$
961 $0.15^\circ)$. Solvers evaluated include TwIST⁴⁰, GAP-TV⁴¹, DGSMP⁴², MST-L⁷, and CST-
962 L⁴³, all of which receive the same operator and differ only in their reconstruction algorithm.
963 The supplementary per-scene analysis additionally includes DeSCI⁴⁴ and HDNet³⁹. *Real-*
964 *data validation.* The TSA real hyperspectral dataset¹² consists of 5 scenes at 660×660
965 spatial resolution with 28 spectral bands and mask-shift step 2. Four solvers are evalu-
966 ated: GAP-TV (200 iterations), HDNet (pre-trained checkpoint, full spatial resolution),
967 MST-S and MST-L (pre-trained checkpoints, centre-cropped to 256×256 due to hardcoded
968 spatial assumptions in the model architecture). The coded aperture mask is perturbed
969 by $dx = 0.5 \text{ px}$, $dy = 0.3 \text{ px}$ to simulate assembly-induced mismatch. No ground truth is
970 available; quality is assessed via the normalised measurement residual $r = \|\mathbf{y} - H\hat{\mathbf{x}}\|^2 / \|\mathbf{y}\|^2$.

971 **CACTI configuration.** The coded aperture compressive temporal imaging system uses
972 binary temporal masks of dimensions 256×256 , encoding 8 video frames into a single
973 snapshot measurement. Mismatch is parameterized as a temporal mask timing offset (sub-
974 frame shift). The default solver is EfficientSCI²¹. *Real-data validation.* The EfficientSCI

975 real temporal dataset²¹ consists of 4 dynamic scenes (duomino, hand, pendulumBall, wa-
 976 terBalloon) at 512×512 with compression ratio 10. The real mask is stored separately
 977 from the measurement data. Two solvers are evaluated: GAP-TV (50 iterations) and PnP-
 978 FFDNet (50 iterations with FFDNet denoiser). Mismatch is induced by shifting the mask
 979 by $dx = 0.5$ px, $dy = 0.3$ px. Quality is assessed via the normalised measurement residual
 980 and total variation of the reconstruction.

981 **SPC configuration.** The single-pixel camera uses random binary measurement patterns
 982 at three compression ratios: 10%, 25%, and 50% ($r = m/n \in \{0.10, 0.25, 0.50\}$). Mismatch
 983 is modeled as an exponential gain drift ($g_i = \exp(-\alpha \cdot i)$) on the measurement matrix. The
 984 default solver is FISTA-TV with total-variation regularization.

985 **MRI configuration.** Cartesian k -space sampling with $4\times$ acceleration (25% of k -space
 986 lines acquired). Mismatch is parameterized as a 5% multiplicative error in the coil sensi-
 987 tivity maps used for parallel imaging reconstruction. The default solver is CG-SENSE¹⁷
 988 with ℓ_1 -wavelet regularization ($\lambda = 10^{-3}$, 30 iterations). *Scenario IV calibration (ESPIRiT-*
 989 *adopted)*. For MRI, Scenario IV adopts ESPIRiT¹⁰ as the calibration method: coil sensi-
 990 tivity maps are estimated directly from the central k -space autocalibration signal (ACS)
 991 region via eigenvalue decomposition of the calibration matrix, then used in the CG-SENSE
 992 reconstruction. When ACS data is sufficient (≥ 48 lines), ESPIRiT achieves 85–95% of
 993 the oracle correction ceiling. For data-limited regimes (< 32 ACS lines) where ESPIRiT
 994 degrades, the fallback is a grid search over per-coil amplitude and phase scaling (beam-
 995 search); this fallback is also the Sc. IV method for all modalities without an established
 996 specialist calibration algorithm (CASSI, CACTI, compressive holography, ultrasound). The
 997 protocol evaluates three conditions—Sc. I (true maps), Sc. II (5% mismatched maps), and
 998 Sc. IV (ESPIRiT-calibrated maps)—on the same CG-SENSE solver to isolate calibration
 999 quality from solver design.

1000 **CT configuration.** Fan-beam geometry with 180 projections over 180° . Mismatch is
 1001 modeled as a center-of-rotation (CoR) offset, which produces characteristic arc artifacts in
 1002 the reconstruction. The default solver is filtered back-projection (FBP)¹⁵ with a Ram-Lak
 1003 filter, supplemented by iterative SART for comparison.

1004 **Controlled hardware experiment protocol.** The software-perturbation protocol above
 1005 applies calibrated mask shifts to existing real measurements. A full hardware-in-the-loop
 1006 validation requires physically displacing the coded aperture mask and re-acquiring data.
 1007 The protocol proceeds as follows: (i) acquire a baseline dataset with the mask at its
 1008 factory-calibrated position; (ii) physically translate the mask by a known displacement
 1009 ($\Delta x \in \{0.25, 0.5, 1.0\}$ px equivalent, verified by micrometer stage) and re-acquire under
 1010 identical illumination; (iii) reconstruct both datasets with the factory mask specification and

1011 compute the PSNR degradation and measurement residual; (iv) apply PWM autonomous
 1012 calibration and measure recovery. This protocol isolates the mismatch effect from all other
 1013 sources of variation (illumination changes, detector drift, scene variation). Additionally, a
 1014 multi-unit variation study comparing 2+ camera units of the same design quantifies the
 1015 inter-unit mismatch baseline—the residual calibration error present in any production sys-
 1016 tem.

1017 **Clinical CT phantom configuration.** For clinical translation, PWM is evaluated on CT
 1018 quality assurance using the ACR CT accreditation phantom (Gammex 464). The phantom
 1019 contains inserts of known attenuation (bone ~ 955 HU, air ~ -1000 HU, acrylic ~ 121 HU,
 1020 polyethylene ~ -96 HU) and geometric targets for measuring spatial resolution, slice thick-
 1021 ness, and low-contrast detectability. Mismatch is parameterized as center-of-rotation offset
 1022 (Δr , mm), beam hardening coefficient drift ($\Delta \mu$, %), and detector gain variation (Δg , %).
 1023 Ten ACR-aligned metrics are computed automatically: CT number accuracy for five mate-
 1024 rials (water, bone, air, acrylic, polyethylene), geometric accuracy (± 2 mm tolerance), slice
 1025 thickness (± 1.5 mm), uniformity (≤ 5 HU), noise standard deviation, and spatial resolution
 1026 (≥ 5 lp/cm).

1027 **Clinical MRI validation configuration.** For MRI clinical validation, PWM processes
 1028 multi-coil k -space data from public datasets (fastMRI⁴⁵). Mismatch is parameterized as
 1029 coil sensitivity map error (5–15% multiplicative deviation from calibrated maps, simulating
 1030 patient-positioning-induced coil coupling changes). The default solver is CG-SENSE with
 1031 ℓ_1 -wavelet regularization at $4\times$ acceleration. Clinical metrics include PSNR, SSIM, and the
 1032 absence of parallel imaging artifacts (GRAPPA/SENSE ghosts).

1033 **Ultrasound configuration.** Pulse-echo B-mode imaging is simulated with a 64-element
 1034 linear array, 5 MHz center frequency, element pitch 0.3 mm, sampling rate 100 MHz with
 1035 1024 time samples, and speed of sound (SoS) $c = 1540$ m/s. The tissue phantom is a
 1036 128×128 reflectivity map with Rayleigh-distributed speckle, point scatterers, and anechoic
 1037 cyst regions. Frequency-dependent attenuation follows $\alpha(f) = 0.5 \text{ dB cm}^{-1} \text{ MHz}^{-1}$. Mis-
 1038 match is parameterized as SoS error $\Delta c \in \{50, 100, 200, 400\}$ m/s, perturbing time-of-flight
 1039 calculations. The default solver is delay-and-sum (DAS) beamforming (operator adjoint).
 1040 Calibration uses grid search over 21 SoS candidates in $[1440, 1640]$ m/s, selecting the value
 1041 that maximizes reconstruction PSNR.

1042 **Cryo-EM configuration.** Cryo-EM imaging is modeled as contrast transfer function
 1043 (CTF) filtering of a 2D projected potential (128×128 , pixel size 0.1 nm). The CTF is param-
 1044 eterized by defocus $\Delta f = -2000$ nm, spherical aberration $C_s = 2.0$ mm, electron wavelength
 1045 $\lambda = 0.00251$ nm (300 kV), with a B-factor envelope ($B = 2 \text{ nm}^2$) and ice-thickness attenua-
 1046 tion (mean free path 350 nm, thickness 50 nm). Mismatch is parameterized as defocus error

1047 $\Delta(\Delta f) \in \{100, 200, 500, 1000\}$ nm. The default solver is Wiener filtering with $\text{SNR} = 50$.
 1048 Calibration uses grid search over 51 defocus values in $[-3000, -500]$ nm, selecting the value
 1049 that maximizes reconstruction PSNR.

1050 **CT detector offset configuration.** Parallel-beam CT is simulated with 360 projection
 1051 angles over $[0, \pi)$ and 128 detector elements. The object is a 128×128 Shepp–Logan
 1052 phantom. Detector offset mismatch is simulated by shifting the sinogram along the detector
 1053 axis by $\Delta s \in \{2, 5, 10, 20\}$ pixels. The default solver is filtered backprojection (FBP) with
 1054 Shepp–Logan filter. Calibration uses grid search over 21 offset candidates in $[-25, 25]$ px,
 1055 selecting the value that maximizes reconstruction PSNR.

1056 **Compressive holography configuration.** Multi-depth inline holography encodes a $(4 \times$
 1057 $64 \times 64)$ 3D object into a single (64×64) hologram via Fresnel propagation at four depth
 1058 planes with spacing $200 \mu\text{m}$, wavelength $\lambda = 532 \text{ nm}$, and pixel pitch $5 \mu\text{m}$. Mismatch is
 1059 parameterized as propagation distance error $\Delta z \in \{10, 50, 100, 200\} \mu\text{m}$ applied uniformly to
 1060 all depth planes. The default solver is FISTA-TV (80 iterations, $\lambda_{\text{TV}} = 0.005$). Calibration
 1061 uses hologram residual minimisation: grid search over distance error candidates with the
 1062 search range clamped to ensure all propagation distances remain positive.

1063 **Fluorescence microscopy configuration.** Widefield fluorescence imaging uses a dual-
 1064 PSF Stokes-shift model with Gaussian excitation PSF ($\sigma_{\text{ex}} = 1.5 \text{ px}$) and emission PSF
 1065 ($\sigma_{\text{em}} = 2.0 \text{ px}$), quantum yield $\eta = 0.7$, and background level $b = 0.02$. The specimen
 1066 is a 64×64 image with fluorescent puncta, diffuse organelles, and filamentary structures.
 1067 Mismatch is parameterized as PSF sigma error $\Delta\sigma \in \{0.1, 0.3, 0.5, 1.0\} \text{ px}$ applied to both
 1068 excitation and emission PSFs. The default solver is Richardson–Lucy (80 iterations). Cal-
 1069 ibration uses a 9×9 grid search over $(\sigma_{\text{ex}}, \sigma_{\text{em}})$ centered at the nominal values, selecting
 1070 the pair that minimizes the measurement residual.

1071 **Real experimental data.** Three of the five multi-phantom modalities are validated on
 1072 real experimental data from established benchmarks: ultrasound uses PICMUS experimen-
 1073 tal phantom recordings and in-vivo carotid scans²⁷ plus one DeepUS CIRS-040GSE acqui-
 1074 sition²⁸; cryo-EM uses five EMDB 3D density maps (EMD-5778, EMD-2984, EMD-6287,
 1075 EMD-11103, EMD-21375) projected at random orientations; CT uses three LoDoPaB-CT
 1076 patient slices²⁴, one FIPS walnut pCT scan²⁵, and one HTC 2022 sample²⁶. Compressive
 1077 holography and fluorescence microscopy retain synthetic multi-phantom benchmarks as no
 1078 public datasets with calibrated ground truth are available for these modalities.

1079 Statistical Analysis

1080 **Per-scene reporting.** All metrics are reported per scene, not merely as dataset averages.
1081 This enables identification of scene-dependent failure modes (*e.g.*, spectrally flat scenes that
1082 are inherently harder for CASSI, or textureless regions that challenge SPC).

1083 **Summary statistics.** For each modality and scenario, we report the mean $\pm$ standard
1084 deviation of PSNR, SSIM, and SAM across all scenes. For CASSI (10 scenes), we addition-
1085 ally report the per-band PSNR to assess spectral uniformity of reconstruction quality.

1086 **Recovery ratio confidence intervals.** The recovery ratio ρ (Equation (5)) is a ratio of
1087 differences and therefore sensitive to noise in the constituent PSNR values. We compute
1088 95% confidence intervals via the bootstrap percentile method with $B = 1,000$ resamples. At
1089 each bootstrap iteration, we resample the scene set with replacement, recompute the mean
1090 PSNR for each scenario, and derive ρ . The 2.5th and 97.5th percentiles of the bootstrap
1091 distribution define the 95% CI.

1092 **Parameter recovery accuracy.** For mismatch correction experiments, we report the
1093 root-mean-square error (RMSE) between the estimated and true mismatch parameters:

$$\text{RMSE}_k = \sqrt{\frac{1}{N_{\text{scene}}} \sum_{i=1}^{N_{\text{scene}}} (\hat{\theta}_{k,i} - \theta_{k,\text{true}})^2} \quad (6)$$

1094 where k indexes the mismatch parameter, i indexes the scene, and N_{scene} is the number of
1095 test scenes. Uncertainty in the RMSE is estimated via bootstrap ($B = 1,000$).

1096 **Ablation significance.** Ablation studies (removal of the photon-budget, recoverability,
1097 or mismatch scoring module, or RunBundle discipline) are evaluated by comparing the
1098 full-pipeline PSNR against each ablated variant. We report the PSNR difference ΔPSNR
1099 per modality and verify that each component contributes ≥ 0.5 dB across all validated
1100 modalities, establishing practical significance.

1101 Code and Data Availability

1102 **Source code.** The complete PWM framework, including all agents, the OperatorGraph
1103 compiler, correction algorithms, YAML registries, and evaluation scripts, is released as
1104 open-source software under the PWM Noncommercial Share-Alike License v1.0 at https://github.com/integritynoble/Physics_World_Model. The codebase is organized into
1105 two Python packages: `pwm_core` (core framework, agents, graph compiler, calibration algo-
1106 rithms) and `pwm_AI_Scientist` (automated experiment generation and analysis).
1107

Reconstruction data. All reconstruction arrays from every experiment—Scenarios I through IV for each modality and solver—are released as NumPy NPZ files. Files are stored using Git LFS and require `allow_pickle=True` for loading. Data ranges are standardized: CASSI and CACTI reconstructions are normalized to $[0, 1]$; SPC reconstructions are in $[0, 255]$.

Experiment manifests. Every experiment is recorded in a RunBundle v0.3.0 manifest containing: the git commit hash at execution time, all random number generator seeds, platform information (Python version, GPU model, CUDA version), SHA-256 hashes of all input data and output artifacts, metric values, and wall-clock timestamps. These manifests enable exact reproduction of every reported result.

Registry data. All 9 YAML registries (7,034 lines total) that drive the agent system—including modality definitions, graph templates, photon models, mismatch databases, compression tables, solver configurations, primitive specifications, dataset paths, and acceptance thresholds—are publicly available in the repository under `packages/pwm_core/contrib/`. The `ExperimentSpec` JSON schemas used for pipeline input validation are included alongside worked examples in `examples/`.

References

- [1] Chengshuai Yang. The finite primitive basis theorem for computational imaging: Formal foundations of the OperatorGraph representation, 2026. URL <https://arxiv.org/abs/2602.20550>.
- [2] Emmanuel J. Candès and Michael B. Wakin. An introduction to compressive sampling. *IEEE Signal Processing Magazine*, 25(2):21–30, 2008. doi: 10.1109/MSP.2007.914731.
- [3] David L. Donoho. Compressed sensing. *IEEE Transactions on Information Theory*, 52(4):1289–1306, 2006. doi: 10.1109/TIT.2006.871582.
- [4] Singanallur V. Venkatakrishnan, Charles A. Bouman, and Brendt Wohlberg. Plug-and-play priors for model based reconstruction. In *Proceedings of the IEEE Global Conference on Signal and Information Processing (GlobalSIP)*, pages 945–948, 2013. doi: 10.1109/GlobalSIP.2013.6737048.
- [5] Xin Yuan, Yang Liu, Jinli Suo, and Qionghai Dai. Plug-and-play algorithms for video snapshot compressive imaging. *IEEE Transactions on Pattern Analysis and Machine Intelligence*, 44(10):7093–7111, 2021.
- [6] Vishal Monga, Yuelong Li, and Yonina C. Eldar. Algorithm unrolling: Interpretable,

- efficient deep learning for signal and image processing. *IEEE Signal Processing Magazine*, 38(2):18–44, 2021. doi: 10.1109/MSP.2020.3016905.
- [7] Yuanhao Cai, Jing Lin, Xiaowan Hu, Haoqian Wang, Xin Yuan, Yulun Zhang, Radu Timofte, and Luc Van Gool. Mask-guided spectral-wise transformer for efficient hyperspectral image reconstruction. In *Proceedings of the IEEE/CVF Conference on Computer Vision and Pattern Recognition (CVPR)*, pages 17502–17511, 2022.
- [8] Xin Yuan, David J. Brady, and Aggelos K. Katsaggelos. Snapshot compressive imaging: Theory, algorithms, and applications. *IEEE Signal Processing Magazine*, 38(2):65–88, 2021.
- [9] Vegard Antun, Francesco Renna, Clarice Poon, Ben Adcock, and Anders C. Hansen. On instabilities of deep learning in image reconstruction and the potential costs of AI. *Proceedings of the National Academy of Sciences*, 117(48):30088–30098, 2020. doi: 10.1073/pnas.1907377117.
- [10] Martin Uecker, Peng Lai, Mark J. Murphy, Patrick Virtue, Michael Elad, John M. Pauly, Shreyas S. Vasanawala, and Michael Lustig. ESPIRiT — an eigenvalue approach to autocalibrating parallel MRI: where SENSE meets GRAPPA. *Magnetic Resonance in Medicine*, 71(3):990–1001, 2014. doi: 10.1002/mrm.24751.
- [11] Andrew M. Maiden and John M. Rodenburg. An improved ptychographical phase retrieval algorithm for diffractive imaging. *Ultramicroscopy*, 109(10):1256–1262, 2009. doi: 10.1016/j.ultramic.2009.05.012.
- [12] Ashwin A. Wagadarikar, Renu John, Rebecca Willett, and David J. Brady. Single disperser design for coded aperture snapshot spectral imaging. *Applied Optics*, 47(10):B44–B51, 2008. doi: 10.1364/AO.47.000B44.
- [13] Patrick Llull, Xuejun Liao, Xin Yuan, Jianbo Yang, David Kittle, Lawrence Carin, Guillermo Sapiro, and David J. Brady. Coded aperture compressive temporal imaging. *Optics Express*, 21(9):10526–10545, 2013. doi: 10.1364/OE.21.010526.
- [14] Marco F. Duarte, Mark A. Davenport, Dharmpal Takhar, Jason N. Laska, Ting Sun, Kevin F. Kelly, and Richard G. Baraniuk. Single-pixel imaging via compressive sampling. *IEEE Signal Processing Magazine*, 25(2):83–91, 2008. doi: 10.1109/MSP.2007.914730.
- [15] L. A. Feldkamp, L. C. Davis, and J. W. Kress. Practical cone-beam algorithm. *Journal of the Optical Society of America A*, 1(6):612–619, 1984. doi: 10.1364/JOSAA.1.000612.
- [16] J. M. Rodenburg and H. M. L. Faulkner. A phase retrieval algorithm for shifting illumination. *Applied Physics Letters*, 85(20):4795–4797, 2004. doi: 10.1063/1.1823034.

- [17] Klaas P. Pruessmann, Markus Weiger, Markus B. Scheidegger, and Peter Boesiger. SENSE: Sensitivity encoding for fast MRI. *Magnetic Resonance in Medicine*, 42(5): 952–962, 1999. doi: 10.1002/(SICI)1522-2594(199911)42:5<952::AID-MRM16>3.0.CO; 2-S.
- [18] Michael Lustig, David Donoho, and John M. Pauly. Sparse MRI: The application of compressed sensing for rapid MR imaging. *Magnetic Resonance in Medicine*, 58(6): 1182–1195, 2007. doi: 10.1002/mrm.21391.
- [19] David J. Brady. *Optical Imaging and Spectroscopy*. John Wiley & Sons, 2009.
- [20] Chengshuai Yang. InverseNet: Benchmarking operator mismatch in snapshot compressive imaging. *arXiv preprint arXiv:2603.04538*, 2026.
- [21] Lishun Wang, Miao Cao, and Xin Yuan. EfficientSCI: Densely connected network with space-time factorization for large-scale video snapshot compressive imaging. In *Proceedings of the IEEE/CVF Conference on Computer Vision and Pattern Recognition (CVPR)*, pages 18477–18486, 2023.
- [22] Nick Antipa, Grace Kuo, Reinhard Heckel, Ben Mildenhall, Emrah Bostan, Ren Ng, and Laura Waller. DiffuserCam: lensless single-exposure 3D imaging. *Optica*, 5(1):1–9, 2018. doi: 10.1364/OPTICA.5.000001. DiffuserCam DLMD dataset: HuggingFace bezzam/DiffuserCam-Lensless-Mirflickr-Dataset.
- [23] David J. Brady, Kerkil Choi, Daniel L. Marks, Ryoichi Horisaki, and Sehoon Lim. Compressive holography. *Optics Express*, 17(15):13040–13049, 2009.
- [24] Johannes Leuschner, Maximilian Schmidt, Daniel Otero Baguer, and Peter Maaß. LoDoPaB-CT, a benchmark dataset for low-dose computed tomography reconstruction. *Scientific Data*, 8:109, 2021. doi: 10.1038/s41597-021-00893-z.
- [25] Keijo Hämäläinen, Lauri Harhanen, Andreas Hauptmann, Aki Kallonen, Esa Niemi, and Samuli Siltanen. Total variation regularization for large-scale X-ray tomography. *International Journal of Tomography and Simulation*, 28(1):1–25, 2015. FIPS walnut dataset: Zenodo 1254206.
- [26] Alexander Meaney, Andreas Hauptmann, and Samuli Siltanen. Helsinki tomography challenge 2022: description, methods, and results. *Applied Mathematics for Modern Challenges*, 1(1):1–28, 2024. doi: 10.3934/ammc.2024001.
- [27] Hervé Liebgott, Alfonso Rodriguez-Molares, Françoise Cervenansky, Jørgen Arendt Jensen, and Olivier Bernard. Plane-wave imaging challenge in medical ultrasound. In *Proc. IEEE International Ultrasonics Symposium (IUS)*, pages 1–4, 2016. doi: 10.1109/ULTSYM.2016.7728908.

- [28] Saif Khan, Jaeyoung Huh, Manish Bhatt, Md. Murad Hossain, and Arun Bhatt. A deep learning framework for ultrasound image formation and image quality assessment. *IEEE Transactions on Ultrasonics, Ferroelectrics, and Frequency Control*, 70(9):972–984, 2023. DeepUS dataset: <https://zenodo.org/doi/10.5281/zenodo.7986407>.
- [29] Hanchen Wang, Tianfan Fu, Yuanqi Du, Wenhao Gao, Kexin Huang, Ziming Liu, Payal Chandak, Shengchao Liu, Peter Van Katwyk, Andreea Deac, et al. Scientific discovery in the age of artificial intelligence. *Nature*, 620:47–60, 2023.
- [30] David J. Brady, Michael E. Gehm, Ronald A. Stack, Daniel L. Marks, David S. Kittle, Dathon R. Golish, E. M. Vera, and Steven D. Feller. Multiscale gigapixel photography. *Nature*, 486(7403):386–389, 2012.
- [31] Yuan Xu, Ti Bai, Hao Yan, Luo Ouyang, Arnold Pompos, Jing Wang, Linghong Zhou, Steve B. Jiang, and Xun Jia. A practical cone-beam CT scatter correction method with optimized Monte Carlo simulations for image-guided radiation therapy. *Physics in Medicine and Biology*, 60(9):3567–3587, 2015. doi: 10.1088/0031-9155/60/9/3567.
- [32] Xiao Liang, Liyuan Chen, Dan Nguyen, Zhiguo Zhou, Xuejun Gu, Ming Yang, Jing Wang, and Steve Jiang. Generating synthesized computed tomography (CT) from cone-beam computed tomography (CBCT) using CycleGAN for adaptive radiation therapy. *Physics in Medicine and Biology*, 64(12):125002, 2019. doi: 10.1088/1361-6560/ab22f9.
- [33] Chenyang Shen, Yesenia Gonzalez, Liyuan Chen, Steve B. Jiang, and Xun Jia. Intelligent parameter tuning in optimization-based iterative CT reconstruction via deep reinforcement learning. *IEEE Transactions on Medical Imaging*, 37(6):1430–1439, 2018. doi: 10.1109/TMI.2018.2823679.
- [34] Eric E. Klein, Joseph Hanley, John Bayouth, Fang-Fang Yin, William Simon, Sean Dresser, Christopher Serago, Francisco Aguirre, Lijun Ma, Bijan Arjomandy, Chihray Liu, Craig Sandin, and Timothy Holmes. Task Group 142 report: Quality assurance of medical accelerators. *Medical Physics*, 36(9):4197–4212, 2009.
- [35] Sasa Mutic, Jatinder R. Palta, Elizabeth K. Butker, Indra J. Das, M. Saiful Huq, Leh-Nien Dick Loo, Bill J. Salter, Cynthia H. McCollough, and Jacob Van Dyk. Quality assurance for computed-tomography simulators and the computed-tomography-simulation process: Report of the AAPM Radiation Therapy Committee Task Group No. 66. *Medical Physics*, 30(10):2762–2792, 2003.
- [36] Inchang Choi, Daniel S. Jeon, Giljoo Nam, Diego Gutierrez, and Min H. Kim. High-quality hyperspectral reconstruction using a spectral prior. *ACM Transactions on Graphics (Proceedings of SIGGRAPH Asia)*, 36(6):218:1–218:13, 2017. doi: 10.1145/3130800.3130810.

- [37] George Cybenko. Approximation by superpositions of a sigmoidal function. *Mathematics of Control, Signals and Systems*, 2(4):303–314, 1989.
- [38] Maziar Raissi, Paris Perdikaris, and George E. Karniadakis. Physics-informed neural networks: A deep learning framework for solving forward and inverse problems involving nonlinear partial differential equations. *Journal of Computational Physics*, 378: 686–707, 2019.
- [39] Xiaowan Hu, Yuanhao Cai, Jing Lin, Haoqian Wang, Xin Yuan, Yulun Zhang, Radu Timofte, and Luc Van Gool. HDNet: High-resolution dual-domain learning for spectral compressive imaging. In *Proceedings of the IEEE/CVF Conference on Computer Vision and Pattern Recognition (CVPR)*, pages 17542–17551, 2022.
- [40] José M. Bioucas-Dias and Mário A. T. Figueiredo. A new TwIST: Two-step iterative shrinkage/thresholding algorithms for image restoration. *IEEE Transactions on Image Processing*, 16(12):2992–3004, 2007. doi: 10.1109/TIP.2007.909319.
- [41] Xin Yuan. Generalized alternating projection based total variation minimization for compressive sensing. In *Proceedings of the IEEE International Conference on Image Processing (ICIP)*, pages 2539–2543, 2016. doi: 10.1109/ICIP.2016.7532817.
- [42] Tao Huang, Weisheng Dong, Xin Yuan, Jinjian Wu, and Guangming Shi. Deep gaussian scale mixture prior for spectral compressive imaging. In *Proceedings of the IEEE/CVF Conference on Computer Vision and Pattern Recognition (CVPR)*, pages 16216–16225, 2021.
- [43] Yuanhao Cai, Jing Lin, Xiaowan Hu, Haoqian Wang, Xin Yuan, Yulun Zhang, Radu Timofte, and Luc Van Gool. CST: Compact spectral transformer for hyperspectral image reconstruction. In *Proceedings of the European Conference on Computer Vision (ECCV)*, 2022.
- [44] Yang Liu, Xin Yuan, Jinli Suo, David J. Brady, and Qionghai Dai. Rank minimization for snapshot compressive imaging. *IEEE Transactions on Pattern Analysis and Machine Intelligence*, 41(12):2990–3006, 2019.
- [45] Jure Zbontar, Florian Knoll, Anuroop Sriram, Tullie Murrell, Zhengnan Huang, Matthew J. Muckley, Aaron Defazio, Ruben Stern, Patricia Johnson, Mary Bruno, Marc Parente, Krzysztof J. Geras, Joe Katsnelson, Hersh Chandarana, Zizhao Zhang, Michal Drozdal, Adriana Romero, Michael Rabbat, Pascal Vincent, Nafissa Yakubova, James Pinkerton, Duo Wang, Erich Owens, C. Lawrence Zitnick, Michael P. Recht, Daniel K. Sodickson, and Yvonne W. Lui. fastMRI: An open dataset and benchmarks for accelerated MRI. *arXiv preprint arXiv:1811.08839*, 2018.




Figure 1: **The universal grammar of computational imaging.** **a**, Any computational imaging system—from coded-aperture cameras to cryo-electron microscopes—serves as input. **b**, The system is composed as a typed directed acyclic graph (DAG) over 11 universal primitives (Figure 2); shown here for CASSI as an example. **c**, The TRIAD DECOMPOSITION diagnoses the DAG through three independent gates: information deficiency (Gate 1), carrier budget (Gate 2), and operator mismatch (Gate 3), each producing a quantitative score (Figure 3). **d**, Based on the dominant gate, a targeted correction is applied: sampling redesign (Gate 1), source/detector improvement (Gate 2), or operator calibration (Gate 3). **e**, The existing solver is re-run with the corrected operator, recovering +0.8 to +13.9 dB reconstruction quality without retraining. Together, the 11 primitives (alphabet) and 3 gates (rules) form a universal grammar for designing, diagnosing, and correcting any imaging system.




Figure 2: 11 primitives: OperatorGraph IR, Physics Fidelity Ladder, and basis completeness. **a**, The 11 universal primitives grouped by physical role: generation (Propagate P), encoding (Modulate M , Project Π , Convolve C , Scatter R , Transform Λ), transform (Encode F , Accumulate Σ), and detection (Sample S , Disperse W , Detect D). Each primitive carries a typed signature mapping its input to its output (e.g., P : free-space propagation; D : detector response). **b**, Example OPERATORGRAPH DAGs for three modalities: CASSI (photon), MRI (spin), and CT (X-ray photon). Each node wraps a primitive operator; edges define data flow. The symbols in panel b correspond directly to the primitive labels in panel a. **c**, Fidelity levels: forward models range from linear shift-invariant (simplest) through nonlinear and full-wave; all 11 primitives apply uniformly across every level. **d**, Basis-growth saturation: the number of distinct primitives K required to express N registered modalities saturates at $K=11$ for $N \geq 35$; no additional primitive has been needed for the 135 subsequent modalities across 23 physical categories.




Figure 3: **Triad Decomposition: all three gates validated with real data.** **a**, Decision tree for the TRIAD DECOMPOSITION: each imaging failure is routed through **Gate 1**, **Gate 2**, and **Gate 3** to produce a TRIADREPORT. **b**, Three-gate degradation and correction heatmap. *G1* (blue): Δ PSNR under extreme compression (Table S12; all 12 modalities). *G2* (amber): Δ PSNR under extreme noise (Table S13; all 12 modalities). *G3* (green): correction gain Δ PSNR under standard mismatch conditions (Table S1; all 12 modalities). *G1/G2* degrade at extreme conditions confirming they are real failure modes; *G3* dominates under standard deployment because Gates 1 and 2 were already optimized at design time. **c**, Recovery ratio ρ distribution across all 12 validated modalities, showing that autonomous calibration recovers a substantial fraction of the oracle correction ceiling across all five carrier families.




Figure 4: **4-Scenario Protocol across 6 representative modalities.** Each panel shows PSNR (dB) for the four scenarios: Sc. I (ideal operator, green), Sc. II (mismatched operator, red), Sc. III (oracle correction, blue), and Sc. IV (autonomous calibration, purple). Six modalities span four of five carrier families: CASSI, CACTI, and SPC (optical photons); CT (X-ray); ptychography (electron); MRI (nuclear spin). For low-dimensional mismatch (CT, ptychography, MRI), Sc. IV $\approx$ Sc. III $\approx$ Sc. I, confirming near-complete recovery. For multi-parameter mismatch (CASSI), recovery reaches 85%; CACTI and SPC achieve 100% and 86%, respectively (Supplementary Table S9). Sc. I PSNR from Table S3; Sc. II/III from Table S1.




Figure 5: **Hardware validation on real CASSI and CACTI instruments.** **a**, CASSI real data: measurement residual ratio (mismatched/calibrated) across 5 TSA scenes. GAP-TV shows $1.8\times$ mean ratio. **b**, CACTI real data: residual ratio across 4 scenes. GAP-TV shows $10.4\times$ mean ratio; PnP-FFDNet shows $2.0\times$. **c**, Simulation-to-hardware gap: comparing mismatch degradation in simulation versus real hardware for CASSI and CACTI. **d**, Autonomous calibration: grid-search parameter recovery for CASSI (85%), CACTI (100%), and SPC (86–92%).

Extended Data Table 1 | Primitive necessity ablation. For each of the 11 primitives, the witness modality whose representation error e_{img} exceeds $\varepsilon = 0.01$ when that primitive is removed from $\mathcal{B}$. Eight primitives ($P, M, \Pi, F, \Sigma, D, R, \Lambda$) are strictly necessary; the remaining three (C, S, W) are necessary under the stated complexity bounds ($N_{\text{max}} = 20, D_{\text{max}} = 10$). Data from ¹, Proposition 1.

1283

Removed primitive	Witness modality	Why irreplaceable	e_{img} without
Propagate P	Ptychography	Distance-dependent phase transfer	> 0.05
Modulate M	CASSI	Arbitrary element-wise mask	> 0.10
Project Π	CT	Line-integral (Radon) geometry	> 0.15
Encode F	MRI	Non-uniform Fourier sampling	> 0.20
1284 Convolve C	Lensless imaging	Arbitrary shift-invariant PSF kernel	> 0.08
Accumulate Σ	SPC	Dimensional summation (not scaling)	> 0.12
Detect D	All modalities	Carrier-to-measurement conversion	> 0.50
Sample S	MRI	Index-set restriction (k -space)	> 0.10
Disperse W	CASSI	λ -parameterized spatial shift	> 0.06
Scatter R	Compton imaging	Direction change + energy shift	0.34
Transform Λ	Polychromatic CT	Non-terminal pointwise nonlinearity	> 0.05

Extended Data Table 2 | Oracle correction ceiling (Scenario III) across 14 validated configurations (12 modalities, 5 carrier families). Full table with per-scene metrics in Supplementary Table S1; autonomous correction recovery (Scenario IV) in Supplementary Table S9. Per-modality supplementary tables: ultrasound (S14), cryo-EM (S15), CBCT (S16), compressive holography (S17), fluorescence microscopy (S18), SPC (S19), lensless imaging (S20).

1290