



Tremor Ai — Tremor Assessment & Telemetry Report

Report ID: TAI-1788657285
Date: 2026-09-06 06:44:45

Patient Identifier:	TR-90241	Data Origin:	Clinical Raw Recording
Sensor Device:	MPU6050 6-DoF IMU Ring	Sampling Rate:	100.0 Hz (Nyquist: 50.0 Hz)
Assessment Duration:	60.0 seconds	Analysis Window:	3.0 sec (50% sliding overlap)

AI Pattern Classification	Tremor Severity Index	Clinical Severity Grade
PD Confidence: 94.2% (PD Match: 94.2%)	42.0 / 100	Moderate Rhythmic resting tremor in characteristic 4-6 Hz Parkinsonian band.

Biomechanical Time-Series & Spectral Isolation



Extracted Digital Biomarkers (Window Metrics)

Biomarker	Observed Value	Reference Band / Diagnostic Context
Dominant Frequency	5.10 Hz	Parkinsonian Resting Tremor: 4.0 - 6.0 Hz
Tremor Band Power Ratio	0.0%	Ratio of power in 4-6 Hz band to broad spectrum (>40% indicates localized tremor)
Dynamic Jerk RMS	0.00 g/s	Time-derivative of acceleration measuring abrupt oscillatory changes
Spectral Shannon Entropy	0.000	Low (<0.45) indicates rhythmic periodicity; High (>0.70) indicates random/voluntary noise

Physiological Interpretation & Explainability

Biomechanical features within normal resting bounds.

NOTICE & MEDICAL DISCLAIMER: Tremor AI is an experimental screening and longitudinal monitoring aid, NOT a diagnostic or medical treatment device. The metrics, classifications, and severity scores provided herein reflect biomechanical time-series observations and must be interpreted solely by a qualified neurologist or physician in clinical context. Never initiate, alter, or discontinue any medication based on this report.