Respiratory Rate Monitor and Analyzer

Project Description:

Principle of Operation:

The Respiratory Rate Monitor and Analyzer is designed to provide a comprehensive assessment of respiratory health by quantifying the respiratory rate and analyzing characteristic breathing sounds. The device operates using two main types of measurements: pressure variation and acoustic signals, which together offer a detailed picture of a person's respiratory function.

The pressure measurement relies on the MPX2010DP differential pressure sensor, which measures the difference between the internal air pressure within the respiratory tract and the external air pressure during the breathing cycle. As the user inhales and exhales, these pressure changes are captured by the sensor. The resulting analog signal, reflecting the dynamic changes in air pressure, is amplified and conditioned through filtering before being converted into a digital signal by the Arduino Nano Every's analog-to-digital converter (ADC). This signal represents the cyclic pattern of breathing and is used to calculate the respiratory rate in real time.

In addition to pressure data, the device incorporates an acoustic measurement system using a sensitive microphone (CMC-4012-40P). This microphone captures the sounds produced by airflow through the respiratory tract, including normal breathing as well as abnormal sounds like wheezes or crackles. These acoustic signals are first amplified by an LM386 audio amplifier, then passed through a band-pass filter that isolates the specific frequencies associated with breathing. Then, a low pass filter will be incorporated as unwanted background noise and irrelevant frequencies need to be filtered out to ensure the cleanest possible signal for analysis. Once digitized, these acoustic signals are processed by the microcontroller to detect patterns or anomalies in the breathing sounds.

The Arduino Nano Every acts as the central processing unit for the device. It receives the digitized signals from both the pressure sensor and microphone, running algorithms to calculate the respiratory rate and analyze the breathing sounds for any irregularities. The results are displayed in real time on an OLED display, offering immediate feedback to the user or healthcare provider. If the system detects abnormalities—such as an irregular respiratory rate or the presence of abnormal sounds like wheezing or crackling—an alert is triggered to indicate a potential respiratory issue.

Physiological Basis:

The operation of the Respiratory Rate Monitor and Analyzer is grounded in the anatomy and physiology of the human respiratory system, focusing on the dynamics of airflow and the sounds generated during breathing. The airways, including the nasal passages, trachea, bronchi, and lungs, play a crucial role in this process. As air is inhaled, it passes through these structures, causing changes in pressure and generating sounds due to turbulence or air movement through narrow or fluid-filled passages.

During normal breathing, the pressure changes are rhythmic and regular, corresponding to the expansion and contraction of the lungs. The microphone captures sounds that correspond to the movement of air through these passages, with the device quantifying the respiratory rate by measuring the pressure and sound cycles. This baseline information helps establish what is considered normal for the patient, making deviations easier to detect.


Abnormal breathing sounds often signal underlying health issues. For example, wheezing is typically caused by the narrowing of airways, often due to conditions like asthma, producing a high-pitched whistling sound. Crackles, on the other hand, may indicate fluid in the lungs, as seen in conditions such as pneumonia or heart failure. These crackling sounds occur when small airways or alveoli pop open during inhalation, signaling compromised lung function.

By continuously monitoring both the rate and sound of breathing, the device provides valuable real-time feedback on respiratory function, making it a useful tool in clinical and home settings for managing conditions like asthma, chronic obstructive pulmonary disease (COPD), or respiratory infections. Its ability to detect irregular breathing patterns or abnormal sounds means that early intervention is possible when respiratory distress occurs, helping to prevent complications and improve patient outcomes.


Overall, the Respiratory Rate Monitor and Analyzer combines pressure and acoustic measurements to create a detailed profile of a patient's respiratory health. This integration of physiological signals into a real-time monitoring system provides healthcare providers with the information necessary to diagnose and manage respiratory conditions more effectively.

Block Diagram:

DC Power Supply.



is


Signal Processing 1a:
Audio Amplifier:
Amplify the breathing sound signals
Non-Inverting Op-Amp
Set gain to approximately 50 to sufficiently amplify breathing sounds without distorting the signal.
Signal Processing 2a:
Microphone Output Filtering:
Type: Band-Pass Filter
Order: Second-order filter.
Cutoff Frequencies: Low cutoff at 100 Hz and high cutoff at 2 kHz. Typical range for breathing sound frequencies.