

Table 1. MRI sequence protocol

Parameter	Sequence						
	Whole-spine sagittal T1-weighted	Whole-spine sagittal STIR	Whole-body axial T2-weighted	Whole body coronal T1-weighted	Whole-body coronal DWIBS	Lumbar spinal DWI	Whole-body coronal mDixon Quant
Sequence type	FSE	STIR	FSE	FSE	STIR	EPI	3D SPGR
Time (ms)							
TR	404	5693	1000	515	5411	8000	5.7
TE	10	70	70	15	70	84	Six evenly spaced echoes
Inversion time	N/A	200	N/A	N/A	250	N/A	N/A
Slice thickness (mm)	4	4	6	5	4	4	6
Number of slices per station	15	15	40	34	50	11	64
In-plane pixel size (mm)	1.4×2.7	1.5×2.2	1.1×1.5	1.7×3.5	2.3×3.5	2.7×2.7	2.4×2.4
Bandwidth/pixel (Hz)	576	625	359	435	2535	2024	128
Acquisition time (min)	6	6	6	6	8	6.5	1.3
b Value (s/mm ²)	N/A	N/A	N/A	N/A	0 and 1000	0, 40, 80, 140, 200, 500, 1000, 1500, and 2000	N/A