
Sensor Technology Limited




P O Box 97, 20 Stewart Road Tel: 705.444.1440
Collingwood, Ontario, Canada L9Y 3Z4 Fax: 705.444.6787

E-Mail:**Techsupport@sensortech.ca**World Wide Web: **www.sensortech.ca**

SQ26-08





SQ26-08

• Operating Ratings

- 1. Supply Voltage 2V to 5V
- 2. Temperature Range -40/C to 85/C



Sensor Technology Limited

P O Box 97, 20 Stewart Road Tel: 705.444.1440 Collingwood, Ontario, Canada L9Y 3Z4 Fax: 705.444.6787

E-Mail:**Techsupport@sensortech.ca**World Wide Web: **www.sensortech.ca**

• Maximum Rating of Supply Voltage: 5.5 V

• Voltage Sensitivity: -168.5 dB re V/μPa or 3.758 mV/Pa

Pressure amplitude (Pa) at the face of the hydrophone = $\frac{V_{rms} (mV)}{3.758}$

• SQ26-08 produces negative response at ring of TRS plug to compressive pressure.

20 years