Date: Version: Status: Page: 1.0 Final 1 of 5

18 July 2018 | Novo Nordisk

#### **Primary Packaging System Drawing**

# Primary Packaging System number: EMK0595

Primary Packaging System name: Syringe 1 ml Long with Staked Needle and RNS

Updated with 0.75 ml PFS for Sema Obesity project

#### Author

Jan Gunnar Jørgensen (JGJ), Development Engineer Dept. 199, Primary Packaging Development

Dept. 199, Prim Pack Development
DFP ManDev
EMK0595-12.01-007
NDID 004276473

INTERNAL USE ONLY

 Date:
 18 July 2018
 Novo Nordisk

 Version:
 1.0

 Status:
 Final


 Page:
 2 of 5

# **Approvals**

This document is electronically approved in NovoDOCS by the following persons or their qualified substitutes:

Name	Organizational unit	Responsibility
JGJ (Jan Gunnar Jørgensen)	199 Primary Packaging Development	Author
MAES (Mads Reedtz Espersen)	199 Primary Packaging Development	Expert reviewer
FIT (Frida Thuresson)	6032, DFP QA ManDev	QA approval

18 July 2018 | Novo Nordisk Date: Version: 1.0 Final Status: Page: 3 of 5



Dept. 199, Prim Pack Development DFP ManDev EMK0595-12.01-007 NDID 004276473

INTERNAL USE ONLY

 Date:
 18 July 2018
 Novo Nordisk

 Version:
 1.0

 Status:
 Final

 Page:
 4 of 5

## Notes regarding plunger position and head space

The specified plunger position [016] is valid at time of release for the filled syringe. The plunger can move outside the specified tolerance window when the syringe is subjected to temperature and pressure changes, e.g. when the head space (amount of air in the syringe) expands during air transportation.

The head space is defined by the filling volume and plunger position at time of release. Based on the specified tolerances for filling volume and plunger position, the head space is calculated in the Tolerance Stack-up [1], [2]:

Head space volume = 0.027-0.155 ml

Head space length = 0-3.9 mm

Air transportation is considered the worst-case situation for plunger movement due to head space expansion at the lower air pressure of higher altitudes (650 mbar in aircraft with pressure cabin, corresponding to normal air pressure at an altitude of 3,500 meter). Plunger movement during air transportation is calculated in the Tolerance Stack-up [1], [2]:

Plunger movement during air transport = 0.5-2.7 mm in outwards direction

## References

- 1. NovoDOCS 003664953, Tolerance stack-up, Syringe 1 ml Long with Staked Needle and RNS, Extractable Volume 0.5 ml, ed 3.0
- 2. NovoDOCS 004276288, Tolerance stack-up, Syringe 1 ml Long with Staked Needle and RNS, Extractable Volume 0.75 ml, ed 1.0

Dept. 199, Prim Pack Development DFP ManDev EMK0595-12.01-007 NDID 004276473

INTERNAL USE ONLY

 $\begin{array}{c|c} 18 \text{ July } 2018 & \textbf{\textit{Novo Nordisk}} \\ 1.0 & \end{array}$ Date: Version: Status: Final 5 of 5 Page:

# **Change log**

Version	Date/Initials	Reason for revision
1.0	2018-07-18/JGJ	New document, prepared to include 0.75 ml PFS in an early phase of the DV3396 Sema Obesity SDD project, will be replaced by drawing in QBIQ later in the project.

### Novo Nordisk A/S Electronic Signature Page 6 Aug 2018 05:59:40

 ${\sf EMK0595-12.01-007\_DWG,\,Syringe\,\,1\,\,ml\,\,Long\,\,with\,\,Staked\,\,Needle\,\,and\,\,RNS,\,\,updated\,\,for\,\,Sema\,\,Obesity\,\,project\_NDID\,\,004276473}$ 

Documentum ObjectID:090239b682ed952a

This is a representation of an electronic record that was signed electronically and this page is the manifestation of the electronic signature.

Document signed by	:	
Initials	Full Name	
Meaning		Date and Time of Signature (Server Time)
jgj	JGJ - Jan Gunnar Jørgensen	_
Author Approval		18-Jul-2018 12:01 GMT+0000
fit	FIT - Frida Thuresson	
Quality Assurance Approval		18-Jul-2018 13:04 GMT+0000
maes	MaEs - Mads Reedtz Espersen	
Expert Approval		06-Aug-2018 05:59 GMT+0000