## SIEMENS MAGNETOM Allegra syngo MR 2004A

 $\verb|\USER\INVESTIGATORS\Castellanos\Awake\_Asleep\_Inscapes\\ \verb|\LFO\_Relax\_offline||$ 

			1001
Routine		R >> L	192 [mm]
Slice group 1		———   F >> H	102 [mm]
Slices	34	Physio	
Dist. factor	0 [%]	1st Signal/Mode	None
Position	Isocenter	Coguence	
Orientation	Transversal	Sequence	0
Phase enc. dir.	R >> L	Introduction	0 2D
Rotation	90 [deg]	Dimension	
Phase oversampling	0 [%]	Averaging mode	Long term
FoV read	240 [mm]	Bandwidth	3906 [Hz/Px]
FoV phase	80.0 [%]	Free echo spacing	0
Slice thickness	3 [mm]	Echo spacing	0.32 [ms]
TR	2000 [ms]	EPI factor	64
TE	30 [ms]	RF pulse type	Normal
Averages	1	Gradient mode	Fast
Concatenations	1		
Filter	None	RF spoiling	1
		Image Reconstruction	Save Raw
Coil elements	TR	Field Map Mode	FMap Off
Contrast		Multi Mode	Multi Shot
MTC	0	No. of Echos	1
Magn. preparation	None	Inner Image Size	16
Flip angle	82 [deg]	No. of Shots	1
Reconstruction	Magnitude	Readout Direction	Normal
Fat suppr.	Fat sat.	Spoil Phase Step	123.0
Measurements	180		
		Spoil Grad Amp	3.0
Pause after meas.	0 [s]		
Multiple series	0		
Resolution			
Base resolution	80		
Phase resolution	100 [%]		
Phase partial Fourier	Off		
Filter 1			
Raw filter	Off		
Filter 2			
Large FoV	Off		
Filter 3			
Normalize	Off		
Filter 4	<b>3</b>		
Elliptical filter	Off		
Interpolation	0		
	•		
Geometry	late de		
Multi-slice mode	Interleaved		
Series	Interleaved		
Special sat.	None		
System			
Save uncombined	0		
Scan at current TP	1		
MSMA	S - C - T		
-	R >> L		
Sagittal Coronal	R >> L A >> P		
Transversal	A >> P F >> H		
Nova_TR / TR	1 		
Shim mode	Standard		
Confirm freq. adjustment	0		
Assume Silicone	0		
Ref. amplitude [1H]	140.000 [V]		
Adjust volume	0.000 [v]		
Position	Isocenter		
Orientation	Transversal		
Rotation			
A >> P	90 [deg] 240 [mm]		