ABIDEII-BNI_1

	Anatomical	Rest
Manufacturer	Philips	Philips
Model	Ingenia	Ingenia
Headcoil	15Ch	15Ch
Field Strength	3T	3T
Sequence	MPRAGE	EPI
Flip Angle [Deg]	9	80
Inversion Time (TI) [ms]	900	-
Echo Time (TE) [ms]	Shortest	25
Repetition Time (TR) [ms]	shortest	3000
Bandwidth per Voxel (Readout) [Hz]	240	3280
Parallel Acquisition	SENSE x 1	SENSE x 2
Partial Fourier	-	-
Number of Slices	170	50
Slice Orientation	Axial	Axial
Slice Phase Encoding Direction	-	-
Slice Acquisition Order	Sequential	-
Slice Thickness [mm]	1.2	4
Slice Gap [mm]	0	0
Field of View [mm]	270X252	240X240
Acquisition Matrix	244X227	64X64
Slice In-Place Resolution [mm^2]	1.1X1.1	3.75X3.75
Number of Measurements	-	120
Acquisition Time [min:sec]	5:34	6.09
Fat Suppression	No	None
Number of Directions	-	-
Number of B Zeros	-	-
B Value(s) [s/mm^2]	-	-
Averages	-	-

Anat Note(s): Reconstructed voxels 1.05X1.05X1.2