

SOP-WT-002 — Dialysis Water Treatment and Testing

Owner: Nikhil Joshi. Applies to the water plants of all three centres. Revision 3.

1. Limits for product water (ISO 23500)

Parameter	Action level	Maximum
Bacteria	50 CFU/mL	100 CFU/ mL
Endotoxin	0.125 EU/mL	0.25 EU/mL

2. Testing

A sample is taken from the end of the distribution loop every month and sent to the external laboratory.

3. Disinfection

Each water plant and its distribution loop receive chemical disinfection every month, recorded in the disinfection log.

4. Results above the maximum

Inform the Medical Director the same day, reduce or stop dialysis at the centre, disinfect the plant and loop, and resample. Dialysis resumes only after a result below the action level.