

Nephrova Renal Care Pvt. Ltd.

City Diagnostic Laboratory — Dialysis Water Report

Client: Nephrova Renal Care Pvt. Ltd., Hadapsar centre. Sample: RO-H1 product water, end of loop.
Collected 18 August 2026. Reported 21 August 2026.

Test	Result	Maximum allowed	Status
Bacteria culture	48 CFU/mL	100 CFU/mL	Within limit
Endotoxin	0.31 EU/mL	0.25 EU/mL	ABOVE LIMIT

The RO-H1 sample showed endotoxin exceedance at 0.31 EU/mL.

Results for RO-H1 are assessed against the ISO 23500 limits for dialysis water.

The bacteria count of 48 CFU/mL is just below the action level of 50 CFU/mL.