

Nephrova Renal Care Pvt. Ltd.

City Diagnostic Laboratory — Dialysis Water Report (Resample)

Client: Nephrova Renal Care Pvt. Ltd., Hadapsar centre. Sample: RO-H1 product water, end of loop.
Collected 23 August 2026. Reported 26 August 2026.

Test	Result	Action level	Status
Bacteria culture	6 CFU/mL	50 CFU/mL	Within limit
Endotoxin	0.04 EU/mL	0.125 EU/mL	Within limit

Both results for RO-H1 are below the action levels. The water is suitable for dialysis.