Checklist for Follow-Up on Sterilization of Towels and Replacement with Disposable Sets for Cardiac Surgery

N	Step	Achieved	Not
0			Achieved
1.	Pre-cleaning and inspection		
	 ✓ Thoroughly cleaned to remove dirt, debris, or organic matter ✓ Towels should be visually inspected to ensure they are free of any tears, holes, or fraying or foreign materials ✓ Identify stains or discoloration ✓ Isolate any towels with defects, mark defective items with a clear label ("Unfit for Use") ✓ Record the quantity, type of defect, and 		
	batch details		
2.	Packaging ✓ Properly folding and wrapping towels ensures uniform sterilization ✓ Items should be packed to allow adequate steam penetration and prevent air pockets		
3.	Sterilization methods ✓ Steam sterilization requires autoclaving at 121°C (250°F) for 15-30 minutes		

ı

November Task: Shirwan Jamil M Amin Doing follow up for sterilization of towels if any defect found eliminate them and bring new disposable sets for cardiac surgery

	✓ Ensuring that sterilization cycles include a	
	drying phase to prevent residual moisture	
4.	• Quality control	
	✓ Each sterilized pack should include	
	indicators (Check sterilization tape	
	indicators for color changes), to confirm	
	that the sterilization parameters were met.	
5.	. Storage	
	✓ Once sterilized, towels should be stored in	
	a clean, dry, designated sterile area	
	✓ Limiting handling to prevent	
	contamination and ensuring that packages	
	remain intact until they are needed for	
	surgical use	
6.	Danla amant with dispassible sets	
0.	Replacement with disposable sets	
	✓ Use high-quality, single-use disposable	
	sets that meet surgical standards (ISO	
	11607)	
	✓ Verify expiration dates and packaging	
	integrity before use	
	✓ Store Properly:	
	- Maintain disposable sets in a clean, dry,	
	and sterile environment	
	- Rotate stock to avoid expired items	