

## Sources Configuration

**NOTE: ADDING OR REMOVING SOURCES WOULD BE BEST LEFT TO SPECTRUM SYSTEMS, INC. AS MISCONFIGURATIONS COULD CAUSE DATA LOSS.**

The *Sources Configuration Tool* in the *Admin Tab* allows the end-user to view, *Add*, or *Retire* sources, as seen in **FIGURE ERROR! No text of specified style in document.-1**. The *Sources Configuration Sub-Tab* also allows for basic confirmation of the *SpectraPak® Prism Servers* and *Modbus* configurations.

Sources Administration

☐ Show Retired Sources

Source Type	Name	Description	Address	Port	Modbus ID	Service Alarm	Database Alarm	Device Alarm	WCF Endpoint	Effective Date
SpectraPak-E	SpectraPak 1	Unit A - Controll	192.168.1.201	1234		SpectraPak	SpectraPak	SpectraPak	net.tcp://172.16.3	2/25/2014 12:00:00
SpectraPak-E	SpectraPak 2	Unit A - Controll	192.168.1.202	1234			SpectraPak	SpectraPak	net.tcp://172.16.3	2/25/2014 12:00:00
Modbus over Eth	ModbusESlave	Test Slave	172.16.3.12	502	1		ModbusESla	ModbusESla	net.tcp://172.16.3	5/22/2014 12:00:00
Modbus over Eth	ModbusEMaster	Test Master	192.168.1.80	502	1		ModbusEMa	ModbusEMa	net.tcp://172.16.3	2/25/2014 12:00:00
Modbus over Eth	ModbusEMaster	Rao	172.16.5.222	502	1				net.tcp://172.16.3	2/25/2014 12:00:00
Modbus over Eth	RAO T1	RAO Test 1	192.168.1.111	502	1				net.tcp://172.16.3	4/15/2014 12:00:00
Modbus over Eth	RAO T2	RAO Test 2	192.168.1.112	502	1				net.tcp://172.16.3	4/15/2014 12:00:00
Modbus over Eth	RAO T3	RAO Test 3	192.168.1.113	502	1				net.tcp://172.16.3	4/15/2014 12:00:00
Modbus over Eth	RAO T4	RAO Test 4	192.168.1.114	502	1				net.tcp://172.16.3	4/15/2014 12:00:00
Modbus over Eth	RAO T5	RAO Test 5	192.168.1.115	502	1				net.tcp://172.16.3	4/15/2014 12:00:00
Modbus over Eth	Hg Analyzer		172.16.5.80	502	80				net.tcp://172.16.3	7/1/2014 12:06:13
Modbus over Eth	Hg Calibrator		172.16.5.81	502	81				net.tcp://172.16.3	7/1/2014 12:35:35

Add

Retire

Back

Save

Figure **ERROR! No text of specified style in document.-1** Sources Configuration