

## Question 9.3

**Topic:** Calibration Gases

**Question:** May I use my calibration gas from daily calibration error tests for a quarterly linearity check?

**Answer:** Yes. The same cylinder of calibration gas used for daily calibration error tests may be used for a quarterly linearity check.

**References:** Appendix A, Section 6.2; Appendix B, Section 2.2.1

**History:** First published in May 1993, Update #1; revised July 1995, Update #6; revised in October 1999 Revised Manual