Question 7.8

Topic: Primary and Backup Designations

Question: Can a primary monitor on one unit be used as a backup monitor on another unit, and vice-versa?

Answer: Yes. Section 75.10(e) provides that a particular monitor may be designated both as a certified primary monitor for one unit and as a certified redundant backup monitor for another unit. An example of this would be an SO_2 analyzer which is *continuously* time-shared between Units 1 and 2. If Unit 2 has its own separate primary SO_2 monitoring system, the time-shared analyzer could then be designated both as the primary SO_2 monitoring system for Unit 1 and as a redundant backup SO_2 monitoring system for Unit 2.

References: § 75.10(e)

History: First published in Original March 1993 Policy Manual; revised in October 1999 Revised Manual