

Question 3.25

Topic: Test Method 2H -- Guniting Stack

Question: To use the 0.9900 default wall effects adjustment factor (WAF) value in Method 2H, does the entire stack have to be brick or mortar or just the lining? What about guniting?

Answer: To use the 0.9900 default WAF, the stack lining must be brick or mortar. Guniting is not considered to be brick or mortar.

References: 40 CFR Part 60, Appendix A-2, Method 2H

History: First published in March 2000, Update #12