

Question 9.13

Topic: Interference Checks and Data Validation

Question: Is there a startup grace period for the daily interference checks of a stack flow monitor?

Answer: Yes. Section 2.1.5.2 of Appendix B provides a startup grace period for both daily calibration error tests and for daily flow monitor interference checks.

References: Appendix A, Section 2.2.2.2; Appendix B, Section 2.1.5.2; Question 9.9

History: First published in November 1995, Update #7; revised in October 1999 Revised Manual; revised in 2013 Manual