

The Therac-25

A Case Study in Medical Device Ethics and Safety

Navigating the intersection of innovation, responsibility, and human life.

PRESENTED BY :

GROUP 7 Members:

1. RA2411026010237 VARTIKA JAMWAL
2. RA2411026010238 ANIRUDDHA MOOKERJEE
3. RA2411026010239 SAFWAN
4. RA2411026010240 JOEL REGI
5. RA2411026010241 RAGHAV MISHRA
6. RA2411026010242 YARRU SATISH BABU
7. RA2411026010244 ARVIND SATHYADAS
8. RA2411026010245 YASH BHARUNTIA

Course Name: Professional Ethics

Date of Presentation: 19-09-2025

Background: The Therac-25 Machine

- Developed by **Atomic Energy of Canada Limited (AECL)**,
- Launched **1982**
- Computer-controlled linear accelerator for cancer radiation therapy
- Replaced critical hardware safety interlocks with complex software controls
- Installed in **11 sites across US and Canada**

The Therac-25 was **an advanced radiation therapy machine** designed to deliver precise doses of radiation to cancer patients. Its innovation lay in its software-driven control system, which unfortunately proved to be its fatal flaw.



Key Ethical Issues Unveiled

Overconfidence in Software

Untested software safety replacing critical hardware safeguards.

Inadequate Practices

Poor software engineering, testing, and quality assurance.

Manufacturer Dismissal

Ignoring patient and operator reports of malfunctions.

Poor Regulatory Oversight

Delayed response and insufficient intervention by regulatory bodies.

These issues created a dangerous environment where technological ambition outpaced ethical responsibility and patient safety.

Stakeholders Affected



The ripple effects of the Therac-25 failures impacted numerous parties, extending far beyond the immediate victims.

- **Patients:** At least six suffered massive radiation overdoses, leading to severe injury or death.
- **Operators/Technicians:** Faced confusing interfaces and inadequate training, often blamed for errors.
- **Manufacturer (AECL):** Responsible for design, testing, and response, facing severe reputational and legal damage.
- **Regulatory Bodies (FDA, Health Canada):** Critiqued for slow response and insufficient oversight.
- **Medical Community & Public:** Trust in advanced medical technology was significantly eroded.

Application of Ethical Theories

Utilitarianism

Failure to minimize harm and maximize the greatest good (patient safety).

Deontological Ethics

Breach of duty to ensure safe medical devices and protect patient rights.

Virtue Ethics

Lack of transparency, accountability, and integrity by the manufacturer.

Professional Ethics

Neglect of rigorous software engineering standards and professional responsibility.

The Therac-25 case highlights how various ethical frameworks were violated, leading to tragic outcomes and underscoring the moral imperative in medical device development.

Ethical Dilemmas Faced

1

Innovation vs. Safety

Balancing rapid innovation and speed-to-market against thorough safety validation.

2

Assigning Responsibility

Who bears the ultimate responsibility for harm: software engineers, management, or regulators?

3

Transparency vs. Reputation

Reporting and addressing errors promptly versus protecting company reputation and avoiding legal repercussions.

4


Patient Safety vs. Costs

Prioritizing patient safety above financial and legal risks associated with recalls or redesigns.

These dilemmas illustrate the complex ethical landscape in high-stakes industries, where decisions have profound human consequences.

Decisions & Consequences

THERAC-25 RADIATION OVERDOSES TIMELINE (1985-POST 1987)



Alternative Solutions for Prevention



Hybrid Safety

Retaining hardware safety interlocks alongside software controls.



Rigorous Testing

Independent and thorough software engineering practices and testing.



Transparent Reporting

Clear incident reporting and swift, decisive corrective action.



User-Centered Design

Comprehensive operator training and intuitive interface design.



Stronger Regulation

Evolving regulatory requirements for software in medical devices.

Had these measures been in place, the Therac-25 tragedy could have been entirely avoided, saving lives and preserving trust.

Lessons Learned: Shaping the Future



Software Rigor

Safety-critical medical device software demands the highest engineering rigor.



No Single Point of Failure

Over-reliance on software without hardware backups is inherently dangerous.



Listen to Users

Ethical responsibility includes actively listening to user feedback and incident reports.



Evolving Regulation

Regulatory frameworks must continuously evolve to address emerging software risks.



Education & Standards

The Therac-25 case profoundly shaped modern medical device safety standards and ethics education.

The tragedy served as a stark reminder of the profound ethical and safety implications in complex technological systems, leading to fundamental changes in the industry.

References & Further Reading

- **Leveson, N.G., & Turner, C.S. (1993).** An Investigation of the Therac-25 Accidents. IEEE Computer.
- **Ethics Unwrapped, University of Texas:** Therac-25 Case Study.
- **Wikipedia:** Therac-25.
- **Online Ethics Center:** Therac-25 Collection.
- **Israelski, E.W. (2004).** A Case Study of the Therac 25 Radiation Therapy System. Journal of Healthcare Engineering.