




- GENERAL NOTES:
1. MATERIAL: **EPS FOAM 1 LB / CUBIC FOOT**
  2. SPECIAL FINISH: COAT CYLINDER WITH KRYLON "Low Odor Clear Matte Finish Aerosol Spray" KRYLON NUMBER 7120. THIS IS A CLEAR LATEX FINISH. SPRAY 2-3 LIGHT COATS TO ENSURE EVEN COVERAGE.
  3. PARTS ARE TO BE CLEAN AND FREE OF OIL, GREASE, AND OTHER CONTAMINANTS
  4. RAW MATERIAL SUGGESTED VENDOR: FOAMBYMAIL.COM
  5. RAW MATERIAL ITEM DESCRIPTION: **EPS Polystyrene Foam 1lb/ft^3 , 4" THK**

REVISION HISTORY				
ZONE	REV	DESCRIPTION	DATE	APPROVED
	-	INITIAL RELEASE	6/6/2023	HHMI\\arnoldj
	A	INCREASE CENTER HOLE TO 0.250 DIA	6/7/2023	HHMI\\arnoldj



Ø7.95 ± .020  
CUT OUTSIDE DIAMETER WITH HOT WIRE CUTTER.  
MINIMIZE DEVIATIONS IN SURFACE FOR BEST RESULTS.  
SEE NOTES FOR COATING DETAILS.

Ø.250 THRU




NOTICE:  
INFORMATION CONTAINED IN THIS DOCUMENT OR ANY REPRODUCTION THEREOF, IS PROPRIETARY INFORMATION AND PROPERTY OF HOWARD HUGHES MEDICAL INSTITUTE. IT SHALL NOT BE DISCLOSED, COPIED, DUPLICATED OR USED FOR MANUFACTURE, PRODUCTION OR PROCUREMENT, WITHOUT THE EXPRESS WRITTEN PERMISSION OF HOWARD HUGHES MEDICAL INSTITUTE.

(UNLESS SPECIFIED OTHERWISE)  
PRIMARY UNITS: INCHES  
[SECONDARY UNITS]: MILLIMETERS



PRIMARY TOLERANCES:  
X.X ± 0.020  
X.XX ± 0.010  
X.XXX ± 0.005  
X.XXXX ± 0.0005  
ANGULAR ± 0.5 DEG

- DO NOT SCALE DRAWING -

THIRD ANGLE PROJECTION:



Howard Hughes  
Medical Institute



Research Campus

CT FOAM WHEEL.ipt

SIZE C	PART NUMBER J010514	REV A	SHEET 1 OF 1
-----------	------------------------	----------	-----------------