

Heaven's Light is Our Guide
Rajshahi University of Engineering and Technology



Course Code
ECE 4144

Course Title
Biomedical Engineering Sessional

Experiment Date: July 16, 2025,
Submission Date: July 23, 2025

Lab Report 1:
Study of ECG Signal Using Digital Electrocardiogram Device

Submitted to
Md Mayenul Islam
Assistant Professor
Dept of EEE, Ruet

Submitted by
Md. Tajim An Noor
Roll: 2010025

Experiment 1

Study of ECG Signal Using Digital Electrocardiogram Device

Objectives

- Acquire and examine human ECG signals utilizing a digital electrocardiograph with 10 leads.
- Detect and distinguish ECG waveform features, including the P wave, QRS complex, and T wave.
- Comprehend the positioning and importance of limb and chest leads for interpreting ECG results.

Theory

An electrocardiogram (ECG) records the heart's electrical signals using electrodes placed on the skin. In clinical practice, a typical ECG setup employs 10 electrodes—4 on the limbs and 6 on the chest—to generate 12 distinct leads. These include 3 standard limb leads (I, II, III), 3 augmented limb leads (aVR, aVL, aVF), and 6 precordial (chest) leads (V1–V6). Each lead offers a different perspective on the heart's electrical activity, aiding in comprehensive cardiac assessment. By analyzing changes in the P wave, QRS complex, and T wave across these leads, clinicians can detect irregularities such as arrhythmias, heart attacks, and other cardiac disorders.

ECG Lead Types and Electrode Placement

- **Lead I** (Limb; Right Arm): LA - RA
- **Lead II** (Limb; Left Arm): LL - RA
- **Lead III** (Limb; Left Leg): LL - LA
- **Lead IV** (Limb; Right Leg): Ground
- **aVR** (Augmented): $RA - (LA + LL)/2$
- **aVL** (Augmented): $LA - (RA + LL)/2$
- **aVF** (Augmented): $LL - (RA + LA)/2$
- **V1** (Chest): 4th intercostal space, right sternal border
- **V2** (Chest): 4th intercostal space, left sternal border

- **V3** (Chest): Between V2 and V4
- **V4** (Chest): 5th intercostal space, midclavicular line
- **V5** (Chest): Same level as V4, anterior axillary line
- **V6** (Chest): Same level as V4, midaxillary line

Required Equipment

- 10-lead Digital Electrocardiograph (HP-1300plus)
- ECG lead wires and disposable electrodes
- Volunteer human subject
- ECG paper or digital display for signal observation

Experimental Observation

- The volunteer was asked to relax and remain still in a supine position to minimize motion artifacts.
- Standard electrode placement was followed: four electrodes were attached to the limbs and six to the chest at designated anatomical locations.
- The digital ECG system successfully recorded signals from all twelve leads, providing clear and consistent waveforms.
- Lead II produced the most prominent and stable trace, which was used to determine the subject's heart rate.
- The characteristic ECG features—P wave, QRS complex, and T wave—were easily identified in the recordings.
- The subject's ECG showed normal sinus rhythm, with no evidence of arrhythmias or abnormal patterns.




Figure 1: ECG

Discussion

Using the HP-1300plus digital ECG device, twelve standard leads were recorded with reliable signal quality. The device's IEC60601-1 safety compliance ensured safe use. Proper electrode placement was crucial for accurate waveforms, with Lead II providing the clearest signal for heart rate measurement. This experiment enhanced understanding of cardiac conduction and the clinical value of different ECG leads.