TRIP Intercomparison Results: October 15, 2001

First Semester QC Sample Results Reference time: 12:00 GMT, April 4, 2001

<u> </u>		· · · ·			
Facility Code	¹²⁵ l (dpm)	Uncertainty (dpm)	Bias	ANSI N13.30 bias criteria	
ANSI N13.30 acceptable range (dpm)		9.06E+05 - 1.81E+06			
Certified value	1.21E+06	3.62E+04			
2-100	1.39E+06	8.87E+04	0.08	Pass	
2-600	1.36E+06	8.99E+04	0.06	Pass	
7	9.03E+05	1.01E+04	-0.30	Fail	
12	4.16E+05	4.30E+03	-0.68	Fail	
25	1.36E+06	9.74E+03	0.06	Pass	
28	1.01E+06	6.06E+04	-0.21	Pass	
37	1.14E+06	5.00E+05	-0.11	Pass	
40	2.01E+06	1.35E+03	0.57	Fail	
52	1.19E+06	1.79E+05	-0.07	Pass	
60	NR	NR	NR	NR	
73	1.04E+06	2.28E+05	-0.19	Pass	
99	7.57E+05	1.45E+05	-0.41	Fail	
Bias = (measured value - certified value) / certified value NR = Not reported					

I-125 Results

Facility Code	¹³¹ I (dpm)	Uncertainty (dpm)	Bias	ANSI N13.30 bias criteria	
ANSI N13.30 acceptable range (dpm)		9.63E+05 - 1.93E+06			
Certified value	1.28E+06	3.85E+04			
2-100	1.59E+06	8.34E+04	0.32	Pass	
2-600	1.71E+06	9.12E+04	0.42	Pass	
7	9.86E+05	4.86E+03	-0.18	Pass	
12	1.33E+06	6.80E+03	0.10	Pass	
25	8.70E+05	2.68E+04	-0.28	Fail	
28	1.15E+06	3.11E+05	-0.05	Pass	
37	1.18E+06	5.19E+05	-0.02	Pass	
40	1.10E+06	9.80E+02	-0.09	Pass	
52	1.18E+06	1.34E+05	-0.02	Pass	
60	NR	NR	NR	NR	
99	1.14E+06	7.25E+04	-0.06	Pass	
Bias = (measured value - certified value) / certified value NR = Not reported					