EDAN

深圳市理邦精密仪器股份有限公司
EDAN INSTRUMENTS, INC.

SONOTRAX(2009)主机出\入库检验报告
SONOTRAX(2009)Series Products Inspection Report

EDAN\QR\F-02.06.260439-1.4　　编号 Number: M2580745-02

检验作业指导书编号/版本 Inspection Work Instructions No./Version : EDAN\WI\QC-02.06.260439-1.4

制令号Order: M2580745

机　型Model: SONOTRAX II Pro

检验设备：自动化胎心测量系统 Auto FHR Measure System: ☒N/A ☐GZZK0311 ☐GZZK0312 超声多普勒胎儿测量仪检测装置FHR sensitivity test device:☐N/A ☐AESC0048☐AEZK0044 ☐AEPG0044 ☒AEZK0047☐AEZK0048 ☐_____ 万用表Multimeter (AEZK0010) 其他 Others:

检验项目 Item / 主机序列号SN	外观、标贴检验 Appearance and Label inspection	探头识别检验 Probe identification	音量调节检验 Volume adjust inspection	*功能按键检验 *Key function inspection	*背光、显示检验 *Backlight、Display function inspection	*充电功能检验 *Charging function inspection	*录音功能检验 *Record function inspection	*综合灵敏度检验 Overall Sensitivity 静噪声 Noise n≤100 (mV)	*综合灵敏度检验 Overall Sensitivity 信号 Signal s≥2n (mV)	*FHR准确度检验 FHR Accuracy (±2BPM) 50	120	240	声音检验 Voice inspection	结论
560054-M258074500 11	√	√	√	√	√	√	√	21	231	50	120	240	√	☒合格Qualified ☐不合格Nonconformity 《不合格品处理记录表Nonconforming Product Handling Record》NO.:
560054-M258074500 12	√	√	√	√	√	√	√	23	256	50	120	240	√	☒合格Qualified ☐不合格Nonconformity 《不合格品处理记录表Nonconforming Product Handling Record》NO.:
560054-M258074500 13	√	√	√	√	√	√	√	20	274	50	120	240	√	☒合格Qualified ☐不合格Nonconformity 《不合格品处理记录表Nonconforming Product Handling Record》NO.:
560054-M258074500 14	√	√	√	√	√	√	√	25	294	50	120	240	√	☒合格Qualified ☐不合格Nonconformity 《不合格品处理记录表Nonconforming Product Handling Record》NO.:
560054-M258074500 15	√	√	√	√	√	√	√	27	255	50	120	240	√	☒合格Qualified ☐不合格Nonconformity 《不合格品处理记录表Nonconforming Product Handling Record》NO.:
560054-M258074500 16	√	√	√	√	√	√	√	29	211	50	120	240	√	☒合格Qualified ☐不合格Nonconformity 《不合格品处理记录表Nonconforming Product Handling Record》NO.:
560054-M258074500 17	√	√	√	√	√	√	√	28	205	50	120	240	√	☒合格Qualified ☐不合格Nonconformity 《不合格品处理记录表Nonconforming Product Handling Record》NO.:
560054-M258074500 18	√	√	√	√	√	√	√	26	273	50	120	240	√	☒合格Qualified ☐不合格Nonconformity 《不合格品处理记录表Nonconforming Product Handling Record》NO.:
√ 560054-M258074500 19	√	√	√	√	√	√	√	22	293	50	120	240	√	☒合格Qualified ☐不合格Nonconformity 《不合格品处理记录表Nonconforming Product Handling Record》NO.:
560054-M258074500 20	√	√	√	√	√	√	√	24	246	50	120	240	√	☒合格Qualified ☐不合格Nonconformity 《不合格品处理记录表Nonconforming Product Handling Record》NO.:
检验员Inspector1/日期Date:	张建龙 2025.8.16													
审核Review /日期Date:	魏正早 2025.8.16													

注：1、检验项目适用，符合要求在记录栏内划“√”；不符合要求在记录栏内划“×”，检验项目不适用，则在检验项打“\”；. If the item checked conforms to the standards, draw “√”, If the item checked does not conform to the standards, draw a “×”. If the item checked does not apply to the product being checked, draw a cross “\”

2、若所有项目合格则在结论栏“☐合格”打“×”，不合格则在“☐不合格”打“×”，并填写上不合格处理记录表NO。If all items is” ACC”, draw a “×” in the ☐ before Qualified in the conclusion。If the item checked does not conform to the standards, and draw a “×” in ☐ before ‘NG’, select Nonconformity, and fill out the Nonconforming Product Handling Record.

3、编号填写方式：”制令号-序号”. 例如M1190580-01，表示制令M1190580的第1份检验报告。The number write as”Order+series”, for example”M1190580-01”, it means the first report of the order.