

EXDA Gas Mixing Export

Project: VH2D-Project

Responsible Researcher: PhD Student Javier I. Camacho

Total Records: 25

Generated: 2026-05-31 16:30:17

MATLAB Verified: Yes

Verification File A: scripts/GasMixingVerificationFiles/H2_MFC_Fill_Calculator_v000.m

Verification File B: scripts/GasMixingVerificationFiles/AuxFcn_H2_MFC_FillCalculator_000.m

Shared Chamber Setup (applies to all runs unless noted otherwise):

L=0.900 m, W=0.900 m, H=0.900 m, Pipes+=VARIES m³, Hotwire+=0.000 m³, Welded-=0.000 m³, Bolts-=0.000 m³, Vcorr=VARIES L

Note: setup values vary across runs; detailed per-run setup table is included below.

Core Gas Mixing Results

Group	Run/Test	H2%	Pch(Pa)	Tch(K)	H2(g)	H2inj(L)	MFC	t(s)	t(min)	Updated
Prep	Prep-01	13.50	101325	292.15	8.276	98.41	20.00	301.3	5.0	2026-05-25 17:27:49
Verif	Verif-01	13.50	101325	294.25	8.217	98.42	20.00	299.2	5.0	2026-05-26 11:46:08
Verif	Verif-02	14.00	101325	293.15	9.384	111.97	20.00	334.8	5.6	2026-05-26 14:03:57
Verif	Verif-03	16.00	101325	294.15	10.638	127.37	20.00	387.3	6.5	2026-05-26 15:53:24
Verif	Verif-04	18.00	101325	291.65	12.070	143.29	20.00	439.5	7.3	2026-05-26 17:31:49
VH2D-01	VH2D-01-01	16.00	101325	293.15	10.674	127.37	20.00	388.6	6.5	2026-05-27 13:51:19
VH2D-01	VH2D-01-02	16.00	101325	292.05	10.714	127.37	20.00	390.1	6.5	2026-05-27 14:47:32
VH2D-01	VH2D-01-03	18.00	101325	291.35	12.083	143.29	20.00	439.9	7.3	2026-05-27 15:24:55
VH2D-01	VH2D-01-04	18.00	101325	290.85	12.103	143.29	20.00	440.7	7.3	2026-05-27 16:30:34
VH2D-01	VH2D-01-05	19.00	101325	290.55	12.789	151.26	20.00	465.6	7.8	2026-05-27 17:50:04
VH2D-01	VH2D-01-07	20.00	101325	290.56	13.462	159.22	20.00	490.1	8.2	2026-05-27 17:57:37
VH2D-02	VH2D-02-01	18.00	101325	289.95	12.141	143.29	20.00	442.0	7.4	2026-05-28 11:41:03
VH2D-02	VH2D-02-02	18.00	101325	291.15	12.091	143.29	20.00	440.2	7.3	2026-05-28 11:43:19
VH2D-02	VH2D-02-03	20.00	101325	293.15	13.343	159.22	20.00	485.8	8.1	2026-05-28 13:22:48
VH2D-02	VH2D-02-04	20.00	101325	292.15	13.388	159.22	20.00	487.5	8.1	2026-05-28 14:16:19
VH2D-03	VH2D-03-01	18.00	101325	292.45	12.037	143.29	20.00	438.3	7.3	2026-05-28 15:53:21
VH2D-03	VH2D-03-02	18.00	101325	292.25	12.045	143.29	20.00	438.6	7.3	2026-05-28 16:35:43
VH2D-03	VH2D-03-03	20.00	101325	292.15	13.388	159.22	20.00	487.5	8.1	2026-05-28 17:13:08
VH2D-03	VH2D-03-04	20.00	101325	291.65	13.411	159.22	20.00	488.3	8.1	2026-05-28 17:59:31
VH2D-03	VH2D-03-05	20.00	101325	291.45	13.421	159.22	20.00	488.6	8.1	2026-05-28 18:37:23
VH2D-04	VH2D-04-01	16.00	101325	291.75	10.725	127.37	20.00	390.5	6.5	2026-05-29 10:32:56
VH2D-04	VH2D-04-02	16.00	101325	292.75	10.689	127.37	20.00	389.2	6.5	2026-05-29 11:36:38
VH2D-04	VH2D-04-03	21.00	101325	293.05	14.015	167.18	20.00	510.3	8.5	2026-05-29 12:40:04
VH2D-04	VH2D-04-04	21.00	101325	293.75	13.981	167.18	20.00	509.0	8.5	2026-05-29 13:31:18

VH2D-04 | VH2D-04-05 | 12.00 | 101325 | 295.05 | 7.954 | 95.53 | 10.00 | 579.2 | 9.7 | 2026-05-29 14:59:44

Per-Run Chamber Setup (because values vary)

Group	Run/Test	L (m)	W (m)	H (m)	Pipes+ (m3)	Hotwire+ (m3)	Welded- (m3)	Bolts- (m3)	Vcorr(L)	Notes
Prep	Prep-01	0.900	0.900	0.900	0.000	0.000	0.000	0.000	729.00	-
Verif	Verif-01	0.900	0.900	0.900	0.000	0.000	0.000	0.000	729.00	-
Verif	Verif-02	0.900	0.900	0.900	0.071	0.000	0.000	0.000	799.80	-
Verif	Verif-03	0.900	0.900	0.900	0.067	0.000	0.000	0.000	796.08	-
Verif	Verif-04	0.900	0.900	0.900	0.067	0.000	0.000	0.000	796.08	-
VH2D-01	VH2D-01-01	0.900	0.900	0.900	0.067	0.000	0.000	0.000	796.08	-
VH2D-01	VH2D-01-02	0.900	0.900	0.900	0.067	0.000	0.000	0.000	796.08	-
VH2D-01	VH2D-01-03	0.900	0.900	0.900	0.067	0.000	0.000	0.000	796.08	-
VH2D-01	VH2D-01-04	0.900	0.900	0.900	0.067	0.000	0.000	0.000	796.08	-
VH2D-01	VH2D-01-05	0.900	0.900	0.900	0.067	0.000	0.000	0.000	796.08	-
VH2D-01	VH2D-01-07	0.900	0.900	0.900	0.067	0.000	0.000	0.000	796.08	-
VH2D-02	VH2D-02-01	0.900	0.900	0.900	0.067	0.000	0.000	0.000	796.08	-
VH2D-02	VH2D-02-02	0.900	0.900	0.900	0.067	0.000	0.000	0.000	796.08	-
VH2D-02	VH2D-02-03	0.900	0.900	0.900	0.067	0.000	0.000	0.000	796.08	-
VH2D-02	VH2D-02-04	0.900	0.900	0.900	0.067	0.000	0.000	0.000	796.08	-
VH2D-03	VH2D-03-01	0.900	0.900	0.900	0.067	0.000	0.000	0.000	796.08	-
VH2D-03	VH2D-03-02	0.900	0.900	0.900	0.067	0.000	0.000	0.000	796.08	-
VH2D-03	VH2D-03-03	0.900	0.900	0.900	0.067	0.000	0.000	0.000	796.08	-
VH2D-03	VH2D-03-04	0.900	0.900	0.900	0.067	0.000	0.000	0.000	796.08	-
VH2D-03	VH2D-03-05	0.900	0.900	0.900	0.067	0.000	0.000	0.000	796.08	-
VH2D-04	VH2D-04-01	0.900	0.900	0.900	0.067	0.000	0.000	0.000	796.08	-
VH2D-04	VH2D-04-02	0.900	0.900	0.900	0.067	0.000	0.000	0.000	796.08	-
VH2D-04	VH2D-04-03	0.900	0.900	0.900	0.067	0.000	0.000	0.000	796.08	-
VH2D-04	VH2D-04-04	0.900	0.900	0.900	0.067	0.000	0.000	0.000	796.08	-
VH2D-04	VH2D-04-05	0.900	0.900	0.900	0.067	0.000	0.000	0.000	796.08	MFC set to 7.954 g (12
										vol.%) but stopped at vol.
										8% at location measured by
										H2CM