December 23, 2019

Boston Scientific Corporation
Kimberly Berg
Sr. Regulatory Affairs Specialist
Three Scimed Place
Maple Grove, Minnesota 55311-1566

Re: K193279
Trade/Device Name: Comet™ II Pressure Guidewire
Regulation Number: 21 CFR 870.2870
Regulation Name: Catheter Tip Pressure Transducer
Regulatory Class: Class II
Product Code: DXO, DQX
Dated: November 25, 2019
Received: November 27, 2019

Dear Kimberly Berg:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance) and CDRH Learn (https://www.fda.gov/training-and-continuing-education/cdrh-learn). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

LT Stephen Browning
Assistant Director
Division of Cardiac Electrophysiology, Diagnostics and Monitoring Devices
Office of Cardiovascular Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug Administration

Indications for Use

Form Approved: OMB No. 0910-0120
Expiration Date: 06/30/2020
See PRA Statement below.

510(k) Number *(if known)*

K193279

Device Name

Comet™ II Pressure Guidewire

Indications for Use *(Describe)*

The Comet II Pressure Guidewire is indicated to direct a catheter through a blood vessel and measure physiological parameters in the coronary and peripheral blood vessels.

Type of Use *(Select one or both, as applicable)*

☒ Prescription Use (Part 21 CFR 801 Subpart D) ☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

300 Boston Scientific Way
Marlborough, MA 01752-1234
508-683-4000
www.bostonscientific.com

510(k) Summary
per 21 CFR §807.92

Sponsor	Boston Scientific Corporation 300 Boston Scientific Way Marlborough, Massachusetts 01752 USA
Contact Name and Information	Kimberly Berg Three Scimed Place Maple Grove, MN 55311-1566 Phone: 763-494-2701 Fax: 763-494-2981 e-mail: Kimberly.Berg@bsci.com
Prepared by	Kimberly Berg November 25, 2019
Proprietary Name	Comet™ II Pressure Guidewire
Common Name	Transducer, Pressure, Catheter Tip
Product Code	DXO, DQX
Classification	Class II, 21 CFR Part 870. 2870
Predicate Device	Comet™ Pressure Guidewire K151610 October 6, 2019

300 Boston Scientific Way
Marlborough, MA 01752-1234
508-683-4000
www.bostonscientific.com

Device Description

Comet™ II is a coronary and peripheral guidewire with a pressure sensor mounted approximately 3 cm proximally to the distal end of the straight tip. The optical pressure sensor is integrated within the distal region of a 0.014" guidewire to measure blood pressure gradient changes across coronary and peripheral lesions during endovascular procedures. The tip is radiopaque to help with guidewire placement and shapeable by the user. The guidewire is attached to an optical cable through a detachable connector. The optical cable has a connector on the proximal end for interface with an ancillary Boston Scientific signal processing module. When connected to the signal-processing module, a Boston Scientific iLab Polaris™ Multi-Modality Guidance System displays various physiological parameters including aortic pressure, distal pressure and a fractional flow reserve (FFR) value calculated from the ratio of distal pressure over aortic pressure. A torque device is preloaded onto the proximal end of the wire. The pressure guidewire, optical cable, and torque device are all connected in the single use sterilized packaging. The intended use of Comet II remains within the scope of the predicate intended use.

Indications for Use / Intended Use

Intended Use

The Comet II Pressure Guidewire measures the blood pressure gradient across coronary and peripheral lesions during endovascular procedures. The Comet II Pressure Guidewire may also be used as a coronary or peripheral guidewire for interventional treatments.

Indications for Use

The Comet II Pressure Guidewire is indicated to direct a catheter through a blood vessel and measure physiological parameters in the coronary and peripheral blood vessels.

Comparison of Technological Characteristics

The Comet II Pressure Guidewire incorporates substantially equivalent design, packaging, fundamental technology, manufacturing processes, sterilization process and intended use as those featured in the predicate, Comet Pressure Guidewire, K151610.

Non-clinical Performance Data

Determination of substantial equivalence is based on an assessment of non-clinical performance bench testing data.

Bench Testing

Bench testing was performed to evaluate physical integrity, functionality, and performance of the Comet II Pressure Guidewire. Performance criteria includes dimensional requirements, pressure gradient measurement (sensor) requirements, and interface with compatible devices.

300 Boston Scientific Way
Marlborough, MA 01752-1234
508-683-4000
www.bostonscientific.com

Clinical Testing
Performance testing from clinical studies is not required to demonstrate substantial equivalence of the Comet II.

Conclusion
Based on the indications for use, technological characteristics, and performance testing the Comet II Pressure Guidewire has been shown to be appropriate for its intended use and is considered to be substantially equivalent to the Comet Pressure Guidewire, K151610.