Medtronic

# Guide catheter

Instructions for use

! USA Caution: Federal Law (USA) restricts this device to sale by or on the order of a physician.

Medtronic and Medtronic logo are trademarks of Medtronic. ™* Third-party brands are trademarks of their respective owners. All other brands are trademarks of a Medtronic company.

| | |
|---|---|
|  | Length |
|  | Inner diameter |
|  | Outer diameter |
|   | Sterilized using ethylene oxide |
|  | Date of manufacture |
|   | Consult instructions for use or consult electronic instructions for use |
|  | Manufactured in |
|  | Manufacturer |
|  | Keep away from sunlight |
|   | Catalog number |
|   | For US audiences only |
|  | Do not reuse |
|  | Use-by date |
|   | Lot number |
|  | Do not use if package is damaged and consult instructions for use |
|  | Quantity |

## Device description

The Medtronic guide catheters are composed of a luer hub, strain relief, shaft, segments, tungsten marker band, sleeve, and soft tip. The catheters have a lubricious inner lumen that runs from the proximal to the distal end via a lubricious liner material or coating on the inner lumen.

## Intended user

Interventional procedures using this device should only be performed by physicians who have experience with interventional techniques in the vascular system.

## Indications

The Medtronic guide catheter is designed to provide a pathway through which therapeutic devices are introduced. The guide catheter is intended to be used in the coronary or peripheral vascular system.

## Contraindications

None.

## Recommended items

The following items are recommended when using this device: appropriately sized angiographic guidewire, hemostasis valve, compatible introducer sheath, sterile saline solution and syringe.

## Warnings

- This device was designed for single use only. Do not reuse, reprocess or resterilize this device. Reuse, reprocessing or resterilization may compromise the structural integrity of the device or create a risk of contamination, which could result in patient injury, illness, or death.
- Visually inspect all sterile-barrier packaging before use. If the integrity of the sterilization barrier has been compromised or the device is damaged, do not use the product. Contact your Medtronic representative for return information.
- Take care to avoid damaging the vessel walls through which this catheter passes.
- When using the catheter in smaller vessels, ensure that it does not completely block blood flow.
- Do not use after the use-by date.
- Use aseptic technique during the entire procedure.
- Failure to abide by these instructions for use may result in damage to the device coating, where applicable, which may necessitate intervention or result in serious adverse events.
- When there is limited clearance between devices and the guide catheter lumen, devices must be advanced and withdrawn slowly with the valve open to reduce the risk of embolism.

## Precautions

- The large internal diameter of the catheter permits the injection of substances with little force being required on the syringe. Inject substances slowly whenever attempting to visualize the vessels under angiography.
- The catheter is supplied sterile and nonpyrogenic.
- Store the device at room temperature in a dry place. Keep the device away from sunlight or UV light.
- Do not force passage if resistance is encountered during the insertion of the interventional device through the lumen of the guide catheter. Determine the cause of resistance before proceeding. If the cause of resistance cannot be determined, it is recommended that all devices potentially contributing to the resistance be carefully removed and replaced, as necessary. Consider withdrawing all devices simultaneously.
- Use caution when handling the guide catheter to reduce the possibility of bending or kinking. If the guide catheter becomes kinked, do not torque the device. Instead, carefully withdraw the catheter, any devices within it, and, if needed, the introducer sheath.
- Dispose of the device in accordance with applicable laws, regulations, and hospital procedures, including those regarding biohazards, microbial hazards, and infectious substances.

## Adverse events

Adverse events that may occur or require intervention include, but are not limited to, the following conditions: hemorrhage, hematoma, infection, embolism, vessel, heart, or heart wall damage, perforation, or dissection, vessel spasm, thrombosis, myocardial infarction, stroke, death, vascular occlusion, or allergic reaction.
If a serious incident related to the device occurs, immediately report the incident to Medtronic and the applicable competent authority or regulatory body.

## Instructions for use

1. **Preparing the catheter**
   a. Using sterile technique, carefully remove the catheter from its package by withdrawing the die-cut card with the mounted catheter from the pouch. Carefully separate the catheter from the card to avoid damaging the product.
   b. Carefully inspect the catheter before use to verify that the size, shape, and condition are suitable for the intended procedure.
   c. Flush catheter with sterile saline solution using a 10cc syringe.
2. **Inserting the catheter**
   a. Prepare the patient for angioplasty using standard techniques. Administer appropriate anticoagulation and vasodilation therapy.
   b. Introduce the catheter into the vascular system using cut-down or percutaneous techniques.
   c. Introduce the catheter over a previously-placed standard angiographic wire. Aspirate and flush the guide catheter after the removal of the angiographic wire.

d. Only manipulate the catheter while it is under fluoroscopic observation.
e. The catheter should address the selected vascular site.

3. **Removing the catheter**
   a. At the completion of the angioplasty procedure, aspiration during withdrawal of the catheter is recommended.

## Disclaimer of warranty

The warnings contained in the product labeling provide more detailed information and are considered an integral part of this disclaimer of warranty. Although the product has been manufactured under carefully controlled conditions, Medtronic has no control over the conditions under which this product is used. Medtronic, therefore, disclaims all warranties, both express and implied, with respect to the product, including, but not limited to, any implied warranty of merchantability or fitness for a particular purpose. Medtronic shall not be liable to any person or entity for any medical expenses or any direct, incidental, or consequential damages caused by any use, defect, failure, or malfunction of the product, whether a claim for such damages is based upon warranty, contract, tort, or otherwise. No person has any authority to bind Medtronic to any representation or warranty with respect to the product.

The exclusions and limitations set out above are not intended to, and should not be construed so as to, contravene mandatory provisions of applicable law. If any part or term of this disclaimer of warranty is held to be illegal, unenforceable, or in conflict with applicable law by a court of competent jurisdiction, the validity of the remaining portions of this disclaimer of warranty shall not be affected, and all rights and obligations shall be construed and enforced as if this disclaimer of warranty did not contain the particular part or term held to be invalid.

Medtronic

**Medtronic, Inc.**
710 Medtronic Parkway
Minneapolis, MN 55432
USA
+1 763 514 4000

**Medtronic Mexico S. de R.L. de CV**
Av. Paseo Cucapah 10510 El Lago
C.P. 22210 Tijuana, Baja
California, Mexico

Copyright © 2023 by Medtronic, Inc.

*M047606C001* Rev. A